

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101424\
 Data File : BP022371.D
 Acq On : 15 Oct 2024 01:32
 Operator : RC/JU
 Sample : PB163830BS
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS830

Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel

10/15/2024
 Supervised By :mohammad
 ahmed

Quant Time: Oct 15 02:04:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 14 11:01:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.699	152	103180	20.000	ng/u1	0.00
20) Naphthalene-d8	10.457	136	407767	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.304	164	330735	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.104	188	817486	20.000	ng/u1	0.00
79) Chrysene-d12	21.551	240	854475	20.000	ng/u1	0.00
88) Perylene-d12	24.833	264	988954	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	10945	4.122	ng/uL	0.00
4) Pyridine-d5	3.634	84	151928	20.843	ng/u1	0.00
7) Phenol-d5	6.887	99	204714	23.570	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.040	67	113273	22.190	ng/u1	0.00
11) 2-Chlorophenol-d4	7.240	132	151804	24.632	ng/u1	0.00
15) 4-Methylphenol-d8	8.399	113	167303	24.137	ng/u1	0.00
21) Nitrobenzene-d5	8.840	128	73715	23.511	ng/u1	0.00
24) 2-Nitrophenol-d4	9.551	143	93643	25.798	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.087	165	191699	24.848	ng/u1	0.00
31) 4-Chloroaniline-d4	10.593	131	210560	23.098	ng/u1	0.00
46) Dimethylphthalate-d6	13.710	166	586428	22.317	ng/u1	0.00
49) Acenaphthylene-d8	13.992	160	661240	23.692	ng/u1	0.00
54) 4-Nitrophenol-d4	14.528	143	106984	24.268	ng/u1	-0.01
60) Fluorene-d10	15.316	176	547442	24.019	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.457	200	122672	22.816	ng/u1	0.01
73) Anthracene-d10	17.210	188	930381	24.193	ng/u1	0.00
81) Pyrene-d10	19.580	212	1198510	26.524	ng/u1	-0.02
92) Benzo(a)pyrene-d12	24.609	264	1315841	24.914	ng/u1	-0.02

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.258	88	22064	7.728	ng/uL	95
5) Pyridine	3.652	79	147352	19.716	ng/u1	98
6) Benzaldehyde	6.857	77	112022	23.867	ng/u1	98
8) Phenol	6.916	94	205951	22.791	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.128	93	149671	22.131	ng/u1	99
12) 2-Chlorophenol	7.275	128	151738	23.810	ng/u1	95
13) 2-Methylphenol	8.140	108	149656	22.821	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.210	45	167340	21.291	ng/u1	99
16) Acetophenone	8.504	105	256330	22.400	ng/u1	94
17) N-Nitroso-di-n-propyla...	8.493	70	126840	21.878	ng/u1	99
18) 4-Methylphenol	8.463	108	165445	22.741	ng/u1	93
19) Hexachloroethane	8.757	117	67965	22.899	ng/u1	96
22) Nitrobenzene	8.881	77	203139	21.706	ng/u1	97
23) Isophorone	9.398	82	367380	21.894	ng/u1	99
25) 2-Nitrophenol	9.581	139	92330	23.602	ng/u1	97
26) 2,4-Dimethylphenol	9.646	107	160390	18.863	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.875	93	207121	21.739	ng/u1	97
29) 2,4-Dichlorophenol	10.116	162	184697	24.311	ng/u1	100
30) Naphthalene	10.504	128	490771	22.597	ng/u1	98
32) 4-Chloroaniline	10.616	127	168069	18.677	ng/u1	98
33) Hexachlorobutadiene	10.787	225	143315	22.382	ng/u1	99
34) Caprolactam	11.393	113	55162	22.521	ng/u1	99
35) 4-Chloro-3-methylphenol	11.745	107	200908	24.237	ng/u1	99
36) 2-Methylnaphthalene	12.110	142	364277	22.849	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101424\
 Data File : BP022371.D
 Acq On : 15 Oct 2024 01:32
 Operator : RC/JU
 Sample : PB163830BS
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS830

Manual Integrations
 APPROVED


Reviewed By :Yogesh
 Patel

10/15/2024
 Supervised By :mohammad
 ahmed

Quant Time: Oct 15 02:04:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 14 11:01:55 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.334	142	365984	22.491	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.481	216	273978	22.404	ng/ul	96
40) Hexachlorocyclopentadiene	12.457	237	87835	11.789	ng/ul	97
41) 2,4,6-Trichlorophenol	12.728	196	175371	22.798	ng/ul	96
42) 2,4,5-Trichlorophenol	12.810	196	193472	23.621	ng/ul	99
43) 1,1'-Biphenyl	13.134	154	523326	21.952	ng/ul	100
44) 2-Chloronaphthalene	13.169	162	437960	22.442	ng/ul	97
45) 2-Nitroaniline	13.387	65	132670	22.768	ng/ul	95
47) Dimethylphthalate	13.763	163	567726	21.682	ng/ul	100
48) 2,6-Dinitrotoluene	13.881	165	119030	24.033	ng/ul	97
50) Acenaphthylene	14.022	152	654088	22.581	ng/ul	100
51) 3-Nitroaniline	14.216	138	112999	24.292	ng/ul	98
52) Acenaphthene	14.375	153	458658	22.392	ng/ul	98
53) 2,4-Dinitrophenol	14.439	184	73203	20.089	ng/ul	95
55) 4-Nitrophenol	14.539	109	115142	23.533	ng/ul	92
56) Dibenzofuran	14.716	168	690735	22.450	ng/ul	98
57) 2,4-Dinitrotoluene	14.675	165	185280	24.074	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.945	232	174472	22.821	ng/ul	98
59) Diethylphthalate	15.145	149	545389	21.710	ng/ul	98
61) Fluorene	15.369	166	566560	22.600	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.357	204	321019	22.392	ng/ul	97
63) 4-Nitroaniline	15.404	138	126972	27.126	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.475	198	125721	21.709	ng/ul	95
67) N-Nitrosodiphenylamine	15.581	169	494286	22.579	ng/ul	99
68) 4-Bromophenyl-phenylether	16.280	248	231872	23.521	ng/ul	98
69) Hexachlorobenzene	16.386	284	286028	23.571	ng/ul	97
70) Atrazine	16.563	200	161611	17.760	ng/ul	95
71) Pentachlorophenol	16.751	266	172119	22.060	ng/ul	97
72) Phenanthrene	17.145	178	1003657	22.948	ng/ul	99
74) Anthracene	17.239	178	996952	22.842	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.092	216	270758	21.753	ng/uL	99
76) Pentachlorobenzene	14.628	250	299963	21.800	ng/uL	99
77) Carbazole	17.527	167	908392	23.521	ng/ul	99
78) Di-n-butylphthalate	18.104	149	977088	22.865	ng/ul	100
80) Fluoranthene	19.233	202	1321442	25.438	ng/ul	100
82) Pyrene	19.616	202	1388478	24.938	ng/ul	98
83) Butylbenzylphthalate	20.563	149	477887	24.653	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.451	252	494063	24.256	ng/ul	98
85) Benzo(a)anthracene	21.533	228	1390147	23.207	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.457	149	652662	23.456	ng/ul	99
87) Chrysene	21.598	228	1299291	23.393	ng/ul	100
89) Di-n-octyl phthalate	22.704	149	1110371	23.604	ng/ul	100
90) Benzo(b)fluoranthene	23.792	252	1492522	23.976	ng/ul	99
91) Benzo(k)fluoranthene	23.857	252	1422402	23.338	ng/ul	99
93) Benzo(a)pyrene	24.680	252	1304610	22.771	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.568	276	1790958m	23.425	ng/ul	
95) Dibenzo(a,h)anthracene	28.639	278	1438457	23.234	ng/ul	99
96) Benzo(g,h,i)perylene	29.733	276	1377947	23.171	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.728 ng/uL
RT: 3.258 min Scan# 40
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion: 88 Resp: 22064
Ion Ratio Lower Upper
88 100
43 26.3 22.3 33.5
58 63.1 54.7 82.1

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#5
Pyridine
Concen: 19.716 ng/uL
RT: 3.652 min Scan# 113
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion: 79 Resp: 147352
Ion Ratio Lower Upper
79 100
52 68.9 53.9 80.9
51 34.3 26.3 39.5





#6
Benzaldehyde
Concen: 23.867 ng/ul
RT: 6.857 min Scan# 611
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830





Tgt Ion: 77 Resp: 11202
Ion Ratio Lower Upper
77 100
105 93.4 72.6 108.8
106 90.3 70.6 106.0

Manual Integrations
APPROVED


Reviewed By :Yogesh
Patel

10/15/2024
Supervised By :mohammad
ahmed



10/22/2024




#8
Phenol
Concen: 22.791 ng/ul
RT: 6.916 min Scan# 668
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32




Tgt Ion: 94 Resp: 205951
Ion Ratio Lower Upper
94 100
65 30.9 25.9 38.9
66 46.1 37.6 56.4





#10
Bis(2-Chloroethyl)ether
Concen: 22.131 ng/ul
RT: 7.128 min Scan# 704
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion: 93 Resp: 14967
Ion Ratio Lower Upper
93 100
63 72.9 59.0 88.6
95 31.8 25.6 38.4

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#12
2-Chlorophenol
Concen: 23.810 ng/ul
RT: 7.275 min Scan# 729
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion:128 Resp: 151738
Ion Ratio Lower Upper
128 100
64 48.5 42.4 63.6
130 32.0 24.8 37.2





#13
2-Methylphenol
Concen: 22.821 ng/ul
RT: 8.140 min Scan# 81
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:108 Resp: 149650
Ion Ratio Lower Upper
108 100
107 89.2 72.5 108.7

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.291 ng/ul
RT: 8.210 min Scan# 888
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion: 45 Resp: 167340
Ion Ratio Lower Upper
45 100
77 19.1 14.5 21.7
79 13.5 11.0 16.4





#16
Acetophenone
Concen: 22.400 ng/ul
RT: 8.504 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:105 Resp: 256330
Ion Ratio Lower Upper
105 100
77 81.1 69.0 103.4
51 24.8 22.6 33.8

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#17
N-Nitroso-di-n-propylamine
Concen: 21.878 ng/ul
RT: 8.493 min Scan# 936
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion: 70 Resp: 126840
Ion Ratio Lower Upper
70 100
42 52.7 42.0 63.0
101 11.3 7.8 11.6
130 20.2 15.0 22.6





#18
4-Methylphenol
Concen: 22.741 ng/ul
RT: 8.463 min Scan# 911
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830





Tgt Ion:108 Resp: 16544
Ion Ratio Lower Upper
108 100
107 110.0 94.4 141.6

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#19
Hexachloroethane
Concen: 22.899 ng/ul
RT: 8.757 min Scan# 981
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion:117 Resp: 67965
Ion Ratio Lower Upper
117 100
201 126.8 106.6 160.0
199 81.9 66.9 100.3





#22
Nitrobenzene
Concen: 21.706 ng/ul
RT: 8.881 min Scan# 1002
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion: 77 Resp: 203139
Ion Ratio Lower Upper
77 100
123 40.1 30.8 46.2
65 13.2 11.4 17.2

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#23
Isophorone
Concen: 21.894 ng/ul
RT: 9.398 min Scan# 1090
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Tgt Ion: 82 Resp: 367380
Ion Ratio Lower Upper
82 100
95 8.6 6.7 10.1
138 16.0 12.3 18.5





#25
2-Nitrophenol
Concen: 23.602 ng/ul
RT: 9.581 min Scan# 1121
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830





Tgt Ion:139 Resp: 92330
Ion Ratio Lower Upper
139 100
65 59.2 48.5 72.7
109 46.4 35.0 52.4

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#26
2,4-Dimethylphenol
Concen: 18.863 ng/ul
RT: 9.646 min Scan# 1132
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Tgt Ion:107 Resp: 160390
Ion Ratio Lower Upper
107 100
121 45.4 34.3 51.5
122 73.1 60.1 90.1





#27

Bis(2-Chloroethoxy)methane

Concen: 21.739 ng/ul

RT: 9.875 min Scan# 1170

Delta R.T. 0.006 min

Lab File: BP022371.D

Acq: 15 Oct 2024 01:32

Instrument :

BNA_P

ClientSampleId :

SLCS830

Tgt Ion: 93 Resp: 20712

Ion Ratio Lower Upper

93 100

95 33.0 25.4 38.0

123 11.3 7.8 11.8

Manual Integrations

APPROVED





Reviewed By :Yogesh

Patel

10/15/2024

Supervised By :mohammad
ahmed

10/22/2024



#29

2,4-Dichlorophenol

Concen: 24.311 ng/ul

RT: 10.116 min Scan# 1212

Delta R.T. 0.006 min

Lab File: BP022371.D

Acq: 15 Oct 2024 01:32




Tgt Ion:162 Resp: 184697


Ion Ratio Lower Upper

162 100

164 67.0 53.5 80.3

98 36.1 28.6 42.8





#30
Naphthalene
Concen: 22.597 ng/ul
RT: 10.504 min Scan# 1278
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830





Tgt Ion:128 Resp: 49077
Ion Ratio Lower Upper
128 100
129 11.3 8.6 12.8
127 13.2 9.9 14.9

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#32
4-Chloroaniline
Concen: 18.677 ng/ul
RT: 10.616 min Scan# 1297
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion:127 Resp: 168069
Ion Ratio Lower Upper
127 100
129 31.9 24.6 36.8





#33
Hexachlorobutadiene
Concen: 22.382 ng/uI
RT: 10.787 min Scan# 1326
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:225 Resp: 143315
Ion Ratio Lower Upper
225 100
223 61.9 50.2 75.4
227 64.9 51.7 77.5

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#34
Caprolactam
Concen: 22.521 ng/uI
RT: 11.393 min Scan# 1429
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion:113 Resp: 55162
Ion Ratio Lower Upper
113 100
55 163.7 130.2 195.2
56 132.0 104.0 156.0





#35
4-Chloro-3-methylphenol
Concen: 24.237 ng/ul
RT: 11.745 min Scan# 1489
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:107 Resp: 200908
Ion Ratio Lower Upper
107 100
144 24.3 18.5 27.7
142 72.9 57.6 86.4

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#36
2-Methylnaphthalene
Concen: 22.849 ng/ul
RT: 12.110 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion:142 Resp: 364277
Ion Ratio Lower Upper
142 100
141 90.0 69.8 104.6





#37
1-Methylnaphthalene
Concen: 22.491 ng/ul
RT: 12.334 min Scan# 1588
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:142 Resp: 365984
Ion Ratio Lower Upper
142 100
141 94.0 74.8 112.2
116 4.4 4.1 6.1

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.404 ng/ul
RT: 12.481 min Scan# 1614
Delta R.T. -0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Tgt Ion:216 Resp: 273978
Ion Ratio Lower Upper
216 100
214 78.6 65.9 98.9
179 20.9 17.8 26.6
108 14.2 12.2 18.4





#40

Hexachlorocyclopentadiene

Concen: 11.789 ng/ul

RT: 12.457 min Scan# 1610

Delta R.T. -0.006 min

Lab File: BP022371.D

Acq: 15 Oct 2024 01:32

Instrument :

BNA_P

ClientSampleId :

SLCS830

Tgt Ion:	237	Resp:	8783
Ion Ratio	Lower	Upper	
237	100		
235	64.9	49.8	74.6
272	15.3	11.4	17.0

Manual Integrations





APPROVED

Reviewed By :Yogesh
Patel

10/15/2024

Supervised By :mohammad
ahmed

10/22/2024



#41

2,4,6-Trichlorophenol

Concen: 22.798 ng/ul




RT: 12.728 min Scan# 1656


Delta R.T. -0.000 min

Lab File: BP022371.D

Acq: 15 Oct 2024 01:32


Tgt Ion:	196	Resp:	175371
Ion Ratio	Lower	Upper	
196	100		
198	99.0	75.8	113.8
200	33.4	25.6	38.4





#42
2,4,5-Trichlorophenol
Concen: 23.621 ng/ul
RT: 12.810 min Scan# 1669
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:196 Resp: 193472
Ion Ratio Lower Upper
196 100
198 96.9 78.0 117.0
200 32.4 27.0 40.4

Manual Integrations
APPROVED


Reviewed By :Yogesh
Patel

10/15/2024
Supervised By :mohammad
ahmed



10/22/2024




#43
1,1'-Biphenyl
Concen: 21.952 ng/ul
RT: 13.134 min Scan# 1725
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32




Tgt Ion:154 Resp: 523326
Ion Ratio Lower Upper
154 100
153 40.7 32.5 48.7
76 11.5 9.0 13.6





#44
2-Chloronaphthalene
Concen: 22.442 ng/ul
RT: 13.169 min Scan# 1732
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:162 Resp: 437960
Ion Ratio Lower Upper
162 100
164 33.3 26.2 39.4
127 38.7 28.8 43.2

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#45
2-Nitroaniline
Concen: 22.768 ng/ul
RT: 13.387 min Scan# 1768
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion: 65 Resp: 132670
Ion Ratio Lower Upper
65 100
92 70.8 55.4 83.2
138 96.2 71.4 107.0





#47
Dimethylphthalate
Concen: 21.682 ng/ul
RT: 13.763 min Scan# 1831
Delta R.T. 0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:163 Resp: 567720
Ion Ratio Lower Upper
163 100
194 4.4 3.5 5.3
164 10.2 8.2 12.4

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#48
2,6-Dinitrotoluene
Concen: 24.033 ng/ul
RT: 13.881 min Scan# 1852
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Tgt Ion:165 Resp: 119030
Ion Ratio Lower Upper
165 100
89 61.0 50.2 75.4
121 19.9 17.1 25.7





#50

Acenaphthylene

Concen: 22.581 ng/ul

RT: 14.022 min Scan# 1877

Delta R.T. -0.000 min

Lab File: BP022371.D

Acq: 15 Oct 2024 01:32

Instrument :

BNA_P

ClientSampleId :

SLCS830

Tgt Ion:152	Resp: 65408
Ion Ratio	Lower Upper
152 100	
151 19.9	15.8 23.6
153 13.1	10.3 15.5





Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel

10/15/2024

Supervised By :mohammad
ahmed

10/22/2024



#51

3-Nitroaniline

Concen: 24.292 ng/ul




RT: 14.216 min Scan# 1909


Delta R.T. -0.006 min

Lab File: BP022371.D

Acq: 15 Oct 2024 01:32


Tgt Ion:138	Resp: 112999
Ion Ratio	Lower Upper
138 100	
108 12.5	9.7 14.5
92 131.1	106.7 160.1





#52
Acenaphthene
Concen: 22.392 ng/ul
RT: 14.375 min Scan# 1936
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:153 Resp: 458658
Ion Ratio Lower Upper
153 100
152 46.9 37.8 56.6
154 89.0 69.5 104.3

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#53
2,4-Dinitrophenol
Concen: 20.089 ng/ul
RT: 14.439 min Scan# 1947
Delta R.T. 0.012 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion:184 Resp: 73203
Ion Ratio Lower Upper
184 100
63 53.6 46.5 69.7
154 70.0 53.3 79.9





#55
4-Nitrophenol
Concen: 23.533 ng/ul
RT: 14.539 min Scan# 1964
Delta R.T. -0.018 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion	Ratio	Lower	Upper
109	100		
139	86.6	76.4	114.6
65	102.2	87.4	131.0

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#56
Dibenzofuran
Concen: 22.450 ng/ul
RT: 14.716 min Scan# 1994
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Tgt Ion	Ratio	Lower	Upper
168	100		
139	36.3	30.2	45.2





#57
2,4-Dinitrotoluene
Concen: 24.074 ng/ul
RT: 14.675 min Scan# 1989
Delta R.T. -0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830





Tgt Ion:165 Resp: 185280
Ion Ratio Lower Upper
165 100
63 41.6 35.3 52.9
182 2.9 2.6 4.0

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.821 ng/ul
RT: 14.945 min Scan# 2033
Delta R.T. -0.012 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Tgt Ion:232 Resp: 174472
Ion Ratio Lower Upper
232 100
131 37.8 29.0 43.4
130 2.1 1.9 2.9
166 28.3 22.6 34.0





#59

Diethylphthalate

Concen: 21.710 ng/ul

RT: 15.145 min Scan# 2067

Delta R.T. -0.006 min

Lab File: BP022371.D

Acq: 15 Oct 2024 01:32

Instrument :

BNA_P

ClientSampleId :

SLCS830

Tgt Ion:	149	Resp:	545389
Ion Ratio	Lower	Upper	
149	100		
177	23.7	19.9	29.9
150	12.9	10.2	15.4

Manual Integrations





APPROVED

Reviewed By :Yogesh
Patel

10/15/2024

Supervised By :mohammad
ahmed

10/22/2024



#61

Fluorene

Concen: 22.600 ng/ul




RT: 15.369 min Scan# 2105


Delta R.T. -0.006 min

Lab File: BP022371.D

Acq: 15 Oct 2024 01:32


Tgt Ion:	166	Resp:	566560
Ion Ratio	Lower	Upper	
166	100		
165	100.5	77.8	116.6
167	14.4	11.1	16.7





#62
4-Chlorophenyl-phenylether
Concen: 22.392 ng/ul
RT: 15.357 min Scan# 2103
Delta R.T. -0.012 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:204 Resp: 321019
Ion Ratio Lower Upper
204 100
206 32.4 26.4 39.6
141 56.0 42.6 63.8

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#63
4-Nitroaniline
Concen: 27.126 ng/ul
RT: 15.404 min Scan# 2111
Delta R.T. 0.012 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Tgt Ion:138 Resp: 126972
Ion Ratio Lower Upper
138 100
92 57.5 49.1 73.7
108 117.4 90.6 136.0





#66
4,6-Dinitro-2-methylphenol
Concen: 21.709 ng/ul
RT: 15.475 min Scan# 2122
Delta R.T. 0.012 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830





Tgt Ion:198 Resp: 12572
Ion Ratio Lower Upper
198 100
182 3.8 3.4 5.2
77 21.7 19.4 29.0

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#67
N-Nitrosodiphenylamine
Concen: 22.579 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.018 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion:169 Resp: 494286
Ion Ratio Lower Upper
169 100
168 66.1 53.4 80.0
167 36.2 29.0 43.6





#68
4-Bromophenyl-phenylether
Concen: 23.521 ng/ul
RT: 16.280 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:248 Resp: 231872
Ion Ratio Lower Upper
248 100
250 97.8 76.0 114.0
141 60.3 49.0 73.6

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#69
Hexachlorobenzene
Concen: 23.571 ng/ul
RT: 16.386 min Scan# 2278
Delta R.T. -0.024 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Tgt Ion:284 Resp: 286028
Ion Ratio Lower Upper
284 100
142 28.3 21.8 32.8
249 32.6 24.8 37.2







#70
 Atrazine
 Concen: 17.760 ng/ul
 RT: 16.563 min Scan# 21
 Delta R.T. -0.018 min
 Lab File: BP022371.D
 Acq: 15 Oct 2024 01:32



Instrument : BNA_P
 ClientSampleId : SLCS830

Tgt Ion:	200	Resp:	16161
Ion Ratio	Lower	Upper	
200	100		
173	23.4	18.0	27.0
215	47.5	35.0	52.4

Manual Integrations
APPROVED
 Reviewed By :Yogesh Patel
 10/15/2024







Supervised By :mohammad ahmed
 10/22/2024



#71
 Pentachlorophenol
 Concen: 22.060 ng/ul
 RT: 16.751 min Scan# 2340
 Delta R.T. -0.006 min
 Lab File: BP022371.D
 Acq: 15 Oct 2024 01:32

Tgt Ion:	266	Resp:	172119
Ion Ratio	Lower	Upper	
266	100		
264	62.9	52.4	78.6
268	64.8	53.0	79.4





#72
 Phenanthrene
 Concen: 22.948 ng/ul
 RT: 17.145 min Scan# 2407
 Delta R.T. -0.006 min
 Lab File: BP022371.D
 Acq: 15 Oct 2024 01:32

Instrument :

BNA_P

ClientSampleId :

SLCS830





Tgt Ion:178 Resp: 100365
 Ion Ratio Lower Upper
 178 100
 179 14.8 12.2 18.4
 176 19.0 15.7 23.5

Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel




10/15/2024
 Supervised By :mohammad
 ahmed


10/22/2024



#74
 Anthracene
 Concen: 22.842 ng/ul
 RT: 17.239 min Scan# 2423
 Delta R.T. -0.006 min
 Lab File: BP022371.D
 Acq: 15 Oct 2024 01:32


Tgt Ion:178 Resp: 996952
 Ion Ratio Lower Upper
 178 100
 179 15.4 12.0 18.0
 176 18.0 15.0 22.4





#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.753 ng/uL
RT: 13.092 min Scan# 1718
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:216 Resp: 270758
Ion Ratio Lower Upper
216 100
214 80.7 64.4 96.6
218 48.4 37.6 56.4

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#76
Pentachlorobenzene
Concen: 21.800 ng/uL
RT: 14.628 min Scan# 1979
Delta R.T. -0.012 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion:250 Resp: 299963
Ion Ratio Lower Upper
250 100
248 63.0 51.1 76.7
252 64.4 51.7 77.5





#77
 Carbazole
 Concen: 23.521 ng/ul
 RT: 17.527 min Scan# 2473
 Delta R.T. -0.000 min
 Lab File: BP022371.D
 Acq: 15 Oct 2024 01:32

Instrument :
 BNA_P
 ClientSampleId :
 SLCS830






Tgt Ion:167 Resp: 908392
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.0 25.6
 139 12.4 10.3 15.5

Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel




10/15/2024
 Supervised By :mohammad
 ahmed


10/22/2024



#78
 Di-n-butylphthalate
 Concen: 22.865 ng/ul
 RT: 18.104 min Scan# 2570
 Delta R.T. -0.012 min
 Lab File: BP022371.D
 Acq: 15 Oct 2024 01:32


Tgt Ion:149 Resp: 977088
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.3 10.9
 104 5.9 4.8 7.2





#80
Fluoranthene
Concen: 25.438 ng/u1
RT: 19.233 min Scan# 2762
Delta R.T. -0.018 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:202 Resp: 132144
Ion Ratio Lower Upper
202 100
101 7.2 5.8 8.8
100 5.5 4.5 6.7

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#82
Pyrene
Concen: 24.938 ng/u1
RT: 19.616 min Scan# 2827
Delta R.T. -0.012 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Tgt Ion:202 Resp: 1388478
Ion Ratio Lower Upper
202 100
101 7.5 6.7 10.1
100 6.5 5.8 8.6





#83

Butylbenzylphthalate

Concen: 24.653 ng/ul

RT: 20.563 min Scan# 2989

Delta R.T. -0.012 min

Lab File: BP022371.D

Acq: 15 Oct 2024 01:32

Instrument :

BNA_P

ClientSampleId :

SLCS830

Tgt Ion:	149	Resp:	47788
Ion Ratio	Lower	Upper	
149	100		
91	72.2	56.6	84.8
206	24.6	18.6	27.8





Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel

10/15/2024

Supervised By :mohammad
ahmed

10/22/2024



#84

3,3'-Dichlorobenzidine

Concen: 24.256 ng/ul




RT: 21.451 min Scan# 3139


Delta R.T. -0.006 min

Lab File: BP022371.D

Acq: 15 Oct 2024 01:32


Tgt Ion:	252	Resp:	494063
Ion Ratio	Lower	Upper	
252	100		
254	64.5	53.0	79.4
126	7.2	5.8	8.6





#85
Benzo(a)anthracene
Concen: 23.207 ng/ul
RT: 21.533 min Scan# 3153
Delta R.T. -0.000 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:228 Resp: 139014
Ion Ratio Lower Upper
228 100
229 19.9 15.4 23.0
226 27.2 21.5 32.3

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#86
Bis(2-ethylhexyl)phthalate
Concen: 23.456 ng/ul
RT: 21.457 min Scan# 3140
Delta R.T. -0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion:149 Resp: 652662
Ion Ratio Lower Upper
149 100
167 29.5 23.0 34.4
279 5.7 4.8 7.2





#87
Chrysene
Concen: 23.393 ng/ul
RT: 21.598 min Scan# 3164
Delta R.T. -0.006 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830





Tgt Ion:228 Resp: 129929
Ion Ratio Lower Upper
228 100
226 29.9 23.8 35.8
229 19.7 15.5 23.3

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#89
Di-n-octyl phthalate
Concen: 23.604 ng/ul
RT: 22.704 min Scan# 3352
Delta R.T. -0.018 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion:149 Resp: 1110371





#90
Benzo(b)fluoranthene
Concen: 23.976 ng/u1
RT: 23.792 min Scan# 3537
Delta R.T. -0.024 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:252 Resp: 149252
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.8
125 5.0 5.0 7.4

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#91
Benzo(k)fluoranthene
Concen: 23.338 ng/u1
RT: 23.857 min Scan# 3548
Delta R.T. -0.024 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion:252 Resp: 1422402
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.0
125 5.0 4.6 6.8





#93
Benzo(a)pyrene
Concen: 22.771 ng/ul
RT: 24.680 min Scan# 30
Delta R.T. -0.024 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32

Instrument :
BNA_P
ClientSampleId :
SLCS830






Tgt Ion:252 Resp: 1304610
Ion Ratio Lower Upper
252 100
253 22.0 17.7 26.5
125 5.3 4.9 7.3

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/15/2024
Supervised By :mohammad
ahmed


10/22/2024



#94
Indeno(1,2,3-cd)pyrene
Concen: 23.425 ng/ul m
RT: 28.568 min Scan# 4349
Delta R.T. -0.041 min
Lab File: BP022371.D
Acq: 15 Oct 2024 01:32


Tgt Ion:276 Resp: 1790958
Ion Ratio Lower Upper
276 100
138 9.5 9.1 13.7
277 23.8 18.6 27.8





#95
 Dibenzo(a,h)anthracene
 Concen: 23.234 ng/ul
 RT: 28.639 min Scan# 41
 Delta R.T. -0.029 min
 Lab File: BP022371.D
 Acq: 15 Oct 2024 01:32

Instrument :
 BNA_P
 ClientSampleId :
 SLCS830






Tgt Ion:278 Resp: 143845
 Ion Ratio Lower Upper
 278 100
 139 7.9 7.3 10.9
 279 24.0 19.0 28.4

Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel

10/15/2024
 Supervised By :mohammad
 ahmed

10/22/2024



#96
 Benzo(g,h,i)perylene
 Concen: 23.171 ng/ul
 RT: 29.733 min Scan# 4547
 Delta R.T. -0.024 min
 Lab File: BP022371.D
 Acq: 15 Oct 2024 01:32

Tgt Ion:276 Resp: 1377947
 Ion Ratio Lower Upper
 276 100
 138 9.9 8.5 12.7
 277 23.3 19.0 28.4

