

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101424\
 Data File : BP022397.D
 Acq On : 15 Oct 2024 20:21
 Operator : RC/JU
 Sample : P4349-16
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EOAG0

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/16/2024
 Supervised By :mohammad ahmed 10/22/2024

Quant Time: Oct 16 02:38:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 14 11:01:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.693	152	77357	20.000	ng/u1	0.00
20) Naphthalene-d8	10.452	136	309475	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.310	164	248872	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.116	188	630446	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.539	240	683283	20.000	ng/u1	#-0.01
88) Perylene-d12	24.827	264	826608	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.217	96	8347	4.193	ng/uL	0.00
4) Pyridine-d5	3.634	84	40331	7.380	ng/u1	0.00
7) Phenol-d5	6.881	99	49143	7.547	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.034	67	109151	28.520	ng/u1	0.00
11) 2-Chlorophenol-d4	7.234	132	124869	27.025	ng/u1	0.00
15) 4-Methylphenol-d8	8.393	113	95326	18.344	ng/u1	0.00
21) Nitrobenzene-d5	8.834	128	72609	30.514	ng/u1	0.00
24) 2-Nitrophenol-d4	9.546	143	85170	30.916	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.081	165	174185	29.749	ng/u1	0.00
31) 4-Chloroaniline-d4	10.587	131	152321	22.016	ng/u1	0.00
46) Dimethylphthalate-d6	13.710	166	614099	31.057	ng/u1	0.00
49) Acenaphthylene-d8	13.998	160	667869	31.801	ng/u1	0.00
54) 4-Nitrophenol-d4	14.534	143	23876m	7.198	ng/u1	0.00
60) Fluorene-d10	15.316	176	561484	32.738	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.457	200	116201	28.025	ng/u1	0.01
73) Anthracene-d10	17.216	188	992771	33.474	ng/u1	0.01
81) Pyrene-d10	19.598	212	1260143	34.875	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.610	264	1483193	33.598	ng/u1	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101424\
Data File : BP022397.D
Acq On : 15 Oct 2024 20:21
Operator : RC/JU
Sample : P4349-16
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
EOAG0

Quant Time: Oct 16 02:38:56 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 14 11:01:55 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/16/2024
Supervised By :mohammad ahmed 10/22/2024

