

Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP101519\
 Data File : BP000732.D
 Acq On : 15 Oct 2019 00:32
 Operator : JU
 Sample : PB123576BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK76

Quant Time: Oct 15 03:56:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP100919MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 15 03:51:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.75	152	267145	20.00	ng/ul	0.00
18) Naphthalene-d8	8.03	136	1008244	20.00	ng/ul	0.00
36) Acenaphthene-d10	9.80	164	561082	20.00	ng/ul	0.00
62) Phenanthrene-d10	11.29	188	1002695	20.00	ng/ul	0.00
78) Chrysene-d12	13.97	240	822355	20.00	ng/ul	0.00
86) Perylene-d12	15.45	264	876546	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.36	96	35722	5.55	ng/uL	0.00
5) Phenol-d5	6.39	99	563775	27.36	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.43	67	281579	27.33	ng/ul	0.00
9) 2-Chlorophenol-d4	6.52	132	471853	27.18	ng/ul	0.00
13) 4-Methylphenol-d8	7.13	113	431307	27.40	ng/ul	0.00
19) Nitrobenzene-d5	7.31	128	221001	28.84	ng/ul	0.00
22) 2-Nitrophenol-d4	7.63	143	239528	29.35	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	7.88	165	412247	26.39	ng/ul	0.00
29) 4-Chloroaniline-d4	8.09	131	587924	33.84	ng/ul	0.00
44) Dimethylphthalate-d6	9.49	166	1079614	27.49	ng/ul	0.00
47) Acenaphthylene-d8	9.64	160	1322136	27.36	ng/ul	0.00
52) 4-Nitrophenol-d4	9.91	143	156391	25.85	ng/ul	0.00
58) Fluorene-d10	10.32	176	905165	26.93	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	10.39	200	97606	20.44	ng/ul	0.00
71) Anthracene-d10	11.35	188	1284361	27.65	ng/ul	0.00
79) Pyrene-d10	12.73	212	1313893	28.25	ng/ul	0.00
90) Benzo(a)pyrene-d12	15.36	264	1199129	27.71	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP101519\
Data File : BP000732.D
Acq On : 15 Oct 2019 00:32
Operator : JU
Sample : PB123576BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK76

Quant Time: Oct 15 03:56:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP100919MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 15 03:51:25 2019
Response via : Initial Calibration

