

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101722\
 Data File : BP012167.D
 Acq On : 17 Oct 2022 16:03
 Operator : CG/JU
 Sample : SST02076
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020640

Quant Time: Oct 18 01:14:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 18 01:11:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	318817	20.000	ng/ul	0.00
20) Naphthalene-d8	10.640	136	1323494	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.487	164	800492	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.245	188	1630688	20.000	ng/ul	0.00
79) Chrysene-d12	21.339	240	1616500	20.000	ng/ul	0.00
88) Perylene-d12	23.745	264	1645442	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	69923	8.647	ng/uL	0.00
4) Pyridine-d5	3.687	84	469321	21.824	ng/ul	0.00
7) Phenol-d5	7.005	99	542128	21.143	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.169	67	332938	22.876	ng/ul	0.00
11) 2-Chlorophenol-d4	7.363	132	419348	20.603	ng/ul	0.00
15) 4-Methylphenol-d8	8.552	113	409090	20.771	ng/ul	0.00
21) Nitrobenzene-d5	9.005	128	190347	20.364	ng/ul	0.00
24) 2-Nitrophenol-d4	9.728	143	196033	20.683	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.263	165	397350	21.632	ng/ul	0.00
31) 4-Chloroaniline-d4	10.787	131	578453	21.370	ng/ul	0.00
46) Dimethylphthalate-d6	13.898	166	1120334	22.099	ng/ul	0.00
49) Acenaphthylene-d8	14.175	160	1322413	21.094	ng/ul	0.00
54) 4-Nitrophenol-d4	14.698	143	203532	20.450	ng/ul	0.00
60) Fluorene-d10	15.481	176	1004682	21.904	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.610	200	168182	19.532	ng/ul	0.00
73) Anthracene-d10	17.345	188	1512575	21.361	ng/ul	0.00
81) Pyrene-d10	19.580	212	1703861	20.574	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.592	264	1691536	21.147	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	73506	8.638	ng/uL	100
5) Pyridine	3.711	79	461353	22.391	ng/ul	100
6) Benzaldehyde	6.981	77	300758	25.000	ng/ul	100
8) Phenol	7.028	94	572978	21.286	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.263	93	465988	21.941	ng/ul	100
12) 2-Chlorophenol	7.399	128	455458	20.619	ng/ul	100
13) 2-Methylphenol	8.287	108	419988	20.919	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.369	45	621385	28.544	ng/ul	100
16) Acetophenone	8.663	105	676833	22.291	ng/ul	100
17) N-Nitroso-di-n-propyla...	8.652	70	323890	23.420	ng/ul	100
18) 4-Methylphenol	8.610	108	461887	21.344	ng/ul	100
19) Hexachloroethane	8.910	117	182272	21.377	ng/ul	100
22) Nitrobenzene	9.046	77	492922	23.005	ng/ul	100
23) Isophorone	9.575	82	905544	23.826	ng/ul	100
25) 2-Nitrophenol	9.757	139	234437	21.054	ng/ul	100
26) 2,4-Dimethylphenol	9.816	107	484174	22.326	ng/ul	100
27) Bis(2-Chloroethoxy)met...	10.057	93	606945	23.271	ng/ul	100
29) 2,4-Dichlorophenol	10.287	162	417939	21.651	ng/ul	100
30) Naphthalene	10.693	128	1467775	20.987	ng/ul	100
32) 4-Chloroaniline	10.810	127	599095	21.467	ng/ul	100
33) Hexachlorobutadiene	10.969	225	259399	21.671	ng/ul	100
34) Caprolactam	11.610	113	114966	21.674	ng/ul	100
35) 4-Chloro-3-methylphenol	11.934	107	434735	23.772	ng/ul	100
36) 2-Methylnaphthalene	12.304	142	970490	21.550	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101722\
 Data File : BP012167.D
 Acq On : 17 Oct 2022 16:03
 Operator : CG/JU
 Sample : SSTD02076
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020640

Quant Time: Oct 18 01:14:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 18 01:11:56 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.522	142	977598	21.676	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.669	216	501782	20.506	ng/ul	100
40) Hexachlorocyclopentadiene	12.646	237	214103	18.014	ng/ul	100
41) 2,4,6-Trichlorophenol	12.916	196	317312	21.527	ng/ul	100
42) 2,4,5-Trichlorophenol	12.987	196	344165	21.484	ng/ul	100
43) 1,1'-Biphenyl	13.316	154	1245682	20.613	ng/ul	100
44) 2-Chloronaphthalene	13.357	162	989290	20.500	ng/ul	100
45) 2-Nitroaniline	13.575	65	243294	23.634	ng/ul	100
47) Dimethylphthalate	13.945	163	1191277	22.303	ng/ul	100
48) 2,6-Dinitrotoluene	14.069	165	208155	22.012	ng/ul	100
50) Acenaphthylene	14.204	152	1499212	21.135	ng/ul	100
51) 3-Nitroaniline	14.398	138	227007	20.615	ng/ul	100
52) Acenaphthene	14.551	153	1047648	20.747	ng/ul	100
53) 2,4-Dinitrophenol	14.610	184	116637	19.092	ng/ul	100
55) 4-Nitrophenol	14.710	109	152558	22.171	ng/ul	100
56) Dibenzofuran	14.887	168	1450488	21.254	ng/ul	100
57) 2,4-Dinitrotoluene	14.857	165	315534	23.217	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.110	232	291507	23.346	ng/ul	100
59) Diethylphthalate	15.310	149	1154285	23.597	ng/ul	100
61) Fluorene	15.534	166	1167247	21.789	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.534	204	568729	21.945	ng/ul	100
63) 4-Nitroaniline	15.569	138	219167	21.798	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.622	198	187102	20.100	ng/ul	100
67) N-Nitrosodiphenylamine	15.745	169	995553	20.626	ng/ul	100
68) 4-Bromophenyl-phenylether	16.434	248	339621	21.403	ng/ul	100
69) Hexachlorobenzene	16.539	284	405981	22.043	ng/ul	100
70) Atrazine	16.710	200	327222	22.046	ng/ul	100
71) Pentachlorophenol	16.892	266	231028	21.410	ng/ul	100
72) Phenanthrene	17.286	178	1878036	21.177	ng/ul	100
74) Anthracene	17.381	178	1881951	21.445	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.281	216	501733	19.041	ng/ul	100
76) Pentachlorobenzene	14.804	250	527223	20.373	ng/ul	100
77) Carbazole	17.651	167	1699827	21.717	ng/ul	100
78) Di-n-butylphthalate	18.204	149	1851166	23.833	ng/ul	100
80) Fluoranthene	19.257	202	2115494	20.847	ng/ul	100
82) Pyrene	19.610	202	2232223	20.786	ng/ul	100
83) Butylbenzylphthalate	20.486	149	806212	23.683	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.257	252	747965	20.998	ng/ul	100
85) Benzo(a)anthracene	21.322	228	2141531	21.047	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.239	149	1157506	25.222	ng/ul	100
87) Chrysene	21.374	228	2057688	21.009	ng/ul	100
89) Di-n-octyl phthalate	22.169	149	1970691	23.922	ng/ul	100
90) Benzo(b)fluoranthene	23.010	252	2207346	21.193	ng/ul	100
91) Benzo(k)fluoranthene	23.057	252	2194127	21.417	ng/ul	100
93) Benzo(a)pyrene	23.639	252	1958113	21.073	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.257	276	2391645	20.356	ng/ul	100
95) Dibenzo(a,h)anthracene	26.268	278	2183506	20.715	ng/ul	100
96) Benzo(g,h,i)perylene	27.021	276	2149998	20.296	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101722\
 Data File : BP012167.D
 Acq On : 17 Oct 2022 16:03
 Operator : CG/JU
 Sample : SSTD02076
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020640

Quant Time: Oct 18 01:14:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 18 01:11:56 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.638 ng/uL
RT: 3.299 min Scan# 51
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion: 88 Resp: 73506
Ion Ratio Lower Upper
88 100
43 24.5 19.6 29.4
58 73.9 59.1 88.7



#5
Pyridine
Concen: 22.391 ng/uL
RT: 3.711 min Scan# 123
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion: 79 Resp: 461353
Ion Ratio Lower Upper
79 100
52 72.4 57.9 86.9
51 33.4 26.7 40.1









#6
Benzaldehyde
Concen: 25.000 ng/ul
RT: 6.981 min Scan# 679
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion: 77 Resp: 300758
Ion Ratio Lower Upper
77 100
105 103.6 82.9 124.3
106 101.0 80.8 121.2



#8
Phenol
Concen: 21.286 ng/ul
RT: 7.028 min Scan# 687
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion: 94 Resp: 572978
Ion Ratio Lower Upper
94 100
65 27.4 21.9 32.9
66 33.3 26.6 40.0









#10
Bis(2-Chloroethyl)ether
Concen: 21.941 ng/ul
RT: 7.263 min Scan# 71
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion: 93 Resp: 465988
Ion Ratio Lower Upper
93 100
63 77.4 61.9 92.9
95 33.7 27.0 40.4



#12
2-Chlorophenol
Concen: 20.619 ng/ul
RT: 7.399 min Scan# 750
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:128 Resp: 455458
Ion Ratio Lower Upper
128 100
64 46.5 37.2 55.8
130 32.3 25.8 38.8





#13

2-Methylphenol

Concen: 20.919 ng/ul

RT: 8.287 min Scan# 901

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Instrument :

BNA_P

ClientSampleId :





SSTD020640

Tgt Ion:108 Resp: 419988

Ion Ratio Lower Upper

108 100

107 87.7 70.2 105.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 28.544 ng/ul

RT: 8.369 min Scan# 915

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03




Tgt Ion: 45 Resp: 621385


Ion Ratio Lower Upper

45 100

77 15.6 12.5 18.7

79 11.8 9.4 14.2









#16
Acetophenone
Concen: 22.291 ng/ul
RT: 8.663 min Scan# 965
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:105 Resp: 676833
Ion Ratio Lower Upper
105 100
77 74.7 59.8 89.6
51 25.5 20.4 30.6



#17
N-Nitroso-di-n-propylamine
Concen: 23.420 ng/ul
RT: 8.652 min Scan# 963
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion: 70 Resp: 323890
Ion Ratio Lower Upper
70 100
42 53.3 42.6 64.0
101 9.7 7.8 11.6
130 21.6 17.3 25.9









#18
4-Methylphenol
Concen: 21.344 ng/ul
RT: 8.610 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:108 Resp: 461887
Ion Ratio Lower Upper
108 100
107 111.1 88.9 133.3



#19
Hexachloroethane
Concen: 21.377 ng/ul
RT: 8.910 min Scan# 1007
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:117 Resp: 182272
Ion Ratio Lower Upper
117 100
201 92.4 73.9 110.9
199 58.8 47.0 70.6









#22
Nitrobenzene
Concen: 23.005 ng/ul
RT: 9.046 min Scan# 1030
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion: 77 Resp: 492922
Ion Ratio Lower Upper
77 100
123 46.0 36.8 55.2
65 12.5 10.0 15.0



#23
Isophorone
Concen: 23.826 ng/ul
RT: 9.575 min Scan# 1120
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion: 82 Resp: 905544
Ion Ratio Lower Upper
82 100
95 6.6 5.3 7.9
138 17.8 14.2 21.4









#25
2-Nitrophenol
Concen: 21.054 ng/ul
RT: 9.757 min Scan# 1151
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:139 Resp: 234437
Ion Ratio Lower Upper
139 100
65 50.5 40.4 60.6
109 27.7 22.2 33.2



#26
2,4-Dimethylphenol
Concen: 22.326 ng/ul
RT: 9.816 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:107 Resp: 484174
Ion Ratio Lower Upper
107 100
121 51.5 41.2 61.8
122 88.1 70.5 105.7





#27

Bis(2-Chloroethoxy)methane

Concen: 23.271 ng/ul

RT: 10.057 min Scan# 1202

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion: 93 Resp: 606945

Ion Ratio Lower Upper

93 100

95 32.7 26.2 39.2

123 10.6 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 21.651 ng/ul

RT: 10.287 min Scan# 1241

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03




Tgt Ion:162 Resp: 417939


Ion Ratio Lower Upper

162 100

164 65.1 52.1 78.1

98 36.7 29.4 44.0









#30
Naphthalene
Concen: 20.987 ng/ul
RT: 10.693 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:128 Resp: 1467775
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 12.9 10.3 15.5



#32
4-Chloroaniline
Concen: 21.467 ng/ul
RT: 10.810 min Scan# 1330
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:127 Resp: 599095
Ion Ratio Lower Upper
127 100
129 32.3 25.8 38.8









#33
Hexachlorobutadiene
Concen: 21.671 ng/ul
RT: 10.969 min Scan# 1357
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:225 Resp: 259399
Ion Ratio Lower Upper
225 100
223 63.0 50.4 75.6
227 64.9 51.9 77.9



#34
Caprolactam
Concen: 21.674 ng/ul
RT: 11.610 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:113 Resp: 114966
Ion Ratio Lower Upper
113 100
55 176.4 141.1 211.7
56 122.9 98.3 147.5









#35
4-Chloro-3-methylphenol
Concen: 23.772 ng/ul
RT: 11.934 min Scan# 1521
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:107 Resp: 434735
Ion Ratio Lower Upper
107 100
144 26.2 21.0 31.4
142 81.4 65.1 97.7



#36
2-Methylnaphthalene
Concen: 21.550 ng/ul
RT: 12.304 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:142 Resp: 970490
Ion Ratio Lower Upper
142 100
141 89.1 71.3 106.9





#37

1-Methylnaphthalene

Concen: 21.676 ng/ul

RT: 12.522 min Scan# 1621

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:142 Resp: 977598

Ion Ratio Lower Upper

142 100

141 91.4 73.1 109.7

116 3.5 2.8 4.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.506 ng/ul

RT: 12.669 min Scan# 1646

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Tgt Ion:216 Resp: 501782




Ion Ratio Lower Upper


216 100

214 78.1 62.5 93.7

179 21.0 16.8 25.2

108 17.6 14.1 21.1









#40
Hexachlorocyclopentadiene
Concen: 18.014 ng/ul
RT: 12.646 min Scan# 1642
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:237 Resp: 214103
Ion Ratio Lower Upper
237 100
235 61.7 49.4 74.0
272 13.9 11.1 16.7



#41
2,4,6-Trichlorophenol
Concen: 21.527 ng/ul
RT: 12.916 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:196 Resp: 317312
Ion Ratio Lower Upper
196 100
198 97.0 77.6 116.4
200 29.8 23.8 35.8





#42

2,4,5-Trichlorophenol

Concen: 21.484 ng/ul

RT: 12.987 min Scan# 1700

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:196 Resp: 344165

Ion Ratio Lower Upper

196 100

198 95.9 76.7 115.1

200 30.5 24.4 36.6



#43

1,1'-Biphenyl

Concen: 20.613 ng/ul

RT: 13.316 min Scan# 1756

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03




Tgt Ion:154 Resp: 1245682


Ion Ratio Lower Upper

154 100

153 40.4 32.3 48.5

76 13.8 11.0 16.6









#44
2-Chloronaphthalene
Concen: 20.500 ng/ul
RT: 13.357 min Scan# 1763
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:162 Resp: 989290
Ion Ratio Lower Upper
162 100
164 33.0 26.4 39.6
127 38.0 30.4 45.6



#45
2-Nitroaniline
Concen: 23.634 ng/ul
RT: 13.575 min Scan# 1800
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion: 65 Resp: 243294
Ion Ratio Lower Upper
65 100
92 71.4 57.1 85.7
138 113.1 90.5 135.7









#47
Dimethylphthalate
Concen: 22.303 ng/ul
RT: 13.945 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:163 Resp: 1191277
Ion Ratio Lower Upper
163 100
194 3.8 3.0 4.6
164 10.2 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 22.012 ng/ul
RT: 14.069 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:165 Resp: 208155
Ion Ratio Lower Upper
165 100
89 60.3 48.2 72.4
121 23.7 19.0 28.4









#50
Acenaphthylene
Concen: 21.135 ng/ul
RT: 14.204 min Scan# 1907
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:152 Resp: 1499212
Ion Ratio Lower Upper
152 100
151 19.9 15.9 23.9
153 13.3 10.6 16.0



#51
3-Nitroaniline
Concen: 20.615 ng/ul
RT: 14.398 min Scan# 1940
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:138 Resp: 227007
Ion Ratio Lower Upper
138 100
108 11.8 9.4 14.2
92 120.7 96.6 144.8









#52
Acenaphthene
Concen: 20.747 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:153 Resp: 1047648
Ion Ratio Lower Upper
153 100
152 48.5 38.8 58.2
154 89.1 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 19.092 ng/ul
RT: 14.610 min Scan# 1976
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:184 Resp: 116637
Ion Ratio Lower Upper
184 100
63 63.1 50.5 75.7
154 57.3 45.8 68.8









#55
4-Nitrophenol
Concen: 22.171 ng/ul
RT: 14.710 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:109 Resp: 152558
Ion Ratio Lower Upper
109 100
139 158.2 126.6 189.8
65 146.4 117.1 175.7



#56
Dibenzofuran
Concen: 21.254 ng/ul
RT: 14.887 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:168 Resp: 1450488
Ion Ratio Lower Upper
168 100
139 35.7 28.6 42.8









#57
2,4-Dinitrotoluene
Concen: 23.217 ng/ul
RT: 14.857 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:165 Resp: 315534
Ion Ratio Lower Upper
165 100
63 44.4 35.5 53.3
182 2.4 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.346 ng/ul
RT: 15.110 min Scan# 2061
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:232 Resp: 291507
Ion Ratio Lower Upper
232 100
131 46.6 37.3 55.9
130 2.2 1.8 2.6
166 30.9 24.7 37.1









#59
Diethylphthalate
Concen: 23.597 ng/ul
RT: 15.310 min Scan# 2095
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:149 Resp: 1154285
Ion Ratio Lower Upper
149 100
177 22.1 17.7 26.5
150 12.5 10.0 15.0



#61
Fluorene
Concen: 21.789 ng/ul
RT: 15.534 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:166 Resp: 1167247
Ion Ratio Lower Upper
166 100
165 97.7 78.2 117.2
167 13.7 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 21.945 ng/ul
RT: 15.534 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:204 Resp: 568729
Ion Ratio Lower Upper
204 100
206 32.9 26.3 39.5
141 56.7 45.4 68.0



#63
4-Nitroaniline
Concen: 21.798 ng/ul
RT: 15.569 min Scan# 2139
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:138 Resp: 219167
Ion Ratio Lower Upper
138 100
92 52.6 42.1 63.1
108 72.5 58.0 87.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.100 ng/ul
 RT: 15.622 min Scan# 2148
 Delta R.T. 0.000 min
 Lab File: BP012167.D
 Acq: 17 Oct 2022 16:03




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020640


Tgt Ion:198 Resp: 187102
 Ion Ratio Lower Upper
 198 100
 182 6.2 5.0 7.4
 77 24.9 19.9 29.9



#67
 N-Nitrosodiphenylamine
 Concen: 20.626 ng/ul
 RT: 15.745 min Scan# 2169
 Delta R.T. 0.000 min
 Lab File: BP012167.D
 Acq: 17 Oct 2022 16:03

Tgt Ion:169 Resp: 995553
 Ion Ratio Lower Upper
 169 100
 168 65.9 52.7 79.1
 167 36.2 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 21.403 ng/ul
RT: 16.434 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:248 Resp: 339621
Ion Ratio Lower Upper
248 100
250 97.6 78.1 117.1
141 68.1 54.5 81.7



#69
Hexachlorobenzene
Concen: 22.043 ng/ul
RT: 16.539 min Scan# 2304
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:284 Resp: 405981
Ion Ratio Lower Upper
284 100
142 37.0 29.6 44.4
249 31.0 24.8 37.2





#70

Atrazine

Concen: 22.046 ng/ul

RT: 16.710 min Scan# 2333

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:200 Resp: 327222

Ion Ratio Lower Upper

200 100

173 26.2 21.0 31.4

215 48.8 39.0 58.6



#71

Pentachlorophenol

Concen: 21.410 ng/ul

RT: 16.892 min Scan# 2364

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03




Tgt Ion:266 Resp: 231028


Ion Ratio Lower Upper

266 100

264 60.9 48.7 73.1

268 63.2 50.6 75.8





#72

Phenanthrene

Concen: 21.177 ng/ul

RT: 17.286 min Scan# 2431

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:178 Resp: 1878036

Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 19.3 15.4 23.2



#74

Anthracene

Concen: 21.445 ng/ul

RT: 17.381 min Scan# 2447

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03




Tgt Ion:178 Resp: 1881951


Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.2

176 18.8 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.041 ng/uL
RT: 13.281 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:216 Resp: 501733
Ion Ratio Lower Upper
216 100
214 79.5 63.6 95.4
218 48.0 38.4 57.6



#76
Pentachlorobenzene
Concen: 20.373 ng/uL
RT: 14.804 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:250 Resp: 527223
Ion Ratio Lower Upper
250 100
248 61.2 49.0 73.4
252 62.2 49.8 74.6





#77

Carbazole

Concen: 21.717 ng/ul

RT: 17.651 min Scan# 2493

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:167 Resp: 1699827

Ion Ratio Lower Upper

167 100

166 20.7 16.6 24.8

139 12.6 10.1 15.1



#78

Di-n-butylphthalate

Concen: 23.833 ng/ul

RT: 18.204 min Scan# 2587

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03




Tgt Ion:149 Resp: 1851166


Ion Ratio Lower Upper

149 100

150 8.8 7.0 10.6

104 4.8 3.8 5.8









#80
Fluoranthene
Concen: 20.847 ng/ul
RT: 19.257 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:202 Resp: 2115494
Ion Ratio Lower Upper
202 100
101 12.6 10.1 15.1
100 9.4 7.5 11.3



#82
Pyrene
Concen: 20.786 ng/ul
RT: 19.610 min Scan# 2826
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:202 Resp: 2232223
Ion Ratio Lower Upper
202 100
101 14.4 11.5 17.3
100 11.5 9.2 13.8





#83

Butylbenzylphthalate

Concen: 23.683 ng/ul

RT: 20.486 min Scan# 2975

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:149 Resp: 806212

Ion Ratio Lower Upper

149 100

91 70.6 56.5 84.7

206 20.3 16.2 24.4



#84

3,3'-Dichlorobenzidine

Concen: 20.998 ng/ul

RT: 21.257 min Scan# 3106

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03




Tgt Ion:252 Resp: 747965


Ion Ratio Lower Upper

252 100

254 64.2 51.4 77.0

126 11.6 9.3 13.9





#85

Benzo(a)anthracene

Concen: 21.047 ng/ul

RT: 21.322 min Scan# 3117

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:228 Resp: 2141531

Ion Ratio Lower Upper

228 100

229 20.1 16.1 24.1

226 27.1 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 25.222 ng/ul

RT: 21.239 min Scan# 3103

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03




Tgt Ion:149 Resp: 1157506


Ion Ratio Lower Upper

149 100

167 25.9 20.7 31.1

279 3.7 3.0 4.4





#87

Chrysene

Concen: 21.009 ng/ul

RT: 21.374 min Scan# 3126

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:228 Resp: 2057688

Ion Ratio Lower Upper

228 100

226 29.8 23.8 35.8

229 19.9 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 23.922 ng/ul




RT: 22.169 min Scan# 3261


Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Tgt Ion:149 Resp: 1970691





#90

Benzo(b)fluoranthene

Concen: 21.193 ng/ul

RT: 23.010 min Scan# 3404

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:252 Resp: 2207346

Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3

125 10.8 8.6 13.0



#91

Benzo(k)fluoranthene

Concen: 21.417 ng/ul

RT: 23.057 min Scan# 3412

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03




Tgt Ion:252 Resp: 2194127


Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 11.0 8.8 13.2









#93
Benzo(a)pyrene
Concen: 21.073 ng/ul
RT: 23.639 min Scan# 3511
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03




Instrument :
BNA_P
ClientSampleId :
SSTD020640


Tgt Ion:252 Resp: 1958113
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 12.0 9.6 14.4



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.356 ng/ul
RT: 26.257 min Scan# 3956
Delta R.T. 0.000 min
Lab File: BP012167.D
Acq: 17 Oct 2022 16:03

Tgt Ion:276 Resp: 2391645
Ion Ratio Lower Upper
276 100
138 25.9 20.7 31.1
277 24.6 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 20.715 ng/ul

RT: 26.268 min Scan# 3958

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Instrument :

BNA_P

ClientSampleId :

SSTD020640





Tgt Ion:278 Resp: 2183506

Ion Ratio Lower Upper

278 100

139 16.9 13.5 20.3

279 24.0 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 20.296 ng/ul

RT: 27.021 min Scan# 4086

Delta R.T. 0.000 min

Lab File: BP012167.D

Acq: 17 Oct 2022 16:03

Tgt Ion:276 Resp: 2149998

Ion Ratio Lower Upper

276 100

138 23.4 18.7 28.1

277 23.9 19.1 28.7

