

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101722\
 Data File : BP012168.D
 Acq On : 17 Oct 2022 16:37
 Operator : CG/JU
 Sample : SSTD04077
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040641

Quant Time: Oct 18 01:15:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 18 01:11:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	365732	20.000	ng/ul	0.00
20) Naphthalene-d8	10.640	136	1578248	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.487	164	881054	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.245	188	1561092	20.000	ng/ul	0.00
79) Chrysene-d12	21.339	240	1256895	20.000	ng/ul	0.00
88) Perylene-d12	23.745	264	1252829	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	153166	16.512	ng/uL	0.00
4) Pyridine-d5	3.687	84	1079348	43.752	ng/ul	0.00
7) Phenol-d5	7.005	99	1321546	44.929	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.169	67	806441	48.301	ng/ul	0.00
11) 2-Chlorophenol-d4	7.363	132	1015903	43.511	ng/ul	0.00
15) 4-Methylphenol-d8	8.552	113	1010451	44.724	ng/ul	0.00
21) Nitrobenzene-d5	9.004	128	496824	44.573	ng/ul	0.00
24) 2-Nitrophenol-d4	9.728	143	533972	47.245	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.263	165	980077	44.744	ng/ul	0.00
31) 4-Chloroaniline-d4	10.787	131	1336131	41.394	ng/ul	0.00
46) Dimethylphthalate-d6	13.898	166	2348127	42.082	ng/ul	0.00
49) Acenaphthylene-d8	14.181	160	2966980	43.000	ng/ul	0.00
54) 4-Nitrophenol-d4	14.698	143	422543	38.573	ng/ul	0.00
60) Fluorene-d10	15.481	176	2053065	40.668	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.610	200	353715	42.911	ng/ul	0.00
73) Anthracene-d10	17.345	188	2861243	42.208	ng/ul	0.00
81) Pyrene-d10	19.580	212	2907615	45.155	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.592	264	2660935	43.691	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	162805	16.678	ng/uL	100
5) Pyridine	3.705	79	1089108	46.077	ng/ul	98
6) Benzaldehyde	6.981	77	703158	50.951	ng/ul	98
8) Phenol	7.034	94	1377787	44.618	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.263	93	1120805	46.003	ng/ul	98
12) 2-Chlorophenol	7.399	128	1087574	42.920	ng/ul	99
13) 2-Methylphenol	8.287	108	1042270	45.254	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.369	45	1500077	60.069	ng/ul	99
16) Acetophenone	8.669	105	1603667	46.040	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.657	70	783536	49.389	ng/ul	100
18) 4-Methylphenol	8.616	108	1125190	45.326	ng/ul	100
19) Hexachloroethane	8.910	117	423420	43.288	ng/ul	93
22) Nitrobenzene	9.046	77	1213243	47.482	ng/ul	98
23) Isophorone	9.581	82	2205116	48.654	ng/ul	99
25) 2-Nitrophenol	9.757	139	607504	45.751	ng/ul	97
26) 2,4-Dimethylphenol	9.822	107	1156695	44.728	ng/ul	100
27) Bis(2-Chloroethoxy)met...	10.057	93	1448282	46.566	ng/ul	99
29) 2,4-Dichlorophenol	10.293	162	997689	43.342	ng/ul	99
30) Naphthalene	10.693	128	3449228	41.357	ng/ul	100
32) 4-Chloroaniline	10.810	127	1390106	41.771	ng/ul	99
33) Hexachlorobutadiene	10.969	225	629911	44.130	ng/ul	99
34) Caprolactam	11.610	113	163048	25.777	ng/ul	96
35) 4-Chloro-3-methylphenol	11.940	107	1014572	46.524	ng/ul	99
36) 2-Methylnaphthalene	12.304	142	2275295	42.368	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101722\
 Data File : BP012168.D
 Acq On : 17 Oct 2022 16:37
 Operator : CG/JU
 Sample : SST04077
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040641

Quant Time: Oct 18 01:15:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 18 01:11:56 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.522	142	2253870	41.908	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.669	216	1163770	43.211	ng/ul	99
40) Hexachlorocyclopentadiene	12.645	237	601817	46.005	ng/ul	98
41) 2,4,6-Trichlorophenol	12.916	196	745542	45.955	ng/ul	98
42) 2,4,5-Trichlorophenol	12.987	196	794823	45.079	ng/ul	99
43) 1,1'-Biphenyl	13.316	154	2781912	41.825	ng/ul	99
44) 2-Chloronaphthalene	13.357	162	2237607	42.128	ng/ul	99
45) 2-Nitroaniline	13.575	65	564918	49.858	ng/ul	97
47) Dimethylphthalate	13.945	163	2463159	41.899	ng/ul	100
48) 2,6-Dinitrotoluene	14.075	165	490099	47.087	ng/ul	93
50) Acenaphthylene	14.210	152	3275344	41.952	ng/ul	100
51) 3-Nitroaniline	14.404	138	506933	41.827	ng/ul	91
52) Acenaphthene	14.551	153	2267267	40.794	ng/ul	99
53) 2,4-Dinitrophenol	14.610	184	275330	40.947	ng/ul	98
55) 4-Nitrophenol	14.710	109	310937	41.056	ng/ul	94
56) Dibenzofuran	14.887	168	3028474	40.318	ng/ul	98
57) 2,4-Dinitrotoluene	14.863	165	662335	44.277	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.116	232	618164	44.981	ng/ul	97
59) Diethylphthalate	15.316	149	2310324	42.911	ng/ul	100
61) Fluorene	15.539	166	2361037	40.043	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.534	204	1156436	40.543	ng/ul	99
63) 4-Nitroaniline	15.569	138	443993	40.120	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.628	198	383849	43.074	ng/ul#	95
67) N-Nitrosodiphenylamine	15.751	169	1978906	42.827	ng/ul	99
68) 4-Bromophenyl-phenylether	16.433	248	706993	46.541	ng/ul	99
69) Hexachlorobenzene	16.545	284	815553	46.255	ng/ul	96
70) Atrazine	16.710	200	614061	43.217	ng/ul	99
71) Pentachlorophenol	16.892	266	469231	45.424	ng/ul	98
72) Phenanthrene	17.286	178	3490350	41.113	ng/ul	100
74) Anthracene	17.381	178	3492930	41.576	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.281	216	1151899	45.663	ng/uL	98
76) Pentachlorobenzene	14.804	250	1132963	45.731	ng/uL	98
77) Carbazole	17.657	167	3041813	40.595	ng/ul	99
78) Di-n-butylphthalate	18.204	149	3459708	46.527	ng/ul	99
80) Fluoranthene	19.257	202	3647981	46.235	ng/ul	100
82) Pyrene	19.610	202	3730879	44.681	ng/ul	99
83) Butylbenzylphthalate	20.486	149	1365844	51.601	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.257	252	1182233	42.685	ng/ul	99
85) Benzo(a)anthracene	21.321	228	3386830	42.809	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.245	149	1854775	51.979	ng/ul	96
87) Chrysene	21.374	228	3172516	41.659	ng/ul	99
89) Di-n-octyl phthalate	22.169	149	3096902	49.375	ng/ul	100
90) Benzo(b)fluoranthene	23.010	252	3372001	42.521	ng/ul	99
91) Benzo(k)fluoranthene	23.063	252	3409437	43.709	ng/ul	99
93) Benzo(a)pyrene	23.639	252	3038362	42.945	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.257	276	3889767	43.481	ng/ul	98
95) Dibenzo(a,h)anthracene	26.274	278	3493647	43.532	ng/ul	99
96) Benzo(g,h,i)perylene	27.027	276	3508845	43.503	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101722\
 Data File : BP012168.D
 Acq On : 17 Oct 2022 16:37
 Operator : CG/JU
 Sample : SSTD04077
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040641

Quant Time: Oct 18 01:15:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 18 01:11:56 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 16.678 ng/uL
RT: 3.299 min Scan# 51
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion: 88 Resp: 162805
Ion Ratio Lower Upper
88 100
43 25.0 19.6 29.4
58 74.1 59.1 88.7



#5
Pyridine
Concen: 46.077 ng/uL
RT: 3.705 min Scan# 122
Delta R.T. -0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion: 79 Resp: 1089108
Ion Ratio Lower Upper
79 100
52 70.2 57.9 86.9
51 32.3 26.7 40.1









#6
Benzaldehyde
Concen: 50.951 ng/ul
RT: 6.981 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion: 77 Resp: 703158
Ion Ratio Lower Upper
77 100
105 102.4 82.9 124.3
106 97.3 80.8 121.2



#8
Phenol
Concen: 44.618 ng/ul
RT: 7.034 min Scan# 688
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion: 94 Resp: 1377787
Ion Ratio Lower Upper
94 100
65 26.1 21.9 32.9
66 32.8 26.6 40.0









#10
Bis(2-Chloroethyl)ether
Concen: 46.003 ng/ul
RT: 7.263 min Scan# 71
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion: 93 Resp: 1120805
Ion Ratio Lower Upper
93 100
63 75.3 61.9 92.9
95 32.8 27.0 40.4



#12
2-Chlorophenol
Concen: 42.920 ng/ul
RT: 7.399 min Scan# 750
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion: 128 Resp: 1087574
Ion Ratio Lower Upper
128 100
64 45.3 37.2 55.8
130 32.3 25.8 38.8





#13

2-Methylphenol

Concen: 45.254 ng/ul

RT: 8.287 min Scan# 901

Delta R.T. -0.000 min

Lab File: BP012168.D

Acq: 17 Oct 2022 16:37

Instrument :

BNA_P

ClientSampleId :





SSTD040641

Tgt Ion:108 Resp: 1042270

Ion Ratio Lower Upper

108 100

107 89.2 70.2 105.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 60.069 ng/ul

RT: 8.369 min Scan# 915

Delta R.T. -0.000 min

Lab File: BP012168.D

Acq: 17 Oct 2022 16:37




Tgt Ion: 45 Resp: 1500077


Ion Ratio Lower Upper

45 100

77 15.0 12.5 18.7

79 11.3 9.4 14.2









#16
Acetophenone
Concen: 46.040 ng/ul
RT: 8.669 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:105 Resp: 1603667
Ion Ratio Lower Upper
105 100
77 73.3 59.8 89.6
51 24.8 20.4 30.6



#17
N-Nitroso-di-n-propylamine
Concen: 49.389 ng/ul
RT: 8.657 min Scan# 964
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion: 70 Resp: 783536
Ion Ratio Lower Upper
70 100
42 53.1 42.6 64.0
101 9.7 7.8 11.6
130 21.7 17.3 25.9









#18
4-Methylphenol
Concen: 45.326 ng/ul
RT: 8.616 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:108 Resp: 1125190
Ion Ratio Lower Upper
108 100
107 110.6 88.9 133.3



#19
Hexachloroethane
Concen: 43.288 ng/ul
RT: 8.910 min Scan# 1007
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:117 Resp: 423420
Ion Ratio Lower Upper
117 100
201 99.9 73.9 110.9
199 63.3 47.0 70.6









#22
Nitrobenzene
Concen: 47.482 ng/ul
RT: 9.046 min Scan# 1030
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion: 77 Resp: 1213243
Ion Ratio Lower Upper
77 100
123 47.2 36.8 55.2
65 12.7 10.0 15.0



#23
Isophorone
Concen: 48.654 ng/ul
RT: 9.581 min Scan# 1121
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion: 82 Resp: 2205116
Ion Ratio Lower Upper
82 100
95 6.5 5.3 7.9
138 18.3 14.2 21.4









#25
2-Nitrophenol
Concen: 45.751 ng/ul
RT: 9.757 min Scan# 1151
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:139 Resp: 607504
Ion Ratio Lower Upper
139 100
65 47.4 40.4 60.6
109 27.1 22.2 33.2



#26
2,4-Dimethylphenol
Concen: 44.728 ng/ul
RT: 9.822 min Scan# 1162
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:107 Resp: 1156695
Ion Ratio Lower Upper
107 100
121 51.4 41.2 61.8
122 88.6 70.5 105.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 46.566 ng/ul
 RT: 10.057 min Scan# 1202
 Delta R.T. -0.000 min
 Lab File: BP012168.D
 Acq: 17 Oct 2022 16:37

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040641

Tgt Ion: 93 Resp: 1448282




Ion	Ratio	Lower	Upper
93	100		
95	32.5	26.2	39.2
123	11.2	8.5	12.7




#29
 2,4-Dichlorophenol
 Concen: 43.342 ng/ul
 RT: 10.293 min Scan# 1242
 Delta R.T. 0.006 min
 Lab File: BP012168.D
 Acq: 17 Oct 2022 16:37

Tgt Ion:162 Resp: 997689

Ion	Ratio	Lower	Upper
162	100		
164	65.8	52.1	78.1
98	35.3	29.4	44.0









#30
Naphthalene
Concen: 41.357 ng/ul
RT: 10.693 min Scan# 1310
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:128 Resp: 3449228
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 13.1 10.3 15.5



#32
4-Chloroaniline
Concen: 41.771 ng/ul
RT: 10.810 min Scan# 1330
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:127 Resp: 1390106
Ion Ratio Lower Upper
127 100
129 31.7 25.8 38.8









#33
Hexachlorobutadiene
Concen: 44.130 ng/ul
RT: 10.969 min Scan# 1357
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:225 Resp: 629911
Ion Ratio Lower Upper
225 100
223 62.3 50.4 75.6
227 63.9 51.9 77.9



#34
Caprolactam
Concen: 25.777 ng/ul
RT: 11.610 min Scan# 1466
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:113 Resp: 163048
Ion Ratio Lower Upper
113 100
55 168.9 141.1 211.7
56 124.5 98.3 147.5









#35
4-Chloro-3-methylphenol
Concen: 46.524 ng/ul
RT: 11.940 min Scan# 1521
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:107 Resp: 1014572
Ion Ratio Lower Upper
107 100
144 26.5 21.0 31.4
142 82.7 65.1 97.7



#36
2-Methylnaphthalene
Concen: 42.368 ng/ul
RT: 12.304 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:142 Resp: 2275295
Ion Ratio Lower Upper
142 100
141 88.9 71.3 106.9









#37
1-Methylnaphthalene
Concen: 41.908 ng/ul
RT: 12.522 min Scan# 1621
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:142 Resp: 2253870
Ion Ratio Lower Upper
142 100
141 91.8 73.1 109.7
116 3.8 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 43.211 ng/ul
RT: 12.669 min Scan# 1646
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:216 Resp: 1163770
Ion Ratio Lower Upper
216 100
214 79.1 62.5 93.7
179 21.1 16.8 25.2
108 17.6 14.1 21.1









#40
Hexachlorocyclopentadiene
Concen: 46.005 ng/ul
RT: 12.645 min Scan# 1642
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:237 Resp: 601817
Ion Ratio Lower Upper
237 100
235 63.4 49.4 74.0
272 13.4 11.1 16.7



#41
2,4,6-Trichlorophenol
Concen: 45.955 ng/ul
RT: 12.916 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:196 Resp: 745542
Ion Ratio Lower Upper
196 100
198 94.6 77.6 116.4
200 30.7 23.8 35.8









#42
2,4,5-Trichlorophenol
Concen: 45.079 ng/ul
RT: 12.987 min Scan# 1700
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:196 Resp: 794823
Ion Ratio Lower Upper
196 100
198 97.1 76.7 115.1
200 30.9 24.4 36.6



#43
1,1'-Biphenyl
Concen: 41.825 ng/ul
RT: 13.316 min Scan# 1756
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:154 Resp: 2781912
Ion Ratio Lower Upper
154 100
153 39.7 32.3 48.5
76 13.6 11.0 16.6









#44
2-Chloronaphthalene
Concen: 42.128 ng/ul
RT: 13.357 min Scan# 1763
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:162 Resp: 2237607
Ion Ratio Lower Upper
162 100
164 32.7 26.4 39.6
127 38.8 30.4 45.6



#45
2-Nitroaniline
Concen: 49.858 ng/ul
RT: 13.575 min Scan# 1800
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion: 65 Resp: 564918
Ion Ratio Lower Upper
65 100
92 74.2 57.1 85.7
138 116.3 90.5 135.7









#47
Dimethylphthalate
Concen: 41.899 ng/ul
RT: 13.945 min Scan# 1863
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:163 Resp: 2463159
Ion Ratio Lower Upper
163 100
194 3.9 3.0 4.6
164 10.2 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 47.087 ng/ul
RT: 14.075 min Scan# 1885
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:165 Resp: 490099
Ion Ratio Lower Upper
165 100
89 54.2 48.2 72.4
121 21.2 19.0 28.4









#50
Acenaphthylene
Concen: 41.952 ng/ul
RT: 14.210 min Scan# 1908
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:152 Resp: 3275344
Ion Ratio Lower Upper
152 100
151 20.0 15.9 23.9
153 13.2 10.6 16.0



#51
3-Nitroaniline
Concen: 41.827 ng/ul
RT: 14.404 min Scan# 1941
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:138 Resp: 506933
Ion Ratio Lower Upper
138 100
108 9.9 9.4 14.2
92 110.8 96.6 144.8









#52
Acenaphthene
Concen: 40.794 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:153 Resp: 2267267
Ion Ratio Lower Upper
153 100
152 48.1 38.8 58.2
154 88.3 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 40.947 ng/ul
RT: 14.610 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:184 Resp: 275330
Ion Ratio Lower Upper
184 100
63 62.6 50.5 75.7
154 59.1 45.8 68.8









#55
4-Nitrophenol
Concen: 41.056 ng/ul
RT: 14.710 min Scan# 1993
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:109 Resp: 310937
Ion Ratio Lower Upper
109 100
139 149.2 126.6 189.8
65 141.2 117.1 175.7



#56
Dibenzofuran
Concen: 40.318 ng/ul
RT: 14.887 min Scan# 2023
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:168 Resp: 3028474
Ion Ratio Lower Upper
168 100
139 36.8 28.6 42.8





#57

2,4-Dinitrotoluene

Concen: 44.277 ng/ul

RT: 14.863 min Scan# 2018

Delta R.T. 0.006 min

Lab File: BP012168.D

Acq: 17 Oct 2022 16:37

Instrument :

BNA_P

ClientSampleId :

SSTD040641





Tgt Ion:165 Resp: 662335

Ion Ratio Lower Upper

165 100

63 39.4 35.5 53.3

182 2.6 1.9 2.9



#58

2,3,4,6-Tetrachlorophenol

Concen: 44.981 ng/ul

RT: 15.116 min Scan# 2062

Delta R.T. 0.006 min

Lab File: BP012168.D

Acq: 17 Oct 2022 16:37

Tgt Ion:232 Resp: 618164




Ion Ratio Lower Upper


232 100

131 44.1 37.3 55.9

130 2.3 1.8 2.6

166 30.2 24.7 37.1









#59
Diethylphthalate
Concen: 42.911 ng/ul
RT: 15.316 min Scan# 2095
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:149 Resp: 2310324
Ion Ratio Lower Upper
149 100
177 22.0 17.7 26.5
150 12.3 10.0 15.0



#61
Fluorene
Concen: 40.043 ng/ul
RT: 15.539 min Scan# 2134
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:166 Resp: 2361037
Ion Ratio Lower Upper
166 100
165 97.7 78.2 117.2
167 13.2 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 40.543 ng/ul
RT: 15.534 min Scan# 2133
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:204 Resp: 1156436
Ion Ratio Lower Upper
204 100
206 33.6 26.3 39.5
141 57.7 45.4 68.0



#63
4-Nitroaniline
Concen: 40.120 ng/ul
RT: 15.569 min Scan# 2139
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:138 Resp: 443993
Ion Ratio Lower Upper
138 100
92 50.8 42.1 63.1
108 72.2 58.0 87.0









#66
4,6-Dinitro-2-methylphenol
Concen: 43.074 ng/ul
RT: 15.628 min Scan# 2148
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:198 Resp: 383849
Ion Ratio Lower Upper
198 100
182 4.5 5.0 7.4#
77 22.4 19.9 29.9



#67
N-Nitrosodiphenylamine
Concen: 42.827 ng/ul
RT: 15.751 min Scan# 2170
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:169 Resp: 1978906
Ion Ratio Lower Upper
169 100
168 66.1 52.7 79.1
167 36.8 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 46.541 ng/ul
RT: 16.433 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:248 Resp: 706993
Ion Ratio Lower Upper
248 100
250 97.1 78.1 117.1
141 67.6 54.5 81.7



#69
Hexachlorobenzene
Concen: 46.255 ng/ul
RT: 16.545 min Scan# 2305
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:284 Resp: 815553
Ion Ratio Lower Upper
284 100
142 33.3 29.6 44.4
249 30.2 24.8 37.2









#70
Atrazine
Concen: 43.217 ng/ul
RT: 16.710 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:200 Resp: 614061
Ion Ratio Lower Upper
200 100
173 26.5 21.0 31.4
215 49.5 39.0 58.6



#71
Pentachlorophenol
Concen: 45.424 ng/ul
RT: 16.892 min Scan# 2364
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:266 Resp: 469231
Ion Ratio Lower Upper
266 100
264 62.3 48.7 73.1
268 64.4 50.6 75.8





#72

Phenanthrene

Concen: 41.113 ng/ul

RT: 17.286 min Scan# 2431

Delta R.T. -0.000 min

Lab File: BP012168.D

Acq: 17 Oct 2022 16:37

Instrument :

BNA_P

ClientSampleId :

SSTD040641





Tgt Ion:178 Resp: 3490350

Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 19.1 15.4 23.2



#74

Anthracene

Concen: 41.576 ng/ul

RT: 17.381 min Scan# 2447

Delta R.T. -0.000 min

Lab File: BP012168.D

Acq: 17 Oct 2022 16:37




Tgt Ion:178 Resp: 3492930


Ion Ratio Lower Upper

178 100

179 15.8 12.2 18.2

176 19.1 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 45.663 ng/uL
RT: 13.281 min Scan# 1750
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:216 Resp: 1151899
Ion Ratio Lower Upper
216 100
214 77.3 63.6 95.4
218 48.6 38.4 57.6



#76
Pentachlorobenzene
Concen: 45.731 ng/uL
RT: 14.804 min Scan# 2009
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:250 Resp: 1132963
Ion Ratio Lower Upper
250 100
248 62.7 49.0 73.4
252 64.0 49.8 74.6









#77
Carbazole
Concen: 40.595 ng/ul
RT: 17.657 min Scan# 2494
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:167 Resp: 3041813
Ion Ratio Lower Upper
167 100
166 21.6 16.6 24.8
139 12.8 10.1 15.1



#78
Di-n-butylphthalate
Concen: 46.527 ng/ul
RT: 18.204 min Scan# 2587
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:149 Resp: 3459708
Ion Ratio Lower Upper
149 100
150 9.1 7.0 10.6
104 4.8 3.8 5.8









#80
Fluoranthene
Concen: 46.235 ng/ul
RT: 19.257 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:202 Resp: 3647981
Ion Ratio Lower Upper
202 100
101 12.6 10.1 15.1
100 9.3 7.5 11.3



#82
Pyrene
Concen: 44.681 ng/ul
RT: 19.610 min Scan# 2826
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:202 Resp: 3730879
Ion Ratio Lower Upper
202 100
101 14.9 11.5 17.3
100 12.1 9.2 13.8





#83

Butylbenzylphthalate

Concen: 51.601 ng/ul

RT: 20.486 min Scan# 2975

Delta R.T. -0.000 min

Lab File: BP012168.D

Acq: 17 Oct 2022 16:37

Instrument :

BNA_P

ClientSampleId :

SSTD040641





Tgt Ion:149 Resp: 1365844

Ion Ratio Lower Upper

149 100

91 70.2 56.5 84.7

206 20.6 16.2 24.4



#84

3,3'-Dichlorobenzidine

Concen: 42.685 ng/ul

RT: 21.257 min Scan# 3106

Delta R.T. -0.000 min

Lab File: BP012168.D

Acq: 17 Oct 2022 16:37




Tgt Ion:252 Resp: 1182233


Ion Ratio Lower Upper

252 100

254 64.8 51.4 77.0

126 12.0 9.3 13.9









#85
Benzo(a)anthracene
Concen: 42.809 ng/ul
RT: 21.321 min Scan# 3117
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:228 Resp: 3386830
Ion Ratio Lower Upper
228 100
229 19.8 16.1 24.1
226 27.4 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 51.979 ng/ul
RT: 21.245 min Scan# 3104
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:149 Resp: 1854775
Ion Ratio Lower Upper
149 100
167 27.9 20.7 31.1
279 4.4 3.0 4.4









#87
Chrysene
Concen: 41.659 ng/ul
RT: 21.374 min Scan# 3126
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:228 Resp: 3172516
Ion Ratio Lower Upper
228 100
226 30.4 23.8 35.8
229 20.1 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 49.375 ng/ul
RT: 22.169 min Scan# 3261
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:149 Resp: 3096902









#90
Benzo(b)fluoranthene
Concen: 42.521 ng/ul
RT: 23.010 min Scan# 3404
Delta R.T. -0.000 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37




Instrument :
BNA_P
ClientSampleId :
SSTD040641


Tgt Ion:252 Resp: 3372001
Ion Ratio Lower Upper
252 100
253 22.4 17.5 26.3
125 11.1 8.6 13.0



#91
Benzo(k)fluoranthene
Concen: 43.709 ng/ul
RT: 23.063 min Scan# 3413
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:252 Resp: 3409437
Ion Ratio Lower Upper
252 100
253 22.3 17.4 26.2
125 10.6 8.8 13.2





#93

Benzo(a)pyrene

Concen: 42.945 ng/ul

RT: 23.639 min Scan# 3511

Delta R.T. -0.000 min

Lab File: BP012168.D

Acq: 17 Oct 2022 16:37

Instrument :

BNA_P

ClientSampleId :

SSTD040641





Tgt Ion:252 Resp: 3038362

Ion Ratio Lower Upper

252 100

253 22.4 17.4 26.0

125 12.3 9.6 14.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 43.481 ng/ul

RT: 26.257 min Scan# 3956

Delta R.T. -0.000 min

Lab File: BP012168.D

Acq: 17 Oct 2022 16:37




Tgt Ion:276 Resp: 3889767


Ion Ratio Lower Upper

276 100

138 24.7 20.7 31.1

277 25.3 19.7 29.5









#95
Dibenzo(a,h)anthracene
Concen: 43.532 ng/ul
RT: 26.274 min Scan# 3958
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Instrument :
BNA_P
ClientSampleId :
SSTD040641

Tgt Ion:278 Resp: 3493647
Ion Ratio Lower Upper
278 100
139 16.7 13.5 20.3
279 24.3 19.2 28.8



#96
Benzo(g,h,i)perylene
Concen: 43.503 ng/ul
RT: 27.027 min Scan# 4087
Delta R.T. 0.006 min
Lab File: BP012168.D
Acq: 17 Oct 2022 16:37

Tgt Ion:276 Resp: 3508845
Ion Ratio Lower Upper
276 100
138 23.2 18.7 28.1
277 23.8 19.1 28.7

