

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101722\
 Data File : BP012171.D
 Acq On : 17 Oct 2022 18:56
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV644

Quant Time: Oct 18 02:11:47 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 18 02:04:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	357270	20.000	ng/u1	0.00
20) Naphthalene-d8	10.640	136	1555827	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.481	164	935939	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.240	188	1750485	20.000	ng/u1	0.00
79) Chrysene-d12	21.333	240	1444762	20.000	ng/u1	0.00
88) Perylene-d12	23.745	264	1465480	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	73149	7.859	ng/uL	0.00
4) Pyridine-d5	3.687	84	499876	19.664	ng/u1	0.00
7) Phenol-d5	7.005	99	614948	20.287	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.169	67	378826	20.609	ng/u1	0.00
11) 2-Chlorophenol-d4	7.364	132	480299	21.059	ng/u1	0.00
15) 4-Methylphenol-d8	8.552	113	490462	21.490	ng/u1	0.00
21) Nitrobenzene-d5	9.005	128	242593	22.448	ng/u1	0.00
24) 2-Nitrophenol-d4	9.722	143	258046	23.045	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.263	165	484429	21.869	ng/u1	0.00
31) 4-Chloroaniline-d4	10.781	131	685135	21.719	ng/u1	0.00
46) Dimethylphthalate-d6	13.899	166	1308635	21.524	ng/u1	0.00
49) Acenaphthylene-d8	14.175	160	1593941	22.126	ng/u1	0.00
54) 4-Nitrophenol-d4	14.693	143	223127	20.141	ng/u1	0.00
60) Fluorene-d10	15.481	176	1146307	21.246	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.604	200	195070	20.833	ng/u1	0.00
73) Anthracene-d10	17.340	188	1630908	21.520	ng/u1	0.00
81) Pyrene-d10	19.581	212	1664389	22.327	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.586	264	1501897	21.014	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	78980	7.883	ng/uL	99
5) Pyridine	3.705	79	505096	20.011	ng/u1	98
6) Benzaldehyde	6.975	77	351110	24.665	ng/u1	97
8) Phenol	7.028	94	647530	20.485	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.264	93	532889	20.878	ng/u1	98
12) 2-Chlorophenol	7.399	128	519609	21.052	ng/u1	97
13) 2-Methylphenol	8.281	108	493120	20.978	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.369	45	711306	20.917	ng/u1	99
16) Acetophenone	8.664	105	791073	22.271	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.652	70	392884	23.014	ng/u1	98
18) 4-Methylphenol	8.611	108	543603	21.487	ng/u1	100
19) Hexachloroethane	8.911	117	202446	20.512	ng/u1	91
22) Nitrobenzene	9.046	77	579219	21.283	ng/u1	99
23) Isophorone	9.575	82	1110868	22.279	ng/u1	100
25) 2-Nitrophenol	9.758	139	296594	22.884	ng/u1	96
26) 2,4-Dimethylphenol	9.816	107	584805	21.733	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.058	93	723008	21.373	ng/u1	100
29) 2,4-Dichlorophenol	10.287	162	502169	21.803	ng/u1	98
30) Naphthalene	10.687	128	1725027	20.937	ng/u1	99
32) 4-Chloroaniline	10.805	127	712042	21.835	ng/u1	100
33) Hexachlorobutadiene	10.969	225	318743	21.421	ng/u1	99
34) Caprolactam	11.599	113	55930	8.869	ng/u1	91
35) 4-Chloro-3-methylphenol	11.934	107	521539	22.344	ng/u1	97
36) 2-Methylnaphthalene	12.299	142	1175418	21.932	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101722\
 Data File : BP012171.D
 Acq On : 17 Oct 2022 18:56
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SICV644


Quant Time: Oct 18 02:11:47 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 18 02:04:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.522	142	1166865	21.845	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.669	216	625114	21.611	ng/ul	99
40) Hexachlorocyclopentadiene	12.646	237	307964	20.420	ng/ul	99
41) 2,4,6-Trichlorophenol	12.916	196	395199	22.117	ng/ul	99
42) 2,4,5-Trichlorophenol	12.987	196	418577	22.085	ng/ul	99
43) 1,1'-Biphenyl	13.316	154	1491214	21.166	ng/ul	99
44) 2-Chloronaphthalene	13.357	162	1182609	21.112	ng/ul	99
45) 2-Nitroaniline	13.569	65	298896	22.901	ng/ul	99
47) Dimethylphthalate	13.946	163	1375097	21.485	ng/ul	99
48) 2,6-Dinitrotoluene	14.069	165	261542	23.570	ng/ul	94
50) Acenaphthylene	14.204	152	1779483	21.808	ng/ul	99
51) 3-Nitroaniline	14.399	138	270523	21.771	ng/ul	93
52) Acenaphthene	14.546	153	1236477	21.176	ng/ul	99
53) 2,4-Dinitrophenol	14.610	184	142230	19.532	ng/ul	93
55) 4-Nitrophenol	14.710	109	168978	20.398	ng/ul	90
56) Dibenzofuran	14.881	168	1683247	21.221	ng/ul	99
57) 2,4-Dinitrotoluene	14.857	165	361329	22.614	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.110	232	340930	22.125	ng/ul	98
59) Diethylphthalate	15.310	149	1293048	21.260	ng/ul	100
61) Fluorene	15.534	166	1318978	21.170	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.528	204	657551	21.623	ng/ul	99
63) 4-Nitroaniline	15.563	138	240986	22.033	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.622	198	210631	20.834	ng/ul	96
67) N-Nitrosodiphenylamine	15.746	169	1107723	21.691	ng/ul	100
68) 4-Bromophenyl-phenylether	16.428	248	401076	22.506	ng/ul	96
69) Hexachlorobenzene	16.540	284	474254	22.314	ng/ul	96
70) Atrazine	16.710	200	350672	21.140	ng/ul	100
71) Pentachlorophenol	16.892	266	258329	20.542	ng/ul	99
72) Phenanthrene	17.287	178	1976796	20.814	ng/ul	100
74) Anthracene	17.375	178	1996397	21.354	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.275	216	621843	22.059	ng/ul	99
76) Pentachlorobenzene	14.798	250	640281	22.525	ng/ul	98
77) Carbazole	17.651	167	1702883	20.715	ng/ul	99
78) Di-n-butylphthalate	18.198	149	1927870	21.529	ng/ul	99
80) Fluoranthene	19.257	202	2076005	22.440	ng/ul	98
82) Pyrene	19.610	202	2159232	22.210	ng/ul	98
83) Butylbenzylphthalate	20.481	149	740479	21.722	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.257	252	672210	21.054	ng/ul	100
85) Benzo(a)anthracene	21.316	228	1928001	21.109	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.239	149	1017548	21.108	ng/ul	98
87) Chrysene	21.369	228	1837407	20.986	ng/ul	100
89) Di-n-octyl phthalate	22.163	149	1648526	19.749	ng/ul	100
90) Benzo(b)fluoranthene	23.010	252	1964546	21.158	ng/ul	99
91) Benzo(k)fluoranthene	23.057	252	1873077	20.363	ng/ul	99
93) Benzo(a)pyrene	23.639	252	1717968	20.933	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.245	276	2220746	21.425	ng/ul	97
95) Dibenzo(a,h)anthracene	26.263	278	2002712	21.402	ng/ul	100
96) Benzo(g,h,i)perylene	27.015	276	2019888	21.547	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 18 02:11:47 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101722.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 18 02:04:26 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.883 ng/uL
RT: 3.299 min Scan# 51
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion: 88	Resp: 78980
Ion Ratio	Lower Upper
88 100	
43 25.4	19.6 29.4
58 73.4	59.1 88.7



#5
Pyridine
Concen: 20.011 ng/uL
RT: 3.705 min Scan# 122
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion: 79	Resp: 505096
Ion Ratio	Lower Upper
79 100	
52 74.5	57.9 86.9
51 32.8	26.7 40.1









#6
Benzaldehyde
Concen: 24.665 ng/ul
RT: 6.975 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion: 77 Resp: 351110
Ion Ratio Lower Upper
77 100
105 101.1 82.9 124.3
106 96.6 80.8 121.2



#8
Phenol
Concen: 20.485 ng/ul
RT: 7.028 min Scan# 687
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion: 94 Resp: 647530
Ion Ratio Lower Upper
94 100
65 26.5 21.9 32.9
66 33.8 26.6 40.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.878 ng/ul
RT: 7.264 min Scan# 71
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion: 93 Resp: 532889
Ion Ratio Lower Upper
93 100
63 74.6 61.9 92.9
95 33.3 27.0 40.4



#12
2-Chlorophenol
Concen: 21.052 ng/ul
RT: 7.399 min Scan# 750
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion: 128 Resp: 519609
Ion Ratio Lower Upper
128 100
64 43.7 37.2 55.8
130 31.6 25.8 38.8









#13
2-Methylphenol
Concen: 20.978 ng/ul
RT: 8.281 min Scan# 901
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:108 Resp: 493120
Ion Ratio Lower Upper
108 100
107 90.4 70.2 105.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.917 ng/ul
RT: 8.369 min Scan# 915
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion: 45 Resp: 711306
Ion Ratio Lower Upper
45 100
77 15.1 12.5 18.7
79 11.8 9.4 14.2









#16
Acetophenone
Concen: 22.271 ng/ul
RT: 8.664 min Scan# 965
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:105 Resp: 791073
Ion Ratio Lower Upper
105 100
77 74.9 59.8 89.6
51 24.6 20.4 30.6



#17
N-Nitroso-di-n-propylamine
Concen: 23.014 ng/ul
RT: 8.652 min Scan# 963
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion: 70 Resp: 392884
Ion Ratio Lower Upper
70 100
42 51.2 42.6 64.0
101 10.3 7.8 11.6
130 22.1 17.3 25.9





#18

4-Methylphenol

Concen: 21.487 ng/ul

RT: 8.611 min Scan# 91

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56

Instrument :

BNA_P

ClientSampleId :





SICV644

Tgt Ion:108 Resp: 543603

Ion Ratio Lower Upper

108 100

107 110.6 88.9 133.3



#19

Hexachloroethane

Concen: 20.512 ng/ul

RT: 8.911 min Scan# 1007

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56




Tgt Ion:117 Resp: 202446


Ion Ratio Lower Upper

117 100

201 101.1 73.9 110.9

199 65.2 47.0 70.6





#22

Nitrobenzene

Concen: 21.283 ng/ul

RT: 9.046 min Scan# 1030

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SICV644





Tgt Ion: 77 Resp: 579219

Ion Ratio Lower Upper

77 100

123 47.0 36.8 55.2

65 12.7 10.0 15.0



#23

Isophorone

Concen: 22.279 ng/ul

RT: 9.575 min Scan# 1120

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56




Tgt Ion: 82 Resp: 1110868


Ion Ratio Lower Upper

82 100

95 6.9 5.3 7.9

138 17.8 14.2 21.4









#25
2-Nitrophenol
Concen: 22.884 ng/ul
RT: 9.758 min Scan# 1151
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:139 Resp: 296594
Ion Ratio Lower Upper
139 100
65 46.9 40.4 60.6
109 27.9 22.2 33.2



#26
2,4-Dimethylphenol
Concen: 21.733 ng/ul
RT: 9.816 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:107 Resp: 584805
Ion Ratio Lower Upper
107 100
121 51.5 41.2 61.8
122 87.4 70.5 105.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.373 ng/ul
 RT: 10.058 min Scan# 1202
 Delta R.T. 0.000 min
 Lab File: BP012171.D
 Acq: 17 Oct 2022 18:56




Instrument :
 BNA_P
 ClientSampleId :
 SICV644


Tgt Ion: 93 Resp: 723008
 Ion Ratio Lower Upper
 93 100
 95 32.8 26.2 39.2
 123 11.0 8.5 12.7



#29
 2,4-Dichlorophenol
 Concen: 21.803 ng/ul
 RT: 10.287 min Scan# 1241
 Delta R.T. 0.000 min
 Lab File: BP012171.D
 Acq: 17 Oct 2022 18:56

Tgt Ion: 162 Resp: 502169
 Ion Ratio Lower Upper
 162 100
 164 65.9 52.1 78.1
 98 35.2 29.4 44.0









#30
Naphthalene
Concen: 20.937 ng/ul
RT: 10.687 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:128 Resp: 1725027
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 13.2 10.3 15.5



#32
4-Chloroaniline
Concen: 21.835 ng/ul
RT: 10.805 min Scan# 1329
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:127 Resp: 712042
Ion Ratio Lower Upper
127 100
129 32.3 25.8 38.8









#33
Hexachlorobutadiene
Concen: 21.421 ng/ul
RT: 10.969 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:225 Resp: 318743
Ion Ratio Lower Upper
225 100
223 62.4 50.4 75.6
227 64.4 51.9 77.9



#34
Caprolactam
Concen: 8.869 ng/ul
RT: 11.599 min Scan# 1464
Delta R.T. -0.012 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:113 Resp: 55930
Ion Ratio Lower Upper
113 100
55 158.2 141.1 211.7
56 119.7 98.3 147.5









#35
4-Chloro-3-methylphenol
Concen: 22.344 ng/ul
RT: 11.934 min Scan# 1521
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:107 Resp: 521539
Ion Ratio Lower Upper
107 100
144 27.4 21.0 31.4
142 84.3 65.1 97.7



#36
2-Methylnaphthalene
Concen: 21.932 ng/ul
RT: 12.299 min Scan# 1583
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:142 Resp: 1175418
Ion Ratio Lower Upper
142 100
141 87.8 71.3 106.9









#37
1-Methylnaphthalene
Concen: 21.845 ng/ul
RT: 12.522 min Scan# 1621
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:142 Resp: 1166865
Ion Ratio Lower Upper
142 100
141 91.1 73.1 109.7
116 3.7 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.611 ng/ul
RT: 12.669 min Scan# 1646
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:216 Resp: 625114
Ion Ratio Lower Upper
216 100
214 77.8 62.5 93.7
179 20.6 16.8 25.2
108 16.1 14.1 21.1









#40
Hexachlorocyclopentadiene
Concen: 20.420 ng/ul
RT: 12.646 min Scan# 1642
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:237 Resp: 307964
Ion Ratio Lower Upper
237 100
235 62.0 49.4 74.0
272 13.2 11.1 16.7



#41
2,4,6-Trichlorophenol
Concen: 22.117 ng/ul
RT: 12.916 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:196 Resp: 395199
Ion Ratio Lower Upper
196 100
198 96.6 77.6 116.4
200 31.2 23.8 35.8





#42

2,4,5-Trichlorophenol

Concen: 22.085 ng/ul

RT: 12.987 min Scan# 1700

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SICV644





Tgt Ion:196 Resp: 418577

Ion Ratio Lower Upper

196 100

198 96.7 76.7 115.1

200 31.6 24.4 36.6



#43

1,1'-Biphenyl

Concen: 21.166 ng/ul

RT: 13.316 min Scan# 1756

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56




Tgt Ion:154 Resp: 1491214


Ion Ratio Lower Upper

154 100

153 39.9 32.3 48.5

76 13.2 11.0 16.6





#44

2-Chloronaphthalene

Concen: 21.112 ng/ul

RT: 13.357 min Scan# 1763

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SICV644





Tgt Ion:162 Resp: 1182609

Ion Ratio Lower Upper

162 100

164 32.8 26.4 39.6

127 38.7 30.4 45.6



#45

2-Nitroaniline

Concen: 22.901 ng/ul

RT: 13.569 min Scan# 1799

Delta R.T. -0.006 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56




Tgt Ion: 65 Resp: 298896


Ion Ratio Lower Upper

65 100

92 71.2 57.1 85.7

138 111.1 90.5 135.7









#47
Dimethylphthalate
Concen: 21.485 ng/ul
RT: 13.946 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:163 Resp: 1375097
Ion Ratio Lower Upper
163 100
194 4.1 3.0 4.6
164 9.9 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 23.570 ng/ul
RT: 14.069 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:165 Resp: 261542
Ion Ratio Lower Upper
165 100
89 55.0 48.2 72.4
121 21.6 19.0 28.4





#50

Acenaphthylene

Concen: 21.808 ng/ul

RT: 14.204 min Scan# 1907

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SICV644





Tgt Ion:152 Resp: 1779483

Ion Ratio Lower Upper

152 100

151 19.6 15.9 23.9

153 12.9 10.6 16.0



#51

3-Nitroaniline

Concen: 21.771 ng/ul

RT: 14.399 min Scan# 1940

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56




Tgt Ion:138 Resp: 270523


Ion Ratio Lower Upper

138 100

108 10.4 9.4 14.2

92 112.4 96.6 144.8









#52
Acenaphthene
Concen: 21.176 ng/ul
RT: 14.546 min Scan# 1965
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:153 Resp: 1236477
Ion Ratio Lower Upper
153 100
152 47.7 38.8 58.2
154 88.9 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 19.532 ng/ul
RT: 14.610 min Scan# 1976
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:184 Resp: 142230
Ion Ratio Lower Upper
184 100
63 58.0 50.5 75.7
154 63.0 45.8 68.8









#55
4-Nitrophenol
Concen: 20.398 ng/ul
RT: 14.710 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:109 Resp: 168978
Ion Ratio Lower Upper
109 100
139 148.2 126.6 189.8
65 130.1 117.1 175.7



#56
Dibenzofuran
Concen: 21.221 ng/ul
RT: 14.881 min Scan# 2022
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:168 Resp: 1683247
Ion Ratio Lower Upper
168 100
139 36.5 28.6 42.8









#57
2,4-Dinitrotoluene
Concen: 22.614 ng/ul
RT: 14.857 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:165 Resp: 361329
Ion Ratio Lower Upper
165 100
63 40.4 35.5 53.3
182 2.8 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.125 ng/ul
RT: 15.110 min Scan# 2061
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:232 Resp: 340930
Ion Ratio Lower Upper
232 100
131 45.1 37.3 55.9
130 2.1 1.8 2.6
166 31.5 24.7 37.1









#59
Diethylphthalate
Concen: 21.260 ng/ul
RT: 15.310 min Scan# 2095
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:149 Resp: 1293048
Ion Ratio Lower Upper
149 100
177 22.3 17.7 26.5
150 12.4 10.0 15.0



#61
Fluorene
Concen: 21.170 ng/ul
RT: 15.534 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:166 Resp: 1318978
Ion Ratio Lower Upper
166 100
165 98.4 78.2 117.2
167 13.6 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 21.623 ng/ul
RT: 15.528 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:204 Resp: 657551
Ion Ratio Lower Upper
204 100
206 32.9 26.3 39.5
141 58.1 45.4 68.0



#63
4-Nitroaniline
Concen: 22.033 ng/ul
RT: 15.563 min Scan# 2138
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:138 Resp: 240986
Ion Ratio Lower Upper
138 100
92 51.5 42.1 63.1
108 71.8 58.0 87.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.834 ng/ul
 RT: 15.622 min Scan# 2148
 Delta R.T. 0.000 min
 Lab File: BP012171.D
 Acq: 17 Oct 2022 18:56




Instrument :
 BNA_P
 ClientSampleId :
 SICV644


Tgt Ion:198 Resp: 210631
 Ion Ratio Lower Upper
 198 100
 182 5.1 5.0 7.4
 77 22.6 19.9 29.9



#67
 N-Nitrosodiphenylamine
 Concen: 21.691 ng/ul
 RT: 15.746 min Scan# 2169
 Delta R.T. 0.000 min
 Lab File: BP012171.D
 Acq: 17 Oct 2022 18:56

Tgt Ion:169 Resp: 1107723
 Ion Ratio Lower Upper
 169 100
 168 65.7 52.7 79.1
 167 36.3 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 22.506 ng/ul
RT: 16.428 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:248 Resp: 401076
Ion Ratio Lower Upper
248 100
250 100.1 78.1 117.1
141 73.3 54.5 81.7



#69
Hexachlorobenzene
Concen: 22.314 ng/ul
RT: 16.540 min Scan# 2304
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:284 Resp: 474254
Ion Ratio Lower Upper
284 100
142 33.4 29.6 44.4
249 29.9 24.8 37.2





#70

Atrazine

Concen: 21.140 ng/ul

RT: 16.710 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SICV644





Tgt Ion:200 Resp: 350672

Ion Ratio Lower Upper

200 100

173 26.0 21.0 31.4

215 48.5 39.0 58.6



#71

Pentachlorophenol

Concen: 20.542 ng/ul

RT: 16.892 min Scan# 2364

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56




Tgt Ion:266 Resp: 258329


Ion Ratio Lower Upper

266 100

264 62.6 48.7 73.1

268 63.3 50.6 75.8





#72

Phenanthrene

Concen: 20.814 ng/ul

RT: 17.287 min Scan# 2431

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SICV644





Tgt Ion:178 Resp: 1976796

Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 19.3 15.4 23.2



#74

Anthracene

Concen: 21.354 ng/ul

RT: 17.375 min Scan# 2446

Delta R.T. -0.006 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56




Tgt Ion:178 Resp: 1996397


Ion Ratio Lower Upper

178 100

179 15.6 12.2 18.2

176 18.2 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.059 ng/uL
RT: 13.275 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:216 Resp: 621843
Ion Ratio Lower Upper
216 100
214 78.0 63.6 95.4
218 48.0 38.4 57.6



#76
Pentachlorobenzene
Concen: 22.525 ng/uL
RT: 14.798 min Scan# 2008
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:250 Resp: 640281
Ion Ratio Lower Upper
250 100
248 63.2 49.0 73.4
252 63.7 49.8 74.6





#77

Carbazole

Concen: 20.715 ng/ul

RT: 17.651 min Scan# 2493

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SICV644





Tgt Ion:167 Resp: 1702883

Ion Ratio Lower Upper

167 100

166 21.3 16.6 24.8

139 12.8 10.1 15.1



#78

Di-n-butylphthalate

Concen: 21.529 ng/ul

RT: 18.198 min Scan# 2586

Delta R.T. -0.006 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56




Tgt Ion:149 Resp: 1927870


Ion Ratio Lower Upper

149 100

150 9.3 7.0 10.6

104 4.8 3.8 5.8









#80
Fluoranthene
Concen: 22.440 ng/ul
RT: 19.257 min Scan# 2766
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:202 Resp: 2076005
Ion Ratio Lower Upper
202 100
101 11.7 10.1 15.1
100 8.9 7.5 11.3



#82
Pyrene
Concen: 22.210 ng/ul
RT: 19.610 min Scan# 2826
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:202 Resp: 2159232
Ion Ratio Lower Upper
202 100
101 13.6 11.5 17.3
100 10.9 9.2 13.8









#83
Butylbenzylphthalate
Concen: 21.722 ng/ul
RT: 20.481 min Scan# 2975
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:149 Resp: 740479
Ion Ratio Lower Upper
149 100
91 72.3 56.5 84.7
206 19.9 16.2 24.4



#84
3,3'-Dichlorobenzidine
Concen: 21.054 ng/ul
RT: 21.257 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:252 Resp: 672210
Ion Ratio Lower Upper
252 100
254 64.3 51.4 77.0
126 11.4 9.3 13.9





#85

Benzo(a)anthracene

Concen: 21.109 ng/ul

RT: 21.316 min Scan# 3116

Delta R.T. -0.006 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SICV644





Tgt Ion:228 Resp: 1928001

Ion Ratio Lower Upper

228 100

229 19.8 16.1 24.1

226 27.1 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.108 ng/ul

RT: 21.239 min Scan# 3103

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56




Tgt Ion:149 Resp: 1017548


Ion Ratio Lower Upper

149 100

167 26.9 20.7 31.1

279 3.8 3.0 4.4









#87
Chrysene
Concen: 20.986 ng/ul
RT: 21.369 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:228 Resp: 1837407
Ion Ratio Lower Upper
228 100
226 30.0 23.8 35.8
229 19.8 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 19.749 ng/ul
RT: 22.163 min Scan# 3260
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:149 Resp: 1648526









#90
Benzo(b)fluoranthene
Concen: 21.158 ng/ul
RT: 23.010 min Scan# 3404
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56




Instrument :
BNA_P
ClientSampleId :
SICV644


Tgt Ion:252 Resp: 1964546
Ion Ratio Lower Upper
252 100
253 21.7 17.5 26.3
125 10.0 8.6 13.0



#91
Benzo(k)fluoranthene
Concen: 20.363 ng/ul
RT: 23.057 min Scan# 3412
Delta R.T. 0.000 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:252 Resp: 1873077
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.2
125 10.3 8.8 13.2





#93

Benzo(a)pyrene

Concen: 20.933 ng/ul

RT: 23.639 min Scan# 3511

Delta R.T. 0.000 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56

Instrument :

BNA_P

ClientSampleId :

SICV644





Tgt Ion:252 Resp: 1717968

Ion Ratio Lower Upper

252 100

253 22.0 17.4 26.0

125 11.0 9.6 14.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.425 ng/ul

RT: 26.245 min Scan# 3954

Delta R.T. -0.012 min

Lab File: BP012171.D

Acq: 17 Oct 2022 18:56




Tgt Ion:276 Resp: 2220746


Ion Ratio Lower Upper

276 100

138 23.8 20.7 31.1

277 25.3 19.7 29.5









#95
Dibenzo(a,h)anthracene
Concen: 21.402 ng/ul
RT: 26.263 min Scan# 3957
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Instrument :
BNA_P
ClientSampleId :
SICV644

Tgt Ion:278 Resp: 2002712

Ion	Ratio	Lower	Upper
278	100		
139	16.8	13.5	20.3
279	24.2	19.2	28.8



#96
Benzo(g,h,i)perylene
Concen: 21.547 ng/ul
RT: 27.015 min Scan# 4085
Delta R.T. -0.006 min
Lab File: BP012171.D
Acq: 17 Oct 2022 18:56

Tgt Ion:276 Resp: 2019888

Ion	Ratio	Lower	Upper
276	100		
138	22.0	18.7	28.1
277	23.7	19.1	28.7

