

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101821\
 Data File : BP007470.D
 Acq On : 18 Oct 2021 18:38
 Operator : CG/JU
 Sample : PB139924BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS924

Manual Integrations
 APPROVED

mohammad
 10/19/2021 12:36:48 PM

Quant Time: Oct 19 01:21:37 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 18 11:26:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	154471	20.00	ng/ul	0.00
20) Naphthalene-d8	10.763	136	621451	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.592	164	391186	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.345	188	828258	20.00	ng/ul	0.00
79) Chrysene-d12	21.439	240	810135	20.00	ng/ul	0.00
88) Perylene-d12	23.892	264	863324	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	27406	7.01	ng/uL	0.00
4) Pyridine-d5	3.793	84	371930	34.74	ng/ul	0.00
7) Phenol-d5	7.116	99	467650	38.77	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	268554	37.32	ng/ul	0.00
11) 2-Chlorophenol-d4	7.487	132	379451	39.98	ng/ul	0.00
15) 4-Methylphenol-d8	8.663	113	374758	40.33	ng/ul	0.00
21) Nitrobenzene-d5	9.110	128	179124	45.55	ng/ul	0.00
24) 2-Nitrophenol-d4	9.840	143	182244	49.61	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	366331	40.88	ng/ul	0.00
31) 4-Chloroaniline-d4	10.893	131	468742	38.69	ng/ul	0.00
46) Dimethylphthalate-d6	14.004	166	1086976	40.16	ng/ul	0.00
49) Acenaphthylene-d8	14.287	160	1326608	39.00	ng/ul	0.00
54) 4-Nitrophenol-d4	14.781	143	200730	46.22	ng/ul	0.00
60) Fluorene-d10	15.587	176	907845	40.08	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.704	200	147394	44.41	ng/ul	0.00
73) Anthracene-d10	17.445	188	1430736	40.68	ng/ul	0.00
81) Pyrene-d10	19.680	212	1715011	42.35	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.733	264	1697493	39.52	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.417	88	56634	13.31	ng/uL	91
5) Pyridine	3.811	79	385326	37.21	ng/ul	99
6) Benzaldehyde	7.087	77	309703	45.69	ng/ul	96
8) Phenol	7.140	94	484395	39.27	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.375	93	390040	39.18	ng/ul	98
12) 2-Chlorophenol	7.522	128	386744	39.69	ng/ul	97
13) 2-Methylphenol	8.399	108	369106	39.73	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.493	45	455781	35.93	ng/ul	98
16) Acetophenone	8.781	105	568049	39.96	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.769	70	279866	38.06	ng/ul	94
18) 4-Methylphenol	8.728	108	396325	40.05	ng/ul	98
19) Hexachloroethane	9.046	117	159383	39.07	ng/ul	98
22) Nitrobenzene	9.157	77	421254	41.72	ng/ul	98
23) Isophorone	9.687	82	820062	38.79	ng/ul	99
25) 2-Nitrophenol	9.869	139	200460	47.08	ng/ul	96
26) 2,4-Dimethylphenol	9.934	107	428939	39.45	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.175	93	504698	38.53	ng/ul	99
29) 2,4-Dichlorophenol	10.410	162	367882	41.77	ng/ul	98
30) Naphthalene	10.816	128	1185033	39.21	ng/ul	100
32) 4-Chloroaniline	10.916	127	492588	40.18	ng/ul	99
33) Hexachlorobutadiene	11.116	225	254931	41.23	ng/ul	97
34) Caprolactam	11.681	113	125778m	54.45	ng/ul	
35) 4-Chloro-3-methylphenol	12.040	107	397622	41.56	ng/ul	98
36) 2-Methylnaphthalene	12.422	142	804416	38.94	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101821\
 Data File : BP007470.D
 Acq On : 18 Oct 2021 18:38
 Operator : CG/JU
 Sample : PB139924BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS924


Manual Integrations
 APPROVED

mohammad
 10/19/2021 12:36:48 PM

Quant Time: Oct 19 01:21:37 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 18 11:26:58 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.640	142	808066	38.64	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.793	216	456082	40.34	ng/ul	99
40) Hexachlorocyclopentadiene	12.781	237	239503	33.59	ng/ul	98
41) 2,4,6-Trichlorophenol	13.028	196	299488	45.04	ng/ul	99
42) 2,4,5-Trichlorophenol	13.098	196	320113	46.24	ng/ul	96
43) 1,1'-Biphenyl	13.428	154	1074936	38.55	ng/ul	99
44) 2-Chloronaphthalene	13.469	162	831129	38.70	ng/ul	99
45) 2-Nitroaniline	13.663	65	237699	47.61	ng/ul	96
47) Dimethylphthalate	14.051	163	1075129	39.99	ng/ul	99
48) 2,6-Dinitrotoluene	14.163	165	213599	47.92	ng/ul	95
50) Acenaphthylene	14.316	152	1340472	39.09	ng/ul	100
51) 3-Nitroaniline	14.487	138	237376	47.79	ng/ul	98
52) Acenaphthene	14.657	153	871072	39.23	ng/ul	100
53) 2,4-Dinitrophenol	14.692	184	98586	47.88	ng/ul	99
55) 4-Nitrophenol	14.792	109	171767	43.78	ng/ul	97
56) Dibenzofuran	14.992	168	1281280	39.74	ng/ul	99
57) 2,4-Dinitrotoluene	14.951	165	305181	48.74	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.216	232	269642	46.94	ng/ul	99
59) Diethylphthalate	15.422	149	1083501	40.42	ng/ul	99
61) Fluorene	15.645	166	953928	39.05	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.639	204	489958	39.39	ng/ul	99
63) 4-Nitroaniline	15.651	138	226736	51.49	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.716	198	150490	43.84	ng/ul	99
67) N-Nitrosodiphenylamine	15.851	169	890633	39.94	ng/ul	98
68) 4-Bromophenyl-phenylether	16.539	248	321581	43.10	ng/ul	96
69) Hexachlorobenzene	16.657	284	374167	41.69	ng/ul	98
70) Atrazine	16.810	200	351936	41.69	ng/ul	99
71) Pentachlorophenol	16.998	266	229486	46.19	ng/ul	98
72) Phenanthrene	17.392	178	1673333	40.41	ng/ul	99
74) Anthracene	17.481	178	1655374	39.97	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.393	216	464307	38.65	ng/uL	99
76) Pentachlorobenzene	14.916	250	436057	38.94	ng/uL	99
77) Carbazole	17.745	167	1547767	42.55	ng/ul	100
78) Di-n-butylphthalate	18.310	149	1890219	42.30	ng/ul	99
80) Fluoranthene	19.357	202	2062002	41.03	ng/ul	99
82) Pyrene	19.710	202	2065652	40.76	ng/ul	98
83) Butylbenzylphthalate	20.586	149	858132	46.12	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.351	252	690804	44.82	ng/ul	98
85) Benzo(a)anthracene	21.422	228	2049562	41.18	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.357	149	1248823	43.90	ng/ul	99
87) Chrysene	21.474	228	1945321	40.34	ng/ul	99
89) Di-n-octyl phthalate	22.304	149	2186499	44.88	ng/ul	100
90) Benzo(b)fluoranthene	23.145	252	2159097	41.09	ng/ul	99
91) Benzo(k)fluoranthene	23.192	252	1996567	40.75	ng/ul	99
93) Benzo(a)pyrene	23.786	252	2066399	41.11	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.462	276	2427532	41.69	ng/ul	99
95) Dibenzo(a,h)anthracene	26.486	278	2054580	41.44	ng/ul	99
96) Benzo(g,h,i)perylene	27.256	276	2096844	41.81	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 13.31 ng/uL
RT: 3.417 min Scan# 56
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924



Tgt Ion: 88 Resp: 56634
Ion Ratio Lower Upper
88 100
43 26.5 24.2 36.2
58 69.7 62.5 93.7




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#5
Pyridine
Concen: 37.21 ng/uL
RT: 3.811 min Scan# 123
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38


Tgt Ion: 79 Resp: 385326
Ion Ratio Lower Upper
79 100
52 71.6 57.0 85.4
51 34.3 27.0 40.6





#6
Benzaldehyde
Concen: 45.69 ng/ul
RT: 7.087 min Scan# 680
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924




Tgt Ion: 77 Resp: 309703
Ion Ratio Lower Upper
77 100
105 101.3 78.2 117.4
106 97.3 74.6 112.0

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#8
Phenol
Concen: 39.27 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38



Tgt Ion: 94 Resp: 484395
Ion Ratio Lower Upper
94 100
65 27.3 21.9 32.9
66 38.6 29.9 44.9





#10
Bis(2-Chloroethyl)ether
Concen: 39.18 ng/ul
RT: 7.375 min Scan# 729
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38





Tgt Ion: 93 Resp: 390040

Ion	Ratio	Lower	Upper
93	100		
63	72.4	59.0	88.4
95	31.4	26.4	39.6

Instrument :
BNA_P
ClientSampleId :
SLCS924

Manual Integrations
APPROVED




mohammad
10/19/2021 12:36:48 PM




#12
2-Chlorophenol
Concen: 39.69 ng/ul
RT: 7.522 min Scan# 754
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38

Tgt Ion: 128 Resp: 386744


Ion	Ratio	Lower	Upper
128	100		
64	44.7	36.9	55.3
130	33.5	25.2	37.8





#13
2-Methylphenol
Concen: 39.73 ng/ul
RT: 8.399 min Scan# 903
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924




Tgt Ion:108 Resp: 369106
Ion Ratio Lower Upper
108 100
107 90.4 72.2 108.2

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#14
2,2'-oxybis(1-Chloropropane)
Concen: 35.93 ng/ul
RT: 8.493 min Scan# 919
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion: 45 Resp: 455781
Ion Ratio Lower Upper
45 100
77 16.3 12.3 18.5
79 12.3 9.4 14.0





#16
Acetophenone
Concen: 39.96 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampled :
SLCS924




Tgt Ion:105 Resp: 568049
Ion Ratio Lower Upper
105 100
77 74.7 61.8 92.6
51 24.5 20.8 31.2

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#17
N-Nitroso-di-n-propylamine
Concen: 38.06 ng/ul
RT: 8.769 min Scan# 966
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38



Tgt Ion: 70 Resp: 279866
Ion Ratio Lower Upper
70 100
42 51.7 46.3 69.5
101 10.3 7.4 11.0
130 21.4 16.2 24.2









#18
4-Methylphenol
Concen: 40.05 ng/ul
RT: 8.728 min Scan# 959
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38

Instrument :
BNA_P
ClientSampleId :
SLCS924

Tgt Ion:108 Resp: 396325
Ion Ratio Lower Upper
108 100
107 111.8 90.9 136.3




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#19
Hexachloroethane
Concen: 39.07 ng/ul
RT: 9.046 min Scan# 1013
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38


Tgt Ion:117 Resp: 159383
Ion Ratio Lower Upper
117 100
201 105.2 82.3 123.5
199 65.4 51.8 77.6





#22
Nitrobenzene
Concen: 41.72 ng/ul
RT: 9.157 min Scan# 1032
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampled :
SLCS924




Tgt Ion: 77 Resp: 421254
Ion Ratio Lower Upper
77 100
123 45.6 34.9 52.3
65 13.0 10.5 15.7

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#23
Isophorone
Concen: 38.79 ng/ul
RT: 9.687 min Scan# 1122
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion: 82 Resp: 820062
Ion Ratio Lower Upper
82 100
95 7.1 5.6 8.4
138 17.4 13.4 20.0





#25
2-Nitrophenol
Concen: 47.08 ng/ul
RT: 9.869 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampled :
SLCS924



Tgt Ion:139 Resp: 200460
Ion Ratio Lower Upper
139 100
65 48.5 40.3 60.5
109 33.1 23.7 35.5




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#26
2,4-Dimethylphenol
Concen: 39.45 ng/ul
RT: 9.934 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38

Tgt Ion:107 Resp: 428939
Ion Ratio Lower Upper
107 100
121 49.4 38.8 58.2
122 82.1 64.0 96.0









#27
Bis(2-Chloroethoxy)methane
Concen: 38.53 ng/ul
RT: 10.175 min Scan# 1205
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38

Tgt Ion: 93 Resp: 504698
Ion Ratio Lower Upper
93 100
95 32.3 26.2 39.4
123 11.1 8.3 12.5

Instrument :
BNA_P
ClientSampled :
SLCS924




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#29
2,4-Dichlorophenol
Concen: 41.77 ng/ul
RT: 10.410 min Scan# 1245
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38


Tgt Ion:162 Resp: 367882
Ion Ratio Lower Upper
162 100
164 65.5 52.9 79.3
98 35.9 30.6 45.8





#30
Naphthalene
Concen: 39.21 ng/ul
RT: 10.816 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924



Tgt Ion:128 Resp: 1185033
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.2 10.5 15.7




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#32
4-Chloroaniline
Concen: 40.18 ng/ul
RT: 10.916 min Scan# 1331
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38


Tgt Ion:127 Resp: 492588
Ion Ratio Lower Upper
127 100
129 31.6 25.8 38.6





#33
Hexachlorobutadiene
Concen: 41.23 ng/ul
RT: 11.116 min Scan# 1365
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampled :
SLCS924




Tgt Ion:225 Resp: 254931
Ion Ratio Lower Upper
225 100
223 60.1 50.6 76.0
227 61.8 51.0 76.6

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#34
Caprolactam
Concen: 54.45 ng/ul m
RT: 11.681 min Scan# 1461
Delta R.T. 0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion:113 Resp: 125778
Ion Ratio Lower Upper
113 100
55 148.3 130.1 195.1
56 116.1 96.2 144.4





#35
4-Chloro-3-methylphenol
Concen: 41.56 ng/ul
RT: 12.040 min Scan# 1522
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924




Tgt Ion:107 Resp: 397622
Ion Ratio Lower Upper
107 100
144 25.8 19.3 28.9
142 78.6 62.2 93.4

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#36
2-Methylnaphthalene
Concen: 38.94 ng/ul
RT: 12.422 min Scan# 1587
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion:142 Resp: 804416
Ion Ratio Lower Upper
142 100
141 87.8 71.8 107.6





#37
1-Methylnaphthalene
Concen: 38.64 ng/ul
RT: 12.640 min Scan# 1624
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampled :
SLCS924




Tgt Ion:142 Resp: 808066
Ion Ratio Lower Upper
142 100
141 92.4 73.5 110.3
116 4.1 3.1 4.7

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#39
1,2,4,5-Tetrachlorobenzene
Concen: 40.34 ng/ul
RT: 12.793 min Scan# 1650
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion:216 Resp: 456082
Ion Ratio Lower Upper
216 100
214 78.6 63.1 94.7
179 21.0 17.4 26.0
108 16.5 15.0 22.6





#40
Hexachlorocyclopentadiene
Concen: 33.59 ng/ul
RT: 12.781 min Scan# 1648
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampled :
SLCS924




Tgt Ion:237 Resp: 239503
Ion Ratio Lower Upper
237 100
235 63.1 49.4 74.2
272 11.5 9.8 14.6

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#41
2,4,6-Trichlorophenol
Concen: 45.04 ng/ul
RT: 13.028 min Scan# 1690
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38



Tgt Ion:196 Resp: 299488
Ion Ratio Lower Upper
196 100
198 94.6 75.8 113.6
200 30.2 23.0 34.4









#42
2,4,5-Trichlorophenol
Concen: 46.24 ng/ul
RT: 13.098 min Scan# 1702
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38

Tgt Ion:196 Resp: 320113
Ion Ratio Lower Upper
196 100
198 97.9 75.1 112.7
200 31.2 25.8 38.8

Instrument :
BNA_P
ClientSampled :
SLCS924




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#43
1,1'-Biphenyl
Concen: 38.55 ng/ul
RT: 13.428 min Scan# 1758
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38

Tgt Ion:154 Resp: 1074936
Ion Ratio Lower Upper
154 100
153 39.7 32.0 48.0
76 12.2 10.3 15.5









#44
2-Chloronaphthalene
Concen: 38.70 ng/ul
RT: 13.469 min Scan# 1765
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38

Tgt Ion:	162	Resp:	831129
Ion	Ratio	Lower	Upper
162	100		
164	32.5	26.5	39.7
127	39.8	31.8	47.6

Instrument :
BNA_P
ClientSampleId :
SLCS924




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#45
2-Nitroaniline
Concen: 47.61 ng/ul
RT: 13.663 min Scan# 1798
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38


Tgt Ion:	65	Resp:	237699
Ion	Ratio	Lower	Upper
65	100		
92	69.0	55.9	83.9
138	110.7	83.4	125.2





#47
Dimethylphthalate
Concen: 39.99 ng/ul
RT: 14.051 min Scan# 1864
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924




Tgt Ion:163 Resp: 1075129
Ion Ratio Lower Upper
163 100
194 4.1 3.2 4.8
164 10.2 7.9 11.9

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#48
2,6-Dinitrotoluene
Concen: 47.92 ng/ul
RT: 14.163 min Scan# 1883
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion:165 Resp: 213599
Ion Ratio Lower Upper
165 100
89 56.9 49.0 73.4
121 21.3 18.4 27.6





#50
Acenaphthylene
Concen: 39.09 ng/ul
RT: 14.316 min Scan# 1909
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924




Tgt Ion:152 Resp: 1340472
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.8
153 12.9 10.6 15.8

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#51
3-Nitroaniline
Concen: 47.79 ng/ul
RT: 14.487 min Scan# 1938
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38



Tgt Ion:138 Resp: 237376
Ion Ratio Lower Upper
138 100
108 13.1 13.0 19.4
92 120.5 95.7 143.5









#52
Acenaphthene
Concen: 39.23 ng/ul
RT: 14.657 min Scan# 1967
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38

Tgt Ion:153 Resp: 871072
Ion Ratio Lower Upper
153 100
152 46.6 37.9 56.9
154 87.7 70.1 105.1

Instrument :
BNA_P
ClientSampleId :
SLCS924




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#53
2,4-Dinitrophenol
Concen: 47.88 ng/ul
RT: 14.692 min Scan# 1973
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38


Tgt Ion:184 Resp: 98586
Ion Ratio Lower Upper
184 100
63 60.1 48.8 73.2
154 70.3 56.2 84.4





#55
4-Nitrophenol
Concen: 43.78 ng/ul
RT: 14.792 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924




Tgt Ion:109 Resp: 171767
Ion Ratio Lower Upper
109 100
139 125.5 100.2 150.2
65 116.1 98.8 148.2

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#56
Dibenzofuran
Concen: 39.74 ng/ul
RT: 14.992 min Scan# 2024
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion:168 Resp: 1281280
Ion Ratio Lower Upper
168 100
139 37.1 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 48.74 ng/ul
RT: 14.951 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924




Tgt Ion:165 Resp: 305181
Ion Ratio Lower Upper
165 100
63 39.3 35.0 52.6
182 2.7 2.3 3.5

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#58
2,3,4,6-Tetrachlorophenol
Concen: 46.94 ng/ul
RT: 15.216 min Scan# 2062
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion:232 Resp: 269642
Ion Ratio Lower Upper
232 100
131 44.5 35.1 52.7
130 2.6 2.0 3.0
166 31.1 25.8 38.8





#59
Diethylphthalate
Concen: 40.42 ng/ul
RT: 15.422 min Scan# 2097
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924



Tgt Ion:149 Resp: 1083501
Ion Ratio Lower Upper
149 100
177 22.4 17.4 26.2
150 12.6 10.0 15.0




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#61
Fluorene
Concen: 39.05 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38


Tgt Ion:166 Resp: 953928
Ion Ratio Lower Upper
166 100
165 96.9 78.4 117.6
167 13.4 10.9 16.3





#62
4-Chlorophenyl-phenylether
Concen: 39.39 ng/ul
RT: 15.639 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampled :
SLCS924



Tgt Ion:204 Resp: 489958
Ion Ratio Lower Upper
204 100
206 33.0 26.5 39.7
141 59.8 47.4 71.0




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#63
4-Nitroaniline
Concen: 51.49 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38


Tgt Ion:138 Resp: 226736
Ion Ratio Lower Upper
138 100
92 50.3 42.1 63.1
108 86.8 70.3 105.5





#66
4,6-Dinitro-2-methylphenol
Concen: 43.84 ng/ul
RT: 15.716 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampled :
SLCS924




Tgt Ion:198 Resp: 150490
Ion Ratio Lower Upper
198 100
182 4.1 3.3 4.9
77 23.6 18.4 27.6

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#67
N-Nitrosodiphenylamine
Concen: 39.94 ng/ul
RT: 15.851 min Scan# 2170
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion:169 Resp: 890633
Ion Ratio Lower Upper
169 100
168 66.8 52.5 78.7
167 36.7 28.4 42.6





#68
4-Bromophenyl-phenylether
Concen: 43.10 ng/ul
RT: 16.539 min Scan# 2287
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924




Tgt Ion:248 Resp: 321581
Ion Ratio Lower Upper
248 100
250 97.4 78.9 118.3
141 73.3 64.3 96.5

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#69
Hexachlorobenzene
Concen: 41.69 ng/ul
RT: 16.657 min Scan# 2307
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38



Tgt Ion:284 Resp: 374167
Ion Ratio Lower Upper
284 100
142 35.4 29.8 44.8
249 32.5 26.6 39.8









#70
Atrazine
Concen: 41.69 ng/ul
RT: 16.810 min Scan# 2333
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38

Tgt Ion:	200	Resp:	351936
Ion Ratio	Lower	Upper	
200	100		
173	26.1	20.9	31.3
215	48.6	38.1	57.1

Instrument :
BNA_P
ClientSampled :
SLCS924




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#71
Pentachlorophenol
Concen: 46.19 ng/ul
RT: 16.998 min Scan# 2365
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38

Tgt Ion:	266	Resp:	229486
Ion Ratio	Lower	Upper	
266	100		
264	60.3	50.6	75.8
268	63.2	51.3	76.9









#72
Phenanthrene
Concen: 40.41 ng/ul
RT: 17.392 min Scan# 2432
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38

Tgt Ion:178 Resp: 1673333
Ion Ratio Lower Upper
178 100
179 15.4 12.1 18.1
176 19.4 15.2 22.8

Instrument :
BNA_P
ClientSampled :
SLCS924




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#74
Anthracene
Concen: 39.97 ng/ul
RT: 17.481 min Scan# 2447
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38


Tgt Ion:178 Resp: 1655374
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.4
176 18.5 14.6 22.0





#75
1,2,3,4-Tetrachlorobenzene
Concen: 38.65 ng/uL
RT: 13.393 min Scan# 1752
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampled :
SLCS924




Tgt Ion:216 Resp: 464307
Ion Ratio Lower Upper
216 100
214 78.1 62.0 93.0
218 48.4 37.6 56.4

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#76
Pentachlorobenzene
Concen: 38.94 ng/uL
RT: 14.916 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion:250 Resp: 436057
Ion Ratio Lower Upper
250 100
248 63.3 50.1 75.1
252 63.9 50.6 76.0





#77
 Carbazole
 Concen: 42.55 ng/ul
 RT: 17.745 min Scan# 2492
 Delta R.T. -0.006 min
 Lab File: BP007470.D
 Acq: 18 Oct 2021 18:38




Instrument :
 BNA_P
 ClientSampled :
 SLCS924




Tgt Ion:167 Resp: 1547767
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.3 25.9
 139 13.1 10.4 15.6

Manual Integrations
 APPROVED



mohammad
 10/19/2021 12:36:48 PM




#78
 Di-n-butylphthalate
 Concen: 42.30 ng/ul
 RT: 18.310 min Scan# 2588
 Delta R.T. -0.006 min
 Lab File: BP007470.D
 Acq: 18 Oct 2021 18:38




Tgt Ion:149 Resp: 1890219
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.3 10.9
 104 5.0 4.2 6.4





#80
Fluoranthene
Concen: 41.03 ng/ul
RT: 19.357 min Scan# 2766
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924




Tgt Ion:202 Resp: 2062002
Ion Ratio Lower Upper
202 100
101 10.1 8.4 12.6
100 8.0 6.3 9.5

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#82
Pyrene
Concen: 40.76 ng/ul
RT: 19.710 min Scan# 2826
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion:202 Resp: 2065652
Ion Ratio Lower Upper
202 100
101 11.9 9.9 14.9
100 9.1 7.9 11.9









#83
Butylbenzylphthalate
Concen: 46.12 ng/ul
RT: 20.586 min Scan# 2975
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38





Tgt Ion:149 Resp: 858132
Ion Ratio Lower Upper
149 100
91 69.2 55.9 83.9
206 21.0 17.4 26.0



#84
3,3'-Dichlorobenzidine
Concen: 44.82 ng/ul
RT: 21.351 min Scan# 3105
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion:252 Resp: 690804
Ion Ratio Lower Upper
252 100
254 64.7 51.1 76.7
126 10.5 10.0 15.0



Instrument :
BNA_P
ClientSampleId :
SLCS924


Manual Integrations
APPROVED

mohammad
10/19/2021 12:36:48 PM



#85
Benzo(a)anthracene
Concen: 41.18 ng/ul
RT: 21.422 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampled :
SLCS924



Tgt Ion:228 Resp: 2049562
Ion Ratio Lower Upper
228 100
229 19.9 15.7 23.5
226 27.1 21.4 32.0




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#86
Bis(2-ethylhexyl)phthalate
Concen: 43.90 ng/ul
RT: 21.357 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38


Tgt Ion:149 Resp: 1248823
Ion Ratio Lower Upper
149 100
167 27.8 22.6 33.8
279 4.3 3.3 4.9





#87
Chrysene
Concen: 40.34 ng/ul
RT: 21.474 min Scan# 3126
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924



Tgt Ion:228 Resp: 1945321
Ion Ratio Lower Upper
228 100
226 30.2 23.7 35.5
229 19.9 16.0 24.0




Manual Integrations
APPROVED


mohammad
10/19/2021 12:36:48 PM



#89
Di-n-octyl phthalate
Concen: 44.88 ng/ul
RT: 22.304 min Scan# 3267
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38


Tgt Ion:149 Resp: 2186499





#90
Benzo(b)fluoranthene
Concen: 41.09 ng/ul
RT: 23.145 min Scan# 3410
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampleId :
SLCS924




Tgt Ion:252 Resp: 2159097
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.2
125 9.0 7.8 11.6

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#91
Benzo(k)fluoranthene
Concen: 40.75 ng/ul
RT: 23.192 min Scan# 3418
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion:252 Resp: 1996567
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 9.1 8.0 12.0






#93
Benzo(a)pyrene
Concen: 41.11 ng/ul
RT: 23.786 min Scan# 3519
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38





Tgt Ion:252 Resp: 2066399
Ion Ratio Lower Upper
252 100
253 21.8 17.3 25.9
125 10.0 8.3 12.5




Instrument :
BNA_P
ClientSampled :
SLCS924

Manual Integrations
APPROVED



mohammad
10/19/2021 12:36:48 PM




#94
Indeno(1,2,3-cd)pyrene
Concen: 41.69 ng/ul
RT: 26.462 min Scan# 3974
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Tgt Ion:276 Resp: 2427532
Ion Ratio Lower Upper
276 100
138 21.1 17.8 26.8
277 24.7 19.8 29.6





#95
Dibenzo(a,h)anthracene
Concen: 41.44 ng/ul
RT: 26.486 min Scan# 3978
Delta R.T. -0.006 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38




Instrument :
BNA_P
ClientSampled :
SLCS924




Tgt Ion:278 Resp: 2054580
Ion Ratio Lower Upper
278 100
139 15.2 12.8 19.2
279 24.4 19.1 28.7

Manual Integrations
APPROVED

mohammad
10/19/2021 12:36:48 PM



#96
Benzo(g,h,i)perylene
Concen: 41.81 ng/ul
RT: 27.256 min Scan# 4109
Delta R.T. 0.000 min
Lab File: BP007470.D
Acq: 18 Oct 2021 18:38



Tgt Ion:276 Resp: 2096844
Ion Ratio Lower Upper
276 100
138 19.7 17.5 26.3
277 22.9 18.5 27.7

