

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101821\
 Data File : BP007476.D
 Acq On : 18 Oct 2021 22:02
 Operator : CG/JU
 Sample : M4204-05
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 JDGG5

Quant Time: Oct 19 01:24:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 18 11:26:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	189201	20.00	ng/ul	0.00
20) Naphthalene-d8	10.763	136	769610	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.593	164	495205	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.345	188	1082837	20.00	ng/ul	0.00
79) Chrysene-d12	21.433	240	1132542	20.00	ng/ul	0.00
88) Perylene-d12	23.886	264	1171376	20.00	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	27841	5.81	ng/uL	0.00
4) Pyridine-d5	3.799	84	132878	10.13	ng/ul	0.00
7) Phenol-d5	7.111	99	98690	6.68	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	274792	31.18	ng/ul	0.00
11) 2-Chlorophenol-d4	7.487	132	285295	24.54	ng/ul	0.00
15) 4-Methylphenol-d8	8.658	113	186641	16.40	ng/ul	0.00
21) Nitrobenzene-d5	9.111	128	181481	37.26	ng/ul	0.00
24) 2-Nitrophenol-d4	9.840	143	175238	38.52	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	331665	29.89	ng/ul	0.00
31) 4-Chloroaniline-d4	10.893	131	477911	31.85	ng/ul	0.00
46) Dimethylphthalate-d6	14.004	166	1216580	35.51	ng/ul	0.00
49) Acenaphthylene-d8	14.287	160	1451647	33.72	ng/ul	0.00
54) 4-Nitrophenol-d4	14.769	143	36466	6.63	ng/ul	0.00
60) Fluorene-d10	15.587	176	1040566	36.29	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.698	200	138564	31.94	ng/ul	0.00
73) Anthracene-d10	17.445	188	1772040	38.53	ng/ul	0.00
81) Pyrene-d10	19.675	212	2114633	37.35	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.727	264	2118126	36.34	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101821\
Data File : BP007476.D
Acq On : 18 Oct 2021 22:02
Operator : CG/JU
Sample : M4204-05
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
JDGG5

Quant Time: Oct 19 01:24:32 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 18 11:26:58 2021
Response via : Initial Calibration

