

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101821\
 Data File : BP007482.D
 Acq On : 19 Oct 2021 03:40
 Operator : CG/JU
 Sample : M4196-20
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 JDGD7

Quant Time: Oct 19 04:08:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 18 11:26:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	212412	20.00	ng/ul	0.00
20) Naphthalene-d8	10.763	136	863990	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.593	164	565350	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.345	188	1221184	20.00	ng/ul	0.00
79) Chrysene-d12	21.427	240	1291424	20.00	ng/ul	-0.01
88) Perylene-d12	23.880	264	1331919	20.00	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	25690	4.78	ng/uL	0.00
4) Pyridine-d5	3.793	84	346235	23.52	ng/ul	0.00
7) Phenol-d5	7.111	99	451209	27.20	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	258183	26.10	ng/ul	0.00
11) 2-Chlorophenol-d4	7.487	132	357544	27.40	ng/ul	0.00
15) 4-Methylphenol-d8	8.658	113	358485	28.06	ng/ul	0.00
21) Nitrobenzene-d5	9.110	128	167204	30.58	ng/ul	0.00
24) 2-Nitrophenol-d4	9.840	143	170682	33.42	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.375	165	345622	27.75	ng/ul	0.00
31) 4-Chloroaniline-d4	10.887	131	489627	29.07	ng/ul	0.00
46) Dimethylphthalate-d6	13.998	166	1088858	27.84	ng/ul	0.00
49) Acenaphthylene-d8	14.281	160	1337090	27.20	ng/ul	0.00
54) 4-Nitrophenol-d4	14.769	143	166013	26.45	ng/ul	0.00
60) Fluorene-d10	15.587	176	947792	28.96	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.692	200	103222	21.09	ng/ul	-0.01
73) Anthracene-d10	17.445	188	1514615	29.20	ng/ul	0.00
81) Pyrene-d10	19.675	212	1807575	28.00	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.721	264	1775100	26.78	ng/ul	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101821\
Data File : BP007482.D
Acq On : 19 Oct 2021 03:40
Operator : CG/JU
Sample : M4196-20
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
JDGD7

Quant Time: Oct 19 04:08:10 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 18 11:26:58 2021
Response via : Initial Calibration

