

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101821\
 Data File : BP007503.D
 Acq On : 19 Oct 2021 22:10
 Operator : CG/JU
 Sample : M4196-15
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 JDGD1

Quant Time: Oct 20 02:05:31 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 18 11:26:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	216465	20.00	ng/ul	0.00
20) Naphthalene-d8	10.757	136	865762	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.587	164	570706	20.00	ng/ul	-0.01
64) Phenanthrene-d10	17.339	188	1226598	20.00	ng/ul	-0.01
79) Chrysene-d12	21.427	240	1295876	20.00	ng/ul	-0.01
88) Perylene-d12	23.874	264	1339160	20.00	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	7852	1.43	ng/uL	0.00
4) Pyridine-d5	3.799	84	52840	3.52	ng/ul	0.00
7) Phenol-d5	7.111	99	134579	7.96	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	81926	8.13	ng/ul	0.00
11) 2-Chlorophenol-d4	7.481	132	105727	7.95	ng/ul	0.00
15) 4-Methylphenol-d8	8.658	113	80062	6.15	ng/ul	0.00
21) Nitrobenzene-d5	9.110	128	48011	8.76	ng/ul	0.00
24) 2-Nitrophenol-d4	9.834	143	45457	8.88	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.375	165	97679	7.83	ng/ul	0.00
31) 4-Chloroaniline-d4	10.887	131	134543	7.97	ng/ul	0.00
46) Dimethylphthalate-d6	13.998	166	348366	8.82	ng/ul	0.00
49) Acenaphthylene-d8	14.281	160	406485	8.19	ng/ul	0.00
54) 4-Nitrophenol-d4	14.769	143	39229	6.19	ng/ul	0.00
60) Fluorene-d10	15.581	176	311477	9.43	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.692	200	20661	4.20	ng/ul	-0.01
73) Anthracene-d10	17.439	188	483622	9.28	ng/ul	-0.01
81) Pyrene-d10	19.669	212	648717	10.01	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.710	264	615444	9.24	ng/ul	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101821\
Data File : BP007503.D
Acq On : 19 Oct 2021 22:10
Operator : CG/JU
Sample : M4196-15
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
JDGD1

Quant Time: Oct 20 02:05:31 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 18 11:26:58 2021
Response via : Initial Calibration

