

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP101821\  
 Data File : BP007510.D  
 Acq On : 20 Oct 2021 04:23  
 Operator : CG/JU  
 Sample : PB140007BL  
 Misc :  
 ALS Vial : 65 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SBLK007

Quant Time: Oct 20 05:08:35 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_P\METHODS\SFAM-EPA-BP101421.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Mon Oct 18 11:26:58 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------------|--------|------|----------|-------|-------|----------|
| -----                         |        |      |          |       |       |          |
| Internal Standards            |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.952  | 152  | 240106   | 20.00 | ng/ul | 0.00     |
| 20) Naphthalene-d8            | 10.758 | 136  | 977135   | 20.00 | ng/ul | 0.00     |
| 38) Acenaphthene-d10          | 14.587 | 164  | 635631   | 20.00 | ng/ul | -0.01    |
| 64) Phenanthrene-d10          | 17.340 | 188  | 1379968  | 20.00 | ng/ul | -0.01    |
| 79) Chrysene-d12              | 21.428 | 240  | 1447811  | 20.00 | ng/ul | -0.01    |
| 88) Perylene-d12              | 23.874 | 264  | 1502331  | 20.00 | ng/ul | -0.02    |
| System Monitoring Compounds   |        |      |          |       |       |          |
| 3) 1,4-Dioxane-d8             | 3.381  | 96   | 32984    | 5.43  | ng/uL | 0.00     |
| 4) Pyridine-d5                | 3.793  | 84   | 462062   | 27.77 | ng/ul | 0.00     |
| 7) Phenol-d5                  | 7.111  | 99   | 560297   | 29.88 | ng/ul | 0.00     |
| 9) Bis-(2-Chloroethyl)eth...  | 7.281  | 67   | 325133   | 29.07 | ng/ul | 0.00     |
| 11) 2-Chlorophenol-d4         | 7.481  | 132  | 442628   | 30.01 | ng/ul | 0.00     |
| 15) 4-Methylphenol-d8         | 8.658  | 113  | 446236   | 30.90 | ng/ul | 0.00     |
| 21) Nitrobenzene-d5           | 9.111  | 128  | 211773   | 34.25 | ng/ul | 0.00     |
| 24) 2-Nitrophenol-d4          | 9.834  | 143  | 220919   | 38.25 | ng/ul | 0.00     |
| 28) 2,4-Dichlorophenol-d3     | 10.375 | 165  | 421759   | 29.94 | ng/ul | 0.00     |
| 31) 4-Chloroaniline-d4        | 10.887 | 131  | 569602   | 29.90 | ng/ul | 0.00     |
| 46) Dimethylphthalate-d6      | 13.999 | 166  | 1349314  | 30.68 | ng/ul | 0.00     |
| 49) Acenaphthylene-d8         | 14.281 | 160  | 1608472  | 29.10 | ng/ul | 0.00     |
| 54) 4-Nitrophenol-d4          | 14.769 | 143  | 227935   | 32.30 | ng/ul | 0.00     |
| 60) Fluorene-d10              | 15.581 | 176  | 1136031  | 30.87 | ng/ul | -0.01    |
| 65) 4,6-Dinitro-2-methylph... | 15.693 | 200  | 142345   | 25.74 | ng/ul | -0.01    |
| 73) Anthracene-d10            | 17.440 | 188  | 1870304  | 31.91 | ng/ul | -0.01    |
| 81) Pyrene-d10                | 19.669 | 212  | 2250646  | 31.10 | ng/ul | -0.01    |
| 92) Benzo(a)pyrene-d12        | 23.716 | 264  | 2277032  | 30.46 | ng/ul | -0.02    |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP101821\  
Data File : BP007510.D  
Acq On : 20 Oct 2021 04:23  
Operator : CG/JU  
Sample : PB140007BL  
Misc :  
ALS Vial : 65 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
SBLK007

Quant Time: Oct 20 05:08:35 2021  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_P\METHODS\SFAM-EPA-BP101421.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Mon Oct 18 11:26:58 2021  
Response via : Initial Calibration

