

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101821\
 Data File : BP007521.D
 Acq On : 20 Oct 2021 10:40
 Operator : CG/JU
 Sample : M4246-04
 Misc :
 ALS Vial : 76 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YAT53

Quant Time: Oct 20 11:08:00 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 18 11:26:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	181969	20.00	ng/ul	0.00
20) Naphthalene-d8	10.757	136	747475	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.586	164	485126	20.00	ng/ul	-0.01
64) Phenanthrene-d10	17.339	188	1044330	20.00	ng/ul	-0.01
79) Chrysene-d12	21.427	240	1095437	20.00	ng/ul	-0.01
88) Perylene-d12	23.874	264	1116720	20.00	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	9319	2.02	ng/uL	0.00
4) Pyridine-d5	3.799	84	112283	8.90	ng/ul	0.00
7) Phenol-d5	7.110	99	107592	7.57	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	281736	33.24	ng/ul	0.00
11) 2-Chlorophenol-d4	7.481	132	301464	26.97	ng/ul	0.00
15) 4-Methylphenol-d8	8.657	113	200727	18.34	ng/ul	0.00
21) Nitrobenzene-d5	9.110	128	186441	39.41	ng/ul	0.00
24) 2-Nitrophenol-d4	9.834	143	189079	42.79	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.375	165	350304	32.50	ng/ul	0.00
31) 4-Chloroaniline-d4	10.887	131	344256	23.62	ng/ul	0.00
46) Dimethylphthalate-d6	13.998	166	1286329	38.33	ng/ul	0.00
49) Acenaphthylene-d8	14.281	160	1454171	34.48	ng/ul	0.00
54) 4-Nitrophenol-d4	14.769	143	40927	7.60	ng/ul	0.00
60) Fluorene-d10	15.581	176	1101746	39.23	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.692	200	154677	36.96	ng/ul	-0.01
73) Anthracene-d10	17.439	188	1925632	43.42	ng/ul	-0.01
81) Pyrene-d10	19.669	212	2304357	42.08	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.715	264	2288254	41.18	ng/ul	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101821\
Data File : BP007521.D
Acq On : 20 Oct 2021 10:40
Operator : CG/JU
Sample : M4246-04
Misc :
ALS Vial : 76 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
YAT53

Quant Time: Oct 20 11:08:00 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 18 11:26:58 2021
Response via : Initial Calibration

