

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP101919\
 Data File : BP000925.D
 Acq On : 19 Oct 2019 19:27
 Operator : JU
 Sample : PB123974BL
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK74

Quant Time: Oct 19 23:05:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP100919MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 19 22:14:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	133725	20.00	ng/ul	0.00
18) Naphthalene-d8	8.02	136	497790	20.00	ng/ul	0.00
36) Acenaphthene-d10	9.79	164	284529	20.00	ng/ul	0.00
62) Phenanthrene-d10	11.28	188	541673	20.00	ng/ul	0.00
78) Chrysene-d12	13.96	240	529412	20.00	ng/ul	0.00
86) Perylene-d12	15.43	264	609020	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	2.33	96	20498	6.36	ng/uL	0.00
5) Phenol-d5	6.38	99	323459	31.36	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.42	67	159425	30.91	ng/ul	0.00
9) 2-Chlorophenol-d4	6.51	132	287014	33.03	ng/ul	0.00
13) 4-Methylphenol-d8	7.13	113	264441	33.56	ng/ul	0.00
19) Nitrobenzene-d5	7.30	128	132450	35.00	ng/ul	0.00
22) 2-Nitrophenol-d4	7.63	143	147858	36.69	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	7.88	165	256122	33.21	ng/ul	0.00
29) 4-Chloroaniline-d4	8.09	131	347036	40.46	ng/ul	0.00
44) Dimethylphthalate-d6	9.48	166	693642	34.83	ng/ul	0.00
47) Acenaphthylene-d8	9.63	160	823082	33.59	ng/ul	0.00
52) 4-Nitrophenol-d4	9.90	143	93674	30.53	ng/ul	-0.01
58) Fluorene-d10	10.31	176	575702	33.78	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	10.39	200	67167	26.04	ng/ul	0.00
71) Anthracene-d10	11.34	188	871961	34.75	ng/ul	0.00
79) Pyrene-d10	12.72	212	950026	31.73	ng/ul	0.00
90) Benzo(a)pyrene-d12	15.35	264	1031822	34.32	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP101919\
Data File : BP000925.D
Acq On : 19 Oct 2019 19:27
Operator : JU
Sample : PB123974BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK74

Quant Time: Oct 19 23:05:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP100919MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 19 22:14:11 2019
Response via : Initial Calibration

