



Benzaldehyde

Response Ratio




$R = -7.270e-002 A^2 + 8.942e-001 A + 1.268e-001$
Coef of Det (r^2) = 0.995999 Curve Fit: Quadratic
Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP101920.M
Calibration Table Last Updated: Mon Oct 19 14:54:24 2020

Benzoic acid



$\text{Response} = 2.092\text{e-}001 * \text{Amt} - 5.273\text{e-}002$
 Coef of Det (r^2) = 0.993133 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP101920.M
 Calibration Table Last Updated: Mon Oct 19 14:54:24 2020

2,4-Dinitrophenol



Response = 2.016e-001 * Amt - 4.865e-002
 Coef of Det (r^2) = 0.995258 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP101920.M
 Calibration Table Last Updated: Mon Oct 19 14:54:24 2020