

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101922\
 Data File : BP012203.D
 Acq On : 19 Oct 2022 13:57
 Operator : CG/JU
 Sample : SST08084
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080649

Quant Time: Oct 20 06:39:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 20 00:51:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	485008	20.000	ng/u1	0.00
20) Naphthalene-d8	10.634	136	2102028	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.481	164	1278938	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.239	188	2533923	20.000	ng/u1	0.00
79) Chrysene-d12	21.339	240	1788472	20.000	ng/u1	0.00
88) Perylene-d12	23.745	264	1661589	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	377872	31.531	ng/uL	0.00
4) Pyridine-d5	3.681	84	2822148	81.722	ng/u1	0.00
7) Phenol-d5	7.005	99	3447975	82.347	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.169	67	1993950	78.522	ng/u1	0.00
11) 2-Chlorophenol-d4	7.363	132	2630870	82.546	ng/u1	0.00
15) 4-Methylphenol-d8	8.557	113	2654484	81.569	ng/u1	0.01
21) Nitrobenzene-d5	9.005	128	1336288	85.883	ng/u1	0.00
24) 2-Nitrophenol-d4	9.722	143	1502347	89.494	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.263	165	2585626	82.113	ng/u1	0.00
31) 4-Chloroaniline-d4	10.787	131	3545969	79.259	ng/u1	0.00
46) Dimethylphthalate-d6	13.898	166	6873622	77.262	ng/u1	0.00
49) Acenaphthylene-d8	14.175	160	7852460	77.327	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	1398184	83.889	ng/u1	0.02
60) Fluorene-d10	15.481	176	5522214	72.151	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.616	200	1248828	86.633	ng/u1	0.01
73) Anthracene-d10	17.345	188	8133133	73.226	ng/u1	0.00
81) Pyrene-d10	19.580	212	8365457	82.154	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.592	264	6657879	79.901	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	405907	31.473	ng/uL	98
5) Pyridine	3.699	79	2742519	80.797	ng/u1	96
6) Benzaldehyde	6.975	77	1449586	73.656	ng/u1	97
8) Phenol	7.034	94	3544873	80.923	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.263	93	2797617	79.126	ng/u1	99
12) 2-Chlorophenol	7.393	128	2788075	81.076	ng/u1	99
13) 2-Methylphenol	8.281	108	2694290	81.696	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.369	45	3664401	77.936	ng/u1	99
16) Acetophenone	8.669	105	3868980	76.965	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.669	70	1944167	76.405	ng/u1	97
18) 4-Methylphenol	8.622	108	2893471	80.561	ng/u1	97
19) Hexachloroethane	8.905	117	1100020	80.982	ng/u1	99
22) Nitrobenzene	9.046	77	3046907	80.372	ng/u1	97
23) Isophorone	9.587	82	5959271	81.900	ng/u1	99
25) 2-Nitrophenol	9.757	139	1625485	85.224	ng/u1	98
26) 2,4-Dimethylphenol	9.816	107	2971367	79.035	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.057	93	3687986	77.425	ng/u1	100
29) 2,4-Dichlorophenol	10.287	162	2602510	79.867	ng/u1	98
30) Naphthalene	10.687	128	8366154	74.417	ng/u1	99
32) 4-Chloroaniline	10.810	127	3608716	78.147	ng/u1	98
33) Hexachlorobutadiene	10.963	225	1608560	78.485	ng/u1	99
34) Caprolactam	11.640	113	471321	46.594	ng/u1	98
35) 4-Chloro-3-methylphenol	11.940	107	2830964	80.468	ng/u1	99
36) 2-Methylnaphthalene	12.298	142	5724511	74.840	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101922\
 Data File : BP012203.D
 Acq On : 19 Oct 2022 13:57
 Operator : CG/JU
 Sample : SST08084
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080649

Quant Time: Oct 20 06:39:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 20 00:51:41 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.522	142	5624970	73.647	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.669	216	3033363	78.057	ng/ul	99
40) Hexachlorocyclopentadiene	12.640	237	1762189	85.292	ng/ul	97
41) 2,4,6-Trichlorophenol	12.916	196	2096094	83.575	ng/ul	99
42) 2,4,5-Trichlorophenol	12.987	196	2266294	82.834	ng/ul	100
43) 1,1'-Biphenyl	13.316	154	7028764	73.893	ng/ul	98
44) 2-Chloronaphthalene	13.357	162	5665029	75.119	ng/ul	98
45) 2-Nitroaniline	13.575	65	1761736	87.871	ng/ul	98
47) Dimethylphthalate	13.951	163	7016385	75.264	ng/ul	99
48) 2,6-Dinitrotoluene	14.075	165	1588010	89.024	ng/ul	99
50) Acenaphthylene	14.204	152	8364353	73.924	ng/ul	98
51) 3-Nitroaniline	14.404	138	1566003	83.169	ng/ul	96
52) Acenaphthene	14.551	153	5842710	73.016	ng/ul	99
53) 2,4-Dinitrophenol	14.610	184	1035804	96.705	ng/ul	97
55) 4-Nitrophenol	14.728	109	1014052	81.892	ng/ul	96
56) Dibenzofuran	14.887	168	7686253	70.165	ng/ul	97
57) 2,4-Dinitrotoluene	14.863	165	2129011	83.629	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.110	232	1863846	80.582	ng/ul	97
59) Diethylphthalate	15.316	149	6942007	75.765	ng/ul	100
61) Fluorene	15.539	166	5622477	64.784	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.528	204	2890229	66.825	ng/ul	97
63) 4-Nitroaniline	15.581	138	1326480	75.046	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.634	198	1292257	85.396	ng/ul#	97
67) N-Nitrosodiphenylamine	15.751	169	5404312	73.483	ng/ul	100
68) 4-Bromophenyl-phenylether	16.428	248	2155528	80.345	ng/ul	98
69) Hexachlorobenzene	16.539	284	2486710	78.632	ng/ul	99
70) Atrazine	16.716	200	2014913	78.480	ng/ul	99
71) Pentachlorophenol	16.892	266	1606329	84.148	ng/ul	99
72) Phenanthrene	17.286	178	9522089	69.918	ng/ul	99
74) Anthracene	17.381	178	9237100	68.551	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.275	216	3070442	81.205	ng/ul	99
76) Pentachlorobenzene	14.804	250	3108133	78.038	ng/ul	99
77) Carbazole	17.657	167	8769513	73.742	ng/ul	98
78) Di-n-butylphthalate	18.198	149	10749075	75.347	ng/ul	98
80) Fluoranthene	19.257	202	10331001	82.105	ng/ul	97
82) Pyrene	19.610	202	9893692	76.854	ng/ul	96
83) Butylbenzylphthalate	20.480	149	4496676	89.808	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.257	252	2968326	76.204	ng/ul	99
85) Benzo(a)anthracene	21.322	228	8721942	74.827	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.239	149	5932715	84.573	ng/ul#	98
87) Chrysene	21.374	228	8117742	74.630	ng/ul	98
89) Di-n-octyl phthalate	22.163	149	9565076	83.229	ng/ul	100
90) Benzo(b)fluoranthene	23.010	252	8327182	75.199	ng/ul	99
91) Benzo(k)fluoranthene	23.063	252	8164803	74.802	ng/ul	99
93) Benzo(a)pyrene	23.645	252	7314021	77.239	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.257	276	9170060	83.654	ng/ul	99
95) Dibenzo(a,h)anthracene	26.280	278	8121602	82.181	ng/ul	99
96) Benzo(g,h,i)perylene	27.033	276	8309074	80.729	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101922\
 Data File : BP012203.D
 Acq On : 19 Oct 2022 13:57
 Operator : CG/JU
 Sample : SSTD08084
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080649

Quant Time: Oct 20 06:39:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 20 00:51:41 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 31.473 ng/uL
RT: 3.293 min Scan# 51
Delta R.T. -0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion: 88 Resp: 405907
Ion Ratio Lower Upper
88 100
43 25.0 20.6 30.8
58 72.4 56.1 84.1



#5
Pyridine
Concen: 80.797 ng/uL
RT: 3.699 min Scan# 121
Delta R.T. 0.000 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion: 79 Resp: 2742519
Ion Ratio Lower Upper
79 100
52 70.3 59.4 89.2
51 32.1 26.5 39.7









#6
Benzaldehyde
Concen: 73.656 ng/ul
RT: 6.975 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion: 77 Resp: 1449586
Ion Ratio Lower Upper
77 100
105 102.6 79.0 118.6
106 98.3 76.4 114.6



#8
Phenol
Concen: 80.923 ng/ul
RT: 7.034 min Scan# 688
Delta R.T. 0.012 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion: 94 Resp: 3544873
Ion Ratio Lower Upper
94 100
65 25.7 20.6 31.0
66 33.0 26.2 39.4









#10
Bis(2-Chloroethyl)ether
Concen: 79.126 ng/ul
RT: 7.263 min Scan# 726
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Instrument :
BNA_P
ClientSampleId :
SSTD080649

Tgt Ion: 93 Resp: 2797617




Ion	Ratio	Lower	Upper
93	100		
63	74.3	60.2	90.2
95	32.8	26.1	39.1




#12
2-Chlorophenol
Concen: 81.076 ng/ul
RT: 7.393 min Scan# 749
Delta R.T. 0.000 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:128 Resp: 2788075

Ion	Ratio	Lower	Upper
128	100		
64	45.1	35.4	53.2
130	32.4	26.1	39.1





#13

2-Methylphenol

Concen: 81.696 ng/ul

RT: 8.281 min Scan# 900

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :





SSTD080649

Tgt Ion:108 Resp: 2694290

Ion Ratio Lower Upper

108 100

107 88.4 71.1 106.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 77.936 ng/ul

RT: 8.369 min Scan# 915

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57




Tgt Ion: 45 Resp: 3664401


Ion Ratio Lower Upper

45 100

77 15.1 12.1 18.1

79 12.2 9.3 13.9









#16
Acetophenone
Concen: 76.965 ng/ul
RT: 8.669 min Scan# 966
Delta R.T. 0.012 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:105 Resp: 3868980
Ion Ratio Lower Upper
105 100
77 74.2 60.0 90.0
51 23.5 20.1 30.1



#17
N-Nitroso-di-n-propylamine
Concen: 76.405 ng/ul
RT: 8.669 min Scan# 966
Delta R.T. 0.024 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion: 70 Resp: 1944167
Ion Ratio Lower Upper
70 100
42 50.4 42.4 63.6
101 10.4 7.6 11.4
130 22.6 17.3 25.9









#18
4-Methylphenol
Concen: 80.561 ng/ul
RT: 8.622 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:108 Resp: 2893471
Ion Ratio Lower Upper
108 100
107 111.3 86.6 130.0



#19
Hexachloroethane
Concen: 80.982 ng/ul
RT: 8.905 min Scan# 1006
Delta R.T. 0.000 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:117 Resp: 1100020
Ion Ratio Lower Upper
117 100
201 101.1 79.7 119.5
199 63.4 50.7 76.1









#22
Nitrobenzene
Concen: 80.372 ng/ul
RT: 9.046 min Scan# 1029
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion: 77 Resp: 3046907
Ion Ratio Lower Upper
77 100
123 48.8 37.1 55.7
65 13.1 10.0 15.0



#23
Isophorone
Concen: 81.900 ng/ul
RT: 9.587 min Scan# 1122
Delta R.T. 0.018 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion: 82 Resp: 5959271
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.0
138 18.4 14.5 21.7









#25
2-Nitrophenol
Concen: 85.224 ng/ul
RT: 9.757 min Scan# 1150
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:139 Resp: 1625485
Ion Ratio Lower Upper
139 100
65 45.9 38.0 57.0
109 27.4 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 79.035 ng/ul
RT: 9.816 min Scan# 1161
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:107 Resp: 2971367
Ion Ratio Lower Upper
107 100
121 51.1 40.6 60.8
122 88.2 71.0 106.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 77.425 ng/ul
 RT: 10.057 min Scan# 1201
 Delta R.T. 0.006 min
 Lab File: BP012203.D
 Acq: 19 Oct 2022 13:57




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080649


Tgt Ion: 93 Resp: 3687986
 Ion Ratio Lower Upper
 93 100
 95 32.8 26.2 39.4
 123 11.5 9.0 13.6



#29
 2,4-Dichlorophenol
 Concen: 79.867 ng/ul
 RT: 10.287 min Scan# 1241
 Delta R.T. 0.006 min
 Lab File: BP012203.D
 Acq: 19 Oct 2022 13:57

Tgt Ion:162 Resp: 2602510
 Ion Ratio Lower Upper
 162 100
 164 65.5 51.2 76.8
 98 36.2 28.3 42.5









#30
Naphthalene
Concen: 74.417 ng/ul
RT: 10.687 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:128 Resp: 8366154
Ion Ratio Lower Upper
128 100
129 11.4 8.8 13.2
127 13.3 10.2 15.4



#32
4-Chloroaniline
Concen: 78.147 ng/ul
RT: 10.810 min Scan# 1330
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:127 Resp: 3608716
Ion Ratio Lower Upper
127 100
129 32.8 25.6 38.4









#33
Hexachlorobutadiene
Concen: 78.485 ng/ul
RT: 10.963 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:225 Resp: 1608560
Ion Ratio Lower Upper
225 100
223 62.6 50.2 75.4
227 64.3 51.9 77.9



#34
Caprolactam
Concen: 46.594 ng/ul
RT: 11.640 min Scan# 1471
Delta R.T. 0.041 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:113 Resp: 471321
Ion Ratio Lower Upper
113 100
55 161.9 133.7 200.5
56 120.0 95.4 143.2









#35
4-Chloro-3-methylphenol
Concen: 80.468 ng/ul
RT: 11.940 min Scan# 1522
Delta R.T. 0.012 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:107 Resp: 2830964
Ion Ratio Lower Upper
107 100
144 27.2 21.2 31.8
142 82.8 65.5 98.3



#36
2-Methylnaphthalene
Concen: 74.840 ng/ul
RT: 12.298 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:142 Resp: 5724511
Ion Ratio Lower Upper
142 100
141 88.5 70.3 105.5









#37
1-Methylnaphthalene
Concen: 73.647 ng/ul
RT: 12.522 min Scan# 1621
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:142 Resp: 5624970
Ion Ratio Lower Upper
142 100
141 91.7 73.4 110.0
116 3.8 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 78.057 ng/ul
RT: 12.669 min Scan# 1646
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:216 Resp: 3033363
Ion Ratio Lower Upper
216 100
214 79.3 62.5 93.7
179 20.9 16.5 24.7
108 16.7 13.4 20.0









#40
Hexachlorocyclopentadiene
Concen: 85.292 ng/ul
RT: 12.640 min Scan# 1641
Delta R.T. 0.000 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:237 Resp: 1762189
Ion Ratio Lower Upper
237 100
235 63.8 49.4 74.0
272 13.7 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 83.575 ng/ul
RT: 12.916 min Scan# 1688
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:196 Resp: 2096094
Ion Ratio Lower Upper
196 100
198 96.4 77.4 116.2
200 31.4 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 82.834 ng/ul
RT: 12.987 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:196 Resp: 2266294
Ion Ratio Lower Upper
196 100
198 95.5 76.2 114.2
200 30.7 24.7 37.1



#43
1,1'-Biphenyl
Concen: 73.893 ng/ul
RT: 13.316 min Scan# 1756
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:154 Resp: 7028764
Ion Ratio Lower Upper
154 100
153 40.7 31.5 47.3
76 12.9 11.0 16.4





#44

2-Chloronaphthalene

Concen: 75.119 ng/ul

RT: 13.357 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :

SSTD080649





Tgt Ion:162 Resp: 5665029

Ion Ratio Lower Upper

162 100

164 33.9 26.1 39.1

127 39.8 30.8 46.2



#45

2-Nitroaniline

Concen: 87.871 ng/ul

RT: 13.575 min Scan# 1800

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57




Tgt Ion: 65 Resp: 1761736


Ion Ratio Lower Upper

65 100

92 74.2 57.5 86.3

138 117.5 92.5 138.7





#47

Dimethylphthalate

Concen: 75.264 ng/ul

RT: 13.951 min Scan# 1864

Delta R.T. 0.012 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :

SSTD080649





Tgt Ion:163 Resp: 7016385

Ion Ratio Lower Upper

163 100

194 4.1 3.0 4.6

164 10.4 8.1 12.1



#48

2,6-Dinitrotoluene

Concen: 89.024 ng/ul

RT: 14.075 min Scan# 1885

Delta R.T. 0.012 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57




Tgt Ion:165 Resp: 1588010


Ion Ratio Lower Upper

165 100

89 54.2 43.9 65.9

121 20.6 17.1 25.7









#50
Acenaphthylene
Concen: 73.924 ng/ul
RT: 14.204 min Scan# 1907
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:152 Resp: 8364353
Ion Ratio Lower Upper
152 100
151 20.7 15.8 23.8
153 13.7 10.6 16.0



#51
3-Nitroaniline
Concen: 83.169 ng/ul
RT: 14.404 min Scan# 1941
Delta R.T. 0.012 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:138 Resp: 1566003
Ion Ratio Lower Upper
138 100
108 10.0 8.2 12.2
92 111.1 93.1 139.7









#52
Acenaphthene
Concen: 73.016 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:153 Resp: 5842710
Ion Ratio Lower Upper
153 100
152 48.3 38.1 57.1
154 90.1 71.4 107.2



#53
2,4-Dinitrophenol
Concen: 96.705 ng/ul
RT: 14.610 min Scan# 1976
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:184 Resp: 1035804
Ion Ratio Lower Upper
184 100
63 60.6 51.1 76.7
154 59.6 47.2 70.8









#55
4-Nitrophenol
Concen: 81.892 ng/ul
RT: 14.728 min Scan# 1996
Delta R.T. 0.024 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:109 Resp: 1014052
Ion Ratio Lower Upper
109 100
139 149.8 118.5 177.7
65 130.0 110.5 165.7



#56
Dibenzofuran
Concen: 70.165 ng/ul
RT: 14.887 min Scan# 2023
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:168 Resp: 7686253
Ion Ratio Lower Upper
168 100
139 38.1 29.0 43.4





#57

2,4-Dinitrotoluene

Concen: 83.629 ng/ul

RT: 14.863 min Scan# 2017

Delta R.T. 0.012 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :

SSTD080649





Tgt Ion:165 Resp: 2129011

Ion Ratio Lower Upper

165 100

63 40.9 34.1 51.1

182 2.6 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 80.582 ng/ul

RT: 15.110 min Scan# 2061

Delta R.T. 0.000 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Tgt Ion:232 Resp: 1863846




Ion Ratio Lower Upper


232 100

131 45.8 35.0 52.4

130 2.2 1.8 2.8

166 31.0 23.5 35.3









#59
Diethylphthalate
Concen: 75.765 ng/ul
RT: 15.316 min Scan# 2095
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:149 Resp: 6942007
Ion Ratio Lower Upper
149 100
177 22.4 17.9 26.9
150 12.8 10.1 15.1



#61
Fluorene
Concen: 64.784 ng/ul
RT: 15.539 min Scan# 2134
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:166 Resp: 5622477
Ion Ratio Lower Upper
166 100
165 97.6 78.0 117.0
167 13.8 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 66.825 ng/ul
RT: 15.528 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:204 Resp: 2890229
Ion Ratio Lower Upper
204 100
206 33.7 26.6 39.8
141 60.0 45.3 67.9



#63
4-Nitroaniline
Concen: 75.046 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. 0.018 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:138 Resp: 1326480
Ion Ratio Lower Upper
138 100
92 49.0 41.6 62.4
108 72.2 57.0 85.6









#66
4,6-Dinitro-2-methylphenol
Concen: 85.396 ng/ul
RT: 15.634 min Scan# 2148
Delta R.T. 0.012 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:198 Resp: 1292257
Ion Ratio Lower Upper
198 100
182 3.5 3.8 5.6#
77 21.0 17.8 26.6



#67
N-Nitrosodiphenylamine
Concen: 73.483 ng/ul
RT: 15.751 min Scan# 2170
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:169 Resp: 5404312
Ion Ratio Lower Upper
169 100
168 66.7 53.4 80.0
167 37.1 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 80.345 ng/ul
RT: 16.428 min Scan# 2285
Delta R.T. 0.000 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:248 Resp: 2155528
Ion Ratio Lower Upper
248 100
250 97.2 79.0 118.4
141 70.4 55.1 82.7



#69
Hexachlorobenzene
Concen: 78.632 ng/ul
RT: 16.539 min Scan# 2304
Delta R.T. 0.000 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:284 Resp: 2486710
Ion Ratio Lower Upper
284 100
142 32.4 25.5 38.3
249 30.5 23.6 35.4





#70

Atrazine

Concen: 78.480 ng/ul

RT: 16.716 min Scan# 21

Delta R.T. 0.012 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :

SSTD080649





Tgt Ion:200 Resp: 2014913

Ion Ratio Lower Upper

200 100

173 27.0 21.4 32.0

215 48.6 38.2 57.2



#71

Pentachlorophenol

Concen: 84.148 ng/ul

RT: 16.892 min Scan# 2364

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57




Tgt Ion:266 Resp: 1606329


Ion Ratio Lower Upper

266 100

264 63.6 50.1 75.1

268 64.6 50.6 76.0









#72
Phenanthrene
Concen: 69.918 ng/ul
RT: 17.286 min Scan# 2430
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57




Instrument :
BNA_P
ClientSampleId :
SSTD080649


Tgt Ion:178 Resp: 9522089
Ion Ratio Lower Upper
178 100
179 16.2 12.6 18.8
176 20.2 15.6 23.4



#74
Anthracene
Concen: 68.551 ng/ul
RT: 17.381 min Scan# 2447
Delta R.T. 0.006 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:178 Resp: 9237100
Ion Ratio Lower Upper
178 100
179 16.4 12.3 18.5
176 19.8 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 81.205 ng/uL

RT: 13.275 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :

SSTD080649





Tgt Ion:216 Resp: 3070442

Ion Ratio Lower Upper

216 100

214 79.5 62.9 94.3

218 48.4 38.6 57.8



#76

Pentachlorobenzene

Concen: 78.038 ng/uL

RT: 14.804 min Scan# 2009

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57




Tgt Ion:250 Resp: 3108133


Ion Ratio Lower Upper

250 100

248 63.0 50.3 75.5

252 63.4 52.0 78.0





#77

Carbazole

Concen: 73.742 ng/ul

RT: 17.657 min Scan# 2494

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :

SSTD080649





Tgt Ion:167 Resp: 8769513

Ion Ratio Lower Upper

167 100

166 22.2 17.3 25.9

139 13.7 10.2 15.4



#78

Di-n-butylphthalate

Concen: 75.347 ng/ul

RT: 18.198 min Scan# 2586

Delta R.T. 0.000 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57




Tgt Ion:149 Resp:10749075


Ion Ratio Lower Upper

149 100

150 10.0 7.3 10.9

104 5.3 3.8 5.6





#80

Fluoranthene

Concen: 82.105 ng/ul

RT: 19.257 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :

SSTD080649





Tgt Ion:202 Resp:10331001

Ion Ratio Lower Upper

202 100

101 11.5 10.4 15.6

100 8.6 7.7 11.5



#82

Pyrene

Concen: 76.854 ng/ul

RT: 19.610 min Scan# 2826

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57




Tgt Ion:202 Resp: 9893692


Ion Ratio Lower Upper

202 100

101 13.6 12.3 18.5

100 11.0 10.0 15.0





#83

Butylbenzylphthalate

Concen: 89.808 ng/ul

RT: 20.480 min Scan# 2974

Delta R.T. 0.000 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :

SSTD080649





Tgt Ion:149 Resp: 4496676

Ion Ratio Lower Upper

149 100

91 70.1 56.1 84.1

206 21.5 16.1 24.1



#84

3,3'-Dichlorobenzidine

Concen: 76.204 ng/ul

RT: 21.257 min Scan# 3106

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57




Tgt Ion:252 Resp: 2968326


Ion Ratio Lower Upper

252 100

254 65.0 52.2 78.4

126 10.7 9.8 14.6





#85

Benzo(a)anthracene

Concen: 74.827 ng/ul

RT: 21.322 min Scan# 3117

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :

SSTD080649





Tgt Ion:228 Resp: 8721942

Ion Ratio Lower Upper

228 100

229 20.5 16.1 24.1

226 28.8 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 84.573 ng/ul

RT: 21.239 min Scan# 3103

Delta R.T. 0.000 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57




Tgt Ion:149 Resp: 5932715


Ion Ratio Lower Upper

149 100

167 27.7 21.6 32.4

279 5.2 3.4 5.0#





#87

Chrysene

Concen: 74.630 ng/ul

RT: 21.374 min Scan# 31

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :

SSTD080649





Tgt Ion:228 Resp: 8117742

Ion Ratio Lower Upper

228 100

226 31.8 24.2 36.4

229 20.4 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 83.229 ng/ul




RT: 22.163 min Scan# 3260


Delta R.T. 0.000 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Tgt Ion:149 Resp: 9565076





#90

Benzo(b)fluoranthene

Concen: 75.199 ng/ul

RT: 23.010 min Scan# 3403

Delta R.T. 0.006 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :

SSTD080649





Tgt Ion:252 Resp: 8327182

Ion Ratio Lower Upper

252 100

253 22.7 17.5 26.3

125 10.4 8.5 12.7



#91

Benzo(k)fluoranthene

Concen: 74.802 ng/ul

RT: 23.063 min Scan# 3413

Delta R.T. 0.012 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57




Tgt Ion:252 Resp: 8164803


Ion Ratio Lower Upper

252 100

253 22.4 17.9 26.9

125 9.9 8.6 13.0





#93

Benzo(a)pyrene

Concen: 77.239 ng/ul

RT: 23.645 min Scan# 3

Delta R.T. 0.012 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57

Instrument :

BNA_P

ClientSampleId :

SSTD080649





Tgt Ion:252 Resp: 7314021

Ion Ratio Lower Upper

252 100

253 22.4 17.8 26.6

125 10.9 9.7 14.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 83.654 ng/ul

RT: 26.257 min Scan# 3956

Delta R.T. 0.012 min

Lab File: BP012203.D

Acq: 19 Oct 2022 13:57




Tgt Ion:276 Resp: 9170060


Ion Ratio Lower Upper

276 100

138 24.3 20.3 30.5

277 25.2 20.2 30.4









#95
Dibenzo(a,h)anthracene
Concen: 82.181 ng/ul
RT: 26.280 min Scan# 3957
Delta R.T. 0.018 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Instrument :
BNA_P
ClientSampleId :
SSTD080649

Tgt Ion:278 Resp: 8121602
Ion Ratio Lower Upper
278 100
139 15.3 12.9 19.3
279 24.0 19.0 28.4



#96
Benzo(g,h,i)perylene
Concen: 80.729 ng/ul
RT: 27.033 min Scan# 4088
Delta R.T. 0.024 min
Lab File: BP012203.D
Acq: 19 Oct 2022 13:57

Tgt Ion:276 Resp: 8309074
Ion Ratio Lower Upper
276 100
138 21.2 17.9 26.9
277 24.2 18.9 28.3

