

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102020\
 Data File : BP003673.D
 Acq On : 20 Oct 2020 10:01
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 20 10:43:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP101920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 19 14:54:24 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	38.234	4.4	103	0.00
3	Pyridine	40.000	38.644	3.4	101	0.00
4	n-Nitrosodimethylamine	40.000	39.234	1.9	103	0.00
5 S	2-Fluorophenol	80.000	78.070	2.4	102	0.00
6	Aniline	40.000	38.873	2.8	101	0.00
7 S	Phenol-d6	80.000	78.393	2.0	102	0.00
8	2-Chlorophenol	40.000	38.607	3.5	100	0.00
9	Benzaldehyde	40.000	40.567	-1.4	102	0.00
10 C	Phenol	40.000	38.865	2.8	101	0.00
11	bis(2-Chloroethyl)ether	40.000	38.991	2.5	102	0.00
12	1,3-Dichlorobenzene	40.000	38.493	3.8	102	0.00
13 C	1,4-Dichlorobenzene	40.000	38.317	4.2	101	0.00
14	1,2-Dichlorobenzene	40.000	38.209	4.5	101	0.00
15	Benzyl Alcohol	40.000	39.562	1.1	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.101	2.2	103	0.00
17	2-Methylphenol	40.000	38.855	2.9	100	0.00
18	Hexachloroethane	40.000	38.690	3.3	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.555	1.1	101	0.00
20	3+4-Methylphenols	40.000	39.270	1.8	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	38.007	5.0	101	0.00
23 S	Nitrobenzene-d5	80.000	78.135	2.3	102	0.00
24	Nitrobenzene	40.000	38.992	2.5	103	0.00
25	Isophorone	40.000	38.960	2.6	101	0.00
26 C	2-Nitrophenol	40.000	39.291	1.8	100	0.00
27	2,4-Dimethylphenol	40.000	38.619	3.5	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.206	4.5	102	0.00
29 C	2,4-Dichlorophenol	40.000	38.882	2.8	102	0.00
30	1,2,4-Trichlorobenzene	40.000	38.262	4.3	102	0.00
31	Naphthalene	40.000	37.842	5.4	101	0.00
32	Benzoic acid	40.000	36.502	8.7	100	0.00
33	4-Chloroaniline	40.000	38.707	3.2	101	0.00
34 C	Hexachlorobutadiene	40.000	38.225	4.4	102	0.00
35	Caprolactam	40.000	40.033	-0.1	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.742	0.6	104	0.00
37	2-Methylnaphthalene	40.000	38.384	4.0	102	0.00
38	1-Methylnaphthalene	40.000	38.371	4.1	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.860	5.4	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.952	2.6	102	0.00
42 S	2,4,6-Tribromophenol	80.000	80.183	-0.2	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.045	2.4	102	0.00
44	2,4,5-Trichlorophenol	40.000	38.506	3.7	101	0.00
45 S	2-Fluorobiphenyl	80.000	76.274	4.7	103	0.00
46	1,1'-Biphenyl	40.000	37.965	5.1	104	0.00
47	2-Chloronaphthalene	40.000	37.741	5.6	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102020\
 Data File : BP003673.D
 Acq On : 20 Oct 2020 10:01
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 20 10:43:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP101920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 19 14:54:24 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.423	-1.1	104	0.00
49	Acenaphthylene	40.000	38.478	3.8	103	0.00
50	Dimethylphthalate	40.000	38.486	3.8	103	0.00
51	2,6-Dinitrotoluene	40.000	39.644	0.9	105	0.00
52 C	Acenaphthene	40.000	38.660	3.4	103	0.00
53	3-Nitroaniline	40.000	39.998	0.0	104	0.00
54 P	2,4-Dinitrophenol	40.000	36.987	7.5	103	0.00
55	Dibenzofuran	40.000	38.513	3.7	104	0.00
56 P	4-Nitrophenol	40.000	39.952	0.1	103	0.00
57	2,4-Dinitrotoluene	40.000	40.643	-1.6	105	0.00
58	Fluorene	40.000	38.551	3.6	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.194	2.0	102	0.00
60	Diethylphthalate	40.000	39.178	2.1	105	0.00
61	4-Chlorophenyl-phenylether	40.000	38.515	3.7	105	0.00
62	4-Nitroaniline	40.000	39.854	0.4	102	0.00
63	Azobenzene	40.000	39.118	2.2	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.949	2.6	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.571	3.6	104	0.00
67	4-Bromophenyl-phenylether	40.000	38.687	3.3	104	0.00
68	Hexachlorobenzene	40.000	38.342	4.1	103	0.00
69	Atrazine	40.000	39.624	0.9	105	0.00
70 C	Pentachlorophenol	40.000	40.223	-0.6	107	0.00
71	Phenanthrene	40.000	38.650	3.4	105	0.00
72	Anthracene	40.000	39.010	2.5	105	0.00
73	Carbazole	40.000	39.059	2.4	105	0.00
74	Di-n-butylphthalate	40.000	40.230	-0.6	105	0.00
75 C	Fluoranthene	40.000	39.599	1.0	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	109	0.00
77	Benzydine	40.000	39.248	1.9	103	0.00
78	Pyrene	40.000	38.006	5.0	107	0.00
79 S	Terphenyl-d14	80.000	76.265	4.7	107	0.00
80	Butylbenzylphthalate	40.000	39.736	0.7	107	0.00
81	Benzo(a)anthracene	40.000	38.940	2.7	108	0.00
82	3,3'-Dichlorobenzidine	40.000	39.645	0.9	108	0.00
83	Chrysene	40.000	38.442	3.9	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.894	0.3	107	0.00
85 c	Di-n-octyl phthalate	40.000	40.652	-1.6	108	0.00
86 I	Perylene-d12	20.000	20.000	0.0	109	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.072	2.3	111	0.00
88	Benzo(b)fluoranthene	40.000	39.616	1.0	110	0.00
89	Benzo(k)fluoranthene	40.000	38.964	2.6	110	0.00
90 C	Benzo(a)pyrene	40.000	39.280	1.8	109	0.00
91	Dibenzo(a,h)anthracene	40.000	39.089	2.3	111	0.00
92	Benzo(g,h,i)perylene	40.000	38.958	2.6	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102020\
Data File : BP003673.D
Acq On : 20 Oct 2020 10:01
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Oct 20 10:43:06 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP101920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 19 14:54:24 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0