

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102320\
 Data File : BP003735.D
 Acq On : 24 Oct 2020 01:06
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 24 02:08:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP101920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 19 14:54:24 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	111	0.00
2	1,4-Dioxane	40.000	35.466	11.3	104	0.00
3	Pyridine	40.000	36.331	9.2	104	0.00
4	n-Nitrosodimethylamine	40.000	33.390	16.5	95	0.00
5 S	2-Fluorophenol	80.000	74.713	6.6	107	0.00
6	Aniline	40.000	37.401	6.5	106	0.00
7 S	Phenol-d6	80.000	74.444	6.9	106	0.00
8	2-Chlorophenol	40.000	37.503	6.2	106	0.00
9	Benzaldehyde	40.000	37.889	5.3	106	0.00
10 C	Phenol	40.000	37.400	6.5	106	0.00
11	bis(2-Chloroethyl)ether	40.000	36.416	9.0	104	0.00
12	1,3-Dichlorobenzene	40.000	36.638	8.4	106	0.00
13 C	1,4-Dichlorobenzene	40.000	36.632	8.4	106	0.00
14	1,2-Dichlorobenzene	40.000	36.190	9.5	104	0.00
15	Benzyl Alcohol	40.000	35.963	10.1	101	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.009	10.0	103	0.00
17	2-Methylphenol	40.000	37.550	6.1	106	0.00
18	Hexachloroethane	40.000	35.769	10.6	101	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.545	11.1	99	0.00
20	3+4-Methylphenols	40.000	37.346	6.6	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	112	0.00
22	Acetophenone	40.000	34.958	12.6	102	0.00
23 S	Nitrobenzene-d5	80.000	70.044	12.4	101	0.00
24	Nitrobenzene	40.000	34.590	13.5	101	0.00
25	Isophorone	40.000	35.213	12.0	101	0.00
26 C	2-Nitrophenol	40.000	39.611	1.0	111	0.00
27	2,4-Dimethylphenol	40.000	36.910	7.7	106	0.00
28	bis(2-Chloroethoxy)methane	40.000	35.369	11.6	104	0.00
29 C	2,4-Dichlorophenol	40.000	36.529	8.7	105	0.00
30	1,2,4-Trichlorobenzene	40.000	35.696	10.8	104	0.00
31	Naphthalene	40.000	35.813	10.5	105	0.00
32	Benzoic acid	40.000	36.477	8.8	110	0.00
33	4-Chloroaniline	40.000	36.314	9.2	104	0.00
34 C	Hexachlorobutadiene	40.000	34.317	14.2	101	0.00
35	Caprolactam	40.000	38.470	3.8	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.527	8.7	105	0.00
37	2-Methylnaphthalene	40.000	35.857	10.4	104	0.00
38	1-Methylnaphthalene	40.000	35.502	11.2	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	113	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	34.700	13.2	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	30.158	24.6	87	0.00
42 S	2,4,6-Tribromophenol	80.000	76.102	4.9	109	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.679	8.3	105	0.00
44	2,4,5-Trichlorophenol	40.000	36.276	9.3	105	0.00
45 S	2-Fluorobiphenyl	80.000	69.468	13.2	103	0.00
46	1,1'-Biphenyl	40.000	34.785	13.0	104	0.00
47	2-Chloronaphthalene	40.000	34.912	12.7	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102320\
 Data File : BP003735.D
 Acq On : 24 Oct 2020 01:06
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 24 02:08:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP101920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 19 14:54:24 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	35.869	10.3	101	0.00
49	Acenaphthylene	40.000	35.593	11.0	105	0.00
50	Dimethylphthalate	40.000	34.796	13.0	103	0.00
51	2,6-Dinitrotoluene	40.000	36.668	8.3	107	0.00
52 C	Acenaphthene	40.000	34.741	13.1	102	0.00
53	3-Nitroaniline	40.000	38.283	4.3	109	0.00
54 P	2,4-Dinitrophenol	40.000	28.623	28.4#	84	0.00
55	Dibenzofuran	40.000	35.324	11.7	105	0.00
56 P	4-Nitrophenol	40.000	38.181	4.5	108	0.01
57	2,4-Dinitrotoluene	40.000	37.464	6.3	106	0.00
58	Fluorene	40.000	35.176	12.1	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.647	8.4	104	0.00
60	Diethylphthalate	40.000	34.367	14.1	101	0.00
61	4-Chlorophenyl-phenylether	40.000	34.570	13.6	103	0.00
62	4-Nitroaniline	40.000	37.794	5.5	107	0.00
63	Azobenzene	40.000	33.829	15.4	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	115	0.00
65	4,6-Dinitro-2-methylphenol	40.000	34.016	15.0	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	35.188	12.0	105	0.00
67	4-Bromophenyl-phenylether	40.000	34.427	13.9	102	0.00
68	Hexachlorobenzene	40.000	35.096	12.3	104	0.00
69	Atrazine	40.000	35.102	12.2	102	0.00
70 C	Pentachlorophenol	40.000	36.413	9.0	107	0.01
71	Phenanthrene	40.000	34.934	12.7	105	0.00
72	Anthracene	40.000	35.448	11.4	106	0.00
73	Carbazole	40.000	35.974	10.1	106	0.00
74	Di-n-butylphthalate	40.000	35.619	11.0	103	0.00
75 C	Fluoranthene	40.000	35.318	11.7	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	117	0.01
77	Benzydine	40.000	37.359	6.6	105	0.00
78	Pyrene	40.000	35.038	12.4	105	0.00
79 S	Terphenyl-d14	80.000	69.836	12.7	105	0.00
80	Butylbenzylphthalate	40.000	36.792	8.0	106	0.00
81	Benzo(a)anthracene	40.000	35.468	11.3	106	0.01
82	3,3'-Dichlorobenzidine	40.000	37.491	6.3	109	0.00
83	Chrysene	40.000	35.540	11.2	106	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.044	9.9	104	0.01
85 c	Di-n-octyl phthalate	40.000	37.241	6.9	106	0.01
86 I	Perylene-d12	20.000	20.000	0.0	122	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	37.627	5.9	120	0.02
88	Benzo(b)fluoranthene	40.000	34.372	14.1	106	0.02
89	Benzo(k)fluoranthene	40.000	35.274	11.8	111	0.02
90 C	Benzo(a)pyrene	40.000	35.664	10.8	111	0.02
91	Dibenzo(a,h)anthracene	40.000	37.647	5.9	119	0.02
92	Benzo(g,h,i)perylene	40.000	38.534	3.7	124	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102320\
Data File : BP003735.D
Acq On : 24 Oct 2020 01:06
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Oct 24 02:08:41 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP101920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 19 14:54:24 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0