

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102522\
 Data File : BP012318.D
 Acq On : 25 Oct 2022 17:01
 Operator : CG/JU
 Sample : SST01087
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST010604

Quant Time: Oct 26 10:06:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 26 10:05:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	382831	20.000	ng/u1	0.00
20) Naphthalene-d8	10.616	136	1723293	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.463	164	1135204	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.222	188	2341327	20.000	ng/u1	0.00
79) Chrysene-d12	21.316	240	2313027	20.000	ng/u1	0.00
88) Perylene-d12	23.710	264	2412683	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	38764	4.107	ng/uL	0.00
4) Pyridine-d5	3.681	84	255695	9.409	ng/u1	0.00
7) Phenol-d5	6.987	99	314633	9.495	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	203027	10.117	ng/u1	0.00
11) 2-Chlorophenol-d4	7.346	132	239575	9.568	ng/u1	0.00
15) 4-Methylphenol-d8	8.534	113	248783	9.643	ng/u1	0.00
21) Nitrobenzene-d5	8.981	128	98217	7.962	ng/u1	0.00
24) 2-Nitrophenol-d4	9.704	143	94783	7.187	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.246	165	242958	9.501	ng/u1	0.00
31) 4-Chloroaniline-d4	10.763	131	373898	10.196	ng/u1	0.00
46) Dimethylphthalate-d6	13.881	166	774030	9.887	ng/u1	0.00
49) Acenaphthylene-d8	14.157	160	895118	10.040	ng/u1	0.00
54) 4-Nitrophenol-d4	14.681	143	114004	7.812	ng/u1	0.00
60) Fluorene-d10	15.463	176	687663	10.222	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.592	200	84525	6.555	ng/u1	0.00
73) Anthracene-d10	17.322	188	1051341	10.364	ng/u1	0.00
81) Pyrene-d10	19.563	212	1181033	9.286	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.557	264	1183346	9.899	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.287	88	41743	4.126	ng/uL	95
5) Pyridine	3.699	79	257569	9.654	ng/u1	98
6) Benzaldehyde	6.963	77	170028	11.041	ng/u1	97
8) Phenol	7.016	94	334846	9.687	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.246	93	281920	10.093	ng/u1	100
12) 2-Chlorophenol	7.381	128	264890	9.802	ng/u1	98
13) 2-Methylphenol	8.263	108	255078	9.767	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.340	45	396383	10.683	ng/u1	99
16) Acetophenone	8.646	105	431436	10.786	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.628	70	196472	9.756	ng/u1	98
18) 4-Methylphenol	8.593	108	284204	9.952	ng/u1	100
19) Hexachloroethane	8.887	117	102872	9.702	ng/u1	99
22) Nitrobenzene	9.028	77	280204	9.166	ng/u1	97
23) Isophorone	9.552	82	600554	10.068	ng/u1	98
25) 2-Nitrophenol	9.734	139	123907	8.159	ng/u1	94
26) 2,4-Dimethylphenol	9.799	107	302045	9.845	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.034	93	395906	10.204	ng/u1	99
29) 2,4-Dichlorophenol	10.269	162	258928	9.787	ng/u1	100
30) Naphthalene	10.669	128	936369	10.237	ng/u1	99
32) 4-Chloroaniline	10.793	127	394372	10.438	ng/u1	98
33) Hexachlorobutadiene	10.946	225	163402	9.878	ng/u1	97
34) Caprolactam	11.598	113	70881	8.516	ng/u1	92
35) 4-Chloro-3-methylphenol	11.916	107	280983	9.773	ng/u1	99
36) 2-Methylnaphthalene	12.281	142	645628	10.364	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102522\
 Data File : BP012318.D
 Acq On : 25 Oct 2022 17:01
 Operator : CG/JU
 Sample : SST01087
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST010604


Quant Time: Oct 26 10:06:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 26 10:05:01 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.498	142	645172	10.352	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.651	216	328033	9.649	ng/ul	99
40) Hexachlorocyclopentadiene	12.622	237	103367	5.925	ng/ul	99
41) 2,4,6-Trichlorophenol	12.898	196	200994	9.177	ng/ul	100
42) 2,4,5-Trichlorophenol	12.969	196	220350	9.219	ng/ul	100
43) 1,1'-Biphenyl	13.292	154	857770	10.259	ng/ul	99
44) 2-Chloronaphthalene	13.340	162	671028	10.160	ng/ul	99
45) 2-Nitroaniline	13.557	65	119166	6.945	ng/ul	96
47) Dimethylphthalate	13.928	163	833263	10.156	ng/ul	100
48) 2,6-Dinitrotoluene	14.051	165	106866	6.973	ng/ul	92
50) Acenaphthylene	14.187	152	1033517	10.386	ng/ul	99
51) 3-Nitroaniline	14.387	138	117365	7.159	ng/ul	99
52) Acenaphthene	14.528	153	724897	10.299	ng/ul	98
53) 2,4-Dinitrophenol	14.598	184	50148	5.400	ng/ul	93
55) 4-Nitrophenol	14.692	109	92356	8.516	ng/ul	96
56) Dibenzofuran	14.863	168	1007696	10.468	ng/ul	100
57) 2,4-Dinitrotoluene	14.839	165	183066	8.278	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.092	232	188183	9.271	ng/ul	98
59) Diethylphthalate	15.292	149	798980	9.905	ng/ul	100
61) Fluorene	15.516	166	817970	10.698	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.510	204	396496	10.444	ng/ul	99
63) 4-Nitroaniline	15.551	138	106869	7.027	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.604	198	96589	7.107	ng/ul#	90
67) N-Nitrosodiphenylamine	15.728	169	698352	10.445	ng/ul	99
68) 4-Bromophenyl-phenylether	16.410	248	233838	9.613	ng/ul	98
69) Hexachlorobenzene	16.522	284	281396	9.808	ng/ul	99
70) Atrazine	16.686	200	211768	9.088	ng/ul	99
71) Pentachlorophenol	16.875	266	143308	8.238	ng/ul	97
72) Phenanthrene	17.263	178	1313438	10.569	ng/ul	100
74) Anthracene	17.357	178	1319600	10.742	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.257	216	336521	9.831	ng/uL	99
76) Pentachlorobenzene	14.781	250	362635	10.035	ng/uL	98
77) Carbazole	17.633	167	1175114	10.779	ng/ul	99
78) Di-n-butylphthalate	18.180	149	1260401	9.661	ng/ul	99
80) Fluoranthene	19.233	202	1467559	9.324	ng/ul	98
82) Pyrene	19.592	202	1566128	9.723	ng/ul	100
83) Butylbenzylphthalate	20.463	149	483761	7.718	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.239	252	444290	8.880	ng/ul	100
85) Benzo(a)anthracene	21.298	228	1527237	10.216	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.221	149	797113	8.845	ng/ul	98
87) Chrysene	21.351	228	1431230	10.283	ng/ul	99
89) Di-n-octyl phthalate	22.139	149	1370842	8.481	ng/ul	100
90) Benzo(b)fluoranthene	22.980	252	1572840	9.959	ng/ul	100
91) Benzo(k)fluoranthene	23.027	252	1511472	9.801	ng/ul	100
93) Benzo(a)pyrene	23.604	252	1387748	10.241	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.204	276	1711780	10.791	ng/ul	99
95) Dibenzo(a,h)anthracene	26.215	278	1533693	10.712	ng/ul	99
96) Benzo(g,h,i)perylene	26.962	276	1499444	10.066	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_P
ClientSampleId :
SSTD010604

Quant Time: Oct 26 10:06:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 26 10:05:01 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 4.126 ng/uL
RT: 3.287 min Scan# 51
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion: 88 Resp: 41743
Ion Ratio Lower Upper
88 100
43 27.2 20.8 31.2
58 76.4 57.5 86.3



#5
Pyridine
Concen: 9.654 ng/uL
RT: 3.699 min Scan# 121
Delta R.T. 0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion: 79 Resp: 257569
Ion Ratio Lower Upper
79 100
52 75.4 59.5 89.3
51 33.9 25.6 38.4









#6
Benzaldehyde
Concen: 11.041 ng/ul
RT: 6.963 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion: 77 Resp: 170028
Ion Ratio Lower Upper
77 100
105 100.1 81.5 122.3
106 94.2 78.7 118.1



#8
Phenol
Concen: 9.687 ng/ul
RT: 7.016 min Scan# 685
Delta R.T. 0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion: 94 Resp: 334846
Ion Ratio Lower Upper
94 100
65 25.9 20.8 31.2
66 32.6 26.4 39.6









#10
Bis(2-Chloroethyl)ether
Concen: 10.093 ng/ul
RT: 7.246 min Scan# 71
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion: 93 Resp: 281920
Ion Ratio Lower Upper
93 100
63 75.9 60.6 90.8
95 32.3 26.0 39.0



#12
2-Chlorophenol
Concen: 9.802 ng/ul
RT: 7.381 min Scan# 747
Delta R.T. 0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:128 Resp: 264890
Ion Ratio Lower Upper
128 100
64 44.5 36.8 55.2
130 32.2 25.4 38.2





#13

2-Methylphenol

Concen: 9.767 ng/ul

RT: 8.263 min Scan# 897

Delta R.T. 0.000 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01

Instrument :

BNA_P

ClientSampleId :





SSTD010604

Tgt Ion:108 Resp: 255078

Ion Ratio Lower Upper

108 100

107 87.0 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 10.683 ng/ul

RT: 8.340 min Scan# 910

Delta R.T. -0.006 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01




Tgt Ion: 45 Resp: 396383


Ion Ratio Lower Upper

45 100

77 14.5 11.9 17.9

79 10.6 9.0 13.6









#16
Acetophenone
Concen: 10.786 ng/ul
RT: 8.646 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:105 Resp: 431436
Ion Ratio Lower Upper
105 100
77 76.5 59.8 89.8
51 25.4 19.8 29.6



#17
N-Nitroso-di-n-propylamine
Concen: 9.756 ng/ul
RT: 8.628 min Scan# 959
Delta R.T. -0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion: 70 Resp: 196472
Ion Ratio Lower Upper
70 100
42 52.3 41.8 62.6
101 8.6 7.5 11.3
130 18.3 16.4 24.6









#18
4-Methylphenol
Concen: 9.952 ng/ul
RT: 8.593 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:108 Resp: 284204
Ion Ratio Lower Upper
108 100
107 111.5 89.1 133.7



#19
Hexachloroethane
Concen: 9.702 ng/ul
RT: 8.887 min Scan# 1003
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:117 Resp: 102872
Ion Ratio Lower Upper
117 100
201 96.7 78.3 117.5
199 60.0 49.0 73.6









#22
Nitrobenzene
Concen: 9.166 ng/ul
RT: 9.028 min Scan# 1027
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion: 77 Resp: 280204
Ion Ratio Lower Upper
77 100
123 45.0 37.7 56.5
65 12.8 10.2 15.4



#23
Isophorone
Concen: 10.068 ng/ul
RT: 9.552 min Scan# 1116
Delta R.T. -0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion: 82 Resp: 600554
Ion Ratio Lower Upper
82 100
95 6.7 5.3 7.9
138 17.1 14.6 21.8









#25
2-Nitrophenol
Concen: 8.159 ng/ul
RT: 9.734 min Scan# 1147
Delta R.T. -0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:139 Resp: 123907
Ion Ratio Lower Upper
139 100
65 49.9 36.6 54.8
109 28.6 20.4 30.6



#26
2,4-Dimethylphenol
Concen: 9.845 ng/ul
RT: 9.799 min Scan# 1158
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:107 Resp: 302045
Ion Ratio Lower Upper
107 100
121 54.3 41.9 62.9
122 92.4 71.8 107.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 10.204 ng/ul
 RT: 10.034 min Scan# 1198
 Delta R.T. 0.000 min
 Lab File: BP012318.D
 Acq: 25 Oct 2022 17:01




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010604


Tgt Ion: 93 Resp: 395906
 Ion Ratio Lower Upper
 93 100
 95 32.8 25.8 38.8
 123 11.1 8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 9.787 ng/ul
 RT: 10.269 min Scan# 1238
 Delta R.T. 0.000 min
 Lab File: BP012318.D
 Acq: 25 Oct 2022 17:01

Tgt Ion: 162 Resp: 258928
 Ion Ratio Lower Upper
 162 100
 164 66.0 52.6 78.8
 98 36.4 29.1 43.7









#30
Naphthalene
Concen: 10.237 ng/ul
RT: 10.669 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:128 Resp: 936369
Ion Ratio Lower Upper
128 100
129 10.7 8.8 13.2
127 13.1 10.3 15.5



#32
4-Chloroaniline
Concen: 10.438 ng/ul
RT: 10.793 min Scan# 1327
Delta R.T. 0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:127 Resp: 394372
Ion Ratio Lower Upper
127 100
129 33.0 25.7 38.5









#33
Hexachlorobutadiene
Concen: 9.878 ng/ul
RT: 10.946 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:225 Resp: 163402
Ion Ratio Lower Upper
225 100
223 61.9 50.8 76.2
227 62.4 52.2 78.2



#34
Caprolactam
Concen: 8.516 ng/ul
RT: 11.598 min Scan# 1464
Delta R.T. 0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:113 Resp: 70881
Ion Ratio Lower Upper
113 100
55 179.6 136.0 204.0
56 128.6 95.0 142.4









#35
4-Chloro-3-methylphenol
Concen: 9.773 ng/ul
RT: 11.916 min Scan# 1518
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:107 Resp: 280983
Ion Ratio Lower Upper
107 100
144 26.7 22.1 33.1
142 83.0 67.0 100.4



#36
2-Methylnaphthalene
Concen: 10.364 ng/ul
RT: 12.281 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:142 Resp: 645628
Ion Ratio Lower Upper
142 100
141 88.3 69.8 104.8









#37
1-Methylnaphthalene
Concen: 10.352 ng/ul
RT: 12.498 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:142 Resp: 645172
Ion Ratio Lower Upper
142 100
141 93.4 72.6 108.8
116 3.7 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.649 ng/ul
RT: 12.651 min Scan# 1643
Delta R.T. 0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:216 Resp: 328033
Ion Ratio Lower Upper
216 100
214 79.1 62.4 93.6
179 20.3 16.7 25.1
108 17.3 14.5 21.7









#40
Hexachlorocyclopentadiene
Concen: 5.925 ng/ul
RT: 12.622 min Scan# 1638
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:237 Resp: 103367
Ion Ratio Lower Upper
237 100
235 62.5 49.6 74.4
272 13.0 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 9.177 ng/ul
RT: 12.898 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:196 Resp: 200994
Ion Ratio Lower Upper
196 100
198 95.4 76.0 114.0
200 30.9 24.7 37.1









#42
2,4,5-Trichlorophenol
Concen: 9.219 ng/ul
RT: 12.969 min Scan# 1697
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:196 Resp: 220350
Ion Ratio Lower Upper
196 100
198 97.2 78.0 117.0
200 31.3 24.9 37.3



#43
1,1'-Biphenyl
Concen: 10.259 ng/ul
RT: 13.292 min Scan# 1752
Delta R.T. -0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:154 Resp: 857770
Ion Ratio Lower Upper
154 100
153 39.8 31.7 47.5
76 14.1 10.7 16.1









#44
2-Chloronaphthalene
Concen: 10.160 ng/ul
RT: 13.340 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:162 Resp: 671028
Ion Ratio Lower Upper
162 100
164 33.4 26.5 39.7
127 38.2 30.8 46.2



#45
2-Nitroaniline
Concen: 6.945 ng/ul
RT: 13.557 min Scan# 1797
Delta R.T. 0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion: 65 Resp: 119166
Ion Ratio Lower Upper
65 100
92 67.7 56.1 84.1
138 100.3 84.0 126.0









#47
Dimethylphthalate
Concen: 10.156 ng/ul
RT: 13.928 min Scan# 1860
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:163 Resp: 833263
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 10.3 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 6.973 ng/ul
RT: 14.051 min Scan# 1881
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:165 Resp: 106866
Ion Ratio Lower Upper
165 100
89 64.3 47.0 70.6
121 26.7 18.1 27.1









#50
Acenaphthylene
Concen: 10.386 ng/ul
RT: 14.187 min Scan# 1904
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:152 Resp: 1033517
Ion Ratio Lower Upper
152 100
151 19.4 16.0 24.0
153 12.7 10.6 15.8



#51
3-Nitroaniline
Concen: 7.159 ng/ul
RT: 14.387 min Scan# 1938
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:138 Resp: 117365
Ion Ratio Lower Upper
138 100
108 13.3 9.8 14.6
92 116.6 92.5 138.7









#52
Acenaphthene
Concen: 10.299 ng/ul
RT: 14.528 min Scan# 1962
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:153 Resp: 724897
Ion Ratio Lower Upper
153 100
152 49.2 38.6 58.0
154 88.0 72.0 108.0



#53
2,4-Dinitrophenol
Concen: 5.400 ng/ul
RT: 14.598 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:184 Resp: 50148
Ion Ratio Lower Upper
184 100
63 70.2 48.8 73.2
154 59.6 46.2 69.4









#55
4-Nitrophenol
Concen: 8.516 ng/ul
RT: 14.692 min Scan# 1991
Delta R.T. -0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:109 Resp: 92356
Ion Ratio Lower Upper
109 100
139 147.3 123.7 185.5
65 144.2 112.3 168.5



#56
Dibenzofuran
Concen: 10.468 ng/ul
RT: 14.863 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:168 Resp: 1007696
Ion Ratio Lower Upper
168 100
139 36.5 29.4 44.0





#57

2,4-Dinitrotoluene

Concen: 8.278 ng/ul

RT: 14.839 min Scan# 2016

Delta R.T. -0.006 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD010604





Tgt Ion:165 Resp: 183066

Ion Ratio Lower Upper

165 100

63 46.6 34.7 52.1

182 2.3 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 9.271 ng/ul

RT: 15.092 min Scan# 2058

Delta R.T. 0.000 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01

Tgt Ion:232 Resp: 188183




Ion Ratio Lower Upper


232 100

131 47.8 36.6 54.8

130 2.3 1.8 2.8

166 31.6 25.4 38.2









#59
Diethylphthalate
Concen: 9.905 ng/ul
RT: 15.292 min Scan# 2092
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:149 Resp: 798980
Ion Ratio Lower Upper
149 100
177 22.1 17.5 26.3
150 12.7 10.0 15.0



#61
Fluorene
Concen: 10.698 ng/ul
RT: 15.516 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:166 Resp: 817970
Ion Ratio Lower Upper
166 100
165 97.2 76.9 115.3
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 10.444 ng/ul
RT: 15.510 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:204 Resp: 396496
Ion Ratio Lower Upper
204 100
206 32.7 26.5 39.7
141 58.1 45.7 68.5



#63
4-Nitroaniline
Concen: 7.027 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:138 Resp: 106869
Ion Ratio Lower Upper
138 100
92 55.2 40.5 60.7
108 76.6 57.0 85.6









#66
4,6-Dinitro-2-methylphenol
Concen: 7.107 ng/ul
RT: 15.604 min Scan# 2145
Delta R.T. -0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:198 Resp: 96589
Ion Ratio Lower Upper
198 100
182 7.8 4.3 6.5#
77 28.9 19.0 28.4#



#67
N-Nitrosodiphenylamine
Concen: 10.445 ng/ul
RT: 15.728 min Scan# 2166
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:169 Resp: 698352
Ion Ratio Lower Upper
169 100
168 66.3 53.4 80.2
167 36.5 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 9.613 ng/ul
RT: 16.410 min Scan# 2282
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:248 Resp: 233838
Ion Ratio Lower Upper
248 100
250 96.4 78.6 118.0
141 70.9 57.6 86.4



#69
Hexachlorobenzene
Concen: 9.808 ng/ul
RT: 16.522 min Scan# 2301
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:284 Resp: 281396
Ion Ratio Lower Upper
284 100
142 33.5 27.6 41.4
249 29.9 23.8 35.6





#70

Atrazine

Concen: 9.088 ng/ul

RT: 16.686 min Scan# 211768

Delta R.T. -0.006 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD010604





Tgt Ion:200 Resp: 211768

Ion Ratio Lower Upper

200 100

173 27.7 21.8 32.6

215 48.3 38.3 57.5



#71

Pentachlorophenol

Concen: 8.238 ng/ul

RT: 16.875 min Scan# 2361

Delta R.T. 0.000 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01




Tgt Ion:266 Resp: 143308


Ion Ratio Lower Upper

266 100

264 64.6 50.6 75.8

268 62.3 52.3 78.5









#72
Phenanthrene
Concen: 10.569 ng/ul
RT: 17.263 min Scan# 2427
Delta R.T. -0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:178 Resp: 1313438
Ion Ratio Lower Upper
178 100
179 15.6 12.4 18.6
176 19.4 15.7 23.5



#74
Anthracene
Concen: 10.742 ng/ul
RT: 17.357 min Scan# 2443
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:178 Resp: 1319600
Ion Ratio Lower Upper
178 100
179 15.6 12.5 18.7
176 18.4 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 9.831 ng/uL
RT: 13.257 min Scan# 1746
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:216 Resp: 336521
Ion Ratio Lower Upper
216 100
214 78.1 61.8 92.6
218 47.6 38.3 57.5



#76
Pentachlorobenzene
Concen: 10.035 ng/uL
RT: 14.781 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:250 Resp: 362635
Ion Ratio Lower Upper
250 100
248 62.1 50.5 75.7
252 62.1 51.5 77.3









#77
 Carbazole
 Concen: 10.779 ng/ul
 RT: 17.633 min Scan# 2490
 Delta R.T. 0.000 min
 Lab File: BP012318.D
 Acq: 25 Oct 2022 17:01




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010604


Tgt Ion:167 Resp: 1175114
 Ion Ratio Lower Upper
 167 100
 166 21.1 17.0 25.6
 139 12.7 10.3 15.5



#78
 Di-n-butylphthalate
 Concen: 9.661 ng/ul
 RT: 18.180 min Scan# 2583
 Delta R.T. 0.000 min
 Lab File: BP012318.D
 Acq: 25 Oct 2022 17:01

Tgt Ion:149 Resp: 1260401
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.4 11.2
 104 5.3 3.9 5.9









#80
Fluoranthene
Concen: 9.324 ng/ul
RT: 19.233 min Scan# 2762
Delta R.T. -0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD010604


Tgt Ion:202 Resp: 1467559
Ion Ratio Lower Upper
202 100
101 13.4 9.9 14.9
100 10.0 7.5 11.3



#82
Pyrene
Concen: 9.723 ng/ul
RT: 19.592 min Scan# 2823
Delta R.T. 0.000 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:202 Resp: 1566128
Ion Ratio Lower Upper
202 100
101 15.2 12.0 18.0
100 12.0 9.5 14.3





#83

Butylbenzylphthalate

Concen: 7.718 ng/ul

RT: 20.463 min Scan# 2971

Delta R.T. 0.000 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD010604





Tgt Ion:149 Resp: 483761

Ion Ratio Lower Upper

149 100

91 74.6 57.8 86.8

206 19.1 15.7 23.5



#84

3,3'-Dichlorobenzidine

Concen: 8.880 ng/ul

RT: 21.239 min Scan# 3103

Delta R.T. 0.000 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01




Tgt Ion:252 Resp: 444290


Ion Ratio Lower Upper

252 100

254 64.6 51.8 77.8

126 12.1 9.4 14.0





#85

Benzo(a)anthracene

Concen: 10.216 ng/ul

RT: 21.298 min Scan# 3114

Delta R.T. -0.006 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD010604





Tgt Ion:228 Resp: 1527237

Ion Ratio Lower Upper

228 100

229 20.1 15.7 23.5

226 27.5 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 8.845 ng/ul

RT: 21.221 min Scan# 3100

Delta R.T. 0.000 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01




Tgt Ion:149 Resp: 797113


Ion Ratio Lower Upper

149 100

167 26.2 21.8 32.8

279 3.8 3.2 4.8





#87

Chrysene

Concen: 10.283 ng/ul

RT: 21.351 min Scan# 3123

Delta R.T. -0.006 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD010604





Tgt Ion:228 Resp: 1431230

Ion Ratio Lower Upper

228 100

226 30.0 24.6 36.8

229 20.0 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 8.481 ng/ul




RT: 22.139 min Scan# 3256


Delta R.T. 0.000 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01

Tgt Ion:149 Resp: 1370842





#90

Benzo(b)fluoranthene

Concen: 9.959 ng/ul

RT: 22.980 min Scan# 3399

Delta R.T. 0.000 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD010604





Tgt Ion:252 Resp: 1572840

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.6

125 11.5 8.9 13.3



#91

Benzo(k)fluoranthene

Concen: 9.801 ng/ul

RT: 23.027 min Scan# 3407

Delta R.T. 0.000 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01




Tgt Ion:252 Resp: 1511472


Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.6

125 11.4 9.0 13.4





#93

Benzo(a)pyrene

Concen: 10.241 ng/ul

RT: 23.604 min Scan# 3506

Delta R.T. -0.006 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD010604





Tgt Ion:252 Resp: 1387748

Ion Ratio Lower Upper

252 100

253 21.5 17.5 26.3

125 12.7 9.4 14.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 10.791 ng/ul

RT: 26.204 min Scan# 3947

Delta R.T. 0.000 min

Lab File: BP012318.D

Acq: 25 Oct 2022 17:01




Tgt Ion:276 Resp: 1711780


Ion Ratio Lower Upper

276 100

138 28.1 21.4 32.2

277 25.2 20.2 30.4









#95
Dibenzo(a,h)anthracene
Concen: 10.712 ng/ul
RT: 26.215 min Scan# 3949
Delta R.T. -0.006 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Instrument :
BNA_P
ClientSampleId :
SSTD010604

Tgt Ion:278 Resp: 1533693

Ion	Ratio	Lower	Upper
278	100		
139	17.4	13.5	20.3
279	24.1	19.2	28.8



#96
Benzo(g,h,i)perylene
Concen: 10.066 ng/ul
RT: 26.962 min Scan# 4076
Delta R.T. -0.012 min
Lab File: BP012318.D
Acq: 25 Oct 2022 17:01

Tgt Ion:276 Resp: 1499444

Ion	Ratio	Lower	Upper
276	100		
138	24.3	18.2	27.4
277	23.5	19.3	28.9

