

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102524\
 Data File : BP022643.D
 Acq On : 26 Oct 2024 15:20
 Operator : RC/JU
 Sample : PB164387BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK387

Quant Time: Oct 28 00:16:24 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 26 05:55:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.652	152	73631	20.000	ng/u1	0.00
20) Naphthalene-d8	10.404	136	280664	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.263	164	215305	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.063	188	531725	20.000	ng/u1	-0.02
79) Chrysene-d12	21.510	240	601141	20.000	ng/u1	0.01
88) Perylene-d12	24.774	264	701097	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.187	96	9521	5.025	ng/uL	-0.01
4) Pyridine-d5	3.599	84	137108	26.359	ng/u1	-0.01
7) Phenol-d5	6.846	99	165526	26.706	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.993	67	97636	26.802	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.193	132	126101	28.673	ng/u1	-0.01
15) 4-Methylphenol-d8	8.352	113	135663	27.427	ng/u1	-0.01
21) Nitrobenzene-d5	8.793	128	61116	28.320	ng/u1	0.00
24) 2-Nitrophenol-d4	9.504	143	75427	30.190	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.046	165	146698	27.626	ng/u1	0.00
31) 4-Chloroaniline-d4	10.546	131	174301	27.779	ng/u1	-0.01
46) Dimethylphthalate-d6	13.663	166	477397	27.907	ng/u1	-0.02
49) Acenaphthylene-d8	13.951	160	517763	28.497	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.516	143	67612	23.560	ng/u1	0.01
60) Fluorene-d10	15.275	176	420543	28.343	ng/u1	-0.02
65) 4,6-Dinitro-2-methylph...	15.416	200	71680	20.497	ng/u1	0.00
73) Anthracene-d10	17.175	188	722394	28.880	ng/u1	0.00
81) Pyrene-d10	19.545	212	950108	29.887	ng/u1	-0.01
92) Benzo(a)pyrene-d12	24.551	264	1127786	30.121	ng/u1	0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102524\
Data File : BP022643.D
Acq On : 26 Oct 2024 15:20
Operator : RC/JU
Sample : PB164387BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK387

Quant Time: Oct 28 00:16:24 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 26 05:55:04 2024
Response via : Initial Calibration

