

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102824\
 Data File : BP022681.D
 Acq On : 28 Oct 2024 15:38
 Operator : RC/JU
 Sample : P4529-04
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 F7H13

Quant Time: Oct 28 16:21:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 26 05:55:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.652	152	68784	20.000	ng/u1	0.00
20) Naphthalene-d8	10.404	136	257448	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.257	164	201051	20.000	ng/u1	-0.02
64) Phenanthrene-d10	17.051	188	495894	20.000	ng/u1	-0.03
79) Chrysene-d12	21.492	240	586557	20.000	ng/u1	0.00
88) Perylene-d12	24.757	264	719966	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.187	96	10954	6.189	ng/uL	-0.01
4) Pyridine-d5	3.599	84	156180	32.141	ng/u1	-0.01
7) Phenol-d5	6.840	99	192787	33.296	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	6.993	67	114990	33.790	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.193	132	146632	35.690	ng/u1	-0.01
15) 4-Methylphenol-d8	8.352	113	158996	34.409	ng/u1	-0.01
21) Nitrobenzene-d5	8.787	128	72182	36.464	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.504	143	86696	37.829	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	168544	34.602	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.546	131	195140	33.905	ng/u1	-0.01
46) Dimethylphthalate-d6	13.663	166	586817	36.736	ng/u1	-0.02
49) Acenaphthylene-d8	13.945	160	618264	36.441	ng/u1	-0.02
54) 4-Nitrophenol-d4	14.492	143	71426	26.653	ng/u1	-0.01
60) Fluorene-d10	15.263	176	504666	36.424	ng/u1	-0.03
65) 4,6-Dinitro-2-methylph...	15.404	200	98175	30.102	ng/u1	-0.01
73) Anthracene-d10	17.157	188	870913	37.333	ng/u1	-0.01
81) Pyrene-d10	19.539	212	1115115	35.950	ng/u1	-0.02
92) Benzo(a)pyrene-d12	24.533	264	1432349	37.252	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102824\
Data File : BP022681.D
Acq On : 28 Oct 2024 15:38
Operator : RC/JU
Sample : P4529-04
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
F7H13

Quant Time: Oct 28 16:21:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 26 05:55:04 2024
Response via : Initial Calibration

