

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103122\
 Data File : BP012383.D
 Acq On : 31 Oct 2022 12:38
 Operator : CG/JU
 Sample : PB148440BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK440

Quant Time: Oct 31 23:49:51 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 28 06:05:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.799	152	284628	20.000	ng/u1	0.00
20) Naphthalene-d8	10.599	136	1230403	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	796920	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.210	188	1811390	20.000	ng/u1	0.00
79) Chrysene-d12	21.304	240	2149284	20.000	ng/u1	0.00
88) Perylene-d12	23.698	264	2175087	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.240	96	42506	6.013	ng/uL	0.00
4) Pyridine-d5	3.658	84	532065	26.202	ng/u1	-0.01
7) Phenol-d5	6.969	99	723752	28.350	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	438920	28.801	ng/u1	0.00
11) 2-Chlorophenol-d4	7.328	132	568922	30.517	ng/u1	0.00
15) 4-Methylphenol-d8	8.511	113	571074	28.284	ng/u1	-0.01
21) Nitrobenzene-d5	8.963	128	281231	35.247	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.687	143	304554	36.775	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.222	165	523858	29.188	ng/u1	0.00
31) 4-Chloroaniline-d4	10.752	131	825849	30.904	ng/u1	0.00
46) Dimethylphthalate-d6	13.863	166	1753309	32.139	ng/u1	0.00
49) Acenaphthylene-d8	14.140	160	1956911	31.390	ng/u1	0.00
54) 4-Nitrophenol-d4	14.669	143	302905	29.337	ng/u1	0.00
60) Fluorene-d10	15.445	176	1545413	33.087	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.581	200	258240	27.382	ng/u1	0.00
73) Anthracene-d10	17.310	188	2601507	33.332	ng/u1	0.00
81) Pyrene-d10	19.551	212	3313661	30.708	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.545	264	3589470	33.774	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103122\
Data File : BP012383.D
Acq On : 31 Oct 2022 12:38
Operator : CG/JU
Sample : PB148440BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK440

Quant Time: Oct 31 23:49:51 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Oct 28 06:05:59 2022
Response via : Initial Calibration

