

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103122\
 Data File : BP012416.D
 Acq On : 01 Nov 2022 08:41
 Operator : CG/JU
 Sample : N5270-08
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BHA93

Quant Time: Nov 01 22:55:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 01 01:03:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	350622	20.000	ng/ul	0.00
20) Naphthalene-d8	10.593	136	1592818	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.440	164	1047146	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.204	188	2313989	20.000	ng/ul	0.00
79) Chrysene-d12	21.298	240	2544910	20.000	ng/ul	0.00
88) Perylene-d12	23.686	264	2511876	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	40672	4.671	ng/uL	0.00
4) Pyridine-d5	3.664	84	213381	8.530	ng/ul	0.00
7) Phenol-d5	6.963	99	266957	8.489	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	680573	36.252	ng/ul	0.00
11) 2-Chlorophenol-d4	7.322	132	697103	30.355	ng/ul	0.00
15) 4-Methylphenol-d8	8.505	113	488842	19.654	ng/ul	0.00
21) Nitrobenzene-d5	8.958	128	465414	45.059	ng/ul	0.00
24) 2-Nitrophenol-d4	9.681	143	492897	45.975	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.216	165	803060	34.564	ng/ul	0.00
31) 4-Chloroaniline-d4	10.740	131	1069526	30.916	ng/ul	0.00
46) Dimethylphthalate-d6	13.857	166	2879829	40.175	ng/ul	0.00
49) Acenaphthylene-d8	14.134	160	3234378	39.483	ng/ul	0.00
54) 4-Nitrophenol-d4	14.669	143	102459	7.552	ng/ul	0.00
60) Fluorene-d10	15.440	176	2526040	41.159	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	456261	37.871	ng/ul	0.00
73) Anthracene-d10	17.304	188	4112632	41.248	ng/ul	0.00
81) Pyrene-d10	19.545	212	4900273	38.352	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.533	264	5060751	41.234	ng/ul	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103122\
Data File : BP012416.D
Acq On : 01 Nov 2022 08:41
Operator : CG/JU
Sample : N5270-08
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BHA93

Quant Time: Nov 01 22:55:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 01 01:03:58 2022
Response via : Initial Calibration

