

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103122\
 Data File : BP012417.D
 Acq On : 01 Nov 2022 09:15
 Operator : CG/JU
 Sample : N5270-09
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BHA94

Quant Time: Nov 01 22:56:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 01 01:03:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	350009	20.000	ng/ul	0.00
20) Naphthalene-d8	10.593	136	1560078	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.439	164	1027496	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.204	188	2249770	20.000	ng/ul	0.00
79) Chrysene-d12	21.298	240	2412562	20.000	ng/ul	0.00
88) Perylene-d12	23.686	264	2273386	20.000	ng/ul	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.240	96	41678	4.795	ng/uL	0.00
4) Pyridine-d5	3.669	84	76065	3.046	ng/ul	0.01
7) Phenol-d5	6.963	99	274661	8.749	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	741693	39.577	ng/ul	0.00
11) 2-Chlorophenol-d4	7.322	132	738329	32.206	ng/ul	0.00
15) 4-Methylphenol-d8	8.504	113	514343	20.715	ng/ul	0.00
21) Nitrobenzene-d5	8.957	128	502645	49.684	ng/ul	0.00
24) 2-Nitrophenol-d4	9.681	143	534784	50.929	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.216	165	870462	38.251	ng/ul	0.00
31) 4-Chloroaniline-d4	10.740	131	770313	22.734	ng/ul	0.00
46) Dimethylphthalate-d6	13.857	166	3250896	46.218	ng/ul	0.00
49) Acenaphthylene-d8	14.134	160	3609582	44.906	ng/ul	0.00
54) 4-Nitrophenol-d4	14.669	143	111522	8.377	ng/ul	0.00
60) Fluorene-d10	15.439	176	2813005	46.711	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	530532	45.293	ng/ul	0.00
73) Anthracene-d10	17.304	188	4629431	47.757	ng/ul	0.00
81) Pyrene-d10	19.545	212	5569381	45.980	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.533	264	5396066	48.578	ng/ul	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103122\
Data File : BP012417.D
Acq On : 01 Nov 2022 09:15
Operator : CG/JU
Sample : N5270-09
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BHA94

Quant Time: Nov 01 22:56:04 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 01 01:03:58 2022
Response via : Initial Calibration

