

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103122\
 Data File : BP012421.D
 Acq On : 01 Nov 2022 11:33
 Operator : CG/JU
 Sample : N5270-14
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Nov 01 22:56:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 01 01:03:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	309689	20.000	ng/ul	0.00
20) Naphthalene-d8	10.593	136	1393248	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.445	164	901002	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.204	188	1973249	20.000	ng/ul	0.00
79) Chrysene-d12	21.298	240	2108176	20.000	ng/ul	0.00
88) Perylene-d12	23.686	264	2069549	20.000	ng/ul	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	38772	5.041	ng/uL	0.00
4) Pyridine-d5	3.664	84	185578	8.399	ng/ul	0.00
7) Phenol-d5	6.963	99	255969	9.215	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	692707	41.775	ng/ul	0.00
11) 2-Chlorophenol-d4	7.322	132	644953	31.796	ng/ul	0.00
15) 4-Methylphenol-d8	8.505	113	476255	21.679	ng/ul	0.00
21) Nitrobenzene-d5	8.963	128	467730	51.769	ng/ul	0.00
24) 2-Nitrophenol-d4	9.681	143	457405	48.776	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.216	165	748218	36.816	ng/ul	0.00
31) 4-Chloroaniline-d4	10.740	131	1080978	35.723	ng/ul	0.00
46) Dimethylphthalate-d6	13.857	166	2980648	48.326	ng/ul	0.00
49) Acenaphthylene-d8	14.134	160	3354609	47.593	ng/ul	0.00
54) 4-Nitrophenol-d4	14.669	143	94250	8.074	ng/ul	0.00
60) Fluorene-d10	15.439	176	2594409	49.129	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	431232	41.975	ng/ul	0.00
73) Anthracene-d10	17.304	188	4224453	49.686	ng/ul	0.00
81) Pyrene-d10	19.545	212	4977091	47.023	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.539	264	5089323	50.329	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103122\
Data File : BP012421.D
Acq On : 01 Nov 2022 11:33
Operator : CG/JU
Sample : N5270-14
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Nov 01 22:56:44 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 01 01:03:58 2022
Response via : Initial Calibration

