

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103122\
 Data File : BP012424.D
 Acq On : 01 Nov 2022 13:16
 Operator : CG/JU
 Sample : PB148496BS
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS496

Quant Time: Nov 01 22:57:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 01 01:03:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	448683	20.000	ng/u1	0.00
20) Naphthalene-d8	10.593	136	2018631	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	1290764	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.204	188	2805380	20.000	ng/u1	0.00
79) Chrysene-d12	21.304	240	2671479	20.000	ng/u1	0.00
88) Perylene-d12	23.692	264	2345999	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.240	96	55112	4.946	ng/uL	0.00
4) Pyridine-d5	3.658	84	429764	13.426	ng/u1	0.00
7) Phenol-d5	6.963	99	1122804	27.900	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	646393	26.906	ng/u1	0.00
11) 2-Chlorophenol-d4	7.322	132	854434	29.074	ng/u1	0.00
15) 4-Methylphenol-d8	8.510	113	909800	28.584	ng/u1	0.00
21) Nitrobenzene-d5	8.963	128	442634	33.814	ng/u1	0.00
24) 2-Nitrophenol-d4	9.681	143	496304	36.528	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.216	165	895911	30.426	ng/u1	0.00
31) 4-Chloroaniline-d4	10.740	131	720811	16.441	ng/u1	0.00
46) Dimethylphthalate-d6	13.863	166	2698187	30.536	ng/u1	0.00
49) Acenaphthylene-d8	14.134	160	3066683	30.370	ng/u1	0.00
54) 4-Nitrophenol-d4	14.669	143	574671	34.364	ng/u1	0.00
60) Fluorene-d10	15.445	176	2313999	30.588	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.581	200	512737	35.105	ng/u1	0.00
73) Anthracene-d10	17.310	188	3674586	30.399	ng/u1	0.00
81) Pyrene-d10	19.545	212	4347774	32.416	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.533	264	3634022	31.703	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.269	88	110433	9.322	ng/uL	99
5) Pyridine	3.675	79	684416	21.756	ng/u1	98
6) Benzaldehyde	6.940	77	613872	32.343	ng/u1	98
8) Phenol	6.993	94	1098009	26.241	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.222	93	853741	25.422	ng/u1	99
12) 2-Chlorophenol	7.357	128	841143	26.292	ng/u1	99
13) 2-Methylphenol	8.240	108	839338	26.084	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.328	45	1068193	23.718	ng/u1	99
16) Acetophenone	8.628	105	1327762	26.859	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.610	70	656277	27.576	ng/u1	99
18) 4-Methylphenol	8.575	108	934948	26.583	ng/u1	99
19) Hexachloroethane	8.863	117	324304	26.357	ng/u1	98
22) Nitrobenzene	9.004	77	974866	28.543	ng/u1	99
23) Isophorone	9.534	82	1924352	26.925	ng/u1	100
25) 2-Nitrophenol	9.716	139	514687	31.592	ng/u1	98
26) 2,4-Dimethylphenol	9.775	107	978283	27.347	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.016	93	1206619	26.382	ng/u1	100
29) 2,4-Dichlorophenol	10.246	162	846729	27.700	ng/u1	100
30) Naphthalene	10.645	128	2793282	26.109	ng/u1	100
32) 4-Chloroaniline	10.769	127	883919	19.594	ng/u1	99
33) Hexachlorobutadiene	10.916	225	497569	26.164	ng/u1	98
34) Caprolactam	11.575	113	155838	14.983	ng/u1	96
35) 4-Chloro-3-methylphenol	11.898	107	974219	29.006	ng/u1	99
36) 2-Methylnaphthalene	12.257	142	1937900	26.421	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103122\
 Data File : BP012424.D
 Acq On : 01 Nov 2022 13:16
 Operator : CG/JU
 Sample : PB148496BS
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS496

Quant Time: Nov 01 22:57:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 01 01:03:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.481	142	1937606	26.452	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.628	216	1013996	26.745	ng/ul	99
40) Hexachlorocyclopentadiene	12.598	237	460287	26.241	ng/ul	99
41) 2,4,6-Trichlorophenol	12.875	196	698084	28.805	ng/ul	99
42) 2,4,5-Trichlorophenol	12.951	196	780676	29.535	ng/ul	99
43) 1,1'-Biphenyl	13.275	154	2548133	26.847	ng/ul	100
44) 2-Chloronaphthalene	13.316	162	2011758	26.974	ng/ul	100
45) 2-Nitroaniline	13.534	65	614285	35.536	ng/ul	99
47) Dimethylphthalate	13.910	163	2611235	28.049	ng/ul	100
48) 2,6-Dinitrotoluene	14.034	165	557853	36.089	ng/ul	99
50) Acenaphthylene	14.163	152	3164270	27.943	ng/ul	100
51) 3-Nitroaniline	14.369	138	632189	33.401	ng/ul	97
52) Acenaphthene	14.510	153	2179593	27.388	ng/ul	99
53) 2,4-Dinitrophenol	14.581	184	353529	33.867	ng/ul	98
55) 4-Nitrophenol	14.681	109	392180	30.994	ng/ul	96
56) Dibenzofuran	14.845	168	3024784	27.841	ng/ul	100
57) 2,4-Dinitrotoluene	14.828	165	815369	35.021	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.075	232	664931	29.638	ng/ul	97
59) Diethylphthalate	15.275	149	2641237	29.158	ng/ul	98
61) Fluorene	15.498	166	2420243	28.036	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.492	204	1171706	27.672	ng/ul	99
63) 4-Nitroaniline	15.539	138	632782	37.005	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.592	198	495377	31.832	ng/ul	99
67) N-Nitrosodiphenylamine	15.710	169	2173783	27.512	ng/ul	100
68) 4-Bromophenyl-phenylether	16.392	248	799383	28.368	ng/ul	100
69) Hexachlorobenzene	16.504	284	955081	28.260	ng/ul	97
70) Atrazine	16.675	200	393226	14.366	ng/ul	99
71) Pentachlorophenol	16.857	266	571081	27.514	ng/ul	99
72) Phenanthrene	17.251	178	4120382	27.802	ng/ul	100
74) Anthracene	17.345	178	4127278	28.032	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.239	216	1069754	26.657	ng/ul	100
76) Pentachlorobenzene	14.763	250	1055939	24.917	ng/ul	99
77) Carbazole	17.622	167	3932376	29.988	ng/ul	100
78) Di-n-butylphthalate	18.163	149	4664437	30.618	ng/ul	100
80) Fluoranthene	19.221	202	4953932	29.915	ng/ul	100
82) Pyrene	19.574	202	5099424	29.730	ng/ul	100
83) Butylbenzylphthalate	20.451	149	2238982	35.012	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.221	252	1674673	28.858	ng/ul	99
85) Benzo(a)anthracene	21.286	228	4880730	28.637	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.204	149	3175008	32.470	ng/ul	99
87) Chrysene	21.339	228	4530276	28.433	ng/ul	100
89) Di-n-octyl phthalate	22.121	149	5281113	37.741	ng/ul	100
90) Benzo(b)fluoranthene	22.962	252	4638774	31.226	ng/ul	99
91) Benzo(k)fluoranthene	23.010	252	4316687	30.490	ng/ul	100
93) Benzo(a)pyrene	23.586	252	3766754	29.036	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.168	276	4276690	26.456	ng/ul	98
95) Dibenzo(a,h)anthracene	26.186	278	3693243	25.516	ng/ul	100
96) Benzo(g,h,i)perylene	26.933	276	3560063	24.913	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 9.322 ng/uL
RT: 3.269 min Scan# 48
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion: 88 Resp: 110433
Ion Ratio Lower Upper
88 100
43 26.4 20.6 31.0
58 72.1 58.0 87.0



#5
Pyridine
Concen: 21.756 ng/uL
RT: 3.675 min Scan# 117
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion: 79 Resp: 684416
Ion Ratio Lower Upper
79 100
52 69.3 57.4 86.0
51 31.7 25.7 38.5









#6
Benzaldehyde
Concen: 32.343 ng/ul
RT: 6.940 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion: 77 Resp: 613872
Ion Ratio Lower Upper
77 100
105 103.5 81.0 121.6
106 98.1 79.2 118.8



#8
Phenol
Concen: 26.241 ng/ul
RT: 6.993 min Scan# 681
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion: 94 Resp: 1098009
Ion Ratio Lower Upper
94 100
65 25.7 20.8 31.2
66 33.4 26.2 39.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 25.422 ng/ul
 RT: 7.222 min Scan# 720
 Delta R.T. -0.006 min
 Lab File: BP012424.D
 Acq: 01 Nov 2022 13:16

Instrument :
 BNA_P
 ClientSampleId :
 SLCS496

Tgt Ion: 93 Resp: 853741




Ion	Ratio	Lower	Upper
93	100		
63	73.7	59.9	89.9
95	32.2	25.7	38.5




#12
 2-Chlorophenol
 Concen: 26.292 ng/ul
 RT: 7.357 min Scan# 743
 Delta R.T. -0.000 min
 Lab File: BP012424.D
 Acq: 01 Nov 2022 13:16

Tgt Ion:128 Resp: 841143

Ion	Ratio	Lower	Upper
128	100		
64	43.7	35.1	52.7
130	33.1	25.8	38.6





#13

2-Methylphenol

Concen: 26.084 ng/ul

RT: 8.240 min Scan# 894

Delta R.T. -0.006 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16

Instrument :

BNA_P

ClientSampleId :





SLCS496

Tgt Ion:108 Resp: 839338

Ion Ratio Lower Upper

108 100

107 88.0 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 23.718 ng/ul

RT: 8.328 min Scan# 908

Delta R.T. -0.006 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16




Tgt Ion: 45 Resp: 1068193


Ion Ratio Lower Upper

45 100

77 15.5 12.1 18.1

79 12.4 9.7 14.5









#16
Acetophenone
Concen: 26.859 ng/ul
RT: 8.628 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:105 Resp: 1327762
Ion Ratio Lower Upper
105 100
77 72.7 60.3 90.5
51 23.1 19.9 29.9



#17
N-Nitroso-di-n-propylamine
Concen: 27.576 ng/ul
RT: 8.610 min Scan# 956
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion: 70 Resp: 656277
Ion Ratio Lower Upper
70 100
42 51.5 41.7 62.5
101 10.3 7.8 11.8
130 21.7 17.4 26.0









#18
4-Methylphenol
Concen: 26.583 ng/ul
RT: 8.575 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:108 Resp: 934948
Ion Ratio Lower Upper
108 100
107 110.3 89.1 133.7



#19
Hexachloroethane
Concen: 26.357 ng/ul
RT: 8.863 min Scan# 999
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:117 Resp: 324304
Ion Ratio Lower Upper
117 100
201 100.4 78.5 117.7
199 62.2 49.9 74.9









#22
Nitrobenzene
Concen: 28.543 ng/ul
RT: 9.004 min Scan# 1023
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion: 77 Resp: 974866
Ion Ratio Lower Upper
77 100
123 47.3 37.4 56.2
65 12.6 10.2 15.2



#23
Isophorone
Concen: 26.925 ng/ul
RT: 9.534 min Scan# 1113
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion: 82 Resp: 1924352
Ion Ratio Lower Upper
82 100
95 6.7 5.2 7.8
138 17.8 14.2 21.2









#25
2-Nitrophenol
Concen: 31.592 ng/ul
RT: 9.716 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:139 Resp: 514687
Ion Ratio Lower Upper
139 100
65 45.3 37.8 56.8
109 27.2 21.4 32.2



#26
2,4-Dimethylphenol
Concen: 27.347 ng/ul
RT: 9.775 min Scan# 1154
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:107 Resp: 978283
Ion Ratio Lower Upper
107 100
121 51.4 40.9 61.3
122 88.7 70.6 105.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 26.382 ng/ul
 RT: 10.016 min Scan# 1195
 Delta R.T. -0.000 min
 Lab File: BP012424.D
 Acq: 01 Nov 2022 13:16




Instrument :
 BNA_P
 ClientSampleId :
 SLCS496


Tgt Ion: 93 Resp: 1206619
 Ion Ratio Lower Upper
 93 100
 95 32.6 26.0 39.0
 123 11.5 9.0 13.4



#29
 2,4-Dichlorophenol
 Concen: 27.700 ng/ul
 RT: 10.246 min Scan# 1234
 Delta R.T. -0.006 min
 Lab File: BP012424.D
 Acq: 01 Nov 2022 13:16

Tgt Ion:162 Resp: 846729
 Ion Ratio Lower Upper
 162 100
 164 64.6 51.9 77.9
 98 35.7 28.6 43.0









#30
Naphthalene
Concen: 26.109 ng/ul
RT: 10.645 min Scan# 1302
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:128 Resp: 2793282
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.1 10.5 15.7



#32
4-Chloroaniline
Concen: 19.594 ng/ul
RT: 10.769 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:127 Resp: 883919
Ion Ratio Lower Upper
127 100
129 31.7 26.0 39.0









#33
Hexachlorobutadiene
Concen: 26.164 ng/ul
RT: 10.916 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:225 Resp: 497569
Ion Ratio Lower Upper
225 100
223 62.2 49.0 73.6
227 62.9 51.9 77.9



#34
Caprolactam
Concen: 14.983 ng/ul
RT: 11.575 min Scan# 1460
Delta R.T. 0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:113 Resp: 155838
Ion Ratio Lower Upper
113 100
55 163.7 135.6 203.4
56 118.2 98.9 148.3









#35
4-Chloro-3-methylphenol
Concen: 29.006 ng/ul
RT: 11.898 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:107 Resp: 974219
Ion Ratio Lower Upper
107 100
144 26.7 21.0 31.4
142 82.2 65.4 98.2



#36
2-Methylnaphthalene
Concen: 26.421 ng/ul
RT: 12.257 min Scan# 1576
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:142 Resp: 1937900
Ion Ratio Lower Upper
142 100
141 87.9 70.6 106.0





#37

1-Methylnaphthalene

Concen: 26.452 ng/ul

RT: 12.481 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16

Instrument :

BNA_P

ClientSampleId :

SLCS496





Tgt Ion:142 Resp: 1937606

Ion Ratio Lower Upper

142 100

141 91.9 73.1 109.7

116 3.7 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 26.745 ng/ul

RT: 12.628 min Scan# 1639

Delta R.T. -0.000 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16

Tgt Ion:216 Resp: 1013996




Ion Ratio Lower Upper


216 100

214 77.5 62.9 94.3

179 20.8 16.6 24.8

108 16.9 13.2 19.8









#40
Hexachlorocyclopentadiene
Concen: 26.241 ng/ul
RT: 12.598 min Scan# 1634
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:237 Resp: 460287
Ion Ratio Lower Upper
237 100
235 63.2 50.1 75.1
272 13.3 10.6 15.8



#41
2,4,6-Trichlorophenol
Concen: 28.805 ng/ul
RT: 12.875 min Scan# 1681
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:196 Resp: 698084
Ion Ratio Lower Upper
196 100
198 97.6 77.2 115.8
200 30.9 25.1 37.7





#42

2,4,5-Trichlorophenol

Concen: 29.535 ng/ul

RT: 12.951 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16

Instrument :

BNA_P

ClientSampleId :

SLCS496





Tgt Ion:196 Resp: 780676

Ion Ratio Lower Upper

196 100

198 96.7 77.8 116.8

200 31.5 25.0 37.4



#43

1,1'-Biphenyl

Concen: 26.847 ng/ul

RT: 13.275 min Scan# 1749

Delta R.T. -0.006 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16




Tgt Ion:154 Resp: 2548133


Ion Ratio Lower Upper

154 100

153 40.0 31.9 47.9

76 13.1 10.6 16.0









#44
2-Chloronaphthalene
Concen: 26.974 ng/ul
RT: 13.316 min Scan# 1756
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:162 Resp: 2011758
Ion Ratio Lower Upper
162 100
164 33.5 26.6 39.8
127 38.7 30.9 46.3



#45
2-Nitroaniline
Concen: 35.536 ng/ul
RT: 13.534 min Scan# 1793
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion: 65 Resp: 614285
Ion Ratio Lower Upper
65 100
92 72.3 58.5 87.7
138 114.5 91.1 136.7









#47
Dimethylphthalate
Concen: 28.049 ng/ul
RT: 13.910 min Scan# 1857
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:163 Resp: 2611235
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.1 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 36.089 ng/ul
RT: 14.034 min Scan# 1878
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:165 Resp: 557853
Ion Ratio Lower Upper
165 100
89 55.6 43.4 65.2
121 21.2 17.1 25.7









#50
Acenaphthylene
Concen: 27.943 ng/ul
RT: 14.163 min Scan# 1900
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:152 Resp: 3164270
Ion Ratio Lower Upper
152 100
151 19.9 16.1 24.1
153 13.3 10.8 16.2



#51
3-Nitroaniline
Concen: 33.401 ng/ul
RT: 14.369 min Scan# 1935
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:138 Resp: 632189
Ion Ratio Lower Upper
138 100
108 10.0 8.3 12.5
92 110.0 90.6 135.8









#52
Acenaphthene
Concen: 27.388 ng/ul
RT: 14.510 min Scan# 1959
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:153 Resp: 2179593
Ion Ratio Lower Upper
153 100
152 48.0 38.6 58.0
154 88.6 71.6 107.4



#53
2,4-Dinitrophenol
Concen: 33.867 ng/ul
RT: 14.581 min Scan# 1971
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:184 Resp: 353529
Ion Ratio Lower Upper
184 100
63 57.2 47.9 71.9
154 59.3 46.9 70.3









#55
4-Nitrophenol
Concen: 30.994 ng/ul
RT: 14.681 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:109 Resp: 392180
Ion Ratio Lower Upper
109 100
139 144.1 123.1 184.7
65 135.3 107.9 161.9



#56
Dibenzofuran
Concen: 27.841 ng/ul
RT: 14.845 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:168 Resp: 3024784
Ion Ratio Lower Upper
168 100
139 36.5 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 35.021 ng/ul
RT: 14.828 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:165 Resp: 815369
Ion Ratio Lower Upper
165 100
63 41.7 33.1 49.7
182 2.6 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 29.638 ng/ul
RT: 15.075 min Scan# 2055
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:232 Resp: 664931
Ion Ratio Lower Upper
232 100
131 44.6 33.6 50.4
130 2.3 1.8 2.6
166 30.9 24.2 36.2









#59
Diethylphthalate
Concen: 29.158 ng/ul
RT: 15.275 min Scan# 2089
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:149 Resp: 2641237
Ion Ratio Lower Upper
149 100
177 21.8 18.2 27.2
150 12.6 9.8 14.6



#61
Fluorene
Concen: 28.036 ng/ul
RT: 15.498 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:166 Resp: 2420243
Ion Ratio Lower Upper
166 100
165 97.2 77.8 116.8
167 13.6 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 27.672 ng/ul
RT: 15.492 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:204 Resp: 1171706
Ion Ratio Lower Upper
204 100
206 33.1 26.5 39.7
141 58.1 47.3 70.9



#63
4-Nitroaniline
Concen: 37.005 ng/ul
RT: 15.539 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:138 Resp: 632782
Ion Ratio Lower Upper
138 100
92 50.4 40.6 61.0
108 72.2 56.6 85.0









#66
4,6-Dinitro-2-methylphenol
Concen: 31.832 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:198 Resp: 495377
Ion Ratio Lower Upper
198 100
182 4.5 3.8 5.6
77 24.8 19.5 29.3



#67
N-Nitrosodiphenylamine
Concen: 27.512 ng/ul
RT: 15.710 min Scan# 2163
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:169 Resp: 2173783
Ion Ratio Lower Upper
169 100
168 65.8 52.9 79.3
167 36.2 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 28.368 ng/ul
RT: 16.392 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:248 Resp: 799383
Ion Ratio Lower Upper
248 100
250 96.8 77.4 116.2
141 70.4 56.2 84.2



#69
Hexachlorobenzene
Concen: 28.260 ng/ul
RT: 16.504 min Scan# 2298
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:284 Resp: 955081
Ion Ratio Lower Upper
284 100
142 31.6 27.1 40.7
249 29.6 24.5 36.7





#70

Atrazine

Concen: 14.366 ng/ul

RT: 16.675 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16

Instrument :

BNA_P

ClientSampleId :

SLCS496





Tgt Ion:200 Resp: 393226

Ion Ratio Lower Upper

200 100

173 27.2 21.8 32.6

215 49.8 39.0 58.6



#71

Pentachlorophenol

Concen: 27.514 ng/ul

RT: 16.857 min Scan# 2358

Delta R.T. -0.006 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16




Tgt Ion:266 Resp: 571081


Ion Ratio Lower Upper

266 100

264 63.8 50.3 75.5

268 64.1 51.4 77.0









#72
Phenanthrene
Concen: 27.802 ng/ul
RT: 17.251 min Scan# 2425
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:178 Resp: 4120382
Ion Ratio Lower Upper
178 100
179 15.6 12.6 18.8
176 19.5 15.8 23.6



#74
Anthracene
Concen: 28.032 ng/ul
RT: 17.345 min Scan# 2441
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:178 Resp: 4127278
Ion Ratio Lower Upper
178 100
179 15.4 12.5 18.7
176 18.8 15.3 22.9









#75
1,2,3,4-Tetrachlorobenzene
Concen: 26.657 ng/uL
RT: 13.239 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:216 Resp: 1069754
Ion Ratio Lower Upper
216 100
214 78.5 62.6 94.0
218 48.2 38.6 58.0



#76
Pentachlorobenzene
Concen: 24.917 ng/uL
RT: 14.763 min Scan# 2002
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:250 Resp: 1055939
Ion Ratio Lower Upper
250 100
248 62.0 50.6 75.8
252 64.1 50.7 76.1





#77

Carbazole

Concen: 29.988 ng/ul

RT: 17.622 min Scan# 2488

Delta R.T. -0.000 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16

Instrument :

BNA_P

ClientSampleId :

SLCS496





Tgt Ion:167 Resp: 3932376

Ion Ratio Lower Upper

167 100

166 21.4 17.1 25.7

139 12.8 10.4 15.6



#78

Di-n-butylphthalate

Concen: 30.618 ng/ul

RT: 18.163 min Scan# 2580

Delta R.T. -0.006 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16




Tgt Ion:149 Resp: 4664437


Ion Ratio Lower Upper

149 100

150 9.6 7.6 11.4

104 4.8 4.0 6.0









#80
Fluoranthene
Concen: 29.915 ng/ul
RT: 19.221 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:202 Resp: 4953932
Ion Ratio Lower Upper
202 100
101 12.3 9.9 14.9
100 9.3 7.4 11.2



#82
Pyrene
Concen: 29.730 ng/ul
RT: 19.574 min Scan# 2820
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:202 Resp: 5099424
Ion Ratio Lower Upper
202 100
101 14.5 11.8 17.6
100 11.6 9.3 13.9





#83

Butylbenzylphthalate

Concen: 35.012 ng/ul

RT: 20.451 min Scan# 2968

Delta R.T. -0.000 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16

Instrument :

BNA_P

ClientSampleId :

SLCS496





Tgt Ion:149 Resp: 2238982

Ion Ratio Lower Upper

149 100

91 67.2 56.9 85.3

206 21.4 16.2 24.4



#84

3,3'-Dichlorobenzidine

Concen: 28.858 ng/ul

RT: 21.221 min Scan# 3100

Delta R.T. -0.006 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16




Tgt Ion:252 Resp: 1674673


Ion Ratio Lower Upper

252 100

254 64.0 51.8 77.8

126 11.5 9.0 13.4





#85

Benzo(a)anthracene

Concen: 28.637 ng/ul

RT: 21.286 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16

Instrument :

BNA_P

ClientSampleId :

SLCS496





Tgt Ion:228 Resp: 4880730

Ion Ratio Lower Upper

228 100

229 20.1 16.2 24.2

226 28.1 22.6 33.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 32.470 ng/ul

RT: 21.204 min Scan# 3097

Delta R.T. -0.006 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16




Tgt Ion:149 Resp: 3175008


Ion Ratio Lower Upper

149 100

167 27.3 22.1 33.1

279 4.3 3.7 5.5









#87
Chrysene
Concen: 28.433 ng/ul
RT: 21.339 min Scan# 3120
Delta R.T. -0.000 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:228 Resp: 4530276
Ion Ratio Lower Upper
228 100
226 31.2 24.9 37.3
229 20.4 16.0 24.0



#89
Di-n-octyl phthalate
Concen: 37.741 ng/ul
RT: 22.121 min Scan# 3253
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:149 Resp: 5281113









#90
Benzo(b)fluoranthene
Concen: 31.226 ng/ul
RT: 22.962 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16




Instrument :
BNA_P
ClientSampleId :
SLCS496


Tgt Ion:252 Resp: 4638774
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.8
125 10.0 8.7 13.1



#91
Benzo(k)fluoranthene
Concen: 30.490 ng/ul
RT: 23.010 min Scan# 3404
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:252 Resp: 4316687
Ion Ratio Lower Upper
252 100
253 22.4 18.0 27.0
125 10.5 8.7 13.1





#93

Benzo(a)pyrene

Concen: 29.036 ng/ul

RT: 23.586 min Scan# 3502

Delta R.T. -0.006 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16

Instrument :

BNA_P

ClientSampleId :

SLCS496





Tgt Ion:252 Resp: 3766754

Ion Ratio Lower Upper

252 100

253 22.1 18.0 27.0

125 11.6 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 26.456 ng/ul

RT: 26.168 min Scan# 3941

Delta R.T. -0.018 min

Lab File: BP012424.D

Acq: 01 Nov 2022 13:16




Tgt Ion:276 Resp: 4276690


Ion Ratio Lower Upper

276 100

138 24.4 20.7 31.1

277 25.3 20.2 30.4









#95
Dibenzo(a,h)anthracene
Concen: 25.516 ng/ul
RT: 26.186 min Scan# 3943
Delta R.T. -0.006 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Instrument :
BNA_P
ClientSampleId :
SLCS496

Tgt Ion:278 Resp: 3693243
Ion Ratio Lower Upper
278 100
139 16.2 13.2 19.8
279 24.2 19.4 29.0



#96
Benzo(g,h,i)perylene
Concen: 24.913 ng/ul
RT: 26.933 min Scan# 4071
Delta R.T. -0.012 min
Lab File: BP012424.D
Acq: 01 Nov 2022 13:16

Tgt Ion:276 Resp: 3560063
Ion Ratio Lower Upper
276 100
138 22.2 17.8 26.8
277 24.2 19.5 29.3

