

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103122\
 Data File : BP012382.D
 Acq On : 31 Oct 2022 12:04
 Operator : CG/JU
 Sample : PB148438BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK438

Quant Time: Oct 31 23:49:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 28 06:05:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Dichlorobenzene-d4	7.799	152	297256	20.000 ng/u1	0.00
20) Naphthalene-d8	10.599	136	1264670	20.000 ng/u1	0.00
38) Acenaphthene-d10	14.451	164	814939	20.000 ng/u1	0.00
64) Phenanthrene-d10	17.210	188	1815127	20.000 ng/u1	0.00
79) Chrysene-d12	21.310	240	1981346	20.000 ng/u1	0.00
88) Perylene-d12	23.698	264	1965941	20.000 ng/u1	0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.240	96	41895	5.675 ng/uL	0.00
4) Pyridine-d5	3.658	84	540884	25.505 ng/u1	-0.01
7) Phenol-d5	6.969	99	724696	27.181 ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	436821	27.445 ng/u1	0.00
11) 2-Chlorophenol-d4	7.328	132	572281	29.393 ng/u1	0.00
15) 4-Methylphenol-d8	8.510	113	568801	26.974 ng/u1	-0.01
21) Nitrobenzene-d5	8.963	128	278867	34.004 ng/u1	-0.01
24) 2-Nitrophenol-d4	9.687	143	309224	36.327 ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.222	165	524954	28.457 ng/u1	0.00
31) 4-Chloroaniline-d4	10.746	131	814487	29.653 ng/u1	0.00
46) Dimethylphthalate-d6	13.863	166	1782677	31.955 ng/u1	0.00
49) Acenaphthylene-d8	14.140	160	1957180	30.700 ng/u1	0.00
54) 4-Nitrophenol-d4	14.669	143	308769	29.244 ng/u1	0.00
60) Fluorene-d10	15.445	176	1542402	32.292 ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.581	200	262039	27.728 ng/u1	0.00
73) Anthracene-d10	17.310	188	2566735	32.819 ng/u1	0.00
81) Pyrene-d10	19.551	212	3181898	31.986 ng/u1	0.00
92) Benzo(a)pyrene-d12	23.545	264	3173355	33.036 ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103122\
Data File : BP012382.D
Acq On : 31 Oct 2022 12:04
Operator : CG/JU
Sample : PB148438BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK438

Quant Time: Oct 31 23:49:41 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Oct 28 06:05:59 2022
Response via : Initial Calibration

