



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110221\
 Data File : BP007803.D
 Acq On : 02 Nov 2021 10:42
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110221.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 02 19:18:55 2021






AutoFind: Scans 2462, 2463, 2464; Background Corrected with Scan 2454

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.0	111576	PASS
68	69	0.00	2	1.9	2160	PASS
69	198	0.00	100	29.2	116337	PASS
70	69	0.00	2	0.5	600	PASS
127	198	10	80	42.0	167142	PASS
197	198	0.00	2	0.1	534	PASS
198	198	100	100	100.0	398037	PASS
199	198	5	9	6.7	26716	PASS
275	198	10	60	29.0	115568	PASS
365	198	1	100	4.4	17655	PASS
441	198	0.01	100	13.8	54989	PASS
442	442	100	100	100.0	349725	PASS
443	442	15	24	19.6	68580	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110221\
Data File : BP007803.D
Acq On : 02 Nov 2021 10:42
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Nov 02 11:56:12 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 02 11:49:47 2021
Response via : Initial Calibration



TIC: BP007803.D\data.ms

(70) Pentachlorophenol (C)

16.969min (-0.012) 3798353.28 ng




response 476229

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.60	63.77
264.00	60.90	64.17
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110221\
Data File : BP007803.D
Acq On : 02 Nov 2021 10:42
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Nov 02 11:56:12 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 02 11:49:47 2021
Response via : Initial Calibration



TIC: BP007803.D\data.ms

(77) Benzidine

19.510min (-0.006) 8179829.50 ng

response 2826170

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	14.06
183.00	11.40	11.89
0.00	0.00	0.00

DDT Breakdown

Instrument :
BNA_P
ClientSampleId :
DFTPP

Date	Instrument Name	DFTPP Data File
11/2/2021	BNA_P	BP007803.D
Compound Name	Response	Retention Time
DDT	1513908	20.71
DDD	3329	20.28
DDE	390	19.78
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3719	1517627	0.25