

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110420\
 Data File : BP003814.D
 Acq On : 04 Nov 2020 19:47
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 00:01:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 03 16:37:10 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.517	0.475	8.1	113	0.00
3	Pyridine	1.516	1.397	7.8	111	0.00
4	n-Nitrosodimethylamine	0.698	0.642	8.0	111	0.00
5 S	2-Fluorophenol	2.397	2.280	4.9	115	0.00
6	Aniline	1.953	1.814	7.1	110	0.00
7 S	Phenol-d6	3.440	3.212	6.6	112	0.00
8	2-Chlorophenol	1.272	1.200	5.7	112	0.00
9	Benzaldehyde	0.873	0.762	12.7	121	0.00
10 C	Phenol	1.660	1.551	6.6	113	0.00
11	bis(2-Chloroethyl)ether	1.334	1.232	7.6	111	0.00
12	1,3-Dichlorobenzene	1.559	1.449	7.1	114	0.00
13 C	1,4-Dichlorobenzene	1.572	1.468	6.6	114	0.00
14	1,2-Dichlorobenzene	1.501	1.394	7.1	113	0.00
15	Benzyl Alcohol	1.191	1.116	6.3	109	0.00
16	2,2'-oxybis(1-Chloropropane	2.247	2.011	10.5	109	0.00
17	2-Methylphenol	1.090	1.017	6.7	111	0.00
18	Hexachloroethane	0.559	0.533	4.7	113	0.00
19 P	n-Nitroso-di-n-propylamine	1.028	0.956	7.0	109	0.00
20	3+4-Methylphenols	1.455	1.361	6.5	110	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
22	Acetophenone	0.537	0.490	8.8	109	0.00
23 S	Nitrobenzene-d5	0.817	0.740	9.4	106	0.00
24	Nitrobenzene	0.406	0.375	7.6	109	0.00
25	Isophorone	0.694	0.644	7.2	108	0.00
26 C	2-Nitrophenol	0.158	0.165	-4.4	117	0.00
27	2,4-Dimethylphenol	0.264	0.248	6.1	112	0.00
28	bis(2-Chloroethoxy)methane	0.476	0.433	9.0	108	0.00
29 C	2,4-Dichlorophenol	0.319	0.304	4.7	111	0.00
30	1,2,4-Trichlorobenzene	0.388	0.365	5.9	113	0.00
31	Naphthalene	1.073	0.978	8.9	110	0.00
32	Benzoic acid	0.161	0.155	3.7	108	0.00
33	4-Chloroaniline	0.456	0.418	8.3	108	0.00
34 C	Hexachlorobutadiene	0.253	0.237	6.3	113	0.00
35	Caprolactam	0.098	0.095	3.1	108	0.00
36 C	4-Chloro-3-methylphenol	0.344	0.322	6.4	110	0.00
37	2-Methylnaphthalene	0.768	0.703	8.5	110	0.00
38	1-Methylnaphthalene	0.733	0.663	9.5	110	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
40	1,2,4,5-Tetrachlorobenzene	0.662	0.617	6.8	109	0.00
41 P	Hexachlorocyclopentadiene	0.334	0.317	5.1	107	0.00
42 S	2,4,6-Tribromophenol	0.500	0.487	2.6	110	0.00
43 C	2,4,6-Trichlorophenol	0.409	0.403	1.5	112	0.00
44	2,4,5-Trichlorophenol	0.431	0.419	2.8	111	0.00
45 S	2-Fluorobiphenyl	2.861	2.627	8.2	108	0.00
46	1,1'-Biphenyl	1.553	1.421	8.5	108	0.00
47	2-Chloronaphthalene	1.274	1.173	7.9	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110420\
 Data File : BP003814.D
 Acq On : 04 Nov 2020 19:47
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 00:01:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 03 16:37:10 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.380	0.377	0.8	109	0.00
49	Acenaphthylene	1.871	1.731	7.5	108	0.00
50	Dimethylphthalate	1.563	1.423	9.0	107	0.00
51	2,6-Dinitrotoluene	0.321	0.302	5.9	107	0.00
52 C	Acenaphthene	1.246	1.139	8.6	107	0.00
53	3-Nitroaniline	0.355	0.337	5.1	108	0.00
54 P	2,4-Dinitrophenol	0.138	0.109	21.0	89	0.00
55	Dibenzofuran	1.851	1.673	9.6	107	0.00
56 P	4-Nitrophenol	0.256	0.249	2.7	108	0.00
57	2,4-Dinitrotoluene	0.431	0.418	3.0	109	0.00
58	Fluorene	1.482	1.347	9.1	108	0.00
59	2,3,4,6-Tetrachlorophenol	0.388	0.371	4.4	108	0.00
60	Diethylphthalate	1.531	1.398	8.7	107	0.00
61	4-Chlorophenyl-phenylether	0.810	0.738	8.9	109	0.00
62	4-Nitroaniline	0.370	0.353	4.6	108	0.00
63	Azobenzene	1.445	1.303	9.8	105	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	0.098	0.088	10.2	100	0.00
66 c	n-Nitrosodiphenylamine	0.612	0.569	7.0	107	0.00
67	4-Bromophenyl-phenylether	0.234	0.219	6.4	108	0.00
68	Hexachlorobenzene	0.267	0.249	6.7	109	0.00
69	Atrazine	0.201	0.193	4.0	108	0.00
70 C	Pentachlorophenol	0.128	0.127	0.8	109	0.00
71	Phenanthrene	1.122	1.028	8.4	108	0.00
72	Anthracene	1.084	0.993	8.4	107	0.00
73	Carbazole	1.006	0.926	8.0	107	0.00
74	Di-n-butylphthalate	1.164	1.118	4.0	107	0.00
75 C	Fluoranthene	1.295	1.191	8.0	106	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
77	Benzydine	0.472	0.426	9.7	95	0.00
78	Pyrene	1.357	1.270	6.4	106	0.00
79 S	Terphenyl-d14	2.070	1.928	6.9	108	0.00
80	Butylbenzylphthalate	0.506	0.516	-2.0	109	0.00
81	Benzo(a)anthracene	1.319	1.231	6.7	107	0.00
82	3,3'-Dichlorobenzidine	0.455	0.439	3.5	109	0.00
83	Chrysene	1.267	1.174	7.3	107	0.00
84	Bis(2-ethylhexyl)phthalate	0.745	0.736	1.2	108	0.00
85 c	Di-n-octyl phthalate	1.242	1.232	0.8	108	0.00
86 I	Perylene-d12	1.000	1.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	1.443	1.306	9.5	108	0.01
88	Benzo(b)fluoranthene	1.309	1.219	6.9	114	0.01
89	Benzo(k)fluoranthene	1.292	1.130	12.5	103	0.00
90 C	Benzo(a)pyrene	1.212	1.100	9.2	109	0.01
91	Dibenzo(a,h)anthracene	1.203	1.088	9.6	107	0.01
92	Benzo(g,h,i)perylene	1.164	1.055	9.4	108	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110420\
Data File : BP003814.D
Acq On : 04 Nov 2020 19:47
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 05 00:01:46 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 03 16:37:10 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0