

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110424\
 Data File : BP022909.D
 Acq On : 07 Nov 2024 01:24
 Operator : RC/JU
 Sample : P4622-10
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH6N6

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/07/2024
 Supervised By :mohammad ahmed 11/08/2024

Quant Time: Nov 07 02:11:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 06 00:31:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.622	152	98954	20.000	ng/u1	0.00
20) Naphthalene-d8	10.381	136	378859	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.240	164	292225	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.045	188	698901	20.000	ng/u1	0.00
79) Chrysene-d12	21.498	240	727556	20.000	ng/u1	# 0.00
88) Perylene-d12	24.763	264	866870	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.175	96	5850	2.297	ng/uL	0.00
4) Pyridine-d5	3.605	84	9054m	1.295	ng/u1	0.02
7) Phenol-d5	6.822	99	130089	15.618	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.969	67	73358	14.984	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	98638	16.689	ng/u1	0.00
15) 4-Methylphenol-d8	8.328	113	107735	16.207	ng/u1	0.00
21) Nitrobenzene-d5	8.763	128	49451	16.976	ng/u1	0.00
24) 2-Nitrophenol-d4	9.481	143	57851	17.153	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.022	165	125444	17.501	ng/u1	0.00
31) 4-Chloroaniline-d4	10.522	131	135857	16.040	ng/u1	0.00
46) Dimethylphthalate-d6	13.651	166	396411	17.073	ng/u1	-0.01
49) Acenaphthylene-d8	13.928	160	480307	19.477	ng/u1	0.00
54) 4-Nitrophenol-d4	14.504	143	45508	11.683	ng/u1	0.00
60) Fluorene-d10	15.251	176	405494	20.135	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.422	200	19772m	4.301	ng/u1	0.00
73) Anthracene-d10	17.163	188	658706	20.035	ng/u1	0.00
81) Pyrene-d10	19.539	212	846798	22.009	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.533	264	950688	20.535	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110424\
Data File : BP022909.D
Acq On : 07 Nov 2024 01:24
Operator : RC/JU
Sample : P4622-10
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BH6N6

Quant Time: Nov 07 02:11:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 06 00:31:03 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/07/2024
Supervised By :mohammad ahmed 11/08/2024

