

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110620\
 Data File : BP003851.D
 Acq On : 07 Nov 2020 00:13
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 07 03:16:13 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 06 11:33:14 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	118	0.00
2	1,4-Dioxane	40.000	34.437	13.9	108	0.00
3	Pyridine	40.000	34.936	12.7	107	0.00
4	n-Nitrosodimethylamine	40.000	35.075	12.3	109	0.00
5 S	2-Fluorophenol	80.000	75.710	5.4	117	0.00
6	Aniline	40.000	36.087	9.8	110	0.00
7 S	Phenol-d6	80.000	73.137	8.6	112	0.00
8	2-Chlorophenol	40.000	37.671	5.8	115	0.00
9	Benzaldehyde	40.000	39.956	0.1	125	0.00
10 C	Phenol	40.000	36.409	9.0	112	0.00
11	bis(2-Chloroethyl)ether	40.000	35.369	11.6	109	0.00
12	1,3-Dichlorobenzene	40.000	37.459	6.4	117	0.00
13 C	1,4-Dichlorobenzene	40.000	37.311	6.7	116	0.00
14	1,2-Dichlorobenzene	40.000	36.968	7.6	115	0.00
15	Benzyl Alcohol	40.000	37.052	7.4	110	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	33.615	16.0	104	0.00
17	2-Methylphenol	40.000	36.550	8.6	111	0.00
18	Hexachloroethane	40.000	33.723	15.7	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.032	9.9	108	0.00
20	3+4-Methylphenols	40.000	36.934	7.7	111	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	111	0.00
22	Acetophenone	40.000	35.952	10.1	107	0.00
23 S	Nitrobenzene-d5	80.000	72.008	10.0	106	0.00
24	Nitrobenzene	40.000	36.369	9.1	108	0.00
25	Isophorone	40.000	37.079	7.3	108	0.00
26 C	2-Nitrophenol	40.000	40.297	-0.7	113	0.00
27	2,4-Dimethylphenol	40.000	37.469	6.3	112	0.00
28	bis(2-Chloroethoxy)methane	40.000	35.096	12.3	105	0.00
29 C	2,4-Dichlorophenol	40.000	39.103	2.2	114	0.00
30	1,2,4-Trichlorobenzene	40.000	37.985	5.0	114	0.00
31	Naphthalene	40.000	36.669	8.3	111	0.00
32	Benzoic acid	40.000	37.462	6.3	114	0.00
33	4-Chloroaniline	40.000	36.225	9.4	107	0.00
34 C	Hexachlorobutadiene	40.000	38.289	4.3	116	0.00
35	Caprolactam	40.000	39.242	1.9	111	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.536	6.2	110	0.00
37	2-Methylnaphthalene	40.000	36.146	9.6	109	0.00
38	1-Methylnaphthalene	40.000	35.772	10.6	109	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.207	4.5	111	0.00
41 P	Hexachlorocyclopentadiene	40.000	9.552	76.1#	27	0.00
42 S	2,4,6-Tribromophenol	80.000	81.907	-2.4	114	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.850	-2.1	114	0.00
44	2,4,5-Trichlorophenol	40.000	40.095	-0.2	113	0.00
45 S	2-Fluorobiphenyl	80.000	73.837	7.7	108	0.00
46	1,1'-Biphenyl	40.000	36.667	8.3	107	0.00
47	2-Chloronaphthalene	40.000	37.136	7.2	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110620\
 Data File : BP003851.D
 Acq On : 07 Nov 2020 00:13
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 07 03:16:13 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 06 11:33:14 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.624	-4.1	113	0.00
49	Acenaphthylene	40.000	37.168	7.1	107	0.00
50	Dimethylphthalate	40.000	36.191	9.5	105	0.00
51	2,6-Dinitrotoluene	40.000	36.893	7.8	104	0.00
52 C	Acenaphthene	40.000	34.831	12.9	101	0.00
53	3-Nitroaniline	40.000	39.762	0.6	112	0.00
54 P	2,4-Dinitrophenol	40.000	9.948	75.1#	18	0.00
55	Dibenzofuran	40.000	36.534	8.7	107	0.00
56 P	4-Nitrophenol	40.000	39.163	2.1	107	0.00
57	2,4-Dinitrotoluene	40.000	36.982	7.5	102	0.00
58	Fluorene	40.000	36.434	8.9	107	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.502	1.2	111	0.00
60	Diethylphthalate	40.000	36.291	9.3	105	0.00
61	4-Chlorophenyl-phenylether	40.000	36.498	8.8	108	0.00
62	4-Nitroaniline	40.000	41.153	-2.9	116	0.00
63	Azobenzene	40.000	35.353	11.6	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	40.000	13.449	66.4#	25	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.921	7.7	107	0.00
67	4-Bromophenyl-phenylether	40.000	38.128	4.7	110	0.00
68	Hexachlorobenzene	40.000	37.503	6.2	110	0.00
69	Atrazine	40.000	37.041	7.4	104	0.00
70 C	Pentachlorophenol	40.000	38.300	4.3	106	0.00
71	Phenanthrene	40.000	36.307	9.2	107	0.00
72	Anthracene	40.000	37.041	7.4	108	0.00
73	Carbazole	40.000	37.156	7.1	108	0.00
74	Di-n-butylphthalate	40.000	39.390	1.5	109	0.00
75 C	Fluoranthene	40.000	36.763	8.1	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	52.824	-32.1#	132	0.00
78	Pyrene	40.000	39.120	2.2	105	0.00
79 S	Terphenyl-d14	80.000	79.432	0.7	109	0.00
80	Butylbenzylphthalate	40.000	44.203	-10.5	112	0.00
81	Benzo(a)anthracene	40.000	37.691	5.8	102	0.00
82	3,3'-Dichlorobenzidine	40.000	41.628	-4.1	111	0.00
83	Chrysene	40.000	36.881	7.8	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.748	-4.4	108	0.00
85 c	Di-n-octyl phthalate	40.000	42.622	-6.6	109	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.148	9.6	104	0.00
88	Benzo(b)fluoranthene	40.000	35.435	11.4	104	0.00
89	Benzo(k)fluoranthene	40.000	34.650	13.4	99	0.00
90 C	Benzo(a)pyrene	40.000	35.693	10.8	103	0.00
91	Dibenzo(a,h)anthracene	40.000	36.223	9.4	104	0.00
92	Benzo(g,h,i)perylene	40.000	36.598	8.5	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110620\
Data File : BP003851.D
Acq On : 07 Nov 2020 00:13
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 07 03:16:13 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 06 11:33:14 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0