

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110620\
 Data File : BP003854.D
 Acq On : 07 Nov 2020 02:29
 Operator : CG/JU
 Sample : L4625-02MS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CONTAINER-001-A-B-COMPMS

Quant Time: Nov 07 03:17:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 06 11:33:14 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.304	152	152082	20.00	ng	0.00
21) Naphthalene-d8	11.134	136	582188	20.00	ng	# 0.00
39) Acenaphthene-d10	14.922	164	349505	20.00	ng	0.00
64) Phenanthrene-d10	17.645	188	722772	20.00	ng	# 0.00
76) Chrysene-d12	21.668	240	660468	20.00	ng	0.00
86) Perylene-d12	24.145	264	693153	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.828	112	1111159	121.94	ng	0.00
7) Phenol-d6	7.475	99	1302111	99.54	ng	0.00
23) Nitrobenzene-d5	9.469	82	996234	83.81	ng	0.00
42) 2,4,6-Tribromophenol	16.398	330	630214	144.17	ng	0.00
45) 2-Fluorobiphenyl	13.545	172	2174829	86.99	ng	0.00
79) Terphenyl-d14	20.133	244	3081049	90.13	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.528	88	143668	36.54	ng	# 1
3) Pyridine	3.964	79	401987	34.86	ng	# 62
4) n-Nitrosodimethylamine	3.858	42	226237	42.62	ng	# 56
6) Aniline	7.610	93	233977	15.75	ng	# 84
8) 2-Chlorophenol	7.875	128	484490	50.08	ng	82
9) Benzaldehyde	7.405	77	116027	16.42	ng	# 81
10) Phenol	7.499	94	536365	42.49	ng	83
11) bis(2-Chloroethyl)ether	7.693	93	481116	47.42	ng	# 81
12) 1,3-Dichlorobenzene	8.199	146	535966	45.20	ng	# 93
13) 1,4-Dichlorobenzene	8.340	146	548506	45.89	ng	97
14) 1,2-Dichlorobenzene	8.675	146	529386	46.37	ng	97
15) Benzyl Alcohol	8.540	79	452175	49.91	ng	89
16) 2,2'-oxybis(1-Chloropr...	8.834	45	722654	42.29	ng	82
17) 2-Methylphenol	8.763	107	397962	48.02	ng	# 93
18) Hexachloroethane	9.422	117	181729	42.78	ng	98
19) n-Nitroso-di-n-propyla...	9.110	70	375356	48.01	ng	# 78
20) 3+4-Methylphenols	9.093	107	522751	47.25	ng	93
22) Acetophenone	9.122	105	744180	47.62	ng	# 86
24) Nitrobenzene	9.510	77	566559	47.96	ng	# 82
25) Isophorone	10.034	82	1015456	50.29	ng	# 90
26) 2-Nitrophenol	10.234	139	253950	55.05	ng	# 75
27) 2,4-Dimethylphenol	10.298	122	424255	55.14	ng	94
28) bis(2-Chloroethoxy)met...	10.504	93	621550	44.86	ng	96
29) 2,4-Dichlorophenol	10.787	162	470256	50.72	ng	96
30) 1,2,4-Trichlorobenzene	11.004	180	528609	46.80	ng	100
31) Naphthalene	11.187	128	1490392	47.71	ng	100
32) Benzoic acid	10.457	122	224050	42.60	ng	# 80
33) 4-Chloroaniline	11.275	127	110178	8.30	ng	# 89
34) Hexachlorobutadiene	11.498	225	336389	45.66	ng	98
35) Caprolactam	12.010	113	107385	37.46	ng	# 22
36) 4-Chloro-3-methylphenol	12.398	107	495346	49.49	ng	90
37) 2-Methylnaphthalene	12.769	142	1070911	47.90	ng	97
38) 1-Methylnaphthalene	12.987	142	994258	46.62	ng	97
40) 1,2,4,5-Tetrachloroben...	13.145	216	590757	51.07	ng	99
41) Hexachlorocyclopentadiene	13.140	237	265361	45.47	ng	99
43) 2,4,6-Trichlorophenol	13.375	196	376498	52.65	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110620\
 Data File : BP003854.D
 Acq On : 07 Nov 2020 02:29
 Operator : CG/JU
 Sample : L4625-02MS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 CONTAINER-001-A-B-COMPMS


Quant Time: Nov 07 03:17:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 06 11:33:14 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.451	196	405027	53.80	ng	96
46) 1,1'-Biphenyl	13.757	154	1338338	49.33	ng	99
47) 2-Chloronaphthalene	13.804	162	1055541	47.42	ng	99
48) 2-Nitroaniline	13.981	65	337683	50.84	ng #	59
49) Acenaphthylene	14.639	152	1616339	49.45	ng	100
50) Dimethylphthalate	14.351	163	1427408	52.25	ng	99
51) 2,6-Dinitrotoluene	14.463	165	280015	49.95	ng #	73
52) Acenaphthene	14.986	154	1018286	46.76	ng	95
53) 3-Nitroaniline	14.792	138	132146	21.30	ng #	76
54) 2,4-Dinitrophenol	14.998	184	64465	29.43	ng #	1
55) Dibenzofuran	15.310	168	1573323	48.63	ng	99
56) 4-Nitrophenol	15.116	139	414670	92.72	ng #	63
57) 2,4-Dinitrotoluene	15.251	165	382628	50.84	ng #	76
58) Fluorene	15.957	166	1273732	49.17	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.545	232	349052	51.43	ng	97
60) Diethylphthalate	15.698	149	1270264	47.49	ng	99
61) 4-Chlorophenyl-phenyle...	15.933	204	675362	47.70	ng	97
62) 4-Nitroaniline	15.951	138	251038	38.81	ng	88
63) Azobenzene	16.228	77	1203242	47.64	ng	86
65) 4,6-Dinitro-2-methylph...	16.022	198	59807	19.68	ng #	80
66) n-Nitrosodiphenylamine	16.145	169	1116928	50.48	ng	97
67) 4-Bromophenyl-phenylether	16.822	248	415919	49.18	ng	97
68) Hexachlorobenzene	16.986	284	464939	48.18	ng	98
69) Atrazine	17.075	200	335043	46.13	ng	98
70) Pentachlorophenol	17.316	266	517283	112.06	ng	100
71) Phenanthrene	17.686	178	1955908	48.22	ng	99
72) Anthracene	17.775	178	2001776	51.10	ng	99
73) Carbazole	18.027	167	1800592	49.54	ng	99
74) Di-n-butylphthalate	18.545	149	2154144	51.20	ng	99
75) Fluoranthene	19.616	202	2285725	48.84	ng	99
77) Benzidine	19.763	184	834402	53.56	ng	99
78) Pyrene	19.963	202	2317256	51.70	ng	98
80) Butylbenzylphthalate	20.780	149	939117	56.23	ng #	86
81) Benzo(a)anthracene	21.651	228	2184960	50.17	ng	98
82) 3,3'-Dichlorobenzidine	21.563	252	680072	45.27	ng	99
83) Chrysene	21.710	228	2067008	49.39	ng	98
84) Bis(2-ethylhexyl)phtha...	21.533	149	1358090	55.17	ng	99
85) Di-n-octyl phthalate	22.463	149	2332730	56.88	ng #	87
87) Indeno(1,2,3-cd)pyrene	26.709	276	2457907	49.16	ng	97
88) Benzo(b)fluoranthene	23.392	252	2188656	48.25	ng	99
89) Benzo(k)fluoranthene	23.439	252	2127604	47.50	ng	98
90) Benzo(a)pyrene	24.039	252	1960336	46.68	ng	99
91) Dibenzo(a,h)anthracene	26.709	278	2090767	50.16	ng	98
92) Benzo(g,h,i)perylene	27.498	276	2076620	51.47	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 07 03:17:19 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 06 11:33:14 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 8.304 min Scan# 921
Delta R.T. -0.001 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:152 Resp: 152082
Ion Ratio Lower Upper
152 100
150 151.7 123.8 185.6
115 58.1 43.8 65.6



#2
1,4-Dioxane
Concen: 36.54 ng
RT: 3.528 min Scan# 109
Delta R.T. 0.012 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion: 88 Resp: 143668
Ion Ratio Lower Upper
88 100
58 83.5 44.2 66.4#
43 151.5 15.2 22.8#









#3
Pyridine
Concen: 34.86 ng
RT: 3.964 min Scan# 183
Delta R.T. 0.012 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion: 79 Resp: 401987
Ion Ratio Lower Upper
79 100
52 82.5 41.3 61.9#
51 40.6 21.2 31.8#



#4
n-Nitrosodimethylamine
Concen: 42.62 ng
RT: 3.858 min Scan# 165
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion: 42 Resp: 226237
Ion Ratio Lower Upper
42 100
74 119.2 148.6 223.0#
44 6.8 5.3 7.9









#5
2-Fluorophenol
Concen: 121.94 ng
RT: 5.828 min Scan# 500
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:112 Resp: 1111159
Ion Ratio Lower Upper
112 100
64 62.0 32.8 49.2#
63 32.1 17.0 25.4#



#6
Aniline
Concen: 15.75 ng
RT: 7.610 min Scan# 803
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion: 93 Resp: 233977
Ion Ratio Lower Upper
93 100
66 41.4 25.4 38.0#
65 21.0 12.5 18.7#









#7
Phenol-d6
Concen: 99.54 ng
RT: 7.475 min Scan# 780
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion: 99 Resp: 1302111
Ion Ratio Lower Upper
99 100
42 19.2 8.6 13.0#
71 35.5 23.4 35.2#



#8
2-Chlorophenol
Concen: 50.08 ng
RT: 7.875 min Scan# 848
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion: 128 Resp: 484490
Ion Ratio Lower Upper
128 100
130 32.1 13.0 53.0
64 50.3 10.4 50.4









#9
Benzaldehyde
Concen: 16.42 ng
RT: 7.405 min Scan# 768
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion: 77 Resp: 116027
Ion Ratio Lower Upper
77 100
105 95.1 97.1 137.1#
106 93.8 93.7 133.7



#10
Phenol
Concen: 42.49 ng
RT: 7.499 min Scan# 784
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion: 94 Resp: 536365
Ion Ratio Lower Upper
94 100
65 29.8 2.3 42.3
66 40.9 10.9 50.9









#11
bis(2-Chloroethyl)ether
Concen: 47.42 ng
RT: 7.693 min Scan# 817
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion: 93 Resp: 481116
Ion Ratio Lower Upper
93 100
63 80.2 39.0 79.0#
95 31.4 12.8 52.8



#12
1,3-Dichlorobenzene
Concen: 45.20 ng
RT: 8.199 min Scan# 903
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion: 146 Resp: 535966
Ion Ratio Lower Upper
146 100
148 63.4 52.5 78.7
75 31.8 18.1 27.1#









#13
1,4-Dichlorobenzene
Concen: 45.89 ng
RT: 8.340 min Scan# 927
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:146 Resp: 548506
Ion Ratio Lower Upper
146 100
148 64.3 51.3 76.9
111 41.1 28.9 43.3



#14
1,2-Dichlorobenzene
Concen: 46.37 ng
RT: 8.675 min Scan# 984
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:146 Resp: 529386
Ion Ratio Lower Upper
146 100
148 64.7 51.6 77.4
111 43.7 31.0 46.6









#15
Benzyl Alcohol
Concen: 49.91 ng
RT: 8.540 min Scan# 961
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS

Tgt Ion: 79 Resp: 452175




Ion	Ratio	Lower	Upper
79	100		
108	74.0	73.5	110.3
77	62.6	50.3	75.5




#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.29 ng
RT: 8.834 min Scan# 1011
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion: 45 Resp: 722654

Ion	Ratio	Lower	Upper
45	100		
77	12.9	2.3	42.3
79	9.6	0.0	36.6









#17
2-Methylphenol
Concen: 48.02 ng
RT: 8.763 min Scan# 999
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:	107	Resp:	397962
Ion	Ratio	Lower	Upper
107	100		
108	110.7	89.6	134.4
77	48.1	31.8	47.6#
79	48.0	31.1	46.7#



#18
Hexachloroethane
Concen: 42.78 ng
RT: 9.422 min Scan# 1111
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:	117	Resp:	181729
Ion	Ratio	Lower	Upper
117	100		
119	95.8	78.8	118.2
201	117.8	92.3	138.5









#19
n-Nitroso-di-n-propylamine
Concen: 48.01 ng
RT: 9.110 min Scan# 1058
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:	70	Resp:	375356
Ion	Ratio	Lower	Upper
70	100		
42	59.9	32.8	49.2#
101	10.3	9.6	14.4
130	21.4	24.3	36.5#



#20
3+4-Methylphenols
Concen: 47.25 ng
RT: 9.093 min Scan# 1055
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:	107	Resp:	522751
Ion	Ratio	Lower	Upper
107	100		
108	91.4	69.0	109.0
77	36.6	8.5	48.5
79	31.3	4.8	44.8









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 11.134 min Scan# 1402
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:	136	Resp:	582188
Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.6	13.0
54	9.3	3.6	5.4#
68	6.1	3.3	4.9#



#22
Acetophenone
Concen: 47.62 ng
RT: 9.122 min Scan# 1060
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:	105	Resp:	744180
Ion	Ratio	Lower	Upper
105	100		
71	3.2	2.6	3.8
51	29.7	13.5	20.3#
120	22.0	19.6	29.4









#23
Nitrobenzene-d5
Concen: 83.81 ng
RT: 9.469 min Scan# 1119
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion: 82 Resp: 996234
Ion Ratio Lower Upper
82 100
128 42.2 45.6 68.4#
54 52.8 30.1 45.1#



#24
Nitrobenzene
Concen: 47.96 ng
RT: 9.510 min Scan# 1126
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion: 77 Resp: 566559
Ion Ratio Lower Upper
77 100
123 44.7 48.6 72.8#
65 14.3 10.0 15.0









#25
Isophorone
Concen: 50.29 ng
RT: 10.034 min Scan# 1215
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion: 82 Resp: 1015456
Ion Ratio Lower Upper
82 100
95 7.6 6.4 9.6
138 16.9 18.7 28.1#



#26
2-Nitrophenol
Concen: 55.05 ng
RT: 10.234 min Scan# 1249
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion: 139 Resp: 253950
Ion Ratio Lower Upper
139 100
109 32.5 22.1 33.1
65 54.2 26.5 39.7#









#27
2,4-Dimethylphenol
Concen: 55.14 ng
RT: 10.298 min Scan# 1260
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:122 Resp: 424255
Ion Ratio Lower Upper
122 100
107 119.1 87.3 130.9
121 59.6 47.3 70.9



#28
bis(2-Chloroethoxy)methane
Concen: 44.86 ng
RT: 10.504 min Scan# 1295
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion: 93 Resp: 621550
Ion Ratio Lower Upper
93 100
95 31.6 26.7 40.1
123 11.2 10.5 15.7









#29
2,4-Dichlorophenol
Concen: 50.72 ng
RT: 10.787 min Scan# 1343
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:162 Resp: 470256
Ion Ratio Lower Upper
162 100
164 66.2 45.8 85.8
98 36.6 10.2 50.2



#30
1,2,4-Trichlorobenzene
Concen: 46.80 ng
RT: 11.004 min Scan# 1380
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:180 Resp: 528609
Ion Ratio Lower Upper
180 100
182 95.8 76.2 114.2
145 29.8 23.8 35.8









#31
Naphthalene
Concen: 47.71 ng
RT: 11.187 min Scan# 1411
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:128 Resp: 1490392
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 13.1 10.6 15.8



#32
Benzoic acid
Concen: 42.60 ng
RT: 10.457 min Scan# 1287
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:122 Resp: 224050
Ion Ratio Lower Upper
122 100
105 127.1 94.1 134.1
77 96.8 50.1 90.1#









#33
4-Chloroaniline
Concen: 8.30 ng
RT: 11.275 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:	127	Resp:	110178
Ion	Ratio	Lower	Upper
127	100		
129	31.7	25.7	38.5
65	33.6	17.3	25.9#
92	21.1	12.9	19.3#



#34
Hexachlorobutadiene
Concen: 45.66 ng
RT: 11.498 min Scan# 1464
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:	225	Resp:	336389
Ion	Ratio	Lower	Upper
225	100		
223	64.1	50.1	75.1
227	65.1	51.4	77.2









#35
Caprolactam
Concen: 37.46 ng
RT: 12.010 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:113 Resp: 107385
Ion Ratio Lower Upper
113 100
55 181.3 77.3 117.3#
56 135.0 55.9 95.9#



#36
4-Chloro-3-methylphenol
Concen: 49.49 ng
RT: 12.398 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:107 Resp: 495346
Ion Ratio Lower Upper
107 100
144 25.4 22.7 34.1
142 77.3 70.7 106.1









#37
2-Methylnaphthalene
Concen: 47.90 ng
RT: 12.769 min Scan# 1680
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:142 Resp: 1070911
Ion Ratio Lower Upper
142 100
141 86.8 71.4 107.2
115 35.0 25.8 38.6



#38
1-Methylnaphthalene
Concen: 46.62 ng
RT: 12.987 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:142 Resp: 994258
Ion Ratio Lower Upper
142 100
141 90.1 74.4 111.6
116 3.7 2.9 4.3









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.922 min Scan# 2046
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:164 Resp: 349505
Ion Ratio Lower Upper
164 100
162 100.0 79.4 119.2
160 43.7 36.1 54.1



#40
1,2,4,5-Tetrachlorobenzene
Concen: 51.07 ng
RT: 13.145 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:216 Resp: 590757
Ion Ratio Lower Upper
216 100
214 78.1 63.0 94.6
179 20.0 16.9 25.3
108 15.9 12.6 19.0









#41
Hexachlorocyclopentadiene
Concen: 45.47 ng
RT: 13.140 min Scan# 1743
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:237 Resp: 265361
Ion Ratio Lower Upper
237 100
235 62.8 43.2 83.2
272 13.1 0.0 33.3



#42
2,4,6-Tribromophenol
Concen: 144.17 ng
RT: 16.398 min Scan# 2297
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:330 Resp: 630214
Ion Ratio Lower Upper
330 100
332 96.6 77.1 115.7
141 31.0 23.4 35.2









#43
2,4,6-Trichlorophenol
Concen: 52.65 ng
RT: 13.375 min Scan# 1783
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:196 Resp: 376498
Ion Ratio Lower Upper
196 100
198 96.5 77.0 115.6
200 31.5 25.1 37.7



#44
2,4,5-Trichlorophenol
Concen: 53.80 ng
RT: 13.451 min Scan# 1796
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:196 Resp: 405027
Ion Ratio Lower Upper
196 100
198 95.4 77.3 115.9
97 46.3 31.0 46.4
132 25.3 20.9 31.3









#45
2-Fluorobiphenyl
Concen: 86.99 ng
RT: 13.545 min Scan# 1812
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:172 Resp: 2174829
Ion Ratio Lower Upper
172 100
171 34.9 28.5 42.7
170 22.2 19.2 28.8



#46
1,1'-Biphenyl
Concen: 49.33 ng
RT: 13.757 min Scan# 1848
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:154 Resp: 1338338
Ion Ratio Lower Upper
154 100
153 40.2 20.4 60.4
76 11.4 0.0 29.3









#47
2-Chloronaphthalene
Concen: 47.42 ng
RT: 13.804 min Scan# 1856
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:162 Resp: 1055541
Ion Ratio Lower Upper
162 100
127 38.1 29.9 44.9
164 33.0 26.1 39.1



#48
2-Nitroaniline
Concen: 50.84 ng
RT: 13.981 min Scan# 1886
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion: 65 Resp: 337683
Ion Ratio Lower Upper
65 100
92 66.7 70.6 106.0#
138 98.4 133.6 200.4#









#49
Acenaphthylene
Concen: 49.45 ng
RT: 14.639 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:152 Resp: 1616339
Ion Ratio Lower Upper
152 100
151 19.7 15.9 23.9
153 13.2 10.5 15.7



#50
Dimethylphthalate
Concen: 52.25 ng
RT: 14.351 min Scan# 1949
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:163 Resp: 1427408
Ion Ratio Lower Upper
163 100
194 4.5 3.3 4.9
164 10.1 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 49.95 ng
RT: 14.463 min Scan# 1968
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:165 Resp: 280015
Ion Ratio Lower Upper
165 100
63 57.3 27.8 41.6#
89 54.9 34.6 52.0#



#52
Acenaphthene
Concen: 46.76 ng
RT: 14.986 min Scan# 2057
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:154 Resp: 1018286
Ion Ratio Lower Upper
154 100
153 108.3 91.0 136.6
152 50.6 43.0 64.6









#53
3-Nitroaniline
Concen: 21.30 ng
RT: 14.792 min Scan# 2024
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:138 Resp: 132146
Ion Ratio Lower Upper
138 100
108 11.4 7.5 11.3#
92 118.6 75.0 112.6#



#54
2,4-Dinitrophenol
Concen: 29.43 ng
RT: 14.998 min Scan# 2059
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:184 Resp: 64465
Ion Ratio Lower Upper
184 100
63 91.9 32.3 48.5#
154 499.2 49.1 73.7#









#55
Dibenzofuran
Concen: 48.63 ng
RT: 15.310 min Scan# 2112
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:168 Resp: 1573323
Ion Ratio Lower Upper
168 100
139 35.8 29.1 43.7
169 13.2 10.4 15.6



#56
4-Nitrophenol
Concen: 92.72 ng
RT: 15.116 min Scan# 2079
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:139 Resp: 414670
Ion Ratio Lower Upper
139 100
109 78.0 44.6 84.6
65 105.7 41.6 81.6#









#57
2,4-Dinitrotoluene
Concen: 50.84 ng
RT: 15.251 min Scan# 2102
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:165 Resp: 382628
Ion Ratio Lower Upper
165 100
63 43.2 20.6 31.0#
89 68.1 42.6 63.8#
182 3.4 2.4 3.6



#58
Fluorene
Concen: 49.17 ng
RT: 15.957 min Scan# 2222
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:166 Resp: 1273732
Ion Ratio Lower Upper
166 100
165 95.1 77.8 116.8
167 13.4 11.0 16.4









#59
2,3,4,6-Tetrachlorophenol
Concen: 51.43 ng
RT: 15.545 min Scan# 2152
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:232 Resp: 349052
Ion Ratio Lower Upper
232 100
131 39.9 34.1 51.1
130 2.3 1.8 2.8
166 29.5 24.6 36.8



#60
Diethylphthalate
Concen: 47.49 ng
RT: 15.698 min Scan# 2178
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:149 Resp: 1270264
Ion Ratio Lower Upper
149 100
177 22.9 18.0 27.0
150 12.7 10.0 15.0









#61
4-Chlorophenyl-phenylether
Concen: 47.70 ng
RT: 15.933 min Scan# 2218
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:204 Resp: 675362
Ion Ratio Lower Upper
204 100
206 33.4 25.9 38.9
141 55.5 46.5 69.7



#62
4-Nitroaniline
Concen: 38.81 ng
RT: 15.951 min Scan# 2221
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:138 Resp: 251038
Ion Ratio Lower Upper
138 100
92 50.4 20.7 60.7
108 76.4 47.8 87.8









#63
Azobenzene
Concen: 47.64 ng
RT: 16.228 min Scan# 2268
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS

Tgt Ion: 77 Resp: 1203242




Ion	Ratio	Lower	Upper
77	100		
182	27.2	16.9	56.9
105	20.7	5.6	45.6
51	30.9	2.9	42.9




#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.645 min Scan# 2509
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion: 188 Resp: 722772

Ion	Ratio	Lower	Upper
188	100		
94	8.2	6.2	9.2
80	10.4	6.2	9.2#









#65
4,6-Dinitro-2-methylphenol
Concen: 19.68 ng
RT: 16.022 min Scan# 2233
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:198 Resp: 59807
Ion Ratio Lower Upper
198 100
51 47.6 6.4 46.4#
105 41.8 18.9 58.9



#66
n-Nitrosodiphenylamine
Concen: 50.48 ng
RT: 16.145 min Scan# 2254
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:169 Resp: 1116928
Ion Ratio Lower Upper
169 100
168 64.2 53.8 80.6
167 35.1 28.9 43.3









#67
4-Bromophenyl-phenylether
Concen: 49.18 ng
RT: 16.822 min Scan# 2369
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:248 Resp: 415919
Ion Ratio Lower Upper
248 100
250 97.0 76.7 115.1
141 70.3 53.0 79.6



#68
Hexachlorobenzene
Concen: 48.18 ng
RT: 16.986 min Scan# 2397
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:284 Resp: 464939
Ion Ratio Lower Upper
284 100
142 30.8 24.2 36.4
249 31.9 23.8 35.8









#69
Atrazine
Concen: 46.13 ng
RT: 17.075 min Scan# 2412
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:200 Resp: 335043
Ion Ratio Lower Upper
200 100
173 25.6 5.9 45.9
215 49.9 28.4 68.4



#70
Pentachlorophenol
Concen: 112.06 ng
RT: 17.316 min Scan# 2453
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:266 Resp: 517283
Ion Ratio Lower Upper
266 100
268 63.9 51.1 76.7
264 63.3 50.6 76.0









#71
Phenanthrene
Concen: 48.22 ng
RT: 17.686 min Scan# 2516
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:178 Resp: 1955908
Ion Ratio Lower Upper
178 100
176 18.6 15.3 22.9
179 15.7 12.5 18.7



#72
Anthracene
Concen: 51.10 ng
RT: 17.775 min Scan# 2531
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:178 Resp: 2001776
Ion Ratio Lower Upper
178 100
176 17.7 15.0 22.4
179 15.5 12.5 18.7









#73
 Carbazole
 Concen: 49.54 ng
 RT: 18.027 min Scan# 2574
 Delta R.T. -0.000 min
 Lab File: BP003854.D
 Acq: 07 Nov 2020 02:29

Instrument :
 BNA_P
 ClientSampleId :
 CONTAINER-001-A-B-COMPMS

Tgt Ion:167 Resp: 1800592




Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.6	26.4
139	12.5	10.2	15.4




#74
 Di-n-butylphthalate
 Concen: 51.20 ng
 RT: 18.545 min Scan# 2662
 Delta R.T. -0.000 min
 Lab File: BP003854.D
 Acq: 07 Nov 2020 02:29

Tgt Ion:149 Resp: 2154144

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.3	10.9
104	4.9	3.4	5.0









#75
Fluoranthene
Concen: 48.84 ng
RT: 19.616 min Scan# 2844
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:202 Resp: 2285725
Ion Ratio Lower Upper
202 100
101 9.7 0.0 29.4
203 17.6 0.0 37.8



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.668 min Scan# 3193
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:240 Resp: 660468
Ion Ratio Lower Upper
240 100
120 8.9 7.8 11.6
236 25.3 20.8 31.2









#77
Benzidine
Concen: 53.56 ng
RT: 19.763 min Scan# 2869
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:184 Resp: 834402
Ion Ratio Lower Upper
184 100
185 13.8 11.5 17.3
183 11.9 9.6 14.4



#78
Pyrene
Concen: 51.70 ng
RT: 19.963 min Scan# 2903
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:202 Resp: 2317256
Ion Ratio Lower Upper
202 100
200 20.3 17.1 25.7
203 17.5 14.2 21.2









#79
Terphenyl-d14
Concen: 90.13 ng
RT: 20.133 min Scan# 2932
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:244 Resp: 3081049
Ion Ratio Lower Upper
244 100
212 7.1 5.8 8.6
122 9.5 7.8 11.8



#80
Butylbenzylphthalate
Concen: 56.23 ng
RT: 20.780 min Scan# 3042
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:149 Resp: 939117
Ion Ratio Lower Upper
149 100
91 70.2 45.4 68.0#
206 22.9 17.1 25.7









#81
Benzo(a)anthracene
Concen: 50.17 ng
RT: 21.651 min Scan# 3190
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:228 Resp: 2184960
Ion Ratio Lower Upper
228 100
226 26.1 22.1 33.1
229 19.6 15.9 23.9



#82
3,3'-Dichlorobenzidine
Concen: 45.27 ng
RT: 21.563 min Scan# 3175
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:252 Resp: 680072
Ion Ratio Lower Upper
252 100
254 64.1 51.0 76.4
126 9.4 7.9 11.9









#83
Chrysene
Concen: 49.39 ng
RT: 21.710 min Scan# 3200
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:228 Resp: 2067008
Ion Ratio Lower Upper
228 100
226 28.9 24.4 36.6
229 19.7 15.8 23.6



#84
Bis(2-ethylhexyl)phthalate
Concen: 55.17 ng
RT: 21.533 min Scan# 3170
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:149 Resp: 1358090
Ion Ratio Lower Upper
149 100
167 28.9 22.6 34.0
279 5.6 4.2 6.2









#85
Di-n-octyl phthalate
Concen: 56.88 ng
RT: 22.463 min Scan# 3328
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:149 Resp: 2332730
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 10.2 3.7 5.5#



#86
Perylene-d12
Concen: 20.00 ng
RT: 24.145 min Scan# 3614
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:264 Resp: 693153
Ion Ratio Lower Upper
264 100
260 22.7 18.8 28.2
265 21.9 17.4 26.2









#87
Indeno(1,2,3-cd)pyrene
Concen: 49.16 ng
RT: 26.709 min Scan# 4050
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29




Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS


Tgt Ion:276 Resp: 2457907
Ion Ratio Lower Upper
276 100
138 20.9 18.7 28.1
277 25.7 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 48.25 ng
RT: 23.392 min Scan# 3486
Delta R.T. -0.000 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:252 Resp: 2188656
Ion Ratio Lower Upper
252 100
253 21.9 17.6 26.4
125 8.3 7.3 10.9





#89

Benzo(k)fluoranthene

Concen: 47.50 ng

RT: 23.439 min Scan# 3494

Delta R.T. -0.000 min

Lab File: BP003854.D

Acq: 07 Nov 2020 02:29

Instrument :





BNA_P

ClientSampleId :

CONTAINER-001-A-B-COMPMS

Tgt Ion:252 Resp: 2127604

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.8	26.8
125	8.2	7.4	11.2



#90

Benzo(a)pyrene

Concen: 46.68 ng

RT: 24.039 min Scan# 3596




Delta R.T. 0.006 min


Lab File: BP003854.D

Acq: 07 Nov 2020 02:29

Tgt Ion:252 Resp: 1960336

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.4	26.0
125	9.4	8.2	12.4









#91
Dibenzo(a,h)anthracene
Concen: 50.16 ng
RT: 26.709 min Scan# 4050
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Instrument :
BNA_P
ClientSampleId :
CONTAINER-001-A-B-COMPMS

Tgt Ion:278 Resp: 2090767
Ion Ratio Lower Upper
278 100
139 13.7 12.4 18.6
279 23.6 19.0 28.6



#92
Benzo(g,h,i)perylene
Concen: 51.47 ng
RT: 27.498 min Scan# 4184
Delta R.T. 0.006 min
Lab File: BP003854.D
Acq: 07 Nov 2020 02:29

Tgt Ion:276 Resp: 2076620
Ion Ratio Lower Upper
276 100
277 24.1 19.0 28.6
138 17.5 16.6 24.8

