



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110620\
 Data File : BP003850.D
 Acq On : 06 Nov 2020 23:39
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Nov 06 11:33:14 2020



AutoFind: Scans 2580, 2581, 2582; Background Corrected with Scan 2567

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.9	38065	PASS
68	69	0.00	2	1.1	429	PASS
69	198	0.00	100	34.4	38584	PASS
70	69	0.00	2	0.7	255	PASS
127	198	10	80	41.1	46066	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	112136	PASS
199	198	5	9	6.9	7763	PASS
275	198	10	60	29.5	33050	PASS
365	198	1	100	3.5	3948	PASS
441	198	0.01	100	16.9	18954	PASS
442	442	100	100	100.0	124064	PASS
443	442	15	24	19.2	23850	PASS