

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110722\
 Data File : BP012535.D
 Acq On : 07 Nov 2022 17:22
 Operator : CG/JU
 Sample : N5353-18
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DBWX8

Quant Time: Nov 07 21:41:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 04 02:53:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	273822	20.000	ng/u1	0.00
20) Naphthalene-d8	10.575	136	1334796	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.428	164	920991	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	2081960	20.000	ng/u1	0.00
79) Chrysene-d12	21.292	240	1984795	20.000	ng/u1	0.00
88) Perylene-d12	23.668	264	1828436	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	25656	3.773	ng/uL	0.00
4) Pyridine-d5	3.663	84	152212	7.792	ng/u1	0.01
7) Phenol-d5	6.957	99	171804	6.995	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.116	67	491682	33.536	ng/u1	0.00
11) 2-Chlorophenol-d4	7.310	132	489154	27.274	ng/u1	0.00
15) 4-Methylphenol-d8	8.499	113	345161	17.769	ng/u1	0.00
21) Nitrobenzene-d5	8.946	128	348490	40.261	ng/u1	0.00
24) 2-Nitrophenol-d4	9.669	143	363182	40.424	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.204	165	641286	32.937	ng/u1	0.00
31) 4-Chloroaniline-d4	10.728	131	690227	23.809	ng/u1	0.00
46) Dimethylphthalate-d6	13.845	166	2399876	38.065	ng/u1	0.00
49) Acenaphthylene-d8	14.122	160	2536218	35.201	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	53195	4.458	ng/u1	0.02
60) Fluorene-d10	15.428	176	2070666	38.360	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.563	200	355472	32.794	ng/u1	0.00
73) Anthracene-d10	17.292	188	3574563	39.847	ng/u1	0.00
81) Pyrene-d10	19.533	212	4249515	42.644	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.515	264	3664790	41.021	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110722\
Data File : BP012535.D
Acq On : 07 Nov 2022 17:22
Operator : CG/JU
Sample : N5353-18
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DBWX8

Quant Time: Nov 07 21:41:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 04 02:53:36 2022
Response via : Initial Calibration

