

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP110724\  
 Data File : BP022918.D  
 Acq On : 07 Nov 2024 14:31  
 Operator : RC/JU  
 Sample : SSTDICCC040  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDICCC040

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 11/08/2024  
 Supervised By :mohammad ahmed 11/08/2024

Quant Time: Nov 07 23:20:25 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP110724.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Nov 07 23:16:12 2024  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| -----                         |        |      |          |        |       |          |
| Internal Standards            |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.593  | 152  | 72496    | 20.000 | ng    | 0.00     |
| 21) Naphthalene-d8            | 10.345 | 136  | 270641   | 20.000 | ng    | 0.00     |
| 39) Acenaphthene-d10          | 14.216 | 164  | 222935   | 20.000 | ng    | 0.00     |
| 64) Phenanthrene-d10          | 17.022 | 188  | 550938   | 20.000 | ng    | 0.00     |
| 76) Chrysene-d12              | 21.463 | 240  | 656894   | 20.000 | ng    | 0.00     |
| 86) Perylene-d12              | 24.698 | 264  | 779446   | 20.000 | ng    | 0.00     |
| System Monitoring Compounds   |        |      |          |        |       |          |
| 5) 2-Fluorophenol             | 5.234  | 112  | 309338   | 80.977 | ng    | 0.00     |
| 7) Phenol-d6                  | 6.793  | 99   | 427244   | 80.198 | ng    | 0.00     |
| 23) Nitrobenzene-d5           | 8.740  | 82   | 462451   | 80.727 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol      | 15.745 | 330  | 341261   | 78.974 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl          | 12.816 | 172  | 1218301  | 78.085 | ng    | 0.00     |
| 79) Terphenyl-d14             | 19.757 | 244  | 2743251  | 78.428 | ng    | 0.00     |
| Target Compounds              |        |      |          |        |       |          |
|                               |        |      |          |        |       | Qvalue   |
| 2) 1,4-Dioxane                | 3.175  | 88   | 61241    | 40.154 | ng    | 100      |
| 3) Pyridine                   | 3.575  | 79   | 168818m  | 40.455 | ng    |          |
| 4) n-Nitrosodimethylamine     | 3.487  | 42   | 80888    | 40.069 | ng    | 100      |
| 6) Aniline                    | 6.940  | 93   | 179698   | 40.464 | ng    | 100      |
| 8) 2-Chlorophenol             | 7.169  | 128  | 161690   | 40.380 | ng    | 100      |
| 9) Benzaldehyde               | 6.757  | 77   | 122491   | 39.794 | ng    | 100      |
| 10) Phenol                    | 6.816  | 94   | 213872   | 40.537 | ng    | 100      |
| 11) bis(2-Chloroethyl)ether   | 7.034  | 93   | 158993   | 39.887 | ng    | 100      |
| 12) 1,3-Dichlorobenzene       | 7.481  | 146  | 202460   | 39.615 | ng    | 100      |
| 13) 1,4-Dichlorobenzene       | 7.634  | 146  | 204335   | 39.157 | ng    | 100      |
| 14) 1,2-Dichlorobenzene       | 7.940  | 146  | 199057   | 39.455 | ng    | 100      |
| 15) Benzyl Alcohol            | 7.840  | 79   | 160616   | 41.050 | ng    | 100      |
| 16) 2,2'-oxybis(1-Chloropr... | 8.110  | 45   | 170032   | 39.316 | ng    | 100      |
| 17) 2-Methylphenol            | 8.046  | 107  | 146377   | 40.872 | ng    | 100      |
| 18) Hexachloroethane          | 8.646  | 117  | 75976    | 39.357 | ng    | 100      |
| 19) n-Nitroso-di-n-propyla... | 8.393  | 70   | 146463   | 40.167 | ng    | 100      |
| 20) 3+4-Methylphenols         | 8.369  | 107  | 201846   | 40.326 | ng    | 100      |
| 22) Acetophenone              | 8.410  | 105  | 284548   | 39.499 | ng    | 100      |
| 24) Nitrobenzene              | 8.781  | 77   | 235280   | 40.433 | ng    | 100      |
| 25) Isophorone                | 9.293  | 82   | 387993   | 40.017 | ng    | 100      |
| 26) 2-Nitrophenol             | 9.481  | 139  | 96718    | 40.475 | ng    | 100      |
| 27) 2,4-Dimethylphenol        | 9.540  | 122  | 108108   | 39.621 | ng    | 100      |
| 28) bis(2-Chloroethoxy)met... | 9.775  | 93   | 219691   | 39.891 | ng    | 100      |
| 29) 2,4-Dichlorophenol        | 10.016 | 162  | 193512   | 40.972 | ng    | 100      |
| 30) 1,2,4-Trichlorobenzene    | 10.216 | 180  | 237828   | 39.388 | ng    | 100      |
| 31) Naphthalene               | 10.398 | 128  | 541529   | 39.934 | ng    | 100      |
| 32) Benzoic acid              | 9.728  | 122  | 137314   | 41.927 | ng    | 100      |
| 33) 4-Chloroaniline           | 10.522 | 127  | 197348   | 40.887 | ng    | 100      |
| 34) Hexachlorobutadiene       | 10.669 | 225  | 177534   | 40.087 | ng    | 100      |
| 35) Caprolactam               | 11.310 | 113  | 57343    | 40.556 | ng    | 100      |
| 36) 4-Chloro-3-methylphenol   | 11.651 | 107  | 214810   | 40.980 | ng    | 100      |
| 37) 2-Methylnaphthalene       | 12.004 | 142  | 404119   | 39.790 | ng    | 100      |
| 38) 1-Methylnaphthalene       | 12.234 | 142  | 406166   | 40.293 | ng    | 100      |
| 40) 1,2,4,5-Tetrachloroben... | 12.381 | 216  | 297343   | 39.397 | ng    | 100      |
| 41) Hexachlorocyclopentadiene | 12.351 | 237  | 89973    | 42.449 | ng    | 100      |
| 43) 2,4,6-Trichlorophenol     | 12.639 | 196  | 195066   | 40.638 | ng    | 100      |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP110724\  
 Data File : BP022918.D  
 Acq On : 07 Nov 2024 14:31  
 Operator : RC/JU  
 Sample : SSTDICCC040  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDICCC040

Manual Integrations  
 APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
 Supervised By :mohammad ahmed 11/08/2024

Quant Time: Nov 07 23:20:25 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP110724.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Nov 07 23:16:12 2024  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol     | 12.722 | 196  | 221187   | 40.479 | ng    | 100      |
| 46) 1,1'-Biphenyl             | 13.034 | 154  | 591369   | 39.904 | ng    | 100      |
| 47) 2-Chloronaphthalene       | 13.075 | 162  | 483833   | 39.340 | ng    | 100      |
| 48) 2-Nitroaniline            | 13.298 | 65   | 145626   | 41.193 | ng    | 100      |
| 49) Acenaphthylene            | 13.939 | 152  | 681794   | 39.695 | ng    | 100      |
| 50) Dimethylphthalate         | 13.686 | 163  | 644601   | 39.492 | ng    | 100      |
| 51) 2,6-Dinitrotoluene        | 13.792 | 165  | 147374   | 39.709 | ng    | 100      |
| 52) Acenaphthene              | 14.286 | 154  | 441144   | 39.942 | ng    | 100      |
| 53) 3-Nitroaniline            | 14.134 | 138  | 123336   | 40.902 | ng    | 100      |
| 54) 2,4-Dinitrophenol         | 14.375 | 184  | 96762    | 42.014 | ng    | 100      |
| 55) Dibenzofuran              | 14.622 | 168  | 755979   | 38.976 | ng    | 100      |
| 56) 4-Nitrophenol             | 14.475 | 139  | 105244   | 40.728 | ng    | 100      |
| 57) 2,4-Dinitrotoluene        | 14.616 | 165  | 216964   | 40.438 | ng    | 100      |
| 58) Fluorene                  | 15.275 | 166  | 636323   | 39.488 | ng    | 100      |
| 59) 2,3,4,6-Tetrachlorophenol | 14.857 | 232  | 211956   | 40.219 | ng    | 100      |
| 60) Diethylphthalate          | 15.069 | 149  | 639683   | 39.404 | ng    | 100      |
| 61) 4-Chlorophenyl-phenyle... | 15.281 | 204  | 369985   | 38.956 | ng    | 100      |
| 62) 4-Nitroaniline            | 15.328 | 138  | 130136   | 42.287 | ng    | 100      |
| 63) Azobenzene                | 15.575 | 77   | 562397   | 40.288 | ng    | 100      |
| 65) 4,6-Dinitro-2-methylph... | 15.398 | 198  | 142572   | 41.191 | ng    | 100      |
| 66) n-Nitrosodiphenylamine    | 15.504 | 169  | 548536   | 40.272 | ng    | 100      |
| 67) 4-Bromophenyl-phenylether | 16.204 | 248  | 264341   | 40.003 | ng    | 100      |
| 68) Hexachlorobenzene         | 16.316 | 284  | 333154   | 39.361 | ng    | 100      |
| 69) Atrazine                  | 16.486 | 200  | 202845   | 44.183 | ng    | 100      |
| 70) Pentachlorophenol         | 16.669 | 266  | 224960   | 41.618 | ng    | 100      |
| 71) Phenanthrene              | 17.063 | 178  | 1064139  | 39.911 | ng    | 100      |
| 72) Anthracene                | 17.163 | 178  | 1031342  | 40.305 | ng    | 100      |
| 73) Carbazole                 | 17.439 | 167  | 981445   | 40.159 | ng    | 100      |
| 74) Di-n-butylphthalate       | 18.027 | 149  | 1104057  | 40.536 | ng    | 100      |
| 75) Fluoranthene              | 19.169 | 202  | 1504601  | 40.679 | ng    | 100      |
| 77) Benzidine                 | 19.369 | 184  | 622338   | 47.949 | ng    | 100      |
| 78) Pyrene                    | 19.539 | 202  | 1530457  | 39.360 | ng    | 100      |
| 80) Butylbenzylphthalate      | 20.486 | 149  | 541540   | 40.847 | ng    | 100      |
| 81) Benzo(a)anthracene        | 21.445 | 228  | 1589858  | 38.894 | ng    | 100      |
| 82) 3,3'-Dichlorobenzidine    | 21.363 | 252  | 651185   | 39.352 | ng    | 100      |
| 83) Chrysene                  | 21.504 | 228  | 1511154  | 39.467 | ng    | 100      |
| 84) Bis(2-ethylhexyl)phtha... | 21.374 | 149  | 747652   | 39.296 | ng    | 100      |
| 85) Di-n-octyl phthalate      | 22.598 | 149  | 1276964  | 40.162 | ng    | 100      |
| 87) Indeno(1,2,3-cd)pyrene    | 28.350 | 276  | 2065299  | 39.290 | ng    | 100      |
| 88) Benzo(b)fluoranthene      | 23.674 | 252  | 1866122  | 40.751 | ng    | 100      |
| 89) Benzo(k)fluoranthene      | 23.745 | 252  | 1678612  | 38.821 | ng    | 100      |
| 90) Benzo(a)pyrene            | 24.533 | 252  | 1564093  | 39.918 | ng    | 100      |
| 91) Dibenzo(a,h)anthracene    | 28.403 | 278  | 1674163  | 38.847 | ng    | 100      |
| 92) Benzo(g,h,i)perylene      | 29.492 | 276  | 1752398  | 39.670 | ng    | 100      |

(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1  
 1,4-Dichlorobenzene-d4  
 Concen: 20.000 ng  
 RT: 7.593 min Scan# 783  
 Delta R.T. -0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31

Instrument :

BNA\_P

ClientSampleId :





SSTDICCC040

Tgt Ion:152 Resp: 72490  
 Ion Ratio Lower Upper  
 152 100  
 150 158.7 127.0 190.4  
 115 63.1 50.5 75.7

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
 Supervised By :mohammad ahmed 11/08/2024



#2  
 1,4-Dioxane  
 Concen: 40.154 ng  
 RT: 3.175 min Scan# 32  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31


Tgt Ion: 88 Resp: 61241  
 Ion Ratio Lower Upper  
 88 100  
 58 66.0 52.8 79.2  
 43 29.2 23.4 35.0





#3  
Pyridine  
Concen: 40.455 ng m  
RT: 3.575 min Scan# 100  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion: 79 Resp: 168818  
Ion Ratio Lower Upper  
79 100  
52 63.0 50.4 75.6  
51 35.8 28.6 43.0




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#4  
n-Nitrosodimethylamine  
Concen: 40.069 ng  
RT: 3.487 min Scan# 85  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion: 42 Resp: 80888  
Ion Ratio Lower Upper  
42 100  
74 122.1 97.7 146.5  
44 5.6 4.5 6.7





#5  
2-Fluorophenol  
Concen: 80.977 ng  
RT: 5.234 min Scan# 382  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 112 Resp: 309338  
Ion Ratio Lower Upper  
112 100  
64 60.9 48.7 73.1  
63 36.2 29.0 43.4

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#6  
Aniline  
Concen: 40.464 ng  
RT: 6.940 min Scan# 672  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Tgt Ion: 93 Resp: 179698  
Ion Ratio Lower Upper  
93 100  
66 41.1 32.9 49.3  
65 20.0 16.0 24.0





#7  
Phenol-d6  
Concen: 80.198 ng  
RT: 6.793 min Scan# 647  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion: 99 Resp: 427244  
Ion Ratio Lower Upper  
99 100  
42 19.7 15.8 23.6  
71 45.8 36.6 55.0




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#8  
2-Chlorophenol  
Concen: 40.380 ng  
RT: 7.169 min Scan# 711  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion:128 Resp: 161690  
Ion Ratio Lower Upper  
128 100  
130 32.6 12.6 52.6  
64 52.7 32.7 72.7





#9  
Benzaldehyde  
Concen: 39.794 ng  
RT: 6.757 min Scan# 641  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 77 Resp: 12249  
Ion Ratio Lower Upper  
77 100  
105 91.4 71.4 111.4  
106 89.9 69.9 109.9

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#10  
Phenol  
Concen: 40.537 ng  
RT: 6.816 min Scan# 651  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31



Tgt Ion: 94 Resp: 213872  
Ion Ratio Lower Upper  
94 100  
65 32.9 12.9 52.9  
66 50.0 30.0 70.0








#11  
bis(2-Chloroethyl)ether  
Concen: 39.887 ng  
RT: 7.034 min Scan# 688  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Instrument :

BNA\_P

ClientSampleId :

SSTDICCC040






Tgt Ion: 93 Resp: 15899  
Ion Ratio Lower Upper  
93 100  
63 73.6 53.6 93.6  
95 32.1 12.1 52.1

Manual Integrations

APPROVED




Reviewed By :Yogesh Patel 11/08/2024


Supervised By :mohammad ahmed 11/08/2024



#12  
1,3-Dichlorobenzene  
Concen: 39.615 ng  
RT: 7.481 min Scan# 764  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Tgt Ion:146 Resp: 202460  
Ion Ratio Lower Upper  
146 100  
148 64.2 51.4 77.0  
75 36.6 29.3 43.9









#13  
1,4-Dichlorobenzene  
Concen: 39.157 ng  
RT: 7.634 min Scan# 790  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

Tgt Ion:146 Resp: 204335  
Ion Ratio Lower Upper  
146 100  
148 66.5 53.2 79.8  
111 43.7 35.0 52.4




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#14  
1,2-Dichlorobenzene  
Concen: 39.455 ng  
RT: 7.940 min Scan# 842  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion:146 Resp: 199057  
Ion Ratio Lower Upper  
146 100  
148 65.5 52.4 78.6  
111 42.8 34.2 51.4





#15  
Benzyl Alcohol  
Concen: 41.050 ng  
RT: 7.840 min Scan# 825  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 79 Resp: 160610  
Ion Ratio Lower Upper  
79 100  
108 64.0 51.2 76.8  
77 64.1 51.3 76.9

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#16  
2,2'-oxybis(1-Chloropropane)  
Concen: 39.316 ng  
RT: 8.110 min Scan# 871  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31



Tgt Ion: 45 Resp: 170032  
Ion Ratio Lower Upper  
45 100  
77 19.5 0.0 39.5  
79 14.9 0.0 34.9






#17  
2-Methylphenol  
Concen: 40.872 ng  
RT: 8.046 min Scan# 860  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Instrument :

BNA\_P

ClientSampleId :

SSTDICCC040






Tgt Ion:107 Resp: 14637  
Ion Ratio Lower Upper  
107 100  
108 104.6 83.7 125.5  
77 56.2 45.0 67.4  
79 56.8 45.4 68.2

Manual Integrations

APPROVED




Reviewed By :Yogesh Patel 11/08/2024


Supervised By :mohammad ahmed 11/08/2024



#18  
Hexachloroethane  
Concen: 39.357 ng  
RT: 8.646 min Scan# 962  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion:117 Resp: 75976  
Ion Ratio Lower Upper  
117 100  
119 95.9 76.7 115.1  
201 137.5 110.0 165.0





#19  
n-Nitroso-di-n-propylamine  
Concen: 40.167 ng  
RT: 8.393 min Scan# 919  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 70 Resp: 14646  
Ion Ratio Lower Upper  
70 100  
42 51.3 41.0 61.6  
101 9.7 7.8 11.6  
130 18.4 14.7 22.1

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#20  
3+4-Methylphenols  
Concen: 40.326 ng  
RT: 8.369 min Scan# 915  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Tgt Ion:107 Resp: 201846  
Ion Ratio Lower Upper  
107 100  
108 87.6 67.6 107.6  
77 42.6 22.6 62.6  
79 36.7 16.7 56.7





#21  
Naphthalene-d8  
Concen: 20.000 ng  
RT: 10.345 min Scan# 1251  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:136 Resp: 27064  
Ion Ratio Lower Upper  
136 100  
137 11.7 9.4 14.0  
54 8.0 6.4 9.6  
68 4.9 3.9 5.9




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#22  
Acetophenone  
Concen: 39.499 ng  
RT: 8.410 min Scan# 922  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion:105 Resp: 284548  
Ion Ratio Lower Upper  
105 100  
71 0.9 0.7 1.1  
51 27.1 21.7 32.5  
120 20.5 16.4 24.6





#23  
Nitrobenzene-d5  
Concen: 80.727 ng  
RT: 8.740 min Scan# 91  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion: 82 Resp: 46245  
Ion Ratio Lower Upper  
82 100  
128 37.3 29.8 44.8  
54 48.7 39.0 58.4




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#24  
Nitrobenzene  
Concen: 40.433 ng  
RT: 8.781 min Scan# 985  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion: 77 Resp: 235280  
Ion Ratio Lower Upper  
77 100  
123 40.3 32.2 48.4  
65 14.6 11.7 17.5





#25  
Isophorone  
Concen: 40.017 ng  
RT: 9.293 min Scan# 1072  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion: 82 Resp: 38799  
Ion Ratio Lower Upper  
82 100  
95 9.3 7.4 11.2  
138 17.1 13.7 20.5




Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#26  
2-Nitrophenol  
Concen: 40.475 ng  
RT: 9.481 min Scan# 1104  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Tgt Ion:139 Resp: 96718  
Ion Ratio Lower Upper  
139 100  
109 47.3 37.8 56.8  
65 60.8 48.6 73.0








#27  
2,4-Dimethylphenol  
Concen: 39.621 ng  
RT: 9.540 min Scan# 1114  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:122 Resp: 108108  
Ion Ratio Lower Upper  
122 100  
107 137.2 109.8 164.6  
121 63.1 50.5 75.7




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#28  
bis(2-Chloroethoxy)methane  
Concen: 39.891 ng  
RT: 9.775 min Scan# 1154  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion: 93 Resp: 219691  
Ion Ratio Lower Upper  
93 100  
95 31.8 25.4 38.2  
123 10.8 8.6 13.0





#29  
2,4-Dichlorophenol  
Concen: 40.972 ng  
RT: 10.016 min Scan# 1195  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:162 Resp: 19351  
Ion Ratio Lower Upper  
162 100  
164 66.0 46.0 86.0  
98 36.7 16.7 56.7

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#30  
1,2,4-Trichlorobenzene  
Concen: 39.388 ng  
RT: 10.216 min Scan# 1229  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31



Tgt Ion:180 Resp: 237828  
Ion Ratio Lower Upper  
180 100  
182 99.3 79.4 119.2  
145 27.0 21.6 32.4









#31  
 Naphthalene  
 Concen: 39.934 ng  
 RT: 10.398 min Scan# 1260  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31

Instrument : BNA\_P  
 ClientSampleId : SSTDICCC040




|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 128   | Resp: | 541529 |
| Ion Ratio | Lower | Upper |        |
| 128       | 100   |       |        |
| 129       | 10.7  | 8.6   | 12.8   |
| 127       | 13.3  | 10.6  | 16.0   |


Manual Integrations  
**APPROVED**  
 Reviewed By :Yogesh Patel 11/08/2024  
 Supervised By :mohammad ahmed 11/08/2024



#32  
 Benzoic acid  
 Concen: 41.927 ng  
 RT: 9.728 min Scan# 1146  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 122   | Resp: | 137314 |
| Ion Ratio | Lower | Upper |        |
| 122       | 100   |       |        |
| 105       | 139.9 | 119.9 | 159.9  |
| 77        | 112.2 | 92.2  | 132.2  |





#33  
4-Chloroaniline  
Concen: 40.887 ng  
RT: 10.522 min Scan# 1281  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 127     | 100   |       |       |
| 129     | 31.5  | 25.2  | 37.8  |
| 65      | 37.1  | 29.7  | 44.5  |
| 92      | 24.5  | 19.6  | 29.4  |

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#34  
Hexachlorobutadiene  
Concen: 40.087 ng  
RT: 10.669 min Scan# 1306  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 225     | 100   |       |       |
| 223     | 62.2  | 49.8  | 74.6  |
| 227     | 63.1  | 50.5  | 75.7  |





#35  
Caprolactam  
Concen: 40.556 ng  
RT: 11.310 min Scan# 1415  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:113 Resp: 5734  
Ion Ratio Lower Upper  
113 100  
55 147.4 127.4 167.4  
56 123.5 103.5 143.5

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#36  
4-Chloro-3-methylphenol  
Concen: 40.980 ng  
RT: 11.651 min Scan# 1473  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31



Tgt Ion:107 Resp: 214810  
Ion Ratio Lower Upper  
107 100  
144 24.2 19.4 29.0  
142 74.4 59.5 89.3









#37  
2-Methylnaphthalene  
Concen: 39.790 ng  
RT: 12.004 min Scan# 1533  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

Tgt Ion:142 Resp: 404119  
Ion Ratio Lower Upper  
142 100  
141 90.3 72.2 108.4  
115 44.9 35.9 53.9




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#38  
1-Methylnaphthalene  
Concen: 40.293 ng  
RT: 12.234 min Scan# 1572  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Tgt Ion:142 Resp: 406166  
Ion Ratio Lower Upper  
142 100  
141 94.8 75.8 113.8  
116 4.7 3.8 5.6





#39  
 Acenaphthene-d10  
 Concen: 20.000 ng  
 RT: 14.216 min Scan# 1909  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31





Instrument : BNA\_P  
 Client Sample Id : SSTDICCC040

Tgt Ion: 164 Resp: 222935

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 164 | 100   |       |       |
| 162 | 102.3 | 81.8  | 122.8 |
| 160 | 44.6  | 35.7  | 53.5  |

Manual Integrations  
 APPROVED




Reviewed By : Yogesh Patel 11/08/2024  
 Supervised By : mohammad ahmed 11/08/2024




#40  
 1,2,4,5-Tetrachlorobenzene  
 Concen: 39.397 ng  
 RT: 12.381 min Scan# 1597  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31

Tgt Ion: 216 Resp: 297343


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 216 | 100   |       |       |
| 214 | 78.1  | 62.5  | 93.7  |
| 179 | 20.6  | 16.5  | 24.7  |
| 108 | 15.1  | 12.1  | 18.1  |





#41  
Hexachlorocyclopentadiene  
Concen: 42.449 ng  
RT: 12.351 min Scan# 1592  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion: 237 Resp: 8997  
Ion Ratio Lower Upper  
237 100  
235 61.2 41.2 81.2  
272 13.7 0.0 33.7




Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#42  
2,4,6-Tribromophenol  
Concen: 78.974 ng  
RT: 15.745 min Scan# 2169  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Tgt Ion: 330 Resp: 341261  
Ion Ratio Lower Upper  
330 100  
332 96.3 77.0 115.6  
141 27.0 21.6 32.4








#43  
2,4,6-Trichlorophenol  
Concen: 40.638 ng  
RT: 12.639 min Scan# 1641  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:196 Resp: 195060  
Ion Ratio Lower Upper  
196 100  
198 95.6 76.5 114.7  
200 30.7 24.6 36.8

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#44  
2,4,5-Trichlorophenol  
Concen: 40.479 ng  
RT: 12.722 min Scan# 1655  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Tgt Ion:196 Resp: 221187  
Ion Ratio Lower Upper  
196 100  
198 97.3 77.8 116.8  
97 44.0 35.2 52.8  
132 19.9 15.9 23.9





#45  
2-Fluorobiphenyl  
Concen: 78.085 ng  
RT: 12.816 min Scan# 1671  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:172 Resp: 121830  
Ion Ratio Lower Upper  
172 100  
171 35.6 28.5 42.7  
170 23.2 18.6 27.8

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#46  
1,1'-Biphenyl  
Concen: 39.904 ng  
RT: 13.034 min Scan# 1708  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Tgt Ion:154 Resp: 591369  
Ion Ratio Lower Upper  
154 100  
153 40.6 20.6 60.6  
76 10.5 0.0 30.5





#47  
2-Chloronaphthalene  
Concen: 39.340 ng  
RT: 13.075 min Scan# 1715  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 162 Resp: 48383  
Ion Ratio Lower Upper  
162 100  
127 35.8 28.6 43.0  
164 33.0 26.4 39.6

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#48  
2-Nitroaniline  
Concen: 41.193 ng  
RT: 13.298 min Scan# 1753  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Tgt Ion: 65 Resp: 145626  
Ion Ratio Lower Upper  
65 100  
92 68.0 54.4 81.6  
138 95.6 76.5 114.7





#49  
Acenaphthylene  
Concen: 39.695 ng  
RT: 13.939 min Scan# 1862  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:152 Resp: 681794  
Ion Ratio Lower Upper  
152 100  
151 21.0 16.8 25.2  
153 13.5 10.8 16.2




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#50  
Dimethylphthalate  
Concen: 39.492 ng  
RT: 13.686 min Scan# 1819  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion:163 Resp: 644601  
Ion Ratio Lower Upper  
163 100  
194 4.4 3.5 5.3  
164 10.0 8.0 12.0





#51  
2,6-Dinitrotoluene  
Concen: 39.709 ng  
RT: 13.792 min Scan# 1837  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:165 Resp: 147374  
Ion Ratio Lower Upper  
165 100  
63 56.4 45.1 67.7  
89 58.6 46.9 70.3

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#52  
Acenaphthene  
Concen: 39.942 ng  
RT: 14.286 min Scan# 1921  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Tgt Ion:154 Resp: 441144  
Ion Ratio Lower Upper  
154 100  
153 112.4 89.9 134.9  
152 52.6 42.1 63.1





#53  
3-Nitroaniline  
Concen: 40.902 ng  
RT: 14.134 min Scan# 1895  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:138 Resp: 123330  
Ion Ratio Lower Upper  
138 100  
108 12.7 10.2 15.2  
92 130.2 104.2 156.2




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#54  
2,4-Dinitrophenol  
Concen: 42.014 ng  
RT: 14.375 min Scan# 1936  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion:184 Resp: 96762  
Ion Ratio Lower Upper  
184 100  
63 57.9 46.3 69.5  
154 70.8 56.6 85.0





#55  
Dibenzenofuran  
Concen: 38.976 ng  
RT: 14.622 min Scan# 1978  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:168 Resp: 755979  
Ion Ratio Lower Upper  
168 100  
139 38.9 31.1 46.7  
169 13.8 11.0 16.6




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#56  
4-Nitrophenol  
Concen: 40.728 ng  
RT: 14.475 min Scan# 1953  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion:139 Resp: 105244  
Ion Ratio Lower Upper  
139 100  
109 123.1 103.1 143.1  
65 116.9 96.9 136.9





#57  
2,4-Dinitrotoluene  
Concen: 40.438 ng  
RT: 14.616 min Scan# 1977  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:165 Resp: 216964  
Ion Ratio Lower Upper  
165 100  
63 58.1 46.5 69.7  
89 80.0 64.0 96.0  
182 2.8 2.2 3.4




Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#58  
Fluorene  
Concen: 39.488 ng  
RT: 15.275 min Scan# 2089  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Tgt Ion:166 Resp: 636323  
Ion Ratio Lower Upper  
166 100  
165 100.1 80.1 120.1  
167 13.9 11.1 16.7








#59  
2,3,4,6-Tetrachlorophenol  
Concen: 40.219 ng  
RT: 14.857 min Scan# 2018  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:232 Resp: 211950  
Ion Ratio Lower Upper  
232 100  
131 35.2 28.2 42.2  
130 2.1 1.7 2.5  
166 26.8 21.4 32.2




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#60  
Diethylphthalate  
Concen: 39.404 ng  
RT: 15.069 min Scan# 2054  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion:149 Resp: 639683  
Ion Ratio Lower Upper  
149 100  
177 23.6 18.9 28.3  
150 12.9 10.3 15.5





#61  
4-Chlorophenyl-phenylether  
Concen: 38.956 ng  
RT: 15.281 min Scan# 2090  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:204 Resp: 36998  
Ion Ratio Lower Upper  
204 100  
206 33.5 26.8 40.2  
141 57.8 46.2 69.4

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#62  
4-Nitroaniline  
Concen: 42.287 ng  
RT: 15.328 min Scan# 2098  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Tgt Ion:138 Resp: 130136  
Ion Ratio Lower Upper  
138 100  
92 61.9 41.9 81.9  
108 129.3 109.3 149.3





#63  
Azobenzene  
Concen: 40.288 ng  
RT: 15.575 min Scan# 2140  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 77 Resp: 56239  
Ion Ratio Lower Upper  
77 100  
182 29.9 9.9 49.9  
105 19.9 0.0 39.9  
51 30.5 10.5 50.5

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#64  
Phenanthrene-d10  
Concen: 20.000 ng  
RT: 17.022 min Scan# 2386  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31



Tgt Ion:188 Resp: 550938  
Ion Ratio Lower Upper  
188 100  
94 6.0 4.8 7.2  
80 7.4 5.9 8.9









#65  
4,6-Dinitro-2-methylphenol  
Concen: 41.191 ng  
RT: 15.398 min Scan# 2110  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

Tgt Ion:198 Resp: 142572  
Ion Ratio Lower Upper  
198 100  
51 37.4 17.4 57.4  
105 35.4 15.4 55.4




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#66  
n-Nitrosodiphenylamine  
Concen: 40.272 ng  
RT: 15.504 min Scan# 2128  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion:169 Resp: 548536  
Ion Ratio Lower Upper  
169 100  
168 67.4 53.9 80.9  
167 36.7 29.4 44.0





#67  
4-Bromophenyl-phenylether  
Concen: 40.003 ng  
RT: 16.204 min Scan# 2247  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:248 Resp: 26434  
Ion Ratio Lower Upper  
248 100  
250 99.7 79.8 119.6  
141 56.5 45.2 67.8




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#68  
Hexachlorobenzene  
Concen: 39.361 ng  
RT: 16.316 min Scan# 2266  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Tgt Ion:284 Resp: 333154  
Ion Ratio Lower Upper  
284 100  
142 23.7 19.0 28.4  
249 28.8 23.0 34.6





#69

Atrazine

Concen: 44.183 ng

RT: 16.486 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP022918.D

Acq: 07 Nov 2024 14:31

Instrument :

BNA\_P

ClientSampleId :

SSTDICCC040





|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 200   | Resp: | 20284 |
| Ion Ratio | Lower | Upper |       |
| 200       | 100   |       |       |
| 173       | 22.7  | 2.7   | 42.7  |
| 215       | 47.1  | 27.1  | 67.1  |

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 11/08/2024

Supervised By :mohammad ahmed 11/08/2024



#70

Pentachlorophenol

Concen: 41.618 ng




RT: 16.669 min Scan# 2326


Delta R.T. 0.000 min

Lab File: BP022918.D

Acq: 07 Nov 2024 14:31


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 266   | Resp: | 224960 |
| Ion Ratio | Lower | Upper |        |
| 266       | 100   |       |        |
| 268       | 62.4  | 49.9  | 74.9   |
| 264       | 62.1  | 49.7  | 74.5   |





#71  
Phenanthrene  
Concen: 39.911 ng  
RT: 17.063 min Scan# 2393  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:178 Resp: 1064139  
Ion Ratio Lower Upper  
178 100  
176 19.5 15.6 23.4  
179 15.0 12.0 18.0




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#72  
Anthracene  
Concen: 40.305 ng  
RT: 17.163 min Scan# 2410  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31


Tgt Ion:178 Resp: 1031342  
Ion Ratio Lower Upper  
178 100  
176 18.3 14.6 22.0  
179 15.4 12.3 18.5





#73  
 Carbazole  
 Concen: 40.159 ng  
 RT: 17.439 min Scan# 2457  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDICC040




Tgt Ion:167 Resp: 981445  
 Ion Ratio Lower Upper  
 167 100  
 166 21.8 17.4 26.2  
 139 12.5 10.0 15.0

Manual Integrations  
 APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
 Supervised By :mohammad ahmed 11/08/2024




#74  
 Di-n-butylphthalate  
 Concen: 40.536 ng  
 RT: 18.027 min Scan# 2557  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31



Tgt Ion:149 Resp: 1104057  
 Ion Ratio Lower Upper  
 149 100  
 150 9.0 7.2 10.8  
 104 6.0 4.8 7.2











#75  
 Fluoranthene  
 Concen: 40.679 ng  
 RT: 19.169 min Scan# 2751  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31

Instrument : BNA\_P  
 ClientSampleId : SSTDICCC040

Tgt Ion: 202 Resp: 150460

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 202 | 100   |       |       |
| 101 | 5.5   | 0.0   | 25.5  |
| 203 | 17.2  | 0.0   | 37.2  |




Manual Integrations  
**APPROVED**  
 Reviewed By : Yogesh Patel 11/08/2024  
 Supervised By : mohammad ahmed 11/08/2024




#76  
 Chrysene-d12  
 Concen: 20.000 ng  
 RT: 21.463 min Scan# 3141  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31

Tgt Ion: 240 Resp: 656894


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 240 | 100   |       |       |
| 120 | 5.2   | 4.2   | 6.2   |
| 236 | 25.5  | 20.4  | 30.6  |





#77  
Benzidine  
Concen: 47.949 ng  
RT: 19.369 min Scan# 21  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:184 Resp: 622338  
Ion Ratio Lower Upper  
184 100  
185 14.1 11.3 16.9  
183 12.1 9.7 14.5




Manual Integrations  
APPROVED


Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024



#78  
Pyrene  
Concen: 39.360 ng  
RT: 19.539 min Scan# 2814  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Tgt Ion:202 Resp: 1530457  
Ion Ratio Lower Upper  
202 100  
200 21.7 17.4 26.0  
203 17.7 14.2 21.2






#79  
 Terphenyl-d14  
 Concen: 78.428 ng  
 RT: 19.757 min Scan# 2851  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31

Instrument :

BNP\_P

ClientSampleId :

SSTDICCC040






Tgt Ion:244 Resp: 274325  
 Ion Ratio Lower Upper  
 244 100  
 212 6.9 5.5 8.3  
 122 5.5 4.4 6.6

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 11/08/2024



Supervised By :mohammad ahmed 11/08/2024




#80  
 Butylbenzylphthalate  
 Concen: 40.847 ng  
 RT: 20.486 min Scan# 2975  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31



Tgt Ion:149 Resp: 541540  
 Ion Ratio Lower Upper  
 149 100  
 91 68.3 54.6 82.0  
 206 26.6 21.3 31.9









#81  
 Benzo(a)anthracene  
 Concen: 38.894 ng  
 RT: 21.445 min Scan# 3138  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31

Instrument : BNA\_P  
 ClientSampleId : SSTDICCC040

Tgt Ion: 228 Resp: 158985

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 228 | 100   |       |       |
| 226 | 27.2  | 21.8  | 32.6  |
| 229 | 20.0  | 16.0  | 24.0  |




Manual Integrations  
**APPROVED**  
 Reviewed By : Yogesh Patel 11/08/2024  
 Supervised By : mohammad ahmed 11/08/2024




#82  
 3,3'-Dichlorobenzidine  
 Concen: 39.352 ng  
 RT: 21.363 min Scan# 3124  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31

Tgt Ion: 252 Resp: 651185

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 252 | 100   |       |       |
| 254 | 64.6  | 51.7  | 77.5  |
| 126 | 6.7   | 5.4   | 8.0   |





#83  
 Chrysene  
 Concen: 39.467 ng  
 RT: 21.504 min Scan# 3148  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31





Instrument : BNA\_P  
 ClientSampleId : SSTDICCC040

Tgt Ion:228 Resp: 1511154

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 228 | 100   |       |       |
| 226 | 29.7  | 23.8  | 35.6  |
| 229 | 19.8  | 15.8  | 23.8  |

Manual Integrations  
 APPROVED




Reviewed By :Yogesh Patel 11/08/2024  
 Supervised By :mohammad ahmed 11/08/2024




#84  
 Bis(2-ethylhexyl)phthalate  
 Concen: 39.296 ng  
 RT: 21.374 min Scan# 3126  
 Delta R.T. 0.000 min  
 Lab File: BP022918.D  
 Acq: 07 Nov 2024 14:31

Tgt Ion:149 Resp: 747652


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 149 | 100   |       |       |
| 167 | 29.6  | 23.7  | 35.5  |
| 279 | 6.6   | 5.3   | 7.9   |





#85  
Di-n-octyl phthalate  
Concen: 40.162 ng  
RT: 22.598 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:149 Resp: 1276964  
Ion Ratio Lower Upper  
149 100  
167 1.7 1.4 2.0  
43 7.9 6.3 9.5

Manual Integrations  
APPROVED



Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#86  
Perylene-d12  
Concen: 20.000 ng  
RT: 24.698 min Scan# 3691  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31



Tgt Ion:264 Resp: 779446  
Ion Ratio Lower Upper  
264 100  
260 23.3 18.6 28.0  
265 21.8 17.4 26.2





#87

Indeno(1,2,3-cd)pyrene

Concen: 39.290 ng

RT: 28.350 min Scan# 4312

Delta R.T. 0.000 min

Lab File: BP022918.D

Acq: 07 Nov 2024 14:31

Instrument :

BNA\_P

ClientSampleId :

SSTDICCC040

Tgt Ion:276 Resp: 2065299

Ion Ratio Lower Upper

276 100

138 10.9 8.7 13.1





277 25.4 20.3 30.5

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 11/08/2024

Supervised By :mohammad ahmed 11/08/2024



#88

Benzo(b)fluoranthene

Concen: 40.751 ng

RT: 23.674 min Scan# 3517

Delta R.T. 0.000 min

Lab File: BP022918.D

Acq: 07 Nov 2024 14:31




Tgt Ion:252 Resp: 1866122


Ion Ratio Lower Upper

252 100

253 21.0 16.8 25.2


125 4.0 3.2 4.8





#89  
Benzo(k)fluoranthene  
Concen: 38.821 ng  
RT: 23.745 min Scan# 3529  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:252 Resp: 167861  
Ion Ratio Lower Upper  
252 100  
253 21.3 17.0 25.6  
125 4.3 3.4 5.2




Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 11/08/2024  
Supervised By :mohammad ahmed 11/08/2024




#90  
Benzo(a)pyrene  
Concen: 39.918 ng  
RT: 24.533 min Scan# 3663  
Delta R.T. 0.000 min  
Lab File: BP022918.D  
Acq: 07 Nov 2024 14:31

Tgt Ion:252 Resp: 1564093  
Ion Ratio Lower Upper  
252 100  
253 22.1 17.7 26.5  
125 5.0 4.0 6.0







#91

Dibenzo(a,h)anthracene

Concen: 38.847 ng

RT: 28.403 min Scan# 4321

Delta R.T. 0.000 min

Lab File: BP022918.D

Acq: 07 Nov 2024 14:31

Instrument :

BNA\_P

ClientSampleId :

SSTDICCC040

Tgt Ion:278 Resp: 167416

Ion Ratio Lower Upper

278 100

139 6.8 5.4 8.2





279 24.1 19.3 28.9

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 11/08/2024

Supervised By :mohammad ahmed 11/08/2024



#92

Benzo(g,h,i)perylene

Concen: 39.670 ng

RT: 29.492 min Scan# 4506

Delta R.T. 0.000 min

Lab File: BP022918.D

Acq: 07 Nov 2024 14:31

Tgt Ion:276 Resp: 1752398

Ion Ratio Lower Upper

276 100

277 23.5 18.8 28.2

138 9.5 7.6 11.4

