

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110822\
 Data File : BP012544.D
 Acq On : 08 Nov 2022 10:04
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020666

Quant Time: Nov 08 21:39:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 04 02:53:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	307117	20.000	ng/ul	0.00
20) Naphthalene-d8	10.569	136	1479696	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.428	164	986304	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.192	188	2168414	20.000	ng/ul	0.00
79) Chrysene-d12	21.286	240	2029763	20.000	ng/ul	0.00
88) Perylene-d12	23.668	264	1900538	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	58104	7.618	ng/uL	0.00
4) Pyridine-d5	3.652	84	427142	19.495	ng/ul	0.00
7) Phenol-d5	6.952	99	572860	20.796	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.110	67	331103	20.135	ng/ul	0.00
11) 2-Chlorophenol-d4	7.305	132	432051	21.478	ng/ul	0.00
15) 4-Methylphenol-d8	8.493	113	472947	21.708	ng/ul	0.00
21) Nitrobenzene-d5	8.946	128	227844	23.745	ng/ul	0.00
24) 2-Nitrophenol-d4	9.663	143	247201	24.821	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.198	165	479225	22.203	ng/ul	0.00
31) 4-Chloroaniline-d4	10.722	131	686621	21.365	ng/ul	0.00
46) Dimethylphthalate-d6	13.845	166	1390085	20.588	ng/ul	0.00
49) Acenaphthylene-d8	14.122	160	1613088	20.906	ng/ul	0.00
54) 4-Nitrophenol-d4	14.657	143	226471	17.723	ng/ul	0.00
60) Fluorene-d10	15.428	176	1221358	21.128	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.563	200	208542	18.472	ng/ul	0.00
73) Anthracene-d10	17.286	188	1961821	20.997	ng/ul	0.00
81) Pyrene-d10	19.533	212	2295922	22.529	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.510	264	1910813	20.577	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.264	88	61210	7.549	ng/uL	97
5) Pyridine	3.669	79	418056	19.415	ng/ul	97
6) Benzaldehyde	6.922	77	311375	23.968	ng/ul	99
8) Phenol	6.975	94	606839	21.187	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.205	93	464727	20.217	ng/ul	100
12) 2-Chlorophenol	7.334	128	469678	21.448	ng/ul	98
13) 2-Methylphenol	8.228	108	464757	21.101	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.310	45	619518	20.096	ng/ul	100
16) Acetophenone	8.604	105	745386	22.028	ng/ul	100
17) N-Nitroso-di-n-propyla...	8.587	70	359696	22.081	ng/ul	98
18) 4-Methylphenol	8.557	108	524138	21.772	ng/ul	99
19) Hexachloroethane	8.840	117	173309	20.578	ng/ul	98
22) Nitrobenzene	8.987	77	561436	22.426	ng/ul	99
23) Isophorone	9.510	82	1043386	19.916	ng/ul	99
25) 2-Nitrophenol	9.693	139	283272	23.721	ng/ul	97
26) 2,4-Dimethylphenol	9.757	107	562599	21.455	ng/ul	99
27) Bis(2-Chloroethoxy)met...	9.993	93	673241	20.082	ng/ul	98
29) 2,4-Dichlorophenol	10.228	162	496543	22.161	ng/ul	99
30) Naphthalene	10.622	128	1583121	20.187	ng/ul	100
32) 4-Chloroaniline	10.751	127	712465	21.545	ng/ul	99
33) Hexachlorobutadiene	10.898	225	280431	20.117	ng/ul	98
34) Caprolactam	11.551	113	148604	19.491	ng/ul	98
35) 4-Chloro-3-methylphenol	11.881	107	552542	22.443	ng/ul	100
36) 2-Methylnaphthalene	12.240	142	1111143	20.667	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110822\
 Data File : BP012544.D
 Acq On : 08 Nov 2022 10:04
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020666

Quant Time: Nov 08 21:39:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 04 02:53:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.457	142	1110705	20.686	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.604	216	588959	20.330	ng/ul	99
40) Hexachlorocyclopentadiene	12.575	237	240541	17.946	ng/ul	99
41) 2,4,6-Trichlorophenol	12.857	196	401778	21.696	ng/ul	99
42) 2,4,5-Trichlorophenol	12.934	196	434485	21.512	ng/ul	99
43) 1,1'-Biphenyl	13.257	154	1465651	20.209	ng/ul	100
44) 2-Chloronaphthalene	13.298	162	1167155	20.480	ng/ul	99
45) 2-Nitroaniline	13.522	65	340553	25.782	ng/ul	99
47) Dimethylphthalate	13.892	163	1463144	20.568	ng/ul	100
48) 2,6-Dinitrotoluene	14.022	165	295496	25.017	ng/ul	99
50) Acenaphthylene	14.145	152	1796644	20.763	ng/ul	99
51) 3-Nitroaniline	14.351	138	313924	21.706	ng/ul	97
52) Acenaphthene	14.492	153	1250013	20.556	ng/ul	98
53) 2,4-Dinitrophenol	14.569	184	122322	15.335	ng/ul	93
55) 4-Nitrophenol	14.669	109	177952	18.405	ng/ul	95
56) Dibenzofuran	14.828	168	1736592	20.918	ng/ul	98
57) 2,4-Dinitrotoluene	14.810	165	433804	24.384	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.057	232	381639	22.262	ng/ul	94
59) Diethylphthalate	15.257	149	1464529	21.158	ng/ul	98
61) Fluorene	15.481	166	1394154	21.135	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.475	204	688636	21.284	ng/ul	98
63) 4-Nitroaniline	15.522	138	284325	21.760	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.575	198	221193	18.389	ng/ul	97
67) N-Nitrosodiphenylamine	15.698	169	1242363	20.342	ng/ul	99
68) 4-Bromophenyl-phenylether	16.375	248	454808	20.881	ng/ul	97
69) Hexachlorobenzene	16.486	284	533446	20.421	ng/ul	99
70) Atrazine	16.657	200	448799	21.213	ng/ul	99
71) Pentachlorophenol	16.839	266	304303	18.968	ng/ul	100
72) Phenanthrene	17.233	178	2366859	20.661	ng/ul	99
74) Anthracene	17.322	178	2377844	20.894	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.216	216	597191	19.252	ng/ul	99
76) Pentachlorobenzene	14.745	250	650808	19.869	ng/ul	100
77) Carbazole	17.604	167	2194747	21.653	ng/ul	100
78) Di-n-butylphthalate	18.151	149	2623071	22.276	ng/ul	99
80) Fluoranthene	19.204	202	2814936	22.372	ng/ul	99
82) Pyrene	19.557	202	2902583	22.272	ng/ul	99
83) Butylbenzylphthalate	20.433	149	1223077	25.173	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.210	252	813710	18.455	ng/ul	99
85) Benzo(a)anthracene	21.269	228	2659644	20.539	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.192	149	1778098	23.933	ng/ul	100
87) Chrysene	21.321	228	2520891	20.824	ng/ul	99
89) Di-n-octyl phthalate	22.110	149	2960322	26.114	ng/ul	100
90) Benzo(b)fluoranthene	22.939	252	2550169	21.190	ng/ul	99
91) Benzo(k)fluoranthene	22.986	252	2457591	21.427	ng/ul	99
93) Benzo(a)pyrene	23.562	252	2164694	20.597	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.133	276	2393601	18.278	ng/ul	98
95) Dibenzo(a,h)anthracene	26.151	278	2131954	18.181	ng/ul	100
96) Benzo(g,h,i)perylene	26.892	276	2127909	18.381	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 7.549 ng/uL
 RT: 3.264 min Scan# 47
 Delta R.T. -0.000 min
 Lab File: BP012544.D
 Acq: 08 Nov 2022 10:04




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020666


Tgt Ion: 88 Resp: 61210
 Ion Ratio Lower Upper
 88 100
 43 23.9 20.6 31.0
 58 70.0 58.0 87.0



#5
 Pyridine
 Concen: 19.415 ng/uL
 RT: 3.669 min Scan# 116
 Delta R.T. -0.000 min
 Lab File: BP012544.D
 Acq: 08 Nov 2022 10:04

Tgt Ion: 79 Resp: 418056
 Ion Ratio Lower Upper
 79 100
 52 74.0 57.4 86.0
 51 33.2 25.7 38.5









#6
Benzaldehyde
Concen: 23.968 ng/ul
RT: 6.922 min Scan# 669
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion: 77 Resp: 311375
Ion Ratio Lower Upper
77 100
105 102.7 81.0 121.6
106 99.2 79.2 118.8



#8
Phenol
Concen: 21.187 ng/ul
RT: 6.975 min Scan# 678
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion: 94 Resp: 606839
Ion Ratio Lower Upper
94 100
65 26.7 20.8 31.2
66 35.3 26.2 39.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.217 ng/ul
 RT: 7.205 min Scan# 717
 Delta R.T. -0.006 min
 Lab File: BP012544.D
 Acq: 08 Nov 2022 10:04




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020666


Tgt Ion: 93 Resp: 464727
 Ion Ratio Lower Upper
 93 100
 63 74.9 59.9 89.9
 95 32.9 25.7 38.5



#12
 2-Chlorophenol
 Concen: 21.448 ng/ul
 RT: 7.334 min Scan# 739
 Delta R.T. -0.012 min
 Lab File: BP012544.D
 Acq: 08 Nov 2022 10:04

Tgt Ion:128 Resp: 469678
 Ion Ratio Lower Upper
 128 100
 64 45.5 35.1 52.7
 130 31.5 25.8 38.6





#13

2-Methylphenol

Concen: 21.101 ng/ul

RT: 8.228 min Scan# 891

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04

Instrument :

BNA_P

ClientSampleId :





SSTD020666

Tgt Ion:108 Resp: 464757

Ion Ratio Lower Upper

108 100

107 89.2 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.096 ng/ul

RT: 8.310 min Scan# 905

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04




Tgt Ion: 45 Resp: 619518


Ion Ratio Lower Upper

45 100

77 15.1 12.1 18.1

79 12.0 9.7 14.5









#16
Acetophenone
Concen: 22.028 ng/ul
RT: 8.604 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:105 Resp: 745386
Ion Ratio Lower Upper
105 100
77 75.7 60.3 90.5
51 24.9 19.9 29.9



#17
N-Nitroso-di-n-propylamine
Concen: 22.081 ng/ul
RT: 8.587 min Scan# 952
Delta R.T. -0.012 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion: 70 Resp: 359696
Ion Ratio Lower Upper
70 100
42 53.2 41.7 62.5
101 10.0 7.8 11.8
130 20.8 17.4 26.0









#18
4-Methylphenol
Concen: 21.772 ng/ul
RT: 8.557 min Scan# 947
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:108 Resp: 524138
Ion Ratio Lower Upper
108 100
107 112.0 89.1 133.7



#19
Hexachloroethane
Concen: 20.578 ng/ul
RT: 8.840 min Scan# 995
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:117 Resp: 173309
Ion Ratio Lower Upper
117 100
201 95.9 78.5 117.7
199 62.8 49.9 74.9









#22
Nitrobenzene
Concen: 22.426 ng/ul
RT: 8.987 min Scan# 1020
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion: 77 Resp: 561436
Ion Ratio Lower Upper
77 100
123 47.4 37.4 56.2
65 12.7 10.2 15.2



#23
Isophorone
Concen: 19.916 ng/ul
RT: 9.510 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion: 82 Resp: 1043386
Ion Ratio Lower Upper
82 100
95 6.7 5.2 7.8
138 17.1 14.2 21.2









#25
2-Nitrophenol
Concen: 23.721 ng/ul
RT: 9.693 min Scan# 1140
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:139 Resp: 283272
Ion Ratio Lower Upper
139 100
65 49.0 37.8 56.8
109 28.3 21.4 32.2



#26
2,4-Dimethylphenol
Concen: 21.455 ng/ul
RT: 9.757 min Scan# 1151
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:107 Resp: 562599
Ion Ratio Lower Upper
107 100
121 51.3 40.9 61.3
122 86.6 70.6 105.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.082 ng/ul
 RT: 9.993 min Scan# 1191
 Delta R.T. -0.006 min
 Lab File: BP012544.D
 Acq: 08 Nov 2022 10:04




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020666


Tgt Ion: 93 Resp: 673241
 Ion Ratio Lower Upper
 93 100
 95 31.6 26.0 39.0
 123 10.7 9.0 13.4



#29
 2,4-Dichlorophenol
 Concen: 22.161 ng/ul
 RT: 10.228 min Scan# 1231
 Delta R.T. -0.006 min
 Lab File: BP012544.D
 Acq: 08 Nov 2022 10:04

Tgt Ion:162 Resp: 496543
 Ion Ratio Lower Upper
 162 100
 164 65.8 51.9 77.9
 98 35.3 28.6 43.0









#30
Naphthalene
Concen: 20.187 ng/ul
RT: 10.622 min Scan# 1298
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:128 Resp: 1583121
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.0 10.5 15.7



#32
4-Chloroaniline
Concen: 21.545 ng/ul
RT: 10.751 min Scan# 1320
Delta R.T. -0.000 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:127 Resp: 712465
Ion Ratio Lower Upper
127 100
129 32.1 26.0 39.0





#33

Hexachlorobutadiene

Concen: 20.117 ng/ul

RT: 10.898 min Scan# 1345

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04

Instrument :

BNA_P

ClientSampleId :

SSTD020666





Tgt Ion:225 Resp: 280431

Ion Ratio Lower Upper

225 100

223 63.9 49.0 73.6

227 64.2 51.9 77.9



#34

Caprolactam

Concen: 19.491 ng/ul

RT: 11.551 min Scan# 1456

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04




Tgt Ion:113 Resp: 148604


Ion Ratio Lower Upper

113 100

55 171.8 135.6 203.4

56 126.2 98.9 148.3









#35
4-Chloro-3-methylphenol
Concen: 22.443 ng/ul
RT: 11.881 min Scan# 1512
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:107 Resp: 552542
Ion Ratio Lower Upper
107 100
144 26.4 21.0 31.4
142 81.7 65.4 98.2



#36
2-Methylnaphthalene
Concen: 20.667 ng/ul
RT: 12.240 min Scan# 1573
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:142 Resp: 1111143
Ion Ratio Lower Upper
142 100
141 87.3 70.6 106.0





#37

1-Methylnaphthalene

Concen: 20.686 ng/ul

RT: 12.457 min Scan# 1610

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04

Instrument :

BNA_P

ClientSampleId :

SSTD020666





Tgt Ion:142 Resp: 1110705

Ion Ratio Lower Upper

142 100

141 92.2 73.1 109.7

116 3.9 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.330 ng/ul

RT: 12.604 min Scan# 1635

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04

Tgt Ion:216 Resp: 588959




Ion Ratio Lower Upper


216 100

214 78.2 62.9 94.3

179 21.4 16.6 24.8

108 17.4 13.2 19.8





#40

Hexachlorocyclopentadiene

Concen: 17.946 ng/ul

RT: 12.575 min Scan# 1630

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04

Instrument :

BNA_P

ClientSampleId :

SSTD020666





Tgt Ion:237 Resp: 240541

Ion Ratio Lower Upper

237 100

235 63.5 50.1 75.1

272 13.8 10.6 15.8



#41

2,4,6-Trichlorophenol

Concen: 21.696 ng/ul

RT: 12.857 min Scan# 1678

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04




Tgt Ion:196 Resp: 401778


Ion Ratio Lower Upper

196 100

198 95.7 77.2 115.8

200 30.9 25.1 37.7









#42
2,4,5-Trichlorophenol
Concen: 21.512 ng/ul
RT: 12.934 min Scan# 1691
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:196 Resp: 434485
Ion Ratio Lower Upper
196 100
198 96.1 77.8 116.8
200 30.4 25.0 37.4



#43
1,1'-Biphenyl
Concen: 20.209 ng/ul
RT: 13.257 min Scan# 1746
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:154 Resp: 1465651
Ion Ratio Lower Upper
154 100
153 39.9 31.9 47.9
76 13.4 10.6 16.0









#44
2-Chloronaphthalene
Concen: 20.480 ng/ul
RT: 13.298 min Scan# 1753
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:162 Resp: 1167155
Ion Ratio Lower Upper
162 100
164 32.3 26.6 39.8
127 38.8 30.9 46.3



#45
2-Nitroaniline
Concen: 25.782 ng/ul
RT: 13.522 min Scan# 1791
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion: 65 Resp: 340553
Ion Ratio Lower Upper
65 100
92 70.9 58.5 87.7
138 113.7 91.1 136.7









#47
Dimethylphthalate
Concen: 20.568 ng/ul
RT: 13.892 min Scan# 1854
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:163 Resp: 1463144
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.2 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 25.017 ng/ul
RT: 14.022 min Scan# 1876
Delta R.T. -0.000 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:165 Resp: 295496
Ion Ratio Lower Upper
165 100
89 55.2 43.4 65.2
121 21.0 17.1 25.7









#50
Acenaphthylene
Concen: 20.763 ng/ul
RT: 14.145 min Scan# 1897
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:152 Resp: 1796644
Ion Ratio Lower Upper
152 100
151 19.5 16.1 24.1
153 13.1 10.8 16.2



#51
3-Nitroaniline
Concen: 21.706 ng/ul
RT: 14.351 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:138 Resp: 313924
Ion Ratio Lower Upper
138 100
108 10.4 8.3 12.5
92 117.2 90.6 135.8









#52
Acenaphthene
Concen: 20.556 ng/ul
RT: 14.492 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:153 Resp: 1250013
Ion Ratio Lower Upper
153 100
152 47.5 38.6 58.0
154 87.6 71.6 107.4



#53
2,4-Dinitrophenol
Concen: 15.335 ng/ul
RT: 14.569 min Scan# 1969
Delta R.T. -0.000 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:184 Resp: 122322
Ion Ratio Lower Upper
184 100
63 66.8 47.9 71.9
154 61.8 46.9 70.3









#55
4-Nitrophenol
Concen: 18.405 ng/ul
RT: 14.669 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:109 Resp: 177952
Ion Ratio Lower Upper
109 100
139 141.3 123.1 184.7
65 134.2 107.9 161.9



#56
Dibenzofuran
Concen: 20.918 ng/ul
RT: 14.828 min Scan# 2013
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:168 Resp: 1736592
Ion Ratio Lower Upper
168 100
139 37.8 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 24.384 ng/ul
RT: 14.810 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:165 Resp: 433804
Ion Ratio Lower Upper
165 100
63 45.3 33.1 49.7
182 2.6 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.262 ng/ul
RT: 15.057 min Scan# 2052
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:232 Resp: 381639
Ion Ratio Lower Upper
232 100
131 47.3 33.6 50.4
130 2.4 1.8 2.6
166 31.5 24.2 36.2









#59
Diethylphthalate
Concen: 21.158 ng/ul
RT: 15.257 min Scan# 2086
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:149 Resp: 1464529
Ion Ratio Lower Upper
149 100
177 21.5 18.2 27.2
150 12.2 9.8 14.6



#61
Fluorene
Concen: 21.135 ng/ul
RT: 15.481 min Scan# 2124
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:166 Resp: 1394154
Ion Ratio Lower Upper
166 100
165 98.3 77.8 116.8
167 13.6 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 21.284 ng/ul
RT: 15.475 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:204 Resp: 688636
Ion Ratio Lower Upper
204 100
206 33.6 26.5 39.7
141 61.7 47.3 70.9



#63
4-Nitroaniline
Concen: 21.760 ng/ul
RT: 15.522 min Scan# 2131
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:138 Resp: 284325
Ion Ratio Lower Upper
138 100
92 50.9 40.6 61.0
108 75.5 56.6 85.0









#66
4,6-Dinitro-2-methylphenol
Concen: 18.389 ng/ul
RT: 15.575 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:198 Resp: 221193
Ion Ratio Lower Upper
198 100
182 5.6 3.8 5.6
77 25.7 19.5 29.3



#67
N-Nitrosodiphenylamine
Concen: 20.342 ng/ul
RT: 15.698 min Scan# 2161
Delta R.T. -0.000 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:169 Resp: 1242363
Ion Ratio Lower Upper
169 100
168 67.3 52.9 79.3
167 36.9 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 20.881 ng/ul
RT: 16.375 min Scan# 2276
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:248 Resp: 454808
Ion Ratio Lower Upper
248 100
250 99.3 77.4 116.2
141 72.8 56.2 84.2



#69
Hexachlorobenzene
Concen: 20.421 ng/ul
RT: 16.486 min Scan# 2295
Delta R.T. -0.000 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:284 Resp: 533446
Ion Ratio Lower Upper
284 100
142 32.8 27.1 40.7
249 30.2 24.5 36.7









#70
Atrazine
Concen: 21.213 ng/ul
RT: 16.657 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:200 Resp: 448799
Ion Ratio Lower Upper
200 100
173 26.4 21.8 32.6
215 48.8 39.0 58.6



#71
Pentachlorophenol
Concen: 18.968 ng/ul
RT: 16.839 min Scan# 2355
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:266 Resp: 304303
Ion Ratio Lower Upper
266 100
264 62.8 50.3 75.5
268 64.1 51.4 77.0









#72
Phenanthrene
Concen: 20.661 ng/ul
RT: 17.233 min Scan# 2422
Delta R.T. -0.000 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:178 Resp: 2366859
Ion Ratio Lower Upper
178 100
179 15.3 12.6 18.8
176 19.2 15.8 23.6



#74
Anthracene
Concen: 20.894 ng/ul
RT: 17.322 min Scan# 2437
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:178 Resp: 2377844
Ion Ratio Lower Upper
178 100
179 15.6 12.5 18.7
176 18.8 15.3 22.9









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.252 ng/uL
RT: 13.216 min Scan# 1739
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:216 Resp: 597191
Ion Ratio Lower Upper
216 100
214 79.3 62.6 94.0
218 47.8 38.6 58.0



#76
Pentachlorobenzene
Concen: 19.869 ng/uL
RT: 14.745 min Scan# 1999
Delta R.T. -0.000 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:250 Resp: 650808
Ion Ratio Lower Upper
250 100
248 63.0 50.6 75.8
252 63.7 50.7 76.1









#77
 Carbazole
 Concen: 21.653 ng/ul
 RT: 17.604 min Scan# 2485
 Delta R.T. -0.006 min
 Lab File: BP012544.D
 Acq: 08 Nov 2022 10:04




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020666


Tgt Ion:167 Resp: 2194747
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.1 25.7
 139 12.9 10.4 15.6



#78
 Di-n-butylphthalate
 Concen: 22.276 ng/ul
 RT: 18.151 min Scan# 2578
 Delta R.T. -0.000 min
 Lab File: BP012544.D
 Acq: 08 Nov 2022 10:04

Tgt Ion:149 Resp: 2623071
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.6 11.4
 104 4.8 4.0 6.0





#80

Fluoranthene

Concen: 22.372 ng/ul

RT: 19.204 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04

Instrument :

BNA_P

ClientSampleId :

SSTD020666





Tgt Ion:202 Resp: 2814936

Ion Ratio Lower Upper

202 100

101 12.1 9.9 14.9

100 9.4 7.4 11.2



#82

Pyrene

Concen: 22.272 ng/ul

RT: 19.557 min Scan# 2817

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04




Tgt Ion:202 Resp: 2902583


Ion Ratio Lower Upper

202 100

101 14.1 11.8 17.6

100 11.5 9.3 13.9





#83

Butylbenzylphthalate

Concen: 25.173 ng/ul

RT: 20.433 min Scan# 2966

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04

Instrument :

BNA_P

ClientSampleId :

SSTD020666





Tgt Ion:149 Resp: 1223077

Ion Ratio Lower Upper

149 100

91 71.3 56.9 85.3

206 20.0 16.2 24.4



#84

3,3'-Dichlorobenzidine

Concen: 18.455 ng/ul

RT: 21.210 min Scan# 3098

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04




Tgt Ion:252 Resp: 813710


Ion Ratio Lower Upper

252 100

254 63.8 51.8 77.8

126 11.4 9.0 13.4





#85

Benzo(a)anthracene

Concen: 20.539 ng/ul

RT: 21.269 min Scan# 3108

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04

Instrument :

BNA_P

ClientSampleId :

SSTD020666





Tgt Ion:228 Resp: 2659644

Ion Ratio Lower Upper

228 100

229 20.1 16.2 24.2

226 27.3 22.6 33.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.933 ng/ul

RT: 21.192 min Scan# 3095

Delta R.T. -0.000 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04




Tgt Ion:149 Resp: 1778098


Ion Ratio Lower Upper

149 100

167 27.3 22.1 33.1

279 4.5 3.7 5.5









#87
Chrysene
Concen: 20.824 ng/ul
RT: 21.321 min Scan# 3117
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:228 Resp: 2520891
Ion Ratio Lower Upper
228 100
226 30.3 24.9 37.3
229 19.5 16.0 24.0



#89
Di-n-octyl phthalate
Concen: 26.114 ng/ul
RT: 22.110 min Scan# 3251
Delta R.T. -0.000 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:149 Resp: 2960322









#90
Benzo(b)fluoranthene
Concen: 21.190 ng/ul
RT: 22.939 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04




Instrument :
BNA_P
ClientSampleId :
SSTD020666


Tgt Ion:252 Resp: 2550169
Ion Ratio Lower Upper
252 100
253 22.9 17.8 26.8
125 10.9 8.7 13.1



#91
Benzo(k)fluoranthene
Concen: 21.427 ng/ul
RT: 22.986 min Scan# 3400
Delta R.T. -0.006 min
Lab File: BP012544.D
Acq: 08 Nov 2022 10:04

Tgt Ion:252 Resp: 2457591
Ion Ratio Lower Upper
252 100
253 22.1 18.0 27.0
125 10.5 8.7 13.1





#93

Benzo(a)pyrene

Concen: 20.597 ng/ul

RT: 23.562 min Scan# 3498

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04

Instrument :

BNA_P

ClientSampleId :

SSTD020666





Tgt Ion:252 Resp: 2164694

Ion Ratio Lower Upper

252 100

253 21.9 18.0 27.0

125 10.9 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.278 ng/ul

RT: 26.133 min Scan# 3935

Delta R.T. -0.006 min

Lab File: BP012544.D

Acq: 08 Nov 2022 10:04




Tgt Ion:276 Resp: 2393601


Ion Ratio Lower Upper

276 100

138 24.5 20.7 31.1

277 24.9 20.2 30.4









#95
 Dibenzo(a,h)anthracene
 Concen: 18.181 ng/ul
 RT: 26.151 min Scan# 3938
 Delta R.T. -0.006 min
 Lab File: BP012544.D
 Acq: 08 Nov 2022 10:04

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020666

Tgt Ion:278 Resp: 2131954

Ion	Ratio	Lower	Upper
278	100		
139	16.4	13.2	19.8
279	24.0	19.4	29.0



#96
 Benzo(g,h,i)perylene
 Concen: 18.381 ng/ul
 RT: 26.892 min Scan# 4064
 Delta R.T. -0.012 min
 Lab File: BP012544.D
 Acq: 08 Nov 2022 10:04

Tgt Ion:276 Resp: 2127909

Ion	Ratio	Lower	Upper
276	100		
138	22.3	17.8	26.8
277	24.0	19.5	29.3

