

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110822\
 Data File : BP012573.D
 Acq On : 09 Nov 2022 03:47
 Operator : CG/JU
 Sample : PB148832BS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS832

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/09/2022
 Supervised By :mohammad ahmed 11/14/2022

Quant Time: Nov 09 04:53:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 08 22:40:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	224547	20.000	ng/u1	0.00
20) Naphthalene-d8	10.569	136	1100661	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.422	164	750743	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	1670458	20.000	ng/u1	0.00
79) Chrysene-d12	21.286	240	1580667	20.000	ng/u1	0.00
88) Perylene-d12	23.656	264	1459328	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	35623	6.388	ng/uL	0.00
4) Pyridine-d5	3.646	84	501025	31.275	ng/u1	0.00
7) Phenol-d5	6.957	99	780002	38.729	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.110	67	428680	35.655	ng/u1	0.00
11) 2-Chlorophenol-d4	7.304	132	576801	39.218	ng/u1	0.00
15) 4-Methylphenol-d8	8.498	113	630805	39.601	ng/u1	0.00
21) Nitrobenzene-d5	8.939	128	299684	41.987	ng/u1	0.00
24) 2-Nitrophenol-d4	9.657	143	340904	46.016	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.204	165	651254	40.564	ng/u1	0.00
31) 4-Chloroaniline-d4	10.722	131	852193	35.649	ng/u1	0.00
46) Dimethylphthalate-d6	13.839	166	1921005	37.379	ng/u1	0.00
49) Acenaphthylene-d8	14.116	160	2184219	37.191	ng/u1	0.00
54) 4-Nitrophenol-d4	14.674	143	332091	34.143	ng/u1	0.00
60) Fluorene-d10	15.421	176	1648150	37.457	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	347853	39.996	ng/u1	0.00
73) Anthracene-d10	17.286	188	2636948	36.636	ng/u1	0.00
81) Pyrene-d10	19.527	212	3060823	38.569	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.509	264	2707827	37.975	ng/u1	0.00
Target Compounds	Qvalue					
2) 1,4-Dioxane	3.257	88	71623	12.081	ng/uL	97
5) Pyridine	3.663	79	512845	32.575	ng/u1	98
6) Benzaldehyde	6.916	77	440424	46.367	ng/u1	97
8) Phenol	6.987	94	772899	36.908	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.198	93	573058	34.097	ng/u1	99
12) 2-Chlorophenol	7.334	128	582231	36.365	ng/u1	97
13) 2-Methylphenol	8.228	108	591557	36.733	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.304	45	750159	33.283	ng/u1	99
16) Acetophenone	8.604	105	915789	37.016	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.586	70	450232	37.802	ng/u1	99
18) 4-Methylphenol	8.563	108	658124	37.390	ng/u1	100
19) Hexachloroethane	8.834	117	215279	34.961	ng/u1	97
22) Nitrobenzene	8.981	77	687075	36.895	ng/u1	100
23) Isophorone	9.510	82	1328067	34.080	ng/u1	99
25) 2-Nitrophenol	9.692	139	358721	40.383	ng/u1	97
26) 2,4-Dimethylphenol	9.757	107	682013	34.966	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.986	93	839072	33.647	ng/u1	99
29) 2,4-Dichlorophenol	10.228	162	629148	37.748	ng/u1	99
30) Naphthalene	10.616	128	1956307	33.537	ng/u1	100
32) 4-Chloroaniline	10.745	127	864549	35.148	ng/u1	98
33) Hexachlorobutadiene	10.886	225	356977	34.427	ng/u1	97
34) Caprolactam	11.551	113	206924m	36.487	ng/u1	
35) 4-Chloro-3-methylphenol	11.886	107	714852	39.034	ng/u1	100
36) 2-Methylnaphthalene	12.233	142	1391720	34.800	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110822\
 Data File : BP012573.D
 Acq On : 09 Nov 2022 03:47
 Operator : CG/JU
 Sample : PB148832BS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS832


Quant Time: Nov 09 04:53:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP102522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 08 22:40:23 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/09/2022
 Supervised By :mohammad ahmed 11/14/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.451	142	1388048	34.754	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.598	216	743637	33.723	ng/u1	99
40) Hexachlorocyclopentadiene	12.569	237	256355	25.128	ng/u1	99
41) 2,4,6-Trichlorophenol	12.857	196	524473	37.208	ng/u1	100
42) 2,4,5-Trichlorophenol	12.939	196	578564	37.633	ng/u1	99
43) 1,1'-Biphenyl	13.251	154	1851307	33.536	ng/u1	99
44) 2-Chloronaphthalene	13.292	162	1472377	33.942	ng/u1	99
45) 2-Nitroaniline	13.516	65	448813	44.640	ng/u1	98
47) Dimethylphthalate	13.886	163	1899064	35.073	ng/u1	100
48) 2,6-Dinitrotoluene	14.016	165	400179	44.511	ng/u1	96
50) Acenaphthylene	14.145	152	2275450	34.548	ng/u1	99
51) 3-Nitroaniline	14.351	138	448410	40.733	ng/u1	98
52) Acenaphthene	14.486	153	1579748	34.129	ng/u1	99
53) 2,4-Dinitrophenol	14.563	184	217578	35.836	ng/u1	98
55) 4-Nitrophenol	14.686	109	248979	33.830	ng/u1	94
56) Dibenzofuran	14.821	168	2182993	34.546	ng/u1	98
57) 2,4-Dinitrotoluene	14.810	165	585316	43.223	ng/u1	91
58) 2,3,4,6-Tetrachlorophenol	15.057	232	505848	38.766	ng/u1	98
59) Diethylphthalate	15.257	149	1938863	36.800	ng/u1	99
61) Fluorene	15.474	166	1751515	34.884	ng/u1	100
62) 4-Chlorophenyl-phenyle...	15.469	204	868889	35.281	ng/u1	97
63) 4-Nitroaniline	15.521	138	451873m	45.434	ng/u1	
66) 4,6-Dinitro-2-methylph...	15.574	198	345993	37.338	ng/u1#	99
67) N-Nitrosodiphenylamine	15.692	169	1574323	33.462	ng/u1	100
68) 4-Bromophenyl-phenylether	16.368	248	603891	35.991	ng/u1	97
69) Hexachlorobenzene	16.480	284	708702	35.217	ng/u1	98
70) Atrazine	16.657	200	577881	35.457	ng/u1	99
71) Pentachlorophenol	16.839	266	426608	34.518	ng/u1	100
72) Phenanthrene	17.227	178	3006342	34.067	ng/u1	100
74) Anthracene	17.321	178	2963711	33.805	ng/u1	100
75) 1,2,3,4-Tetrachloroben...	13.210	216	754519	31.575	ng/uL	99
76) Pentachlorobenzene	14.739	250	757835	30.033	ng/uL	99
77) Carbazole	17.598	167	2788171	35.708	ng/u1	100
78) Di-n-butylphthalate	18.145	149	3466009	38.208	ng/u1	100
80) Fluoranthene	19.204	202	3632817	37.076	ng/u1	97
82) Pyrene	19.557	202	3658332	36.047	ng/u1	97
83) Butylbenzylphthalate	20.433	149	1664140	43.981	ng/u1	96
84) 3,3'-Dichlorobenzidine	21.203	252	1253463	36.506	ng/u1	99
85) Benzo(a)anthracene	21.268	228	3457151	34.283	ng/u1	99
86) Bis(2-ethylhexyl)phtha...	21.186	149	2438215	42.142	ng/u1	99
87) Chrysene	21.321	228	3242845	34.398	ng/u1	99
89) Di-n-octyl phthalate	22.103	149	4126427	47.406	ng/u1	100
90) Benzo(b)fluoranthene	22.933	252	3369426	36.462	ng/u1	99
91) Benzo(k)fluoranthene	22.980	252	3260064	37.018	ng/u1	99
93) Benzo(a)pyrene	23.556	252	2852317	35.346	ng/u1	98
94) Indeno(1,2,3-cd)pyrene	26.127	276	3322098	33.037	ng/u1	97
95) Dibenzo(a,h)anthracene	26.144	278	2944101	32.698	ng/u1	97
96) Benzo(g,h,i)perylene	26.886	276	2883326	32.437	ng/u1	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.081 ng/uL
RT: 3.257 min Scan# 40
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion: 88 Resp: 71623
Ion Ratio Lower Upper
88 100
43 27.0 20.6 31.0
58 70.2 58.0 87.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#5
Pyridine
Concen: 32.575 ng/uL
RT: 3.663 min Scan# 115
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion: 79 Resp: 512845
Ion Ratio Lower Upper
79 100
52 74.2 57.4 86.0
51 32.5 25.7 38.5





#6
Benzaldehyde
Concen: 46.367 ng/ul
RT: 6.916 min Scan# 60
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion: 77 Resp: 440424
Ion Ratio Lower Upper
77 100
105 100.0 81.0 121.6
106 95.1 79.2 118.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#8
Phenol
Concen: 36.908 ng/ul
RT: 6.987 min Scan# 680
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion: 94 Resp: 772899
Ion Ratio Lower Upper
94 100
65 26.7 20.8 31.2
66 35.2 26.2 39.2





#10
Bis(2-Chloroethyl)ether
Concen: 34.097 ng/ul
RT: 7.198 min Scan# 716
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832




Tgt Ion: 93 Resp: 573058
Ion Ratio Lower Upper
93 100
63 75.9 59.9 89.9
95 31.9 25.7 38.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022




#12
2-Chlorophenol
Concen: 36.365 ng/ul
RT: 7.334 min Scan# 739
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47



Tgt Ion:128 Resp: 582231
Ion Ratio Lower Upper
128 100
64 46.7 35.1 52.7
130 32.4 25.8 38.6





#13

2-Methylphenol

Concen: 36.733 ng/ul

RT: 8.228 min Scan# 891

Delta R.T. -0.006 min

Lab File: BP012573.D

Acq: 09 Nov 2022 03:47

Instrument :

BNA_P

ClientSampleId :

SLCS832

Tgt Ion:108 Resp: 59155

Ion Ratio Lower Upper





108 100

107 89.3 70.8 106.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 33.283 ng/ul

RT: 8.304 min Scan# 904

Delta R.T. -0.000 min

Lab File: BP012573.D

Acq: 09 Nov 2022 03:47




Tgt Ion: 45 Resp: 750159


Ion Ratio Lower Upper

45 100

77 15.6 12.1 18.1


79 12.0 9.7 14.5





#16
Acetophenone
Concen: 37.016 ng/ul
RT: 8.604 min Scan# 915
Delta R.T. -0.000 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion:105 Resp: 915789
Ion Ratio Lower Upper
105 100
77 74.8 60.3 90.5
51 24.0 19.9 29.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#17
N-Nitroso-di-n-propylamine
Concen: 37.802 ng/ul
RT: 8.586 min Scan# 952
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion: 70 Resp: 450232
Ion Ratio Lower Upper
70 100
42 51.7 41.7 62.5
101 9.7 7.8 11.8
130 22.0 17.4 26.0





#18
4-Methylphenol
Concen: 37.390 ng/ul
RT: 8.563 min Scan# 94
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion:108 Resp: 658124
Ion Ratio Lower Upper
108 100
107 111.8 89.1 133.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#19
Hexachloroethane
Concen: 34.961 ng/ul
RT: 8.834 min Scan# 994
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion:117 Resp: 215279
Ion Ratio Lower Upper
117 100
201 101.7 78.5 117.7
199 63.9 49.9 74.9





#22
Nitrobenzene
Concen: 36.895 ng/ul
RT: 8.981 min Scan# 1019
Delta R.T. -0.000 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832




Tgt Ion: 77 Resp: 687075
Ion Ratio Lower Upper
77 100
123 46.6 37.4 56.2
65 13.1 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022




#23
Isophorone
Concen: 34.080 ng/ul
RT: 9.510 min Scan# 1109
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47



Tgt Ion: 82 Resp: 1328067
Ion Ratio Lower Upper
82 100
95 7.3 5.2 7.8
138 18.2 14.2 21.2









#25
2-Nitrophenol
Concen: 40.383 ng/ul
RT: 9.692 min Scan# 1140
Delta R.T. -0.000 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47

Instrument :
BNA_P
ClientSampleId :
SLCS832

Tgt Ion:139 Resp: 358721
Ion Ratio Lower Upper
139 100
65 45.7 37.8 56.8
109 28.5 21.4 32.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#26
2,4-Dimethylphenol
Concen: 34.966 ng/ul
RT: 9.757 min Scan# 1151
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47

Tgt Ion:107 Resp: 682013
Ion Ratio Lower Upper
107 100
121 49.8 40.9 61.3
122 87.1 70.6 105.8





#27

Bis(2-Chloroethoxy)methane

Concen: 33.647 ng/ul

RT: 9.986 min Scan# 1190

Delta R.T. -0.006 min

Lab File: BP012573.D

Acq: 09 Nov 2022 03:47

Instrument :

BNA_P

ClientSampleId :





SLCS832

Tgt Ion: 93	Resp: 839072
Ion Ratio	Lower Upper
93 100	
95 32.3	26.0 39.0
123 10.9	9.0 13.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#29

2,4-Dichlorophenol

Concen: 37.748 ng/ul




RT: 10.228 min Scan# 1231


Delta R.T. -0.006 min

Lab File: BP012573.D

Acq: 09 Nov 2022 03:47


Tgt Ion:162	Resp: 629148
Ion Ratio	Lower Upper
162 100	
164 65.8	51.9 77.9
98 36.2	28.6 43.0





#30
Naphthalene
Concen: 33.537 ng/ul
RT: 10.616 min Scan# 1298
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion:128 Resp: 1956307
Ion Ratio Lower Upper
128 100
129 11.3 8.9 13.3
127 13.1 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#32
4-Chloroaniline
Concen: 35.148 ng/ul
RT: 10.745 min Scan# 1319
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion:127 Resp: 864549
Ion Ratio Lower Upper
127 100
129 31.4 26.0 39.0





#33
Hexachlorobutadiene
Concen: 34.427 ng/ul
RT: 10.886 min Scan# 1344
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion:225 Resp: 356972
Ion Ratio Lower Upper
225 100
223 63.0 49.0 73.6
227 62.5 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#34
Caprolactam
Concen: 36.487 ng/ul m
RT: 11.551 min Scan# 1456
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion:113 Resp: 206924
Ion Ratio Lower Upper
113 100
55 164.3 135.6 203.4
56 123.4 98.9 148.3





#35
4-Chloro-3-methylphenol
Concen: 39.034 ng/ul
RT: 11.886 min Scan# 1513
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832




Tgt Ion:107 Resp: 714852
Ion Ratio Lower Upper
107 100
144 26.7 21.0 31.4
142 81.8 65.4 98.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022




#36
2-Methylnaphthalene
Concen: 34.800 ng/ul
RT: 12.233 min Scan# 1572
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Tgt Ion:142 Resp: 1391720
Ion Ratio Lower Upper
142 100
141 89.4 70.6 106.0





#37
1-Methylnaphthalene
Concen: 34.754 ng/ul
RT: 12.451 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion:142 Resp: 1388048
Ion Ratio Lower Upper
142 100
141 92.5 73.1 109.7
116 3.8 3.0 4.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 33.723 ng/ul
RT: 12.598 min Scan# 1634
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion:216 Resp: 743637
Ion Ratio Lower Upper
216 100
214 78.6 62.9 94.3
179 21.0 16.6 24.8
108 17.6 13.2 19.8





#40
Hexachlorocyclopentadiene
Concen: 25.128 ng/ul
RT: 12.569 min Scan# 1629
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion:237 Resp: 256355
Ion Ratio Lower Upper
237 100
235 62.0 50.1 75.1
272 13.0 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#41
2,4,6-Trichlorophenol
Concen: 37.208 ng/ul
RT: 12.857 min Scan# 1678
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion:196 Resp: 524473
Ion Ratio Lower Upper
196 100
198 96.5 77.2 115.8
200 31.6 25.1 37.7





#42
2,4,5-Trichlorophenol
Concen: 37.633 ng/ul
RT: 12.939 min Scan# 1692
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion:196 Resp: 578564
Ion Ratio Lower Upper
196 100
198 96.7 77.8 116.8
200 31.0 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#43
1,1'-Biphenyl
Concen: 33.536 ng/ul
RT: 13.251 min Scan# 1745
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion:154 Resp: 1851307
Ion Ratio Lower Upper
154 100
153 39.6 31.9 47.9
76 12.7 10.6 16.0





#44
2-Chloronaphthalene
Concen: 33.942 ng/ul
RT: 13.292 min Scan# 1753
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832




Tgt Ion:162 Resp: 147237
Ion Ratio Lower Upper
162 100
164 32.8 26.6 39.8
127 39.8 30.9 46.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022




#45
2-Nitroaniline
Concen: 44.640 ng/ul
RT: 13.516 min Scan# 1790
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Tgt Ion: 65 Resp: 448813
Ion Ratio Lower Upper
65 100
92 72.0 58.5 87.7
138 112.0 91.1 136.7





#47
Dimethylphthalate
Concen: 35.073 ng/ul
RT: 13.886 min Scan# 1853
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832




Tgt Ion:163 Resp: 1899064
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.3 8.2 12.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022




#48
2,6-Dinitrotoluene
Concen: 44.511 ng/ul
RT: 14.016 min Scan# 1875
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Tgt Ion:165 Resp: 400179
Ion Ratio Lower Upper
165 100
89 58.3 43.4 65.2
121 21.8 17.1 25.7





#50
Acenaphthylene
Concen: 34.548 ng/ul
RT: 14.145 min Scan# 1897
Delta R.T. -0.000 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion:152 Resp: 2275450
Ion Ratio Lower Upper
152 100
151 19.8 16.1 24.1
153 13.0 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#51
3-Nitroaniline
Concen: 40.733 ng/ul
RT: 14.351 min Scan# 1932
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion:138 Resp: 448410
Ion Ratio Lower Upper
138 100
108 9.9 8.3 12.5
92 111.1 90.6 135.8





#52
Acenaphthene
Concen: 34.129 ng/ul
RT: 14.486 min Scan# 1955
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832




Tgt Ion:153 Resp: 1579748
Ion Ratio Lower Upper
153 100
152 47.1 38.6 58.0
154 88.4 71.6 107.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022




#53
2,4-Dinitrophenol
Concen: 35.836 ng/ul
RT: 14.563 min Scan# 1968
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Tgt Ion:184 Resp: 217578
Ion Ratio Lower Upper
184 100
63 60.2 47.9 71.9
154 61.3 46.9 70.3





#55
4-Nitrophenol
Concen: 33.830 ng/ul
RT: 14.686 min Scan# 1990
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion:109 Resp: 248979
Ion Ratio Lower Upper
109 100
139 140.1 123.1 184.7
65 132.9 107.9 161.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#56
Dibenzofuran
Concen: 34.546 ng/ul
RT: 14.821 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion:168 Resp: 2182993
Ion Ratio Lower Upper
168 100
139 37.8 29.1 43.7





#57
2,4-Dinitrotoluene
Concen: 43.223 ng/ul
RT: 14.810 min Scan# 2010
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832




Tgt Ion:165 Resp: 585316
Ion Ratio Lower Upper
165 100
63 47.7 33.1 49.7
182 2.7 2.1 3.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 38.766 ng/ul
RT: 15.057 min Scan# 2052
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Tgt Ion:232 Resp: 505848
Ion Ratio Lower Upper
232 100
131 44.7 33.6 50.4
130 2.3 1.8 2.6
166 30.2 24.2 36.2





#59
Diethylphthalate
Concen: 36.800 ng/ul
RT: 15.257 min Scan# 2086
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion:149 Resp: 1938863
Ion Ratio Lower Upper
149 100
177 22.0 18.2 27.2
150 12.3 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#61
Fluorene
Concen: 34.884 ng/ul
RT: 15.474 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion:166 Resp: 1751515
Ion Ratio Lower Upper
166 100
165 97.2 77.8 116.8
167 13.4 10.6 15.8





#62
4-Chlorophenyl-phenylether
Concen: 35.281 ng/ul
RT: 15.469 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832




Tgt Ion:204 Resp: 868889
Ion Ratio Lower Upper
204 100
206 33.1 26.5 39.7
141 62.3 47.3 70.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022




#63
4-Nitroaniline
Concen: 45.434 ng/ul m
RT: 15.521 min Scan# 2131
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Tgt Ion:138 Resp: 451873
Ion Ratio Lower Upper
138 100
92 51.5 40.6 61.0
108 76.5 56.6 85.0





#66
4,6-Dinitro-2-methylphenol
Concen: 37.338 ng/ul
RT: 15.574 min Scan# 2140
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion:198 Resp: 345993
Ion Ratio Lower Upper
198 100
182 3.6 3.8 5.6
77 24.0 19.5 29.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#67
N-Nitrosodiphenylamine
Concen: 33.462 ng/ul
RT: 15.692 min Scan# 2160
Delta R.T. -0.000 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion:169 Resp: 1574323
Ion Ratio Lower Upper
169 100
168 65.9 52.9 79.3
167 36.2 29.1 43.7





#68
4-Bromophenyl-phenylether
Concen: 35.991 ng/ul
RT: 16.368 min Scan# 2275
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832




Tgt Ion:248 Resp: 603891
Ion Ratio Lower Upper
248 100
250 98.5 77.4 116.2
141 73.3 56.2 84.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022




#69
Hexachlorobenzene
Concen: 35.217 ng/ul
RT: 16.480 min Scan# 2294
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47



Tgt Ion:284 Resp: 708702
Ion Ratio Lower Upper
284 100
142 32.6 27.1 40.7
249 29.9 24.5 36.7





#70

Atrazine

Concen: 35.457 ng/ul

RT: 16.657 min Scan# 2324

Delta R.T. -0.000 min

Lab File: BP012573.D

Acq: 09 Nov 2022 03:47

Instrument :

BNA_P

ClientSampleId :

SLCS832

Tgt Ion:200 Resp: 57788

Ion Ratio Lower Upper

200 100





173 26.0 21.8 32.6

215 48.1 39.0 58.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#71

Pentachlorophenol

Concen: 34.518 ng/ul

RT: 16.839 min Scan# 2355

Delta R.T. -0.001 min

Lab File: BP012573.D

Acq: 09 Nov 2022 03:47




Tgt Ion:266 Resp: 426608


Ion Ratio Lower Upper

266 100

264 63.0 50.3 75.5

268 64.5 51.4 77.0









#72
Phenanthrene
Concen: 34.067 ng/ul
RT: 17.227 min Scan# 2422
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47

Instrument :
BNA_P
ClientSampleId :
SLCS832

Tgt Ion:178 Resp: 3006342
Ion Ratio Lower Upper
178 100
179 15.5 12.6 18.8
176 19.8 15.8 23.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#74
Anthracene
Concen: 33.805 ng/ul
RT: 17.321 min Scan# 2437
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47

Tgt Ion:178 Resp: 2963711
Ion Ratio Lower Upper
178 100
179 15.5 12.5 18.7
176 18.8 15.3 22.9









#75
1,2,3,4-Tetrachlorobenzene
Concen: 31.575 ng/uL
RT: 13.210 min Scan# 1738
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47

Instrument :
BNA_P
ClientSampleId :
SLCS832

Tgt Ion:216 Resp: 754519
Ion Ratio Lower Upper
216 100
214 79.0 62.6 94.0
218 47.9 38.6 58.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#76
Pentachlorobenzene
Concen: 30.033 ng/uL
RT: 14.739 min Scan# 1998
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47

Tgt Ion:250 Resp: 757835
Ion Ratio Lower Upper
250 100
248 63.2 50.6 75.8
252 64.8 50.7 76.1









#77
 Carbazole
 Concen: 35.708 ng/ul
 RT: 17.598 min Scan# 2484
 Delta R.T. -0.006 min
 Lab File: BP012573.D
 Acq: 09 Nov 2022 03:47

Instrument :
 BNA_P
 ClientSampleId :
 SLCS832

Tgt Ion:167 Resp: 278817
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.1 25.7
 139 13.5 10.4 15.6




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
 Supervised By :mohammad ahmed 11/14/2022



#78
 Di-n-butylphthalate
 Concen: 38.208 ng/ul
 RT: 18.145 min Scan# 2577
 Delta R.T. -0.006 min
 Lab File: BP012573.D
 Acq: 09 Nov 2022 03:47


Tgt Ion:149 Resp: 3466009
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.6 11.4
 104 5.0 4.0 6.0





#80
Fluoranthene
Concen: 37.076 ng/ul
RT: 19.204 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832



Tgt Ion:202 Resp: 363281
Ion Ratio Lower Upper
202 100
101 11.1 9.9 14.9
100 8.4 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#82
Pyrene
Concen: 36.047 ng/ul
RT: 19.557 min Scan# 2817
Delta R.T. -0.000 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47


Tgt Ion:202 Resp: 3658332
Ion Ratio Lower Upper
202 100
101 13.1 11.8 17.6
100 10.6 9.3 13.9





#83
Butylbenzylphthalate
Concen: 43.981 ng/ul
RT: 20.433 min Scan# 2966
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47




Instrument :
BNA_P
ClientSampleId :
SLCS832




Tgt Ion:149 Resp: 1664140
Ion Ratio Lower Upper
149 100
91 67.5 56.9 85.3
206 21.2 16.2 24.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022




#84
3,3'-Dichlorobenzidine
Concen: 36.506 ng/ul
RT: 21.203 min Scan# 3097
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47



Tgt Ion:252 Resp: 1253463
Ion Ratio Lower Upper
252 100
254 65.2 51.8 77.8
126 11.3 9.0 13.4









#85
Benzo(a)anthracene
Concen: 34.283 ng/ul
RT: 21.268 min Scan# 3108
Delta R.T. -0.001 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47

Instrument :
BNA_P
ClientSampleId :
SLCS832

Tgt Ion:228 Resp: 345715
Ion Ratio Lower Upper
228 100
229 19.9 16.2 24.2
226 27.4 22.6 33.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 42.142 ng/ul
RT: 21.186 min Scan# 3094
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47

Tgt Ion:149 Resp: 2438215
Ion Ratio Lower Upper
149 100
167 26.8 22.1 33.1
279 4.4 3.7 5.5





#87

Chrysene

Concen: 34.398 ng/ul

RT: 21.321 min Scan# 3117

Delta R.T. -0.001 min

Lab File: BP012573.D

Acq: 09 Nov 2022 03:47

Instrument :

BNA_P

ClientSampleId :

SLCS832

Tgt Ion:228 Resp: 3242845

Ion Ratio Lower Upper

228 100





226 30.1 24.9 37.3

229 19.8 16.0 24.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#89

Di-n-octyl phthalate

Concen: 47.406 ng/ul




RT: 22.103 min Scan# 3250


Delta R.T. -0.001 min

Lab File: BP012573.D

Acq: 09 Nov 2022 03:47

Tgt Ion:149 Resp: 4126427









#90
Benzo(b)fluoranthene
Concen: 36.462 ng/ul
RT: 22.933 min Scan# 3392
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47

Instrument :
BNA_P
ClientSampleId :
SLCS832

Tgt Ion:252 Resp: 3369426
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.8
125 9.8 8.7 13.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/09/2022
Supervised By :mohammad ahmed 11/14/2022



#91
Benzo(k)fluoranthene
Concen: 37.018 ng/ul
RT: 22.980 min Scan# 3399
Delta R.T. -0.006 min
Lab File: BP012573.D
Acq: 09 Nov 2022 03:47

Tgt Ion:252 Resp: 3260064
Ion Ratio Lower Upper
252 100
253 22.2 18.0 27.0
125 9.8 8.7 13.1





#93

Benzo(a)pyrene

Concen: 35.346 ng/ul

RT: 23.556 min Scan# 3497

Delta R.T. -0.000 min

Lab File: BP012573.D

Acq: 09 Nov 2022 03:47

Instrument :

BNA_P

ClientSampleId :





SLCS832

Tgt Ion:252 Resp: 2852317
 Ion Ratio Lower Upper
 252 100
 253 21.7 18.0 27.0
 125 10.7 9.4 14.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/09/2022
 Supervised By :mohammad ahmed 11/14/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 33.037 ng/ul




RT: 26.127 min Scan# 3934


Delta R.T. -0.006 min

Lab File: BP012573.D

Acq: 09 Nov 2022 03:47

Tgt Ion:276 Resp: 3322098
 Ion Ratio Lower Upper
 276 100
 138 23.5 20.7 31.1
 277 25.0 20.2 30.4





#95

Dibenzo(a,h)anthracene

Concen: 32.698 ng/ul

RT: 26.144 min Scan# 3938

Delta R.T. -0.006 min

Lab File: BP012573.D

Acq: 09 Nov 2022 03:47

Instrument :

BNA_P

ClientSampleId :

SLCS832

Tgt Ion:278 Resp: 294410

Ion Ratio Lower Upper

278 100

139 14.6 13.2 19.8





279 23.5 19.4 29.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/09/2022

Supervised By :mohammad ahmed 11/14/2022



#96

Benzo(g,h,i)perylene

Concen: 32.437 ng/ul

RT: 26.886 min Scan# 4063

Delta R.T. -0.006 min

Lab File: BP012573.D

Acq: 09 Nov 2022 03:47

Tgt Ion:276 Resp: 2883326

Ion Ratio Lower Upper

276 100

138 20.3 17.8 26.8

277 24.1 19.5 29.3

