

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110824\
 Data File : BP022938.D
 Acq On : 08 Nov 2024 17:32
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 08 17:56:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 07 23:36:11 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	84	0.00
2	1,4-Dioxane	40.000	40.480	-1.2	85	0.00
3	Pyridine	40.000	41.378	-3.4	86	0.00
4	n-Nitrosodimethylamine	40.000	47.107	-17.8	99	0.00
5 S	2-Fluorophenol	80.000	80.839	-1.0	84	0.00
6	Aniline	40.000	40.474	-1.2	84	0.00
7 S	Phenol-d6	80.000	80.535	-0.7	84	0.00
8	2-Chlorophenol	40.000	40.622	-1.6	85	0.00
9	Benzaldehyde	40.000	40.836	-2.1	86	0.00
10 C	Phenol	40.000	40.565	-1.4	84	0.00
11	bis(2-Chloroethyl)ether	40.000	40.278	-0.7	85	0.00
12	1,3-Dichlorobenzene	40.000	40.752	-1.9	87	0.00
13 C	1,4-Dichlorobenzene	40.000	40.872	-2.2	88	0.00
14	1,2-Dichlorobenzene	40.000	40.264	-0.7	86	0.00
15	Benzyl Alcohol	40.000	42.377	-5.9	90	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.410	-1.0	86	0.00
17	2-Methylphenol	40.000	41.536	-3.8	86	0.00
18	Hexachloroethane	40.000	42.284	-5.7	90	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.740	-6.9	90	0.00
20	3+4-Methylphenols	40.000	42.413	-6.0	88	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	91	0.00
22	Acetophenone	40.000	37.666	5.8	87	0.00
23 S	Nitrobenzene-d5	80.000	80.874	-1.1	91	0.00
24	Nitrobenzene	40.000	40.688	-1.7	92	0.00
25	Isophorone	40.000	39.539	1.2	90	0.00
26 C	2-Nitrophenol	40.000	39.045	2.4	88	0.00
27	2,4-Dimethylphenol	40.000	39.224	1.9	90	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.818	3.0	89	0.00
29 C	2,4-Dichlorophenol	40.000	40.691	-1.7	90	0.00
30	1,2,4-Trichlorobenzene	40.000	40.374	-0.9	93	0.00
31	Naphthalene	40.000	39.534	1.2	90	0.00
32	Benzoic acid	40.000	40.516	-1.3	88	0.00
33	4-Chloroaniline	40.000	40.739	-1.8	91	0.00
34 C	Hexachlorobutadiene	40.000	41.748	-4.4	95	0.00
35	Caprolactam	40.000	40.839	-2.1	92	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.393	-3.5	92	0.00
37	2-Methylnaphthalene	40.000	41.455	-3.6	95	0.00
38	1-Methylnaphthalene	40.000	41.209	-3.0	93	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.692	5.8	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.478	-3.7	97	0.00
42 S	2,4,6-Tribromophenol	80.000	87.657	-9.6	110	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.968	2.6	95	0.00
44	2,4,5-Trichlorophenol	40.000	38.877	2.8	95	0.00
45 S	2-Fluorobiphenyl	80.000	77.810	2.7	99	0.00
46	1,1'-Biphenyl	40.000	38.342	4.1	95	0.00
47	2-Chloronaphthalene	40.000	37.990	5.0	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110824\
 Data File : BP022938.D
 Acq On : 08 Nov 2024 17:32
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 08 17:56:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 07 23:36:11 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.847	-4.6	102	0.00
49	Acenaphthylene	40.000	39.795	0.5	99	0.00
50	Dimethylphthalate	40.000	38.527	3.7	97	-0.01
51	2,6-Dinitrotoluene	40.000	39.659	0.9	99	0.02
52 C	Acenaphthene	40.000	39.244	1.9	97	0.00
53	3-Nitroaniline	40.000	40.238	-0.6	98	0.01
54 P	2,4-Dinitrophenol	40.000	41.739	-4.3	99	-0.02
55	Dibenzofuran	40.000	40.112	-0.3	102	0.00
56 P	4-Nitrophenol	40.000	43.420	-8.6	102	0.01
57	2,4-Dinitrotoluene	40.000	41.930	-4.8	103	-0.01
58	Fluorene	40.000	41.723	-4.3	105	0.02
59	2,3,4,6-Tetrachlorophenol	40.000	41.396	-3.5	102	0.02
60	Diethylphthalate	40.000	41.914	-4.8	106	-0.01
61	4-Chlorophenyl-phenylether	40.000	41.604	-4.0	106	0.00
62	4-Nitroaniline	40.000	47.782	-19.5	113	0.00
63	Azobenzene	40.000	43.314	-8.3	107	0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	112	0.01
65	4,6-Dinitro-2-methylphenol	40.000	39.575	1.1	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.122	7.2	103	0.00
67	4-Bromophenyl-phenylether	40.000	38.663	3.3	108	-0.01
68	Hexachlorobenzene	40.000	38.476	3.8	110	0.00
69	Atrazine	40.000	40.980	-2.4	104	0.00
70 C	Pentachlorophenol	40.000	39.706	0.7	107	0.02
71	Phenanthrene	40.000	39.726	0.7	112	0.02
72	Anthracene	40.000	39.810	0.5	111	0.00
73	Carbazole	40.000	39.849	0.4	111	0.01
74	Di-n-butylphthalate	40.000	41.164	-2.9	114	0.01
75 C	Fluoranthene	40.000	39.417	1.5	109	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	114	0.00
77	Benzydine	40.000	32.783	18.0	78	0.00
78	Pyrene	40.000	38.566	3.6	112	0.00
79 S	Terphenyl-d14	80.000	77.819	2.7	113	0.00
80	Butylbenzylphthalate	40.000	40.025	-0.1	112	0.01
81	Benzo(a)anthracene	40.000	38.781	3.0	114	0.00
82	3,3'-Dichlorobenzidine	40.000	39.397	1.5	114	0.01
83	Chrysene	40.000	39.512	1.2	114	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.478	-3.7	120	0.00
85 c	Di-n-octyl phthalate	40.000	40.770	-1.9	116	0.00
86 I	Perylene-d12	20.000	20.000	0.0	115	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.137	-0.3	117	0.04
88	Benzo(b)fluoranthene	40.000	41.362	-3.4	117	0.01
89	Benzo(k)fluoranthene	40.000	40.764	-1.9	121	0.00
90 C	Benzo(a)pyrene	40.000	40.619	-1.5	117	0.02
91	Dibenzo(a,h)anthracene	40.000	39.948	0.1	118	0.03
92	Benzo(g,h,i)perylene	40.000	39.880	0.3	115	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110824\
Data File : BP022938.D
Acq On : 08 Nov 2024 17:32
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 08 17:56:13 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 07 23:36:11 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0