



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110920\
 Data File : BP003880.D
 Acq On : 09 Nov 2020 23:37
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 10 13:41:24 2020



AutoFind: Scans 2581, 2582, 2583; Background Corrected with Scan 2574

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.7	56819	PASS
68	69	0.00	2	1.8	987	PASS
69	198	0.00	100	34.1	55897	PASS
70	69	0.00	2	0.7	366	PASS
127	198	10	80	40.8	66862	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	163891	PASS
199	198	5	9	6.9	11385	PASS
275	198	10	60	28.9	47418	PASS
365	198	1	100	3.6	5926	PASS
441	198	0.01	100	16.2	26587	PASS
442	442	100	100	100.0	180693	PASS
443	442	15	24	19.2	34718	PASS