

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111021\
 Data File : BP007929.D
 Acq On : 10 Nov 2021 20:32
 Operator : CG/JU
 Sample : M4578-03
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 FB

Quant Time: Nov 11 10:17:49 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 09 10:45:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.851	152	112437	20.000	ng	-0.01
21) Naphthalene-d8	10.651	136	380438	20.000	ng	-0.01
39) Acenaphthene-d10	14.492	164	245890	20.000	ng	-0.01
64) Phenanthrene-d10	17.245	188	568694	20.000	ng	-0.01
76) Chrysene-d12	21.339	240	717266	20.000	ng	#-0.01
86) Perylene-d12	23.750	264	812834	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.440	112	641057	96.693	ng	0.00
7) Phenol-d6	7.028	99	502579	52.374	ng	-0.01
23) Nitrobenzene-d5	9.010	82	718884	101.505	ng	-0.02
42) 2,4,6-Tribromophenol	15.986	330	592640	168.882	ng	-0.01
45) 2-Fluorobiphenyl	13.116	172	1693070	100.271	ng	-0.01
79) Terphenyl-d14	19.809	244	3281386	93.157	ng	-0.01


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111021\
Data File : BP007929.D
Acq On : 10 Nov 2021 20:32
Operator : CG/JU
Sample : M4578-03
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
FB

Quant Time: Nov 11 10:17:49 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 09 10:45:24 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.851 min Scan# 71
Delta R.T. -0.013 min
Lab File: BP007929.D
Acq: 10 Nov 2021 20:32




Instrument :
BNA_P
ClientSampleId :
FB

Tgt Ion:152 Resp: 112437
Ion Ratio Lower Upper
152 100
150 152.5 124.1 186.1
115 61.3 48.1 72.1



#5
2-Fluorophenol
Concen: 96.693 ng
RT: 5.440 min Scan# 366
Delta R.T. -0.006 min
Lab File: BP007929.D
Acq: 10 Nov 2021 20:32

Tgt Ion:112 Resp: 641057
Ion Ratio Lower Upper
112 100
64 54.5 46.6 70.0
63 30.1 25.0 37.6









#7
Phenol-d6
Concen: 52.374 ng
RT: 7.028 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP007929.D
Acq: 10 Nov 2021 20:32




Instrument :
BNA_P
ClientSampleId :
FB


Tgt Ion: 99 Resp: 502579
Ion Ratio Lower Upper
99 100
42 17.3 14.9 22.3
71 39.8 29.2 43.8



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.651 min Scan# 1252
Delta R.T. -0.012 min
Lab File: BP007929.D
Acq: 10 Nov 2021 20:32

Tgt Ion: 136 Resp: 380438
Ion Ratio Lower Upper
136 100
137 11.3 8.6 13.0
54 7.7 6.2 9.4
68 5.0 4.7 7.1









#23
Nitrobenzene-d5
Concen: 101.505 ng
RT: 9.010 min Scan# 91
Delta R.T. -0.018 min
Lab File: BP007929.D
Acq: 10 Nov 2021 20:32




Instrument :
BNA_P
ClientSampleId :
FB


Tgt Ion: 82 Resp: 718884
Ion Ratio Lower Upper
82 100
128 42.1 32.6 48.8
54 49.6 40.4 60.6



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.492 min Scan# 1905
Delta R.T. -0.012 min
Lab File: BP007929.D
Acq: 10 Nov 2021 20:32

Tgt Ion: 164 Resp: 245890
Ion Ratio Lower Upper
164 100
162 103.5 81.1 121.7
160 45.6 36.0 54.0









#42
2,4,6-Tribromophenol
Concen: 168.882 ng
RT: 15.986 min Scan# 2171
Delta R.T. -0.012 min
Lab File: BP007929.D
Acq: 10 Nov 2021 20:32




Instrument :
BNA_P
ClientSampleId :
FB


Tgt Ion:330 Resp: 592640
Ion Ratio Lower Upper
330 100
332 97.1 78.9 118.3
141 32.6 30.7 46.1



#45
2-Fluorobiphenyl
Concen: 100.271 ng
RT: 13.116 min Scan# 1671
Delta R.T. -0.012 min
Lab File: BP007929.D
Acq: 10 Nov 2021 20:32

Tgt Ion:172 Resp: 1693070
Ion Ratio Lower Upper
172 100
171 35.7 28.2 42.2
170 23.9 19.0 28.4









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.245 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP007929.D
Acq: 10 Nov 2021 20:32




Instrument :
BNA_P
ClientSampleId :
FB


Tgt Ion:188 Resp: 568694
Ion Ratio Lower Upper
188 100
94 7.5 7.1 10.7
80 8.3 7.8 11.6



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.339 min Scan# 3069
Delta R.T. -0.012 min
Lab File: BP007929.D
Acq: 10 Nov 2021 20:32

Tgt Ion:240 Resp: 717266
Ion Ratio Lower Upper
240 100
120 7.8 7.8 11.8#
236 26.4 20.6 30.8









#79
 Terphenyl-d14
 Concen: 93.157 ng
 RT: 19.809 min Scan# 21
 Delta R.T. -0.012 min
 Lab File: BP007929.D
 Acq: 10 Nov 2021 20:32

Instrument :
 BNA_P
 ClientSampleId :
 FB

Tgt Ion:244 Resp: 3281386
 Ion Ratio Lower Upper
 244 100
 212 7.7 6.2 9.2
 122 8.0 7.8 11.8



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 23.750 min Scan# 3479
 Delta R.T. -0.018 min
 Lab File: BP007929.D
 Acq: 10 Nov 2021 20:32

Tgt Ion:264 Resp: 812834
 Ion Ratio Lower Upper
 264 100
 260 24.0 19.1 28.7
 265 21.9 16.7 25.1

