

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111124\
 Data File : BP023007.D
 Acq On : 11 Nov 2024 10:12
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 11 11:07:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 07 23:36:11 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	97	-0.01
2	1,4-Dioxane	40.000	41.285	-3.2	100	0.00
3	Pyridine	40.000	40.719	-1.8	98	0.00
4	n-Nitrosodimethylamine	40.000	46.461	-16.2	112	0.00
5 S	2-Fluorophenol	80.000	81.538	-1.9	97	0.00
6	Aniline	40.000	40.352	-0.9	97	0.00
7 S	Phenol-d6	80.000	79.335	0.8	96	-0.01
8	2-Chlorophenol	40.000	40.020	-0.1	96	-0.01
9	Benzaldehyde	40.000	38.735	3.2	94	-0.01
10 C	Phenol	40.000	40.114	-0.3	96	-0.01
11	bis(2-Chloroethyl)ether	40.000	40.058	-0.1	97	-0.01
12	1,3-Dichlorobenzene	40.000	40.772	-1.9	100	-0.01
13 C	1,4-Dichlorobenzene	40.000	40.459	-1.1	100	-0.02
14	1,2-Dichlorobenzene	40.000	40.164	-0.4	99	-0.02
15	Benzyl Alcohol	40.000	40.440	-1.1	98	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	40.190	-0.5	99	-0.01
17	2-Methylphenol	40.000	40.345	-0.9	96	-0.01
18	Hexachloroethane	40.000	41.736	-4.3	103	-0.02
19 P	n-Nitroso-di-n-propylamine	40.000	40.971	-2.4	99	-0.01
20	3+4-Methylphenols	40.000	41.062	-2.7	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	-0.01
22	Acetophenone	40.000	37.832	5.4	99	0.00
23 S	Nitrobenzene-d5	80.000	79.949	0.1	102	-0.01
24	Nitrobenzene	40.000	39.688	0.8	101	-0.01
25	Isophorone	40.000	38.406	4.0	99	-0.01
26 C	2-Nitrophenol	40.000	39.217	2.0	100	-0.01
27	2,4-Dimethylphenol	40.000	38.704	3.2	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.079	2.3	101	-0.01
29 C	2,4-Dichlorophenol	40.000	40.912	-2.3	103	-0.01
30	1,2,4-Trichlorobenzene	40.000	40.330	-0.8	105	-0.02
31	Naphthalene	40.000	39.026	2.4	101	-0.01
32	Benzoic acid	40.000	39.905	0.2	98	0.00
33	4-Chloroaniline	40.000	40.867	-2.2	103	-0.01
34 C	Hexachlorobutadiene	40.000	41.291	-3.2	106	-0.01
35	Caprolactam	40.000	40.581	-1.5	103	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.805	-4.5	105	0.00
37	2-Methylnaphthalene	40.000	41.442	-3.6	107	-0.01
38	1-Methylnaphthalene	40.000	41.217	-3.0	105	-0.02
39 I	Acenaphthene-d10	20.000	20.000	0.0	114	-0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	38.044	4.9	110	-0.01
41 P	Hexachlorocyclopentadiene	40.000	41.449	-3.6	111	-0.01
42 S	2,4,6-Tribromophenol	80.000	85.584	-7.0	123	-0.02
43 C	2,4,6-Trichlorophenol	40.000	38.578	3.6	108	-0.02
44	2,4,5-Trichlorophenol	40.000	38.293	4.3	107	-0.01
45 S	2-Fluorobiphenyl	80.000	75.747	5.3	110	-0.01
46	1,1'-Biphenyl	40.000	38.003	5.0	108	-0.01
47	2-Chloronaphthalene	40.000	38.364	4.1	111	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111124\
 Data File : BP023007.D
 Acq On : 11 Nov 2024 10:12
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 11 11:07:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 07 23:36:11 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.270	-3.2	115	-0.01
49	Acenaphthylene	40.000	39.199	2.0	112	-0.02
50	Dimethylphthalate	40.000	38.631	3.4	111	-0.01
51	2,6-Dinitrotoluene	40.000	39.184	2.0	112	0.00
52 C	Acenaphthene	40.000	39.102	2.2	111	-0.02
53	3-Nitroaniline	40.000	41.405	-3.5	115	0.00
54 P	2,4-Dinitrophenol	40.000	40.973	-2.4	111	-0.02
55	Dibenzofuran	40.000	41.122	-2.8	120	-0.02
56 P	4-Nitrophenol	40.000	42.825	-7.1	115	0.00
57	2,4-Dinitrotoluene	40.000	41.430	-3.6	116	-0.02
58	Fluorene	40.000	40.706	-1.8	117	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	41.175	-2.9	116	0.00
60	Diethylphthalate	40.000	41.070	-2.7	118	-0.02
61	4-Chlorophenyl-phenylether	40.000	41.107	-2.8	120	-0.02
62	4-Nitroaniline	40.000	43.878	-9.7	118	-0.02
63	Azobenzene	40.000	43.581	-9.0	123	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	125	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.631	-1.6	120	-0.02
66 c	n-Nitrosodiphenylamine	40.000	38.540	3.7	120	-0.01
67	4-Bromophenyl-phenylether	40.000	38.642	3.4	121	-0.02
68	Hexachlorobenzene	40.000	38.831	2.9	124	-0.01
69	Atrazine	40.000	41.347	-3.4	117	-0.02
70 C	Pentachlorophenol	40.000	40.173	-0.4	121	0.00
71	Phenanthrene	40.000	39.510	1.2	124	0.00
72	Anthracene	40.000	40.383	-1.0	126	-0.02
73	Carbazole	40.000	40.247	-0.6	126	0.00
74	Di-n-butylphthalate	40.000	41.432	-3.6	128	-0.01
75 C	Fluoranthene	40.000	40.083	-0.2	123	-0.01
76 I	Chrysene-d12	20.000	20.000	0.0	126	-0.01
77	Benzydine	40.000	43.346	-8.4	114	-0.02
78	Pyrene	40.000	40.133	-0.3	128	0.00
79 S	Terphenyl-d14	80.000	80.999	-1.2	130	-0.01
80	Butylbenzylphthalate	40.000	41.033	-2.6	127	0.00
81	Benzo(a)anthracene	40.000	39.563	1.1	128	-0.01
82	3,3'-Dichlorobenzidine	40.000	39.979	0.1	128	0.00
83	Chrysene	40.000	40.381	-1.0	129	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.105	-2.8	132	-0.02
85 c	Di-n-octyl phthalate	40.000	41.800	-4.5	131	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	127	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	40.147	-0.4	129	-0.02
88	Benzo(b)fluoranthene	40.000	40.349	-0.9	125	-0.02
89	Benzo(k)fluoranthene	40.000	40.751	-1.9	133	-0.02
90 C	Benzo(a)pyrene	40.000	40.273	-0.7	128	0.00
91	Dibenzo(a,h)anthracene	40.000	40.404	-1.0	132	-0.01
92	Benzo(g,h,i)perylene	40.000	39.194	2.0	125	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111124\
Data File : BP023007.D
Acq On : 11 Nov 2024 10:12
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 11 11:07:29 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 07 23:36:11 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0