

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111124\
 Data File : BP023014.D
 Acq On : 11 Nov 2024 15:01
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 11 15:28:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 07 23:36:11 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	88	-0.02
2	1,4-Dioxane	40.000	38.882	2.8	85	-0.01
3	Pyridine	40.000	39.791	0.5	87	-0.02
4	n-Nitrosodimethylamine	40.000	46.904	-17.3	103	-0.02
5 S	2-Fluorophenol	80.000	77.624	3.0	85	-0.03
6	Aniline	40.000	38.929	2.7	85	-0.03
7 S	Phenol-d6	80.000	77.249	3.4	85	-0.02
8	2-Chlorophenol	40.000	38.973	2.6	85	-0.02
9	Benzaldehyde	40.000	39.858	0.4	88	-0.03
10 C	Phenol	40.000	38.698	3.3	84	-0.02
11	bis(2-Chloroethyl)ether	40.000	37.750	5.6	84	-0.04
12	1,3-Dichlorobenzene	40.000	38.667	3.3	86	-0.02
13 C	1,4-Dichlorobenzene	40.000	39.124	2.2	88	-0.03
14	1,2-Dichlorobenzene	40.000	39.544	1.1	88	-0.02
15	Benzyl Alcohol	40.000	40.231	-0.6	89	-0.02
16	2,2'-oxybis(1-Chloropropane	40.000	39.679	0.8	89	-0.03
17	2-Methylphenol	40.000	39.953	0.1	86	-0.04
18	Hexachloroethane	40.000	41.385	-3.5	93	-0.02
19 P	n-Nitroso-di-n-propylamine	40.000	40.669	-1.7	89	-0.03
20	3+4-Methylphenols	40.000	40.164	-0.4	88	-0.03
21 I	Naphthalene-d8	20.000	20.000	0.0	93	-0.03
22	Acetophenone	40.000	38.167	4.6	89	-0.03
23 S	Nitrobenzene-d5	80.000	82.723	-3.4	95	-0.03
24	Nitrobenzene	40.000	40.802	-2.0	93	-0.03
25	Isophorone	40.000	39.710	0.7	92	-0.02
26 C	2-Nitrophenol	40.000	39.395	1.5	90	-0.02
27	2,4-Dimethylphenol	40.000	38.293	4.3	89	-0.02
28	bis(2-Chloroethoxy)methane	40.000	38.728	3.2	90	-0.02
29 C	2,4-Dichlorophenol	40.000	40.765	-1.9	92	-0.02
30	1,2,4-Trichlorobenzene	40.000	39.758	0.6	93	-0.03
31	Naphthalene	40.000	39.929	0.2	92	-0.03
32	Benzoic acid	40.000	40.035	-0.1	88	-0.01
33	4-Chloroaniline	40.000	41.461	-3.7	94	-0.02
34 C	Hexachlorobutadiene	40.000	41.388	-3.5	96	-0.02
35	Caprolactam	40.000	39.396	1.5	90	-0.02
36 C	4-Chloro-3-methylphenol	40.000	42.358	-5.9	96	-0.02
37	2-Methylnaphthalene	40.000	42.025	-5.1	98	-0.02
38	1-Methylnaphthalene	40.000	41.813	-4.5	96	-0.02
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	-0.02
40	1,2,4,5-Tetrachlorobenzene	40.000	36.999	7.5	98	-0.02
41 P	Hexachlorocyclopentadiene	40.000	42.031	-5.1	104	-0.02
42 S	2,4,6-Tribromophenol	80.000	84.101	-5.1	111	-0.01
43 C	2,4,6-Trichlorophenol	40.000	38.358	4.1	99	-0.02
44	2,4,5-Trichlorophenol	40.000	37.979	5.1	98	-0.02
45 S	2-Fluorobiphenyl	80.000	75.238	6.0	101	-0.02
46	1,1'-Biphenyl	40.000	38.561	3.6	101	-0.02
47	2-Chloronaphthalene	40.000	37.801	5.5	100	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111124\
 Data File : BP023014.D
 Acq On : 11 Nov 2024 15:01
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 11 15:28:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 07 23:36:11 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.228	-3.1	106	-0.01
49	Acenaphthylene	40.000	39.481	1.3	104	-0.02
50	Dimethylphthalate	40.000	38.589	3.5	102	-0.02
51	2,6-Dinitrotoluene	40.000	38.948	2.6	103	-0.01
52 C	Acenaphthene	40.000	38.958	2.6	102	-0.02
53	3-Nitroaniline	40.000	41.033	-2.6	105	-0.01
54 P	2,4-Dinitrophenol	40.000	39.953	0.1	99	-0.01
55	Dibenzofuran	40.000	40.382	-1.0	108	-0.02
56 P	4-Nitrophenol	40.000	41.432	-3.6	102	-0.01
57	2,4-Dinitrotoluene	40.000	41.627	-4.1	108	-0.02
58	Fluorene	40.000	41.138	-2.8	109	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	41.184	-3.0	107	-0.01
60	Diethylphthalate	40.000	41.309	-3.3	110	-0.02
61	4-Chlorophenyl-phenylether	40.000	41.265	-3.2	111	-0.02
62	4-Nitroaniline	40.000	41.173	-2.9	102	-0.01
63	Azobenzene	40.000	44.380	-11.0	115	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	119	-0.01
65	4,6-Dinitro-2-methylphenol	40.000	38.051	4.9	107	-0.01
66 c	n-Nitrosodiphenylamine	40.000	37.289	6.8	111	-0.01
67	4-Bromophenyl-phenylether	40.000	37.728	5.7	113	-0.01
68	Hexachlorobenzene	40.000	37.171	7.1	113	-0.02
69	Atrazine	40.000	38.881	2.8	105	0.00
70 C	Pentachlorophenol	40.000	37.752	5.6	108	0.00
71	Phenanthrene	40.000	38.367	4.1	115	0.00
72	Anthracene	40.000	38.530	3.7	114	-0.01
73	Carbazole	40.000	38.734	3.2	115	0.00
74	Di-n-butylphthalate	40.000	39.158	2.1	115	0.00
75 C	Fluoranthene	40.000	39.312	1.7	115	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	117	0.00
77	Benzydine	40.000	53.982	-35.0#	132	0.00
78	Pyrene	40.000	39.674	0.8	118	0.00
79 S	Terphenyl-d14	80.000	78.994	1.3	118	0.00
80	Butylbenzylphthalate	40.000	39.912	0.2	115	0.00
81	Benzo(a)anthracene	40.000	39.210	2.0	118	0.00
82	3,3'-Dichlorobenzidine	40.000	38.792	3.0	116	0.00
83	Chrysene	40.000	39.597	1.0	118	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.003	-2.5	123	0.00
85 c	Di-n-octyl phthalate	40.000	40.386	-1.0	118	0.00
86 I	Perylene-d12	20.000	20.000	0.0	118	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.862	0.3	120	0.00
88	Benzo(b)fluoranthene	40.000	39.884	0.3	115	0.00
89	Benzo(k)fluoranthene	40.000	39.766	0.6	121	-0.01
90 C	Benzo(a)pyrene	40.000	39.977	0.1	118	0.00
91	Dibenzo(a,h)anthracene	40.000	40.241	-0.6	122	0.02
92	Benzo(g,h,i)perylene	40.000	39.615	1.0	118	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111124\
Data File : BP023014.D
Acq On : 11 Nov 2024 15:01
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 11 15:28:00 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 07 23:36:11 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0