

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111224\
 Data File : BP023050.D
 Acq On : 12 Nov 2024 17:16
 Operator : RC/JU
 Sample : PB164832BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK832

Quant Time: Nov 13 02:07:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 12 23:59:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.587	152	61218	20.000	ng/u1	0.00
20) Naphthalene-d8	10.334	136	236361	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.198	164	197139	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.010	188	488506	20.000	ng/u1	0.00
79) Chrysene-d12	21.445	240	555417	20.000	ng/u1	0.00
88) Perylene-d12	24.651	264	631827	20.000	ng/u1	-0.04

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.140	96	8779	6.893	ng/uL	0.00
4) Pyridine-d5	3.552	84	116838	30.878	ng/u1	0.00
7) Phenol-d5	6.787	99	139811	30.891	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.934	67	86483	32.523	ng/u1	0.00
11) 2-Chlorophenol-d4	7.134	132	111983	33.783	ng/u1	0.00
15) 4-Methylphenol-d8	8.299	113	115379	30.978	ng/u1	0.00
21) Nitrobenzene-d5	8.728	128	54713	33.250	ng/u1	0.00
24) 2-Nitrophenol-d4	9.440	143	66911	34.109	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.981	165	128022	31.163	ng/u1	0.00
31) 4-Chloroaniline-d4	10.487	131	150161	30.638	ng/u1	0.00
46) Dimethylphthalate-d6	13.604	166	486980	34.274	ng/u1	-0.01
49) Acenaphthylene-d8	13.887	160	504648	34.039	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.475	143	58702	27.492	ng/u1	0.02
60) Fluorene-d10	15.222	176	425845	34.185	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.363	200	80501	27.509	ng/u1	-0.01
73) Anthracene-d10	17.104	188	741214	35.835	ng/u1	-0.02
81) Pyrene-d10	19.498	212	902891	35.293	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.445	264	1100044	36.551	ng/u1	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111224\
Data File : BP023050.D
Acq On : 12 Nov 2024 17:16
Operator : RC/JU
Sample : PB164832BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK832

Quant Time: Nov 13 02:07:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 12 23:59:19 2024
Response via : Initial Calibration

