

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111224\
 Data File : BP023069.D
 Acq On : 13 Nov 2024 06:45
 Operator : RC/JU
 Sample : P4687-23
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0CC7

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/14/2024
 Supervised By :mohammad ahmed 11/14/2024

Quant Time: Nov 13 07:14:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 13 04:49:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.581	152	58038	20.000	ng/u1	0.00
20) Naphthalene-d8	10.328	136	235805	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.187	164	193946	20.000	ng/u1	-0.02
64) Phenanthrene-d10	16.998	188	502391	20.000	ng/u1	-0.02
79) Chrysene-d12	21.433	240	606039	20.000	ng/u1	-0.01
88) Perylene-d12	24.633	264	776668	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.140	96	6174	5.113	ng/uL	0.00
4) Pyridine-d5	3.575	84	10003m	2.788	ng/u1	0.03
7) Phenol-d5	6.799	99	21066m	4.909	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	6.928	67	66963	26.562	ng/u1	0.00
11) 2-Chlorophenol-d4	7.134	132	43811	13.941	ng/u1	0.00
15) 4-Methylphenol-d8	8.293	113	50547	14.315	ng/u1	0.00
21) Nitrobenzene-d5	8.722	128	39167	23.859	ng/u1	0.00
24) 2-Nitrophenol-d4	9.446	143	27520	14.062	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	9.987	165	58281m	14.220	ng/u1	0.02
31) 4-Chloroaniline-d4	10.487	131	99370	20.323	ng/u1	0.00
46) Dimethylphthalate-d6	13.610	166	372065	26.617	ng/u1	0.00
49) Acenaphthylene-d8	13.881	160	345299	23.674	ng/u1	-0.02
54) 4-Nitrophenol-d4	14.534	143	6126m	2.916	ng/u1	0.08
60) Fluorene-d10	15.204	176	321204	26.209	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.369	200	31416m	10.439	ng/u1	0.00
73) Anthracene-d10	17.092	188	523572	24.614	ng/u1	-0.02
81) Pyrene-d10	19.486	212	734676	26.319	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.421	264	708580	19.153	ng/u1	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111224\
Data File : BP023069.D
Acq On : 13 Nov 2024 06:45
Operator : RC/JU
Sample : P4687-23
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0CC7

Quant Time: Nov 13 07:14:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 13 04:49:01 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/14/2024
Supervised By :mohammad ahmed 11/14/2024

