


Benzaldehyde

Response Ratio



$$\text{Response} = 8.253\text{e-}001 * \text{Amt} + 1.661\text{e-}001$$


Coef of Det (r^2) = 0.994663 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP111321.M

Calibration Table Last Updated: Mon Nov 15 05:36:49 2021

2,2'-oxybis(1-Chloropropane)

Response Ratio



$$\text{Response} = 1.282\text{e}+000 * \text{Amt} + 2.025\text{e}-001$$


Coef of Det (r^2) = 0.998263 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP111321.M

Calibration Table Last Updated: Mon Nov 15 05:36:49 2021

Hexachlorocyclopentadiene

Response Ratio



$$\text{Response} = 3.496\text{e-}001 * \text{Amt} - 5.366\text{e-}002$$

Coef of Det (r^2) = 0.993231 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP111321.M

Calibration Table Last Updated: Mon Nov 15 05:36:49 2021