




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111422\
 Data File : BP012645.D
 Acq On : 14 Nov 2022 10:02
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 14 17:01:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 14 04:00:17 2022
 Response via : Initial Calibration



TIC: BP012645.D\data.ms




(70) Pentachlorophenol (C)

16.786min (-0.000) 345585.52 ng

response 643594

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.80	63.55
264.00	64.30	61.34
0.00	0.00	0.00

Quant Time: Nov 14 17:01:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 14 04:00:17 2022
Response via : Initial Calibration



TIC: BP012645.D\data.ms

(77) Benzidine

19.351min (+ 0.006) 965558.99 ng

response **3301860**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	14.01
183.00	11.60	11.56
0.00	0.00	0.00