


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111422\
 Data File : BP012645.D
 Acq On : 14 Nov 2022 10:02
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111222.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Nov 14 04:00:17 2022






AutoFind: Scans 2479, 2480, 2481; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.0	152520	PASS
68	69	0.00	2	1.6	2614	PASS
69	198	0.00	100	33.8	161126	PASS
70	69	0.00	2	0.5	863	PASS
127	198	10	80	46.0	219178	PASS
197	198	0.00	2	0.5	2264	PASS
198	198	100	100	100.0	476650	PASS
199	198	5	9	6.8	32371	PASS
275	198	10	60	28.6	136131	PASS
365	198	1	100	4.3	20397	PASS
441	198	0.01	100	14.7	69916	PASS
442	442	100	100	100.0	456955	PASS
443	442	15	24	19.3	88371	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111422\
Data File : BP012645.D
Acq On : 14 Nov 2022 10:02
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Nov 14 17:01:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 14 04:00:17 2022
Response via : Initial Calibration



TIC: BP012645.D\data.ms

(70) Pentachlorophenol (C)

16.786min (-0.000) 345585.52 ng




response 643594

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.80	63.55
264.00	64.30	61.34
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111422\
 Data File : BP012645.D
 Acq On : 14 Nov 2022 10:02
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Nov 14 17:01:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 14 04:00:17 2022
 Response via : Initial Calibration



TIC: BP012645.D\data.ms

(77) Benzidine

19.351min (+ 0.006) 965558.99 ng

response 3301860

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	14.01
183.00	11.60	11.56
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
11/14/2022	BNA_P	<u>BP012645.D</u>
Compound Name	Response	Retention Time
DDT	1788020	20.539
DDD	34568	20.116
DDE	1430	19.616
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
35998	1824018	1.97

Instrument :
BNA_P
ClientSampleId :
DFTPP