

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111521\
 Data File : BP007978.D
 Acq On : 15 Nov 2021 10:18
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 15 11:44:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 05:36:49 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.519	0.482	7.1	86	0.00
3	Pyridine	1.488	1.446	2.8	88	0.00
4	n-Nitrosodimethylamine	0.550	0.574	-4.4	89	0.00
5 S	2-Fluorophenol	1.233	1.190	3.5	87	0.00
6	Aniline	1.852	1.886	-1.8	87	0.00
7 S	Phenol-d6	1.640	1.619	1.3	87	0.00
8	2-Chlorophenol	1.303	1.309	-0.5	85	0.00
9	Benzaldehyde	1.058	0.952	10.0	85	0.00
10 C	Phenol	1.591	1.620	-1.8	86	0.00
11	bis(2-Chloroethyl)ether	1.308	1.245	4.8	85	0.00
12	1,3-Dichlorobenzene	1.466	1.469	-0.2	85	0.00
13 C	1,4-Dichlorobenzene	1.490	1.495	-0.3	86	0.00
14	1,2-Dichlorobenzene	1.505	1.451	3.6	86	0.00
15	Benzyl Alcohol	1.241	1.311	-5.6	88	0.00
16	2,2'-oxybis(1-Chloropropane	1.580	1.464	7.3	87	0.00
17	2-Methylphenol	1.096	1.124	-2.6	86	0.00
18	Hexachloroethane	0.505	0.487	3.6	81	0.00
19 P	n-Nitroso-di-n-propylamine	1.021	0.977	4.3	84	0.00
20	3+4-Methylphenols	1.519	1.530	-0.7	84	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	81	0.00
22	Acetophenone	0.509	0.532	-4.5	84	0.00
23 S	Nitrobenzene-d5	0.374	0.392	-4.8	83	0.00
24	Nitrobenzene	0.357	0.371	-3.9	82	0.00
25	Isophorone	0.650	0.669	-2.9	81	0.00
26 C	2-Nitrophenol	0.182	0.188	-3.3	79	0.00
27	2,4-Dimethylphenol	0.269	0.274	-1.9	80	0.00
28	bis(2-Chloroethoxy)methane	0.431	0.433	-0.5	81	0.00
29 C	2,4-Dichlorophenol	0.332	0.347	-4.5	82	0.00
30	1,2,4-Trichlorobenzene	0.379	0.385	-1.6	82	0.00
31	Naphthalene	1.021	1.023	-0.2	81	0.00
32	Benzoic acid	0.233	0.245	-5.2	78	0.00
33	4-Chloroaniline	0.439	0.447	-1.8	81	0.00
34 C	Hexachlorobutadiene	0.259	0.269	-3.9	82	0.00
35	Caprolactam	0.073	0.074	-1.4	79	0.00
36 C	4-Chloro-3-methylphenol	0.375	0.391	-4.3	83	0.00
37	2-Methylnaphthalene	0.746	0.748	-0.3	81	0.00
38	1-Methylnaphthalene	0.728	0.734	-0.8	81	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	81	0.00
40	1,2,4,5-Tetrachlorobenzene	0.666	0.669	-0.5	81	0.00
41 P	Hexachlorocyclopentadiene	0.306	0.319	-4.2	80	0.00
42 S	2,4,6-Tribromophenol	0.315	0.339	-7.6	88	0.00
43 C	2,4,6-Trichlorophenol	0.455	0.463	-1.8	81	0.00
44	2,4,5-Trichlorophenol	0.495	0.499	-0.8	81	0.00
45 S	2-Fluorobiphenyl	1.426	1.404	1.5	83	0.00
46	1,1'-Biphenyl	1.494	1.464	2.0	81	0.00
47	2-Chloronaphthalene	1.169	1.157	1.0	82	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111521\
 Data File : BP007978.D
 Acq On : 15 Nov 2021 10:18
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 15 11:44:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 05:36:49 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.332	0.352	-6.0	83	0.00
49	Acenaphthylene	1.778	1.808	-1.7	81	0.00
50	Dimethylphthalate	1.582	1.568	0.9	82	0.00
51	2,6-Dinitrotoluene	0.329	0.328	0.3	80	0.00
52 C	Acenaphthene	1.064	1.079	-1.4	82	0.00
53	3-Nitroaniline	0.335	0.347	-3.6	81	0.00
54 P	2,4-Dinitrophenol	0.198	0.215	-8.6	80	0.00
55	Dibenzofuran	1.809	1.849	-2.2	88	0.00
56 P	4-Nitrophenol	0.273	0.291	-6.6	83	0.00
57	2,4-Dinitrotoluene	0.438	0.471	-7.5	87	0.00
58	Fluorene	1.440	1.417	1.6	88	0.00
59	2,3,4,6-Tetrachlorophenol	0.451	0.480	-6.4	89	0.00
60	Diethylphthalate	1.525	1.570	-3.0	87	0.00
61	4-Chlorophenyl-phenylether	0.779	0.795	-2.1	88	0.00
62	4-Nitroaniline	0.344	0.372	-8.1	88	0.00
63	Azobenzene	1.284	1.355	-5.5	88	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
65	4,6-Dinitro-2-methylphenol	0.131	0.142	-8.4	85	0.00
66 c	n-Nitrosodiphenylamine	0.572	0.579	-1.2	88	0.00
67	4-Bromophenyl-phenylether	0.234	0.238	-1.7	87	0.00
68	Hexachlorobenzene	0.263	0.274	-4.2	89	0.00
69	Atrazine	0.149	0.152	-2.0	86	0.00
70 C	Pentachlorophenol	0.161	0.175	-8.7	87	0.00
71	Phenanthrene	1.066	1.065	0.1	87	0.00
72	Anthracene	1.045	1.060	-1.4	88	0.00
73	Carbazole	1.029	1.046	-1.7	88	0.00
74	Di-n-butylphthalate	1.164	1.219	-4.7	88	0.00
75 C	Fluoranthene	1.349	1.373	-1.8	88	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
77	Benzydine	0.307	0.370	-20.5	98	0.00
78	Pyrene	1.196	1.176	1.7	89	0.00
79 S	Terphenyl-d14	0.986	0.973	1.3	91	0.00
80	Butylbenzylphthalate	0.494	0.503	-1.8	89	0.00
81	Benzo(a)anthracene	1.317	1.313	0.3	91	0.00
82	3,3'-Dichlorobenzidine	0.608	0.646	-6.3	93	0.00
83	Chrysene	1.262	1.260	0.2	91	0.00
84	Bis(2-ethylhexyl)phthalate	0.705	0.720	-2.1	93	0.00
85 c	Di-n-octyl phthalate	1.299	1.353	-4.2	91	0.00
86 I	Perylene-d12	1.000	1.000	0.0	93	0.00
87	Indeno(1,2,3-cd)pyrene	1.444	1.509	-4.5	97	0.00
88	Benzo(b)fluoranthene	1.191	1.174	1.4	94	0.00
89	Benzo(k)fluoranthene	1.189	1.209	-1.7	93	0.00
90 C	Benzo(a)pyrene	1.018	1.028	-1.0	93	0.00
91	Dibenzo(a,h)anthracene	1.198	1.253	-4.6	97	0.00
92	Benzo(g,h,i)perylene	1.181	1.247	-5.6	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111521\
Data File : BP007978.D
Acq On : 15 Nov 2021 10:18
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 15 11:44:07 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 15 05:36:49 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0