

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111721\
 Data File : BP008001.D
 Acq On : 17 Nov 2021 09:36
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 17 10:39:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 05:36:49 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	34.951	12.6	95	0.00
3	Pyridine	40.000	36.783	8.0	97	0.00
4	n-Nitrosodimethylamine	40.000	39.484	1.3	99	0.00
5 S	2-Fluorophenol	80.000	74.572	6.8	98	0.00
6	Aniline	40.000	39.819	0.5	100	0.00
7 S	Phenol-d6	80.000	76.540	4.3	99	0.00
8	2-Chlorophenol	40.000	39.326	1.7	98	0.00
9	Benzaldehyde	40.000	38.976	2.6	93	0.00
10 C	Phenol	40.000	39.895	0.3	99	0.00
11	bis(2-Chloroethyl)ether	40.000	37.341	6.6	98	0.00
12	1,3-Dichlorobenzene	40.000	39.140	2.1	98	0.00
13 C	1,4-Dichlorobenzene	40.000	39.370	1.6	99	0.00
14	1,2-Dichlorobenzene	40.000	37.747	5.6	98	0.00
15	Benzyl Alcohol	40.000	40.493	-1.2	99	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.685	-1.7	98	0.00
17	2-Methylphenol	40.000	39.603	1.0	97	0.00
18	Hexachloroethane	40.000	40.149	-0.4	98	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.369	4.1	99	0.00
20	3+4-Methylphenols	40.000	40.310	-0.8	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	90	0.00
22	Acetophenone	40.000	44.055	-10.1	98	0.00
23 S	Nitrobenzene-d5	80.000	89.727	-12.2	100	0.00
24	Nitrobenzene	40.000	44.382	-11.0	98	0.00
25	Isophorone	40.000	44.395	-11.0	98	0.00
26 C	2-Nitrophenol	40.000	45.473	-13.7	97	0.00
27	2,4-Dimethylphenol	40.000	43.918	-9.8	96	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.122	-5.3	94	0.00
29 C	2,4-Dichlorophenol	40.000	42.421	-6.1	93	0.00
30	1,2,4-Trichlorobenzene	40.000	41.557	-3.9	94	0.00
31	Naphthalene	40.000	40.638	-1.6	92	0.00
32	Benzoic acid	40.000	44.883	-12.2	93	0.00
33	4-Chloroaniline	40.000	40.632	-1.6	91	0.00
34 C	Hexachlorobutadiene	40.000	43.155	-7.9	96	0.00
35	Caprolactam	40.000	40.563	-1.4	87	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.546	-3.9	92	-0.01
37	2-Methylnaphthalene	40.000	40.438	-1.1	91	0.00
38	1-Methylnaphthalene	40.000	40.538	-1.3	92	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	90	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.803	-2.0	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.403	-6.0	97	0.00
42 S	2,4,6-Tribromophenol	80.000	86.551	-8.2	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.708	-1.8	90	0.00
44	2,4,5-Trichlorophenol	40.000	40.480	-1.2	91	0.00
45 S	2-Fluorobiphenyl	80.000	79.505	0.6	93	0.00
46	1,1'-Biphenyl	40.000	39.170	2.1	91	0.00
47	2-Chloronaphthalene	40.000	39.668	0.8	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111721\
 Data File : BP008001.D
 Acq On : 17 Nov 2021 09:36
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 17 10:39:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 05:36:49 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.076	-5.2	92	0.00
49	Acenaphthylene	40.000	40.198	-0.5	90	0.00
50	Dimethylphthalate	40.000	39.145	2.1	90	0.00
51	2,6-Dinitrotoluene	40.000	40.093	-0.2	89	0.00
52 C	Acenaphthene	40.000	39.820	0.4	90	0.00
53	3-Nitroaniline	40.000	40.720	-1.8	89	0.00
54 P	2,4-Dinitrophenol	40.000	42.799	-7.0	89	0.00
55	Dibenzofuran	40.000	40.438	-1.1	97	0.00
56 P	4-Nitrophenol	40.000	40.576	-1.4	89	0.00
57	2,4-Dinitrotoluene	40.000	42.042	-5.1	95	0.00
58	Fluorene	40.000	39.163	2.1	98	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.608	-4.0	97	0.00
60	Diethylphthalate	40.000	40.654	-1.6	96	0.00
61	4-Chlorophenyl-phenylether	40.000	40.832	-2.1	99	0.00
62	4-Nitroaniline	40.000	41.869	-4.7	95	0.00
63	Azobenzene	40.000	41.198	-3.0	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.402	-11.0	94	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.864	-2.2	95	0.00
67	4-Bromophenyl-phenylether	40.000	42.053	-5.1	97	0.00
68	Hexachlorobenzene	40.000	42.644	-6.6	98	0.00
69	Atrazine	40.000	41.384	-3.5	93	0.00
70 C	Pentachlorophenol	40.000	45.306	-13.3	98	0.00
71	Phenanthrene	40.000	40.428	-1.1	95	0.00
72	Anthracene	40.000	40.965	-2.4	96	0.00
73	Carbazole	40.000	40.441	-1.1	94	0.00
74	Di-n-butylphthalate	40.000	41.748	-4.4	95	0.00
75 C	Fluoranthene	40.000	40.591	-1.5	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	96	0.00
77	Benzydine	40.000	39.181	2.0	84	0.00
78	Pyrene	40.000	39.926	0.2	95	0.00
79 S	Terphenyl-d14	80.000	81.263	-1.6	99	0.00
80	Butylbenzylphthalate	40.000	40.333	-0.8	93	0.00
81	Benzo(a)anthracene	40.000	40.147	-0.4	96	0.00
82	3,3'-Dichlorobenzidine	40.000	41.093	-2.7	94	0.00
83	Chrysene	40.000	40.023	-0.1	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.291	-3.2	99	0.00
85 c	Di-n-octyl phthalate	40.000	41.108	-2.8	95	0.00
86 I	Perylene-d12	20.000	20.000	0.0	97	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	40.762	-1.9	98	-0.01
88	Benzo(b)fluoranthene	40.000	39.665	0.8	98	-0.01
89	Benzo(k)fluoranthene	40.000	40.172	-0.4	95	0.00
90 C	Benzo(a)pyrene	40.000	40.294	-0.7	97	0.00
91	Dibenzo(a,h)anthracene	40.000	41.082	-2.7	99	-0.01
92	Benzo(g,h,i)perylene	40.000	40.952	-2.4	99	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111721\
Data File : BP008001.D
Acq On : 17 Nov 2021 09:36
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 17 10:39:11 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 15 05:36:49 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0