

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111824\
 Data File : BP023196.D
 Acq On : 18 Nov 2024 13:51
 Operator : RC/JU
 Sample : P4829-17
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E29S8

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/19/2024
 Supervised By :mohammad ahmed 11/19/2024

Quant Time: Nov 18 14:46:18 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 13 04:49:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.575	152	52144	20.000	ng/u1	0.00
20) Naphthalene-d8	10.316	136	204048	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.187	164	166112	20.000	ng/u1	-0.02
64) Phenanthrene-d10	17.004	188	417156	20.000	ng/u1	#-0.01
79) Chrysene-d12	21.433	240	452555	20.000	ng/u1	-0.01
88) Perylene-d12	24.662	264	525525	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.134	96	6148	5.667	ng/uL	0.00
4) Pyridine-d5	3.569	84	14837	4.603	ng/u1	0.02
7) Phenol-d5	6.787	99	40625m	10.538	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.922	67	69913	30.867	ng/u1	0.00
11) 2-Chlorophenol-d4	7.122	132	81782	28.966	ng/u1	0.00
15) 4-Methylphenol-d8	8.287	113	68290	21.526	ng/u1	0.00
21) Nitrobenzene-d5	8.716	128	43396	30.549	ng/u1	0.00
24) 2-Nitrophenol-d4	9.428	143	53728	31.726	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.975	165	109649	30.917	ng/u1	0.00
31) 4-Chloroaniline-d4	10.487	131	50247m	11.876	ng/u1	0.00
46) Dimethylphthalate-d6	13.587	166	413716	34.556	ng/u1	-0.02
49) Acenaphthylene-d8	13.875	160	397766	31.841	ng/u1	-0.02
54) 4-Nitrophenol-d4	14.522	143	15875m	8.823	ng/u1	0.07
60) Fluorene-d10	15.216	176	372625	35.500	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.363	200	68771	27.520	ng/u1	0.00
73) Anthracene-d10	17.098	188	673458	38.129	ng/u1	-0.02
81) Pyrene-d10	19.492	212	861999	41.353	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.427	264	949322	37.923	ng/u1	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111824\
Data File : BP023196.D
Acq On : 18 Nov 2024 13:51
Operator : RC/JU
Sample : P4829-17
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E29S8

Quant Time: Nov 18 14:46:18 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 13 04:49:01 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/19/2024
Supervised By :mohammad ahmed 11/19/2024

