

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111924\
 Data File : BP023218.D
 Acq On : 19 Nov 2024 15:17
 Operator : RC/JU
 Sample : P4875-11
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E29S9

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/20/2024
 Supervised By :mohammad ahmed 11/22/2024

Quant Time: Nov 19 15:59:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 13 04:49:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.569	152	83882	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.322	136	300578	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.193	164	217464	20.000	ng/u1	-0.01
64) Phenanthrene-d10	16.992	188	546617	20.000	ng/u1	-0.02
79) Chrysene-d12	21.439	240	637710	20.000	ng/u1	0.00
88) Perylene-d12	24.657	264	753800	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.140	96	9072	5.199	ng/uL	0.00
4) Pyridine-d5	3.564	84	40381	7.788	ng/u1	0.02
7) Phenol-d5	6.787	99	46157m	7.443	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.922	67	108568	29.797	ng/u1	0.00
11) 2-Chlorophenol-d4	7.122	132	120741	26.584	ng/u1	0.00
15) 4-Methylphenol-d8	8.287	113	89347	17.507	ng/u1	0.00
21) Nitrobenzene-d5	8.716	128	63604	30.396	ng/u1	0.00
24) 2-Nitrophenol-d4	9.422	143	75592	30.301	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	9.975	165	164667	31.519	ng/u1	0.00
31) 4-Chloroaniline-d4	10.481	131	141817	22.754	ng/u1	0.00
46) Dimethylphthalate-d6	13.604	166	571958	36.493	ng/u1	0.00
49) Acenaphthylene-d8	13.881	160	550239	33.646	ng/u1	-0.02
54) 4-Nitrophenol-d4	14.516	143	16786m	7.127	ng/u1	0.06
60) Fluorene-d10	15.187	176	486485	35.403	ng/u1	-0.02
65) 4,6-Dinitro-2-methylph...	15.369	200	100869	30.804	ng/u1	0.00
73) Anthracene-d10	17.104	188	905566	39.127	ng/u1	-0.01
81) Pyrene-d10	19.486	212	1192346	40.593	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.433	264	1452961	40.465	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111924\
Data File : BP023218.D
Acq On : 19 Nov 2024 15:17
Operator : RC/JU
Sample : P4875-11
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E29S9

Quant Time: Nov 19 15:59:22 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 13 04:49:01 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/20/2024
Supervised By :mohammad ahmed 11/22/2024

